

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051921\
 Data File : BN014674.D
 Acq On : 19 May 2021 12:55
 Operator : CG/JU
 Sample : PB136498BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS498

Quant Time: May 19 14:41:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 15:50:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	7.730	152	526	0.40 ng/ul	0.00
4) Naphthalene-d8	10.516	136	1731	0.40 ng/ul	0.00
9) Acenaphthene-d10	14.375	164	1215	0.40 ng/ul	0.00
13) Phenanthrene-d10	17.127	188	2964	0.40 ng/ul	0.00
17) Chrysene-d12	21.329	240	3394	0.40 ng/ul	0.00
23) Perylene-d12	23.597	264	3032	0.40 ng/ul	0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.254	96	355	0.66 ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.111	152	964	0.32 ng/ul	0.00
18) Fluoranthene-d10	19.163	212	3213	0.38 ng/ul	0.00
Target Compounds					
					Qvalue
2) 1,4-Dioxane	3.287	88	906	1.61 ng/ul#	86
5) Naphthalene	10.565	128	1599	0.32 ng/ul	96
7) 2-Methylnaphthalene	12.182	142	1113	0.31 ng/ul	98
8) 1-Methylnaphthalene	12.408	142	1077	0.31 ng/ul	99
10) Acenaphthylene	14.093	152	1348	0.31 ng/ul	99
11) Acenaphthene	14.435	153	1285	0.31 ng/ul	98
12) Fluorene	15.430	166	1557	0.30 ng/ul	99
14) Pentachlorophenol	16.781	266	291	0.70 ng/ul	95
15) Phenanthrene	17.169	178	2891	0.31 ng/ul	98
16) Anthracene	17.262	178	2358	0.33 ng/ul	99
19) Fluoranthene	19.190	202	3811	0.36 ng/ul	97
20) Pyrene	19.558	202	3825	0.36 ng/ul	98
21) Benzo(a)anthracene	21.312	228	3692	0.34 ng/ul	99
22) Chrysene	21.364	228	4646	0.38 ng/ul	99
24) Benzo(b)fluoranthene	22.913	252	4675	0.35 ng/ul	99
25) Benzo(k)fluoranthene	22.957	252	4931	0.35 ng/ul	96
26) Benzo(a)pyrene	23.495	252	3653	0.36 ng/ul	97
27) Indeno(1,2,3-cd)pyrene	25.886	276	5368	0.37 ng/ul	99
28) Dibenzo(a,h)anthracene	25.903	278	4366	0.38 ng/ul	99
29) Benzo(g,h,i)perylene	26.583	276	4581	0.37 ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051921\
Data File : BN014674.D
Acq On : 19 May 2021 12:55
Operator : CG/JU
Sample : PB136498BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS498

Quant Time: May 19 14:41:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 15:50:11 2021
Response via : Initial Calibration

