

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051923\
 Data File : BN025505.D
 Acq On : 20 May 2023 02:15
 Operator : CG/JU
 Sample : 02834-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE122D3-20230516

Quant Time: May 20 02:58:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 19 02:17:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.077	152	4372	0.400	ng	0.00
7) Naphthalene-d8	10.899	136	12555	0.400	ng	0.00
13) Acenaphthene-d10	14.705	164	7420	0.400	ng	0.00
19) Phenanthrene-d10	17.447	188	17247	0.400	ng	# 0.00
29) Chrysene-d12	21.620	240	13343	0.400	ng	#-0.01
35) Perylene-d12	24.142	264	13097	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.607	112	1818	0.184	ng	0.00
5) Phenol-d6	7.218	99	1436	0.113	ng	0.00
8) Nitrobenzene-d5	9.234	82	3967	0.396	ng	0.00
11) 2-Methylnaphthalene-d10	12.472	152	7503	0.430	ng	0.00
14) 2,4,6-Tribromophenol	16.181	330	1244	0.430	ng	-0.01
15) 2-Fluorobiphenyl	13.336	172	11976	0.422	ng	0.00
27) Fluoranthene-d10	19.464	212	18732	0.470	ng	0.00
31) Terphenyl-d14	20.053	244	12176	0.513	ng	-0.01
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.522	149	2771	0.097	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051923\
Data File : BN025505.D
Acq On : 20 May 2023 02:15
Operator : CG/JU
Sample : 02834-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE122D3-20230516

Quant Time: May 20 02:58:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 19 02:17:42 2023
Response via : Initial Calibration

