

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051925\
 Data File : BN037062.D
 Acq On : 20 May 2025 01:54
 Operator : RC/JU
 Sample : Q2063-03
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW5-SP303-20250515

Quant Time: May 20 04:20:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	

Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.611	152	1885	0.400	ng	0.00	
7) Naphthalene-d8	10.394	136	5104	0.400	ng	-0.01	
13) Acenaphthene-d10	14.256	164	2926	0.400	ng	-0.01	
19) Phenanthrene-d10	17.009	188	6019	0.400	ng	# 0.00	
29) Chrysene-d12	21.207	240	5040	0.400	ng	# 0.00	
35) Perylene-d12	23.404	264	4381	0.400	ng	-0.01	
System Monitoring Compounds							
4) 2-Fluorophenol	5.206	112	623	0.126	ng	0.00	
5) Phenol-d6	6.788	99	454	0.073	ng	0.00	
8) Nitrobenzene-d5	8.760	82	1491	0.268	ng	-0.01	
11) 2-Methylnaphthalene-d10	11.991	152	2027	0.282	ng	0.00	
14) 2,4,6-Tribromophenol	15.755	330	440	0.342	ng	-0.01	
15) 2-Fluorobiphenyl	12.883	172	3558	0.266	ng	0.00	
27) Fluoranthene-d10	19.045	212	6074	0.368	ng	0.00	
31) Terphenyl-d14	19.649	244	4808	0.446	ng	0.00	
Target Compounds							
34) Bis(2-ethylhexyl)phtha...	21.126	149	3328	0.285	ng		Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051925\
Data File : BN037062.D
Acq On : 20 May 2025 01:54
Operator : RC/JU
Sample : Q2063-03
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW5-SP303-20250515

Quant Time: May 20 04:20:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 14 11:26:32 2025
Response via : Initial Calibration

