

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052121\
 Data File : BN014697.D
 Acq On : 21 May 2021 17:24
 Operator : CG/JU
 Sample : M2383-14
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 FB-02-05122021

Quant Time: May 21 17:53:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 21 11:30:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	944	0.40	ng	# 0.00
7) Naphthalene-d8	10.505	136	3400	0.40	ng	# 0.00
13) Acenaphthene-d10	14.372	164	2449	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	5173	0.40	ng	# 0.00
29) Chrysene-d12	21.324	240	4934	0.40	ng	# 0.01
35) Perylene-d12	23.588	264	4404	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.325	112	301	0.12	ng	0.00
5) Phenol-d6	6.895	99	238	0.07	ng	0.00
8) Nitrobenzene-d5	8.870	82	1007	0.34	ng	0.00
11) 2-Methylnaphthalene-d10	12.106	152	1752	0.21	ng	0.00
14) 2,4,6-Tribromophenol	15.869	330	452	0.40	ng	0.00
15) 2-Fluorobiphenyl	12.989	172	2831	0.32	ng	0.00
27) Fluoranthene-d10	19.159	212	5730	0.25	ng	0.00
31) Terphenyl-d14	19.764	244	6361	0.55	ng	0.00
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	21.239	149	769	0.19	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052121\
Data File : BN014697.D
Acq On : 21 May 2021 17:24
Operator : CG/JU
Sample : M2383-14
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
FB-02-05122021

Quant Time: May 21 17:53:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 21 11:30:28 2021
Response via : Initial Calibration

