

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052125\
 Data File : BN037074.D
 Acq On : 20 May 2025 15:09
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 21 01:37:13 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	129	0.00
2	1,4-Dioxane	0.400	0.398	0.5	123	0.00
3	n-Nitrosodimethylamine	0.400	0.384	4.0	133	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	110	0.00
5 S	Phenol-d6	0.400	0.366	8.5	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.395	1.3	133	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	136	-0.01
8 S	Nitrobenzene-d5	0.400	0.369	7.8	137	-0.01
9	Naphthalene	0.400	0.373	6.8	131	-0.01
10	Hexachlorobutadiene	0.400	0.388	3.0	134	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.407	-1.7	141	0.00
12	2-Methylnaphthalene	0.400	0.377	5.8	132	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	135	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.349	12.8	116	-0.01
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	128	0.00
16	Acenaphthylene	0.400	0.377	5.8	131	-0.01
17	Acenaphthene	0.400	0.380	5.0	132	-0.01
18	Fluorene	0.400	0.382	4.5	132	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	135	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.347	13.3	122	-0.01
21	4-Bromophenyl-phenylether	0.400	0.392	2.0	134	-0.01
22	Hexachlorobenzene	0.400	0.402	-0.5	131	-0.01
23	Atrazine	0.400	0.351	12.3	124	-0.01
24	Pentachlorophenol	0.400	0.353	11.8	126	-0.01
25	Phenanthrene	0.400	0.373	6.8	127	-0.01
26	Anthracene	0.400	0.360	10.0	124	-0.01
27 SURR	Fluoranthene-d10	0.400	0.368	8.0	127	0.00
28	Fluoranthene	0.400	0.341	14.8	120	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	104	0.00
30	Pyrene	0.400	0.448	-12.0	116	0.00
31 S	Terphenyl-d14	0.400	0.440	-10.0	112	0.00
32	Benzo(a)anthracene	0.400	0.373	6.8	99	0.00
33	Chrysene	0.400	0.379	5.3	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	95	0.00
35 I	Perylene-d12	0.400	0.400	0.0	84	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	80	-0.02
37	Benzo(b)fluoranthene	0.400	0.382	4.5	84	0.00
38	Benzo(k)fluoranthene	0.400	0.404	-1.0	85	0.00
39 C	Benzo(a)pyrene	0.400	0.373	6.8	80	0.00
40	Dibenzo(a,h)anthracene	0.400	0.372	7.0	79	-0.02
41	Benzo(g,h,i)perylene	0.400	0.391	2.3	80	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4