

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052324\
 Data File : BN031623.D
 Acq On : 23 May 2024 08:48
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020258

Quant Time: May 23 22:49:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 21 23:42:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.693	152	582986	20.000	ng/u1	0.00
20) Naphthalene-d8	10.475	136	2559029	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.328	164	1593728	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.075	188	3073307	20.000	ng/u1	0.00
79) Chrysene-d12	21.304	240	2104193	20.000	ng/u1	0.00
88) Perylene-d12	24.233	264	1893923	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.205	96	116703	7.934	ng/uL	0.00
4) Pyridine-d5	3.605	84	849636	20.700	ng/u1	0.00
7) Phenol-d5	6.852	99	995859	20.751	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.028	67	609982	21.192	ng/u1	0.00
11) 2-Chlorophenol-d4	7.222	132	747251	20.207	ng/u1	0.00
15) 4-Methylphenol-d8	8.387	113	786643	21.174	ng/u1	0.00
21) Nitrobenzene-d5	8.846	128	380872	19.666	ng/u1	0.00
24) 2-Nitrophenol-d4	9.563	143	358524	19.386	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.093	165	738938	20.435	ng/u1	0.00
31) 4-Chloroaniline-d4	10.610	131	1101697	21.136	ng/u1	0.00
46) Dimethylphthalate-d6	13.746	166	2132379	21.231	ng/u1	0.00
49) Acenaphthylene-d8	14.022	160	2602447	20.674	ng/u1	0.00
54) 4-Nitrophenol-d4	14.522	143	405088	21.316	ng/u1	0.00
60) Fluorene-d10	15.322	176	1839507	21.134	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.440	200	252602	19.011	ng/u1	0.00
73) Anthracene-d10	17.169	188	2667839	20.506	ng/u1	0.00
81) Pyrene-d10	19.469	212	2745958	20.805	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.039	264	1783089	20.001	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	124734	8.048	ng/uL	95
5) Pyridine	3.623	79	849150	21.020	ng/u1	99
6) Benzaldehyde	6.840	77	577728	23.503	ng/u1	98
8) Phenol	6.881	94	1040903	21.104	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.122	93	864833	21.130	ng/u1	98
12) 2-Chlorophenol	7.252	128	791351	20.405	ng/u1	99
13) 2-Methylphenol	8.128	108	779436	21.096	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.216	45	1024191	21.663	ng/u1	99
16) Acetophenone	8.511	105	1290856	22.276	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.499	70	641329	21.542	ng/u1	100
18) 4-Methylphenol	8.452	108	842430	21.454	ng/u1	97
19) Hexachloroethane	8.763	117	340648	20.558	ng/u1	96
22) Nitrobenzene	8.887	77	953005	20.087	ng/u1	98
23) Isophorone	9.410	82	1856493	20.571	ng/u1	97
25) 2-Nitrophenol	9.593	139	415925	20.178	ng/u1	99
26) 2,4-Dimethylphenol	9.652	107	854813	20.301	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.893	93	1183384	20.569	ng/u1	98
29) 2,4-Dichlorophenol	10.116	162	721412	20.124	ng/u1	98
30) Naphthalene	10.522	128	2631790	19.984	ng/u1	99
32) 4-Chloroaniline	10.634	127	1088961	21.410	ng/u1	98
33) Hexachlorobutadiene	10.805	225	435586	19.856	ng/u1	94
34) Caprolactam	11.399	113	165401	14.246	ng/u1	91
35) 4-Chloro-3-methylphenol	11.757	107	844923	21.703	ng/u1	98
36) 2-Methylnaphthalene	12.140	142	1784025	20.700	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052324\
 Data File : BN031623.D
 Acq On : 23 May 2024 08:48
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020258

Quant Time: May 23 22:49:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 21 23:42:49 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.357	142	1778493	20.847	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.504	216	863420	19.143	ng/ul	98
40) Hexachlorocyclopentadiene	12.487	237	451721	16.707	ng/ul	93
41) 2,4,6-Trichlorophenol	12.751	196	550121	19.737	ng/ul	99
42) 2,4,5-Trichlorophenol	12.816	196	615128	20.388	ng/ul	94
43) 1,1'-Biphenyl	13.163	154	2288358	19.916	ng/ul	99
44) 2-Chloronaphthalene	13.198	162	1797062	19.753	ng/ul	99
45) 2-Nitroaniline	13.404	65	503650	21.147	ng/ul	97
47) Dimethylphthalate	13.793	163	2186656	21.336	ng/ul	98
48) 2,6-Dinitrotoluene	13.910	165	429198	21.974	ng/ul	99
50) Acenaphthylene	14.051	152	2960362	20.372	ng/ul	99
51) 3-Nitroaniline	14.234	138	478814	21.597	ng/ul	99
52) Acenaphthene	14.393	153	1948276	20.693	ng/ul	89
53) 2,4-Dinitrophenol	14.440	184	157850	17.599	ng/ul	100
55) 4-Nitrophenol	14.540	109	310693	21.698	ng/ul	96
56) Dibenzofuran	14.728	168	2575460	20.490	ng/ul	99
57) 2,4-Dinitrotoluene	14.693	165	616530	23.561	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.951	232	467941	21.350	ng/ul	96
59) Diethylphthalate	15.163	149	2191369	22.692	ng/ul	98
61) Fluorene	15.375	166	2099016	21.319	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.381	204	994025	20.971	ng/ul	99
63) 4-Nitroaniline	15.398	138	459986	23.134	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.451	198	273921	18.798	ng/ul	99
67) N-Nitrosodiphenylamine	15.592	169	1774030	19.609	ng/ul	97
68) 4-Bromophenyl-phenylether	16.269	248	615930	19.905	ng/ul	91
69) Hexachlorobenzene	16.369	284	671954	20.120	ng/ul	99
70) Atrazine	16.545	200	505759	20.771	ng/ul	100
71) Pentachlorophenol	16.716	266	385567	20.241	ng/ul	96
72) Phenanthrene	17.116	178	3214043	20.607	ng/ul	99
74) Anthracene	17.204	178	3241754	20.716	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.116	216	853510	17.470	ng/ul	96
76) Pentachlorobenzene	14.640	250	807239	18.247	ng/ul	94
77) Carbazole	17.475	167	2867139	21.471	ng/ul	99
78) Di-n-butylphthalate	18.051	149	3508792	22.884	ng/ul	99
80) Fluoranthene	19.133	202	3319156	20.345	ng/ul	99
82) Pyrene	19.498	202	3421948	20.453	ng/ul	99
83) Butylbenzylphthalate	20.410	149	1303732	22.961	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.216	252	797821	19.719	ng/ul	99
85) Benzo(a)anthracene	21.286	228	2856714	20.189	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.222	149	1795285	23.091	ng/ul	99
87) Chrysene	21.345	228	2622541	20.065	ng/ul	98
89) Di-n-octyl phthalate	22.339	149	2518970	23.270	ng/ul	100
90) Benzo(b)fluoranthene	23.310	252	2254080	20.395	ng/ul	100
91) Benzo(k)fluoranthene	23.369	252	2305151	20.461	ng/ul	98
93) Benzo(a)pyrene	24.104	252	2172975	20.468	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	27.527	276	2136786	16.770	ng/ul	99
95) Dibenzo(a,h)anthracene	27.598	278	2056721	19.601	ng/ul	98
96) Benzo(g,h,i)perylene	28.562	276	2062086	19.588	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052324\
Data File : BN031623.D
Acq On : 23 May 2024 08:48
Operator : MA/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD020258

Quant Time: May 23 22:49:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 21 23:42:49 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.048 ng/uL
 RT: 3.240 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020258


Tgt Ion: 88 Resp: 124734
 Ion Ratio Lower Upper
 88 100
 43 21.2 18.3 27.5
 58 64.6 55.5 83.3



#5
 Pyridine
 Concen: 21.020 ng/uL
 RT: 3.623 min Scan# 108
 Delta R.T. 0.000 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Tgt Ion: 79 Resp: 849150
 Ion Ratio Lower Upper
 79 100
 52 65.9 52.3 78.5
 51 30.7 24.4 36.6









#6
Benzaldehyde
Concen: 23.503 ng/ul
RT: 6.840 min Scan# 655
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion: 77 Resp: 577728
Ion Ratio Lower Upper
77 100
105 99.2 82.1 123.1
106 95.0 75.5 113.3



#8
Phenol
Concen: 21.104 ng/ul
RT: 6.881 min Scan# 662
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion: 94 Resp: 1040903
Ion Ratio Lower Upper
94 100
65 26.3 20.8 31.2
66 32.8 26.7 40.1









#10
Bis(2-Chloroethyl)ether
Concen: 21.130 ng/ul
RT: 7.122 min Scan# 703
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion: 93 Resp: 864833
Ion Ratio Lower Upper
93 100
63 72.3 56.3 84.5
95 32.4 25.2 37.8



#12
2-Chlorophenol
Concen: 20.405 ng/ul
RT: 7.252 min Scan# 725
Delta R.T. -0.006 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:128 Resp: 791351
Ion Ratio Lower Upper
128 100
64 45.2 36.1 54.1
130 33.3 26.0 39.0









#13
2-Methylphenol
Concen: 21.096 ng/ul
RT: 8.128 min Scan# 874
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:108 Resp: 779436
Ion Ratio Lower Upper
108 100
107 85.5 71.0 106.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.663 ng/ul
RT: 8.216 min Scan# 889
Delta R.T. -0.006 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion: 45 Resp: 1024191
Ion Ratio Lower Upper
45 100
77 16.3 13.2 19.8
79 12.5 10.6 16.0









#16
Acetophenone
Concen: 22.276 ng/ul
RT: 8.511 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:105 Resp: 1290856
Ion Ratio Lower Upper
105 100
77 76.4 60.7 91.1
51 22.3 18.6 28.0



#17
N-Nitroso-di-n-propylamine
Concen: 21.542 ng/ul
RT: 8.499 min Scan# 937
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion: 70 Resp: 641329
Ion Ratio Lower Upper
70 100
42 47.9 38.2 57.4
101 10.1 8.1 12.1
130 21.9 17.7 26.5









#18
4-Methylphenol
Concen: 21.454 ng/ul
RT: 8.452 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:108 Resp: 842430
Ion Ratio Lower Upper
108 100
107 112.9 87.6 131.4



#19
Hexachloroethane
Concen: 20.558 ng/ul
RT: 8.763 min Scan# 982
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:117 Resp: 340648
Ion Ratio Lower Upper
117 100
201 86.2 72.6 109.0
199 54.6 45.6 68.4









#22
Nitrobenzene
Concen: 20.087 ng/ul
RT: 8.887 min Scan# 1003
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion: 77 Resp: 953005
Ion Ratio Lower Upper
77 100
123 44.9 34.7 52.1
65 12.4 10.1 15.1



#23
Isophorone
Concen: 20.571 ng/ul
RT: 9.410 min Scan# 1092
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion: 82 Resp: 1856493
Ion Ratio Lower Upper
82 100
95 7.1 5.4 8.2
138 18.4 13.4 20.2





#25

2-Nitrophenol

Concen: 20.178 ng/ul

RT: 9.593 min Scan# 1123

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:139 Resp: 415925

Ion Ratio Lower Upper

139 100

65 49.4 38.6 58.0

109 29.3 23.2 34.8



#26

2,4-Dimethylphenol

Concen: 20.301 ng/ul

RT: 9.652 min Scan# 1133

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion:107 Resp: 854813


Ion Ratio Lower Upper

107 100

121 49.9 40.9 61.3

122 84.7 69.5 104.3





#27

Bis(2-Chloroethoxy)methane

Concen: 20.569 ng/ul

RT: 9.893 min Scan# 1174

Delta R.T. -0.006 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion: 93 Resp: 1183384

Ion Ratio Lower Upper

93 100

95 31.0 26.2 39.2

123 11.2 9.1 13.7



#29

2,4-Dichlorophenol

Concen: 20.124 ng/ul

RT: 10.116 min Scan# 1212

Delta R.T. -0.006 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion: 162 Resp: 721412


Ion Ratio Lower Upper

162 100

164 65.7 51.2 76.8

98 38.0 29.0 43.4









#30
 Naphthalene
 Concen: 19.984 ng/ul
 RT: 10.522 min Scan# 1281
 Delta R.T. -0.006 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020258

Tgt Ion:128 Resp: 2631790




Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.8	13.2
127	13.3	10.2	15.2




#32
 4-Chloroaniline
 Concen: 21.410 ng/ul
 RT: 10.634 min Scan# 1300
 Delta R.T. 0.000 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Tgt Ion:127 Resp: 1088961

Ion	Ratio	Lower	Upper
127	100		
129	31.1	25.8	38.6









#33
Hexachlorobutadiene
Concen: 19.856 ng/ul
RT: 10.805 min Scan# 1329
Delta R.T. -0.006 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:225 Resp: 435586
Ion Ratio Lower Upper
225 100
223 67.6 49.5 74.3
227 66.9 50.2 75.2



#34
Caprolactam
Concen: 14.246 ng/ul
RT: 11.399 min Scan# 1430
Delta R.T. -0.006 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:113 Resp: 165401
Ion Ratio Lower Upper
113 100
55 154.5 114.6 172.0
56 114.2 84.3 126.5









#35
4-Chloro-3-methylphenol
Concen: 21.703 ng/ul
RT: 11.757 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:107 Resp: 844923
Ion Ratio Lower Upper
107 100
144 24.9 21.4 32.0
142 79.5 65.0 97.6



#36
2-Methylnaphthalene
Concen: 20.700 ng/ul
RT: 12.140 min Scan# 1556
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:142 Resp: 1784025
Ion Ratio Lower Upper
142 100
141 89.8 70.5 105.7









#37
 1-Methylnaphthalene
 Concen: 20.847 ng/ul
 RT: 12.357 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020258

Tgt Ion:142 Resp: 1778493




Ion	Ratio	Lower	Upper
142	100		
141	92.4	77.0	115.6
116	3.8	3.0	4.6




#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 19.143 ng/ul
 RT: 12.504 min Scan# 1618
 Delta R.T. -0.006 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Tgt Ion:216 Resp: 863420

Ion	Ratio	Lower	Upper
216	100		
214	77.3	64.0	96.0
179	21.1	16.8	25.2
108	18.5	15.5	23.3









#40
Hexachlorocyclopentadiene
Concen: 16.707 ng/ul
RT: 12.487 min Scan# 1615
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:237 Resp: 451721
Ion Ratio Lower Upper
237 100
235 57.8 51.3 76.9
272 14.3 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 19.737 ng/ul
RT: 12.751 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:196 Resp: 550121
Ion Ratio Lower Upper
196 100
198 100.3 79.4 119.2
200 31.0 24.0 36.0





#42

2,4,5-Trichlorophenol

Concen: 20.388 ng/ul

RT: 12.816 min Scan# 1671

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:196 Resp: 615128

Ion Ratio Lower Upper

196 100

198 96.5 71.9 107.9

200 31.1 23.2 34.8



#43

1,1'-Biphenyl

Concen: 19.916 ng/ul

RT: 13.163 min Scan# 1730

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion:154 Resp: 2288358


Ion Ratio Lower Upper

154 100

153 39.9 32.6 48.8

76 13.6 11.3 16.9





#44

2-Chloronaphthalene

Concen: 19.753 ng/ul

RT: 13.198 min Scan# 1736

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:162 Resp: 1797062

Ion Ratio Lower Upper

162 100

164 32.5 26.6 39.8

127 38.8 30.8 46.2



#45

2-Nitroaniline

Concen: 21.147 ng/ul

RT: 13.404 min Scan# 1771

Delta R.T. -0.006 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion: 65 Resp: 503650


Ion Ratio Lower Upper

65 100

92 72.9 60.2 90.4

138 106.8 88.6 132.8





#47

Dimethylphthalate

Concen: 21.336 ng/ul

RT: 13.793 min Scan# 1837

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:163 Resp: 2186656

Ion Ratio Lower Upper

163 100

194 4.2 3.2 4.8

164 10.4 7.8 11.6



#48

2,6-Dinitrotoluene

Concen: 21.974 ng/ul

RT: 13.910 min Scan# 1857

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion:165 Resp: 429198


Ion Ratio Lower Upper

165 100

89 59.2 47.0 70.6

121 21.9 18.7 28.1





#50

Acenaphthylene

Concen: 20.372 ng/ul

RT: 14.051 min Scan# 1881

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:152 Resp: 2960362

Ion Ratio Lower Upper

152 100

151 20.2 15.6 23.4

153 13.2 10.8 16.2



#51

3-Nitroaniline

Concen: 21.597 ng/ul

RT: 14.234 min Scan# 1912

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion:138 Resp: 478814


Ion Ratio Lower Upper

138 100

108 12.2 10.7 16.1

92 119.1 94.6 141.8





#52

Acenaphthene

Concen: 20.693 ng/ul

RT: 14.393 min Scan# 1939

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:153 Resp: 1948276

Ion Ratio Lower Upper

153 100

152 47.2 40.5 60.7

154 82.3 76.6 114.8



#53

2,4-Dinitrophenol

Concen: 17.599 ng/ul

RT: 14.440 min Scan# 1947

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion:184 Resp: 157850


Ion Ratio Lower Upper

184 100

63 63.0 50.7 76.1

154 61.7 49.1 73.7









#55
4-Nitrophenol
Concen: 21.698 ng/ul
RT: 14.540 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:109 Resp: 310693
Ion Ratio Lower Upper
109 100
139 136.4 113.8 170.6
65 129.0 101.1 151.7



#56
Dibenzofuran
Concen: 20.490 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:168 Resp: 2575460
Ion Ratio Lower Upper
168 100
139 37.4 29.2 43.8





#57

2,4-Dinitrotoluene

Concen: 23.561 ng/ul

RT: 14.693 min Scan# 1990

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:165 Resp: 616530

Ion Ratio Lower Upper

165 100

63 41.4 35.1 52.7

182 2.9 2.4 3.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.350 ng/ul

RT: 14.951 min Scan# 2034

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Tgt Ion:232 Resp: 467941




Ion Ratio Lower Upper


232 100

131 45.5 39.2 58.8

130 2.3 2.1 3.1

166 31.1 26.3 39.5









#59
 Diethylphthalate
 Concen: 22.692 ng/ul
 RT: 15.163 min Scan# 2070
 Delta R.T. 0.000 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020258

Tgt Ion:149 Resp: 2191369




Ion	Ratio	Lower	Upper
149	100		
177	21.5	16.3	24.5
150	12.4	10.0	15.0




#61
 Fluorene
 Concen: 21.319 ng/ul
 RT: 15.375 min Scan# 2106
 Delta R.T. -0.006 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Tgt Ion:166 Resp: 2099016

Ion	Ratio	Lower	Upper
166	100		
165	99.1	77.0	115.6
167	13.4	10.9	16.3









#62
4-Chlorophenyl-phenylether
Concen: 20.971 ng/ul
RT: 15.381 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:204 Resp: 994025
Ion Ratio Lower Upper
204 100
206 32.5 25.8 38.8
141 57.7 46.5 69.7



#63
4-Nitroaniline
Concen: 23.134 ng/ul
RT: 15.398 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:138 Resp: 459986
Ion Ratio Lower Upper
138 100
92 53.9 40.7 61.1
108 79.1 57.1 85.7









#66
4,6-Dinitro-2-methylphenol
Concen: 18.798 ng/ul
RT: 15.451 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:198 Resp: 273921
Ion Ratio Lower Upper
198 100
182 4.5 3.9 5.9
77 30.5 24.2 36.2



#67
N-Nitrosodiphenylamine
Concen: 19.609 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:169 Resp: 1774030
Ion Ratio Lower Upper
169 100
168 68.4 52.5 78.7
167 35.6 29.8 44.8









#68
4-Bromophenyl-phenylether
Concen: 19.905 ng/ul
RT: 16.269 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:248 Resp: 615930
Ion Ratio Lower Upper
248 100
250 85.8 79.6 119.4
141 74.5 61.4 92.0



#69
Hexachlorobenzene
Concen: 20.120 ng/ul
RT: 16.369 min Scan# 2275
Delta R.T. -0.006 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:284 Resp: 671954
Ion Ratio Lower Upper
284 100
142 41.2 33.0 49.4
249 31.7 26.4 39.6









#70
 Atrazine
 Concen: 20.771 ng/ul
 RT: 16.545 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48




Instrument :
 BNA_N
 ClientSampleId :
 SST020258


Tgt Ion:	200	Resp:	505759
Ion Ratio	Lower	Upper	
200	100		
173	27.7	22.1	33.1
215	49.3	39.4	59.0



#71
 Pentachlorophenol
 Concen: 20.241 ng/ul
 RT: 16.716 min Scan# 2334
 Delta R.T. -0.006 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Tgt Ion:	266	Resp:	385567
Ion Ratio	Lower	Upper	
266	100		
264	60.8	45.8	68.6
268	61.1	47.4	71.0









#72
Phenanthrene
Concen: 20.607 ng/ul
RT: 17.116 min Scan# 2402
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:178 Resp: 3214043
Ion Ratio Lower Upper
178 100
179 15.4 12.6 18.8
176 19.3 15.2 22.8



#74
Anthracene
Concen: 20.716 ng/ul
RT: 17.204 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:178 Resp: 3241754
Ion Ratio Lower Upper
178 100
179 15.7 11.9 17.9
176 19.6 15.4 23.2









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 17.470 ng/uL
 RT: 13.116 min Scan# 1722
 Delta R.T. -0.006 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Instrument :
 BNA_N
 ClientSampleId :
 SST020258

Tgt Ion:216 Resp: 853510




Ion	Ratio	Lower	Upper
216	100		
214	79.6	61.7	92.5
218	45.4	39.0	58.6




#76
 Pentachlorobenzene
 Concen: 18.247 ng/uL
 RT: 14.640 min Scan# 1981
 Delta R.T. -0.006 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Tgt Ion:250 Resp: 807239

Ion	Ratio	Lower	Upper
250	100		
248	66.3	47.3	70.9
252	65.4	51.0	76.4









#77
 Carbazole
 Concen: 21.471 ng/ul
 RT: 17.475 min Scan# 2463
 Delta R.T. 0.000 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020258

Tgt Ion:167 Resp: 2867139




Ion	Ratio	Lower	Upper
167	100		
166	21.6	16.7	25.1
139	13.0	10.5	15.7




#78
 Di-n-butylphthalate
 Concen: 22.884 ng/ul
 RT: 18.051 min Scan# 2561
 Delta R.T. 0.000 min
 Lab File: BN031623.D
 Acq: 23 May 2024 08:48

Tgt Ion:149 Resp: 3508792

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.0	10.6
104	5.0	4.2	6.2









#80
Fluoranthene
Concen: 20.345 ng/ul
RT: 19.133 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48




Instrument :
BNA_N
ClientSampleId :
SSTD020258


Tgt Ion:202 Resp: 3319156
Ion Ratio Lower Upper
202 100
101 13.4 10.1 15.1
100 10.0 8.0 12.0



#82
Pyrene
Concen: 20.453 ng/ul
RT: 19.498 min Scan# 2807
Delta R.T. 0.000 min
Lab File: BN031623.D
Acq: 23 May 2024 08:48

Tgt Ion:202 Resp: 3421948
Ion Ratio Lower Upper
202 100
101 14.7 11.4 17.2
100 12.0 9.3 13.9





#83

Butylbenzylphthalate

Concen: 22.961 ng/ul

RT: 20.410 min Scan# 2962

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:149 Resp: 1303732

Ion Ratio Lower Upper

149 100

91 72.0 60.8 91.2

206 20.3 15.8 23.8



#84

3,3'-Dichlorobenzidine

Concen: 19.719 ng/ul

RT: 21.216 min Scan# 3099

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion:252 Resp: 797821


Ion Ratio Lower Upper

252 100

254 63.3 51.4 77.0

126 12.4 10.6 15.8





#85

Benzo(a)anthracene

Concen: 20.189 ng/ul

RT: 21.286 min Scan# 3111

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:228 Resp: 2856714

Ion Ratio Lower Upper

228 100

229 20.3 15.7 23.5

226 27.5 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.091 ng/ul

RT: 21.222 min Scan# 3100

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion:149 Resp: 1795285


Ion Ratio Lower Upper

149 100

167 27.7 21.6 32.4

279 3.8 2.8 4.2





#87

Chrysene

Concen: 20.065 ng/ul

RT: 21.345 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:228 Resp: 2622541

Ion Ratio Lower Upper

228 100

226 29.5 22.9 34.3

229 19.7 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 23.270 ng/ul




RT: 22.339 min Scan# 3290


Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Tgt Ion:149 Resp: 2518970





#90

Benzo(b)fluoranthene

Concen: 20.395 ng/ul

RT: 23.310 min Scan# 3455

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:252 Resp: 2254080

Ion Ratio Lower Upper

252 100

253 21.2 16.8 25.2

125 11.4 9.3 13.9



#91

Benzo(k)fluoranthene

Concen: 20.461 ng/ul

RT: 23.369 min Scan# 3465

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion:252 Resp: 2305151


Ion Ratio Lower Upper

252 100

253 22.8 17.4 26.0

125 12.4 9.3 13.9





#93

Benzo(a)pyrene

Concen: 20.468 ng/u1

RT: 24.104 min Scan# 3

Delta R.T. 0.000 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:252 Resp: 2172975

Ion Ratio Lower Upper

252 100

253 20.5 16.6 24.8

125 12.3 9.8 14.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.770 ng/u1

RT: 27.527 min Scan# 4172

Delta R.T. -0.006 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48




Tgt Ion:276 Resp: 2136786


Ion Ratio Lower Upper

276 100

138 24.9 19.8 29.6

277 23.8 18.2 27.4





#95

Dibenzo(a,h)anthracene

Concen: 19.601 ng/ul

RT: 27.598 min Scan# 4184

Delta R.T. 0.006 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Instrument :

BNA_N

ClientSampleId :

SSTD020258





Tgt Ion:278 Resp: 2056721

Ion Ratio Lower Upper

278 100

139 18.7 15.2 22.8

279 24.9 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 19.588 ng/ul

RT: 28.562 min Scan# 4348

Delta R.T. 0.006 min

Lab File: BN031623.D

Acq: 23 May 2024 08:48

Tgt Ion:276 Resp: 2062086

Ion Ratio Lower Upper

276 100

138 25.8 19.5 29.3

277 22.3 17.6 26.4

