




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052423\
Data File : BN025580.D
Acq On : 24 May 2023 09:02
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 25 07:13:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 25 07:13:42 2023
Response via : Initial Calibration



TIC: BN025580.D\data.ms

(70) Pentachlorophenol (C)




17.080min (0.000) 18301.18 ng

response 393240

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	67.06
264.00	61.60	67.40
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052423\
 Data File : BN025580.D
 Acq On : 24 May 2023 09:02
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 25 07:13:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 25 07:13:42 2023
 Response via : Initial Calibration



TIC: BN025580.D\data.ms

(77) Benzidine**19.674min (0.000) 665829.33 ng****response 3242861**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	12.89
183.00	13.20	11.52
0.00	0.00	0.00