

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052523\
 Data File : BN025596.D
 Acq On : 25 May 2023 21:10
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: May 26 00:05:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 26 00:02:08 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.048	152	6397	0.400	ng	0.00
7) Naphthalene-d8	10.867	136	15181	0.400	ng	0.00
13) Acenaphthene-d10	14.673	164	7815	0.400	ng	-0.01
19) Phenanthrene-d10	17.422	188	15020	0.400	ng	0.00
29) Chrysene-d12	21.607	240	10956	0.400	ng	# 0.00
35) Perylene-d12	24.115	264	10461	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.585	112	77873	4.731	ng	0.00
5) Phenol-d6	7.196	99	94999	4.581	ng	0.00
8) Nitrobenzene-d5	9.212	82	71041	5.083	ng	0.00
11) 2-Methylnaphthalene-d10	12.445	152	111769	5.018	ng	0.00
14) 2,4,6-Tribromophenol	16.169	330	10844	3.321	ng	-0.01
15) 2-Fluorobiphenyl	13.304	172	182268	5.379	ng	-0.01
27) Fluoranthene-d10	19.443	212	181400	5.096	ng	0.00
31) Terphenyl-d14	20.031	244	112432	4.956	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.426	88	35365	4.777	ng	99
3) n-Nitrosodimethylamine	3.751	42	42689	4.861	ng	# 98
6) bis(2-Chloroethyl)ether	7.463	93	87310	4.593	ng	# 80
9) Naphthalene	10.910	128	216701	5.148	ng	99
10) Hexachlorobutadiene	11.209	225	38242	5.207	ng	# 99
12) 2-Methylnaphthalene	12.517	142	146660	4.925	ng	99
16) Acenaphthylene	14.395	152	210884	5.499	ng	100
17) Acenaphthene	14.737	154	130191	4.921	ng	99
18) Fluorene	15.722	166	158868	4.808	ng	98
20) 4,6-Dinitro-2-methylph...	15.796	198	14892	3.747	ng	# 12
21) 4-Bromophenyl-phenylether	16.603	248	48861	5.492	ng	# 73
22) Hexachlorobenzene	16.727	284	43301	5.459	ng	98
23) Atrazine	16.876	200	36951	4.715	ng	# 94
25) Phenanthrene	17.460	178	247467	5.123	ng	100
26) Anthracene	17.547	178	236831	5.396	ng	100
28) Fluoranthene	19.473	202	266759	5.000	ng	99
30) Pyrene	19.835	202	269007	5.437	ng	100
32) Benzo(a)anthracene	21.589	228	229518	5.295	ng	98
33) Chrysene	21.643	228	231407	5.145	ng	99
34) Bis(2-ethylhexyl)phtha...	21.500	149	137574	5.105	ng	99
36) Indeno(1,2,3-cd)pyrene	26.731	276	256659	5.372	ng	98
37) Benzo(b)fluoranthene	23.343	252	222169	5.276	ng	# 85
38) Benzo(k)fluoranthene	23.395	252	226766	5.291	ng	# 85
39) Benzo(a)pyrene	23.992	252	214286	5.531	ng	# 80
40) Dibenzo(a,h)anthracene	26.749	278	207624	5.352	ng	# 89
41) Benzo(g,h,i)perylene	27.538	276	217752	5.336	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052523\
Data File : BN025596.D
Acq On : 25 May 2023 21:10
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Quant Time: May 26 00:05:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 26 00:02:08 2023
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 8.048 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN025596.D
 Acq: 25 May 2023 21:10

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0







Tgt Ion:152 Resp: 6397
 Ion Ratio Lower Upper
 152 100
 150 155.5 124.6 187.0
 115 60.8 49.2 73.8



#2
 1,4-Dioxane
 Concen: 4.777 ng
 RT: 3.426 min Scan# 34
 Delta R.T. -0.007 min
 Lab File: BN025596.D
 Acq: 25 May 2023 21:10

Tgt Ion: 88 Resp: 35365
 Ion Ratio Lower Upper
 88 100
 43 29.8 23.5 35.3
 58 69.8 56.7 85.1









#3
n-Nitrosodimethylamine
Concen: 4.861 ng
RT: 3.751 min Scan# 79
Delta R.T. -0.022 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 42689
Ion Ratio Lower Upper
42 100
74 121.3 97.6 146.4
44 5.5 8.3 12.5#



#4
2-Fluorophenol
Concen: 4.731 ng
RT: 5.585 min Scan# 333
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion: 112 Resp: 77873
Ion Ratio Lower Upper
112 100
64 56.9 45.3 67.9
63 31.4 25.1 37.7









#5
Phenol-d6
Concen: 4.581 ng
RT: 7.196 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 94999
Ion Ratio Lower Upper
99 100
42 20.8 16.9 25.3
71 36.1 29.6 44.4



#6
bis(2-Chloroethyl)ether
Concen: 4.593 ng
RT: 7.463 min Scan# 593
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion: 93 Resp: 87310
Ion Ratio Lower Upper
93 100
63 71.2 43.4 65.2#
95 31.3 19.5 29.3#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.867 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	15181
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.1	13.7
54	7.7	6.8	10.2
68	5.9	5.4	8.2



#8
Nitrobenzene-d5
Concen: 5.083 ng
RT: 9.212 min Scan# 795
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:	82	Resp:	71041
Ion Ratio	Lower	Upper	
82	100		
128	40.8	30.5	45.7
54	46.9	44.1	66.1









#9
Naphthalene
Concen: 5.148 ng
RT: 10.910 min Scan# 910
Delta R.T. -0.011 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 216701
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.3 10.3 15.5



#10
Hexachlorobutadiene
Concen: 5.207 ng
RT: 11.209 min Scan# 982
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:225 Resp: 38242
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.4 50.0 75.0









#11
2-Methylnaphthalene-d10
Concen: 5.018 ng
RT: 12.445 min Scan# 11
Delta R.T. -0.004 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 111769
Ion Ratio Lower Upper
152 100
151 20.5 16.1 24.1



#12
2-Methylnaphthalene
Concen: 4.925 ng
RT: 12.517 min Scan# 1277
Delta R.T. -0.008 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:142 Resp: 146660
Ion Ratio Lower Upper
142 100
141 89.4 70.2 105.4
115 39.1 31.1 46.7









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.673 min Scan# 1562
Delta R.T. -0.011 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 7815
Ion Ratio Lower Upper
164 100
162 106.2 82.7 124.1
160 51.6 38.8 58.2



#14
2,4,6-Tribromophenol
Concen: 3.321 ng
RT: 16.169 min Scan# 1696
Delta R.T. -0.013 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:330 Resp: 10844
Ion Ratio Lower Upper
330 100
332 96.2 78.5 117.7
141 56.2 50.9 76.3









#15
2-Fluorobiphenyl
Concen: 5.379 ng
RT: 13.304 min Scan# 1435
Delta R.T. -0.011 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 182268
Ion Ratio Lower Upper
172 100
171 34.9 27.2 40.8
170 23.9 17.9 26.9



#16
Acenaphthylene
Concen: 5.499 ng
RT: 14.395 min Scan# 1536
Delta R.T. -0.011 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:152 Resp: 210884
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 4.921 ng
RT: 14.737 min Scan# 1568
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 130191
Ion Ratio Lower Upper
154 100
153 113.6 91.4 137.0
152 55.4 45.0 67.4



#18
Fluorene
Concen: 4.808 ng
RT: 15.722 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:166 Resp: 158868
Ion Ratio Lower Upper
166 100
165 98.5 81.0 121.4
167 13.3 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.422 min Scan# 1797
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 15020
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.8 7.5 11.3



#20
4,6-Dinitro-2-methylphenol
Concen: 3.747 ng
RT: 15.796 min Scan# 1666
Delta R.T. -0.037 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:198 Resp: 14892
Ion Ratio Lower Upper
198 100
51 36.6 139.5 209.3#
105 36.7 91.2 136.8#









#21
4-Bromophenyl-phenylether
Concen: 5.492 ng
RT: 16.603 min Scan# 1731
Delta R.T. -0.013 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:248 Resp: 48861
Ion Ratio Lower Upper
248 100
250 94.0 84.7 127.1
141 86.7 36.5 54.7#



#22
Hexachlorobenzene
Concen: 5.459 ng
RT: 16.727 min Scan# 1741
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:284 Resp: 43301
Ion Ratio Lower Upper
284 100
142 49.6 41.2 61.8
249 34.7 28.7 43.1









#23
 Atrazine
 Concen: 4.715 ng
 RT: 16.876 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN025596.D
 Acq: 25 May 2023 21:10




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:200 Resp: 36951
 Ion Ratio Lower Upper
 200 100
 173 22.1 23.3 34.9#
 215 45.4 37.5 56.3



#25
 Phenanthrene
 Concen: 5.123 ng
 RT: 17.460 min Scan# 1800
 Delta R.T. -0.000 min
 Lab File: BN025596.D
 Acq: 25 May 2023 21:10

Tgt Ion:178 Resp: 247467
 Ion Ratio Lower Upper
 178 100
 176 19.3 15.4 23.2
 179 15.2 12.3 18.5





#26

Anthracene

Concen: 5.396 ng

RT: 17.547 min Scan# 1807

Delta R.T. -0.013 min

Lab File: BN025596.D

Acq: 25 May 2023 21:10

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:178 Resp: 236831

Ion Ratio Lower Upper

178 100

176 18.7 14.9 22.3

179 15.2 12.0 18.0



#27

Fluoranthene-d10

Concen: 5.096 ng

RT: 19.443 min Scan# 2185

Delta R.T. -0.000 min

Lab File: BN025596.D

Acq: 25 May 2023 21:10




Tgt Ion:212 Resp: 181400


Ion Ratio Lower Upper

212 100

106 15.5 12.2 18.4

104 8.9 7.0 10.6









#28
Fluoranthene
Concen: 5.000 ng
RT: 19.473 min Scan# 2193
Delta R.T. -0.004 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 266759
Ion Ratio Lower Upper
202 100
101 11.7 9.1 13.7
203 17.2 13.6 20.4



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.607 min Scan# 2497
Delta R.T. 0.009 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:240 Resp: 10956
Ion Ratio Lower Upper
240 100
120 13.8 14.2 21.2#
236 28.6 23.8 35.6









#30
 Pyrene
 Concen: 5.437 ng
 RT: 19.835 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN025596.D
 Acq: 25 May 2023 21:10




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:202 Resp: 269007
 Ion Ratio Lower Upper
 202 100
 200 21.1 16.8 25.2
 203 17.8 13.9 20.9



#31
 Terphenyl-d14
 Concen: 4.956 ng
 RT: 20.031 min Scan# 2321
 Delta R.T. -0.000 min
 Lab File: BN025596.D
 Acq: 25 May 2023 21:10

Tgt Ion:244 Resp: 112432
 Ion Ratio Lower Upper
 244 100
 212 9.8 8.6 12.8
 122 16.0 14.2 21.2









#32
Benzo(a)anthracene
Concen: 5.295 ng
RT: 21.589 min Scan# 2495
Delta R.T. 0.009 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 229518
Ion Ratio Lower Upper
228 100
226 26.9 23.0 34.4
229 19.6 15.7 23.5



#33
Chrysene
Concen: 5.145 ng
RT: 21.643 min Scan# 2501
Delta R.T. -0.000 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:228 Resp: 231407
Ion Ratio Lower Upper
228 100
226 30.6 24.6 37.0
229 19.6 16.0 24.0









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 5.105 ng
 RT: 21.500 min Scan# 2485
 Delta R.T. -0.000 min
 Lab File: BN025596.D
 Acq: 25 May 2023 21:10




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:149 Resp: 137574
 Ion Ratio Lower Upper
 149 100
 167 26.9 21.2 31.8
 279 2.7 2.4 3.6



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 24.115 min Scan# 3171
 Delta R.T. 0.006 min
 Lab File: BN025596.D
 Acq: 25 May 2023 21:10

Tgt Ion:264 Resp: 10461
 Ion Ratio Lower Upper
 264 100
 260 27.9 22.2 33.4
 265 68.2 52.7 79.1









#36
Indeno(1,2,3-cd)pyrene
Concen: 5.372 ng
RT: 26.731 min Scan# 4066
Delta R.T. -0.006 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 256659
Ion Ratio Lower Upper
276 100
138 30.7 23.7 35.5
277 24.8 19.0 28.4



#37
Benzo(b)fluoranthene
Concen: 5.276 ng
RT: 23.343 min Scan# 2907
Delta R.T. 0.003 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:252 Resp: 222169
Ion Ratio Lower Upper
252 100
253 22.5 25.8 38.6#
125 12.2 13.1 19.7#









#38
Benzo(k)fluoranthene
Concen: 5.291 ng
RT: 23.395 min Scan# 2924
Delta R.T. 0.003 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 226766
Ion Ratio Lower Upper
252 100
253 22.0 25.0 37.6#
125 11.7 12.7 19.1#



#39
Benzo(a)pyrene
Concen: 5.531 ng
RT: 23.992 min Scan# 3129
Delta R.T. -0.006 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:252 Resp: 214286
Ion Ratio Lower Upper
252 100
253 22.4 28.6 43.0#
125 13.8 16.9 25.3#









#40
Dibenzo(a,h)anthracene
Concen: 5.352 ng
RT: 26.749 min Scan# 4072
Delta R.T. -0.009 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:278 Resp: 207624
Ion Ratio Lower Upper
278 100
139 20.1 18.4 27.6
279 23.6 25.8 38.6#



#41
Benzo(g,h,i)perylene
Concen: 5.336 ng
RT: 27.538 min Scan# 4342
Delta R.T. -0.003 min
Lab File: BN025596.D
Acq: 25 May 2023 21:10

Tgt Ion:276 Resp: 217752
Ion Ratio Lower Upper
276 100
277 23.1 20.0 30.0
138 26.1 22.8 34.2

