

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052622\
 Data File : BN019972.D
 Acq On : 26 May 2022 14:17
 Operator : CG/JU
 Sample : N2978-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-201-20220519

Quant Time: May 27 03:07:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 18 07:42:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3768	0.400	ng	0.00
7) Naphthalene-d8	10.637	136	13080	0.400	ng	# 0.00
13) Acenaphthene-d10	14.463	164	7966	0.400	ng	0.00
19) Phenanthrene-d10	17.205	188	16073	0.400	ng	# 0.00
29) Chrysene-d12	21.395	240	9912	0.400	ng	0.00
35) Perylene-d12	23.738	264	8224	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.435	112	1492	0.158	ng	0.00
5) Phenol-d6	7.024	99	1034	0.088	ng	0.01
8) Nitrobenzene-d5	9.003	82	2876	0.291	ng	0.01
11) 2-Methylnaphthalene-d10	12.227	152	7187	0.316	ng	0.00
14) 2,4,6-Tribromophenol	15.954	330	849	0.270	ng	0.00
15) 2-Fluorobiphenyl	13.095	172	10073	0.290	ng	0.00
27) Fluoranthene-d10	19.240	212	16137	0.353	ng	0.00
31) Terphenyl-d14	19.839	244	12065	0.516	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.315	149	2010	0.134	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052622\
Data File : BN019972.D
Acq On : 26 May 2022 14:17
Operator : CG/JU
Sample : N2978-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-201-20220519

Quant Time: May 27 03:07:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 18 07:42:52 2022
Response via : Initial Calibration

