




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052622\
Data File : BN019965.D
Acq On : 26 May 2022 08:54
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 27 07:44:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019965.D\data.ms

(77) Benzidine




19.451min (-0.015) 0.00 ng

response 1806366

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.83
183.00	13.20	11.52
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052622\
Data File : BN019965.D
Acq On : 26 May 2022 08:54
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 27 07:44:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019965.D\data.ms

(70) Pentachlorophenol (C)

16.863min (-0.009) 21084.04 ng

response 317853

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.04
264.00	61.60	57.09
0.00	0.00	0.00