




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052622\
Data File : BN019982.D
Acq On : 26 May 2022 20:53
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 27 07:46:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019982.D\data.ms

(70) Pentachlorophenol (C)




16.863min (-0.009) 19489.97 ng

response 192567

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.89
264.00	61.60	63.74
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052622\
Data File : BN019982.D
Acq On : 26 May 2022 20:53
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 27 07:46:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019982.D\data.ms

(77) Benzidine

19.451min (-0.015) 0.00 ng

response 1214955

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.48
183.00	13.20	11.92
0.00	0.00	0.00