

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052622\
 Data File : BN019981.D
 Acq On : 26 May 2022 19:41
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 27 03:07:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 18 07:42:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	70	0.00
2	1,4-Dioxane	0.400	0.353	11.8	63	0.00
3	n-Nitrosodimethylamine	0.400	0.358	10.5	64	0.00
4 S	2-Fluorophenol	0.400	0.426	-6.5	75	0.00
5 S	Phenol-d6	0.400	0.434	-8.5	79	0.00
6	bis(2-Chloroethyl)ether	0.400	0.363	9.3	64	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	79	0.00
8 S	Nitrobenzene-d5	0.400	0.390	2.5	83	0.00
9	Naphthalene	0.400	0.371	7.3	76	0.00
10	Hexachlorobutadiene	0.400	0.376	6.0	76	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.375	6.3	77	0.00
12	2-Methylnaphthalene	0.400	0.372	7.0	76	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	76	0.00
14 S	2,4,6-Tribromophenol	0.400	0.295	26.3#	64	0.00
15 S	2-Fluorobiphenyl	0.400	0.376	6.0	73	0.00
16	Acenaphthylene	0.400	0.389	2.8	82	0.00
17	Acenaphthene	0.400	0.357	10.8	71	0.00
18	Fluorene	0.400	0.354	11.5	70	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	66	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.088	78.0#	16	0.01
21	4-Bromophenyl-phenylether	0.400	0.392	2.0	68	0.00
22	Hexachlorobenzene	0.400	0.405	-1.3	70	0.00
23	Atrazine	0.400	0.386	3.5	72	0.00
24	Pentachlorophenol	0.400	0.171	57.3#	35	0.00
25	Phenanthrene	0.400	0.365	8.8	63	0.00
26	Anthracene	0.400	0.358	10.5	64	0.00
27 SURR	Fluoranthene-d10	0.400	0.346	13.5	61	0.00
28	Fluoranthene	0.400	0.338	15.5	59	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	48	0.00
30	Pyrene	0.400	0.469	-17.2	58	0.00
31 S	Terphenyl-d14	0.400	0.469	-17.2	58	0.00
32	Benzo(a)anthracene	0.400	0.372	7.0	47	0.00
33	Chrysene	0.400	0.370	7.5	45	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.515	-28.7#	70	0.00
35 I	Perylene-d12	0.400	0.400	0.0	50	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.361	9.8	48	0.00
37	Benzo(b)fluoranthene	0.400	0.356	11.0	46	0.00
38	Benzo(k)fluoranthene	0.400	0.345	13.8	45	0.00
39 C	Benzo(a)pyrene	0.400	0.372	7.0	49	0.00
40	Dibenzo(a,h)anthracene	0.400	0.350	12.5	47	0.00
41	Benzo(g,h,i)perylene	0.400	0.357	10.8	48	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4