

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052722\
 Data File : BN019994.D
 Acq On : 27 May 2022 19:08
 Operator : CG/JU
 Sample : N3065-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-100-20220525

Quant Time: May 28 04:38:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 18 07:42:52 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Christian Giraldo 05/28/2022
 Supervised By :Jagrut Upadhyay 05/31/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3341	0.400	ng	0.00
7) Naphthalene-d8	10.637	136	11552	0.400	ng	# 0.00
13) Acenaphthene-d10	14.474	164	6971	0.400	ng	0.01
19) Phenanthrene-d10	17.215	188	14111	0.400	ng	0.01
29) Chrysene-d12	21.395	240	8609	0.400	ng	0.00
35) Perylene-d12	23.744	264	6834	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.442	112	1024	0.123	ng	0.00
5) Phenol-d6	7.024	99	637	0.061	ng	0.01
8) Nitrobenzene-d5	9.003	82	1993m	0.228	ng	0.01
11) 2-Methylnaphthalene-d10	12.227	152	5340	0.265	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	345	0.125	ng	0.01
15) 2-Fluorobiphenyl	13.095	172	7456	0.245	ng	0.00
27) Fluoranthene-d10	19.244	212	11581	0.289	ng	0.00
31) Terphenyl-d14	19.843	244	8958	0.441	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.348	88	13761m	3.470	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.315	149	947	0.072	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052722\
Data File : BN019994.D
Acq On : 27 May 2022 19:08
Operator : CG/JU
Sample : N3065-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-100-20220525

Quant Time: May 28 04:38:06 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 18 07:42:52 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 05/28/2022
Supervised By :Jagrut Upadhyay 05/31/2022

