

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052722\
 Data File : BN019995.D
 Acq On : 27 May 2022 19:45
 Operator : CG/JU
 Sample : N3065-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-201-20220525

Quant Time: May 28 04:38:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 18 07:42:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3681	0.400	ng	0.00
7) Naphthalene-d8	10.637	136	12716	0.400	ng	# 0.00
13) Acenaphthene-d10	14.474	164	7600	0.400	ng	0.01
19) Phenanthrene-d10	17.215	188	15358	0.400	ng	0.01
29) Chrysene-d12	21.395	240	8590	0.400	ng	0.00
35) Perylene-d12	23.741	264	6823	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.442	112	1176	0.128	ng	0.00
5) Phenol-d6	7.024	99	699	0.061	ng	0.01
8) Nitrobenzene-d5	9.003	82	2532	0.264	ng	0.01
11) 2-Methylnaphthalene-d10	12.227	152	6527	0.295	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	484	0.161	ng	0.01
15) 2-Fluorobiphenyl	13.105	172	9022	0.272	ng	0.01
27) Fluoranthene-d10	19.244	212	13403	0.307	ng	0.00
31) Terphenyl-d14	19.839	244	9053	0.447	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.315	149	1657	0.127	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052722\
Data File : BN019995.D
Acq On : 27 May 2022 19:45
Operator : CG/JU
Sample : N3065-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-201-20220525

Quant Time: May 28 04:38:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 18 07:42:52 2022
Response via : Initial Calibration

