

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN053123\
 Data File : BN025743.D
 Acq On : 31 May 2023 16:56
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN053123

Quant Time: Jun 01 01:06:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 01 01:04:04 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2	1,4-Dioxane	0.400	0.457	-14.2	113	0.00
3	n-Nitrosodimethylamine	0.400	0.392	2.0	99	0.00
4 S	2-Fluorophenol	0.400	0.416	-4.0	103	0.00
5 S	Phenol-d6	0.400	0.409	-2.2	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.428	-7.0	104	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	104	0.00
8 S	Nitrobenzene-d5	0.400	0.400	0.0	102	0.01
9	Naphthalene	0.400	0.413	-3.2	101	0.00
10	Hexachlorobutadiene	0.400	0.439	-9.7	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.427	-6.7	103	0.00
12	2-Methylnaphthalene	0.400	0.430	-7.5	104	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.400	0.353	11.8	91	0.00
15 S	2-Fluorobiphenyl	0.400	0.421	-5.2	103	0.00
16	Acenaphthylene	0.400	0.405	-1.3	100	0.00
17	Acenaphthene	0.400	0.419	-4.7	103	0.00
18	Fluorene	0.400	0.424	-6.0	100	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.362	9.5	88	0.01
21	4-Bromophenyl-phenylether	0.400	0.430	-7.5	99	0.00
22	Hexachlorobenzene	0.400	0.438	-9.5	101	0.00
23	Atrazine	0.400	0.399	0.3	89	0.00
24	Pentachlorophenol	0.400	0.301	24.8	73	0.01
25	Phenanthrene	0.400	0.429	-7.2	97	0.00
26	Anthracene	0.400	0.415	-3.7	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.410	-2.5	89	0.00
28	Fluoranthene	0.400	0.412	-3.0	89	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	82	0.00
30	Pyrene	0.400	0.455	-13.7	88	0.00
31 S	Terphenyl-d14	0.400	0.434	-8.5	86	0.00
32	Benzo(a)anthracene	0.400	0.402	-0.5	80	0.00
33	Chrysene	0.400	0.441	-10.2	83	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.388	3.0	68	0.00
35 I	Perylene-d12	0.400	0.400	0.0	85	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.428	-7.0	91	0.00
37	Benzo(b)fluoranthene	0.400	0.413	-3.2	80	0.00
38	Benzo(k)fluoranthene	0.400	0.416	-4.0	83	0.00
39 C	Benzo(a)pyrene	0.400	0.425	-6.2	86	0.00
40	Dibenzo(a,h)anthracene	0.400	0.428	-7.0	92	0.00
41	Benzo(g,h,i)perylene	0.400	0.441	-10.2	94	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN053123