

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
 Data File : BN014815.D
 Acq On : 01 Jun 2021 15:53
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Jun 01 16:30:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 14:44:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	619	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	2103	0.40	ng	# 0.00
13) Acenaphthene-d10	14.346	164	1527	0.40	ng	0.00
19) Phenanthrene-d10	17.104	188	3477	0.40	ng	0.00
29) Chrysene-d12	21.303	240	5173	0.40	ng	0.00
35) Perylene-d12	23.564	264	4541	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.308	112	5003	3.02	ng	0.00
5) Phenol-d6	6.879	99	7011	3.09	ng	0.00
8) Nitrobenzene-d5	8.857	82	6262	3.28	ng	0.00
11) 2-Methylnaphthalene-d10	12.084	152	12869	3.03	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	3037	3.44	ng	-0.01
15) 2-Fluorobiphenyl	12.963	172	18416	3.28	ng	-0.01
27) Fluoranthene-d10	19.144	212	42007	3.05	ng	0.00
31) Terphenyl-d14	19.744	244	39324	3.21	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.263	88	2478	3.43	ng	# 93
3) n-Nitrosodimethylamine	3.593	42	830	3.66	ng	# 4
6) bis(2-Chloroethyl)ether	7.136	93	5858	3.07	ng	98
9) Naphthalene	10.543	128	18663	3.09	ng	# 89
10) Hexachlorobutadiene	10.834	225	4980	3.14	ng	# 100
12) 2-Methylnaphthalene	12.160	142	13670	3.08	ng	99
16) Acenaphthylene	14.067	152	20352	3.33	ng	98
17) Acenaphthene	14.410	154	14212	3.18	ng	96
18) Fluorene	15.399	166	18785	2.97	ng	99
20) 4,6-Dinitro-2-methylph...	15.488	198	1870	5.06	ng	# 22
21) 4-Bromophenyl-phenylether	16.301	248	8007	3.34	ng	92
22) Hexachlorobenzene	16.410	284	8943	3.42	ng	99
23) Atrazine	16.569	200	5014	3.25	ng	# 91
24) Pentachlorophenol	16.763	266	2597	4.62	ng	90
25) Phenanthrene	17.140	178	33049	3.16	ng	99
26) Anthracene	17.238	178	29206	3.32	ng	97
28) Fluoranthene	19.173	202	46140	3.17	ng	99
30) Pyrene	19.536	202	46265	3.12	ng	99
32) Benzo(a)anthracene	21.292	228	49416	3.14	ng	97
33) Chrysene	21.335	228	55784	3.22	ng	98
34) Bis(2-ethylhexyl)phtha...	21.229	149	12453	2.79	ng	96
36) Indeno(1,2,3-cd)pyrene	25.841	276	65755	3.51	ng	100
37) Benzo(b)fluoranthene	22.883	252	57657	3.46	ng	# 91
38) Benzo(k)fluoranthene	22.928	252	59058	3.37	ng	# 90
39) Benzo(a)pyrene	23.465	252	49666	3.35	ng	# 85
40) Dibenzo(a,h)anthracene	25.851	278	56029	3.60	ng	# 92
41) Benzo(g,h,i)perylene	26.529	276	58010	3.47	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
 Data File : BN014815.D
 Acq On : 01 Jun 2021 15:53
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Jun 01 16:30:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 14:44:32 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.708 min Scan# 612
 Delta R.T. 0.000 min
 Lab File: BN014815.D
 Acq: 01 Jun 2021 15:53

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC3.2

Tgt Ion: 152 Resp: 619




Ion	Ratio	Lower	Upper
152	100		
150	154.4	123.4	185.2
115	86.9	73.2	109.8




#2
 1,4-Dioxane
 Concen: 3.43 ng
 RT: 3.263 min Scan# 60
 Delta R.T. -0.008 min
 Lab File: BN014815.D
 Acq: 01 Jun 2021 15:53

Tgt Ion: 88 Resp: 2478

Ion	Ratio	Lower	Upper
88	100		
58	54.8	44.7	67.1
43	16.6	21.0	31.6









#3
n-Nitrosodimethylamine
Concen: 3.66 ng
RT: 3.593 min Scan# 101
Delta R.T. -0.040 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 830
Ion Ratio Lower Upper
42 100
74 206.7 0.0 0.0#
44 4.5 80.0 120.0#



#4
2-Fluorophenol
Concen: 3.02 ng
RT: 5.308 min Scan# 314
Delta R.T. -0.008 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion: 112 Resp: 5003
Ion Ratio Lower Upper
112 100
64 56.2 43.6 65.4
63 29.6 25.0 37.6









#5
Phenol-d6
Concen: 3.09 ng
RT: 6.879 min Scan# 509
Delta R.T. -0.008 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	99	Resp:	7011
Ion	Ratio	Lower	Upper
99	100		
42	11.4	10.2	15.2
71	44.5	32.4	48.6



#6
bis(2-Chloroethyl)ether
Concen: 3.07 ng
RT: 7.136 min Scan# 541
Delta R.T. -0.008 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:	93	Resp:	5858
Ion	Ratio	Lower	Upper
93	100		
63	60.1	47.5	71.3
95	34.0	28.5	42.7









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	2103
Ion	Ratio	Lower	Upper
136	100		
137	14.2	13.2	19.8
54	8.1	9.0	13.4#
68	8.2	8.8	13.2#



#8
Nitrobenzene-d5
Concen: 3.28 ng
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:	82	Resp:	6262
Ion	Ratio	Lower	Upper
82	100		
128	35.2	34.2	51.2
54	34.5	38.4	57.6#









#9
Naphthalene
Concen: 3.09 ng
RT: 10.543 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 18663
Ion Ratio Lower Upper
128 100
129 11.3 13.0 19.4#
127 14.0 15.0 22.6#



#10
Hexachlorobutadiene
Concen: 3.14 ng
RT: 10.834 min Scan# 886
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:225 Resp: 4980
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 50.8 76.2









#11
2-Methylnaphthalene-d10
Concen: 3.03 ng
RT: 12.084 min Scan# 1069
Delta R.T. -0.009 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 12869
Ion Ratio Lower Upper
152 100
151 21.0 16.8 25.2



#12
2-Methylnaphthalene
Concen: 3.08 ng
RT: 12.160 min Scan# 1086
Delta R.T. -0.004 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:142 Resp: 13670
Ion Ratio Lower Upper
142 100
141 92.0 72.9 109.3
115 50.5 41.8 62.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 1527
Ion Ratio Lower Upper
164 100
162 102.2 83.2 124.8
160 48.0 41.8 62.6



#14
2,4,6-Tribromophenol
Concen: 3.44 ng
RT: 15.843 min Scan# 1427
Delta R.T. -0.013 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:330 Resp: 3037
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 24.6 21.8 32.6









#15
2-Fluorobiphenyl
Concen: 3.28 ng
RT: 12.963 min Scan# 1200
Delta R.T. -0.013 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 18416
Ion Ratio Lower Upper
172 100
171 38.2 30.9 46.3
170 25.4 21.4 32.0



#16
Acenaphthylene
Concen: 3.33 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:152 Resp: 20352
Ion Ratio Lower Upper
152 100
151 20.7 17.1 25.7
153 13.1 9.9 14.9









#17
Acenaphthene
Concen: 3.18 ng
RT: 14.410 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 14212
Ion Ratio Lower Upper
154 100
153 111.3 91.8 137.8
152 54.2 46.0 69.0



#18
Fluorene
Concen: 2.97 ng
RT: 15.399 min Scan# 1392
Delta R.T. -0.013 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:166 Resp: 18785
Ion Ratio Lower Upper
166 100
165 98.9 79.9 119.9
167 13.8 10.5 15.7









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 3477
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.5 8.2 12.4



#20
4,6-Dinitro-2-methylphenol
Concen: 5.06 ng
RT: 15.488 min Scan# 1399
Delta R.T. -0.013 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:198 Resp: 1870
Ion Ratio Lower Upper
198 100
51 21.7 96.4 144.6#
105 29.2 72.3 108.5#









#21
4-Bromophenyl-phenylether
Concen: 3.34 ng
RT: 16.301 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 8007
Ion Ratio Lower Upper
248 100
250 101.7 77.5 116.3
141 53.6 50.2 75.4



#22
Hexachlorobenzene
Concen: 3.42 ng
RT: 16.410 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:284 Resp: 8943
Ion Ratio Lower Upper
284 100
142 25.6 20.3 30.5
249 32.2 26.4 39.6









#23
Atrazine
Concen: 3.25 ng
RT: 16.569 min Scan# 1486
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 5014
Ion Ratio Lower Upper
200 100
173 28.1 29.1 43.7#
215 51.8 45.1 67.7



#24
Pentachlorophenol
Concen: 4.62 ng
RT: 16.763 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:266 Resp: 2597
Ion Ratio Lower Upper
266 100
264 61.3 44.2 66.2
268 63.9 44.2 66.2









#25
Phenanthrene
Concen: 3.16 ng
RT: 17.140 min Scan# 1533
Delta R.T. -0.012 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 33049
Ion Ratio Lower Upper
178 100
176 19.3 16.0 24.0
179 15.4 12.6 19.0



#26
Anthracene
Concen: 3.32 ng
RT: 17.238 min Scan# 1541
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:178 Resp: 29206
Ion Ratio Lower Upper
178 100
176 18.9 14.2 21.2
179 15.5 13.7 20.5









#27
Fluoranthene-d10
Concen: 3.05 ng
RT: 19.144 min Scan# 1832
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 42007
Ion Ratio Lower Upper
212 100
106 9.5 7.4 11.2
104 5.5 4.1 6.1



#28
Fluoranthene
Concen: 3.17 ng
RT: 19.173 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:202 Resp: 46140
Ion Ratio Lower Upper
202 100
101 9.0 7.0 10.6
203 17.4 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 5173
Ion Ratio Lower Upper
240 100
120 7.2 5.2 7.8
236 27.4 21.6 32.4



#30
Pyrene
Concen: 3.12 ng
RT: 19.536 min Scan# 1911
Delta R.T. -0.005 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:202 Resp: 46265
Ion Ratio Lower Upper
202 100
200 21.1 16.6 24.8
203 17.6 13.9 20.9









#31
Terphenyl-d14
Concen: 3.21 ng
RT: 19.744 min Scan# 1953
Delta R.T. -0.005 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 39324
Ion Ratio Lower Upper
244 100
212 7.4 6.8 10.2
122 7.9 6.6 9.8



#32
Benzo(a)anthracene
Concen: 3.14 ng
RT: 21.292 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN014815.D
Acq: 01 Jun 2021 15:53

Tgt Ion:228 Resp: 49416
Ion Ratio Lower Upper
228 100
226 26.4 22.9 34.3
229 20.1 16.2 24.2





#33

Chrysene

Concen: 3.22 ng

RT: 21.335 min Scan# 2131

Delta R.T. -0.011 min

Lab File: BN014815.D

Acq: 01 Jun 2021 15:53





Tgt Ion:228 Resp: 55784

Ion Ratio Lower Upper

228 100

226 30.9 24.3 36.5

229 19.1 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 2.79 ng

RT: 21.229 min Scan# 2121

Delta R.T. 0.000 min

Lab File: BN014815.D

Acq: 01 Jun 2021 15:53




Tgt Ion:149 Resp: 12453


Ion Ratio Lower Upper

149 100

167 38.1 28.3 42.5

279 7.0 5.5 8.3





#35

Perylene-d12

Concen: 0.40 ng

RT: 23.564 min Scan# 2570

Delta R.T. -0.003 min

Lab File: BN014815.D

Acq: 01 Jun 2021 15:53





Tgt Ion:264 Resp: 4541

Ion Ratio Lower Upper

264 100

260 26.2 21.4 32.2

265 51.1 40.9 61.3



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.51 ng

RT: 25.841 min Scan# 3235

Delta R.T. -0.003 min

Lab File: BN014815.D

Acq: 01 Jun 2021 15:53




Tgt Ion:276 Resp: 65755


Ion Ratio Lower Upper

276 100

138 12.6 10.2 15.4

277 25.6 20.4 30.6





#37

Benzo(b)fluoranthene

Concen: 3.46 ng

RT: 22.883 min Scan# 2371

Delta R.T. -0.003 min

Lab File: BN014815.D

Acq: 01 Jun 2021 15:53

Tgt Ion:252 Resp: 57657





Ion	Ratio	Lower	Upper
252	100		
253	22.1	21.5	32.3
125	5.8	6.6	10.0#

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2



#38

Benzo(k)fluoranthene

Concen: 3.37 ng

RT: 22.928 min Scan# 2384




Delta R.T. -0.003 min


Lab File: BN014815.D

Acq: 01 Jun 2021 15:53

Tgt Ion:252 Resp: 59058

Ion	Ratio	Lower	Upper
252	100		
253	22.4	22.7	34.1#
125	5.8	6.4	9.6#





#39

Benzo(a)pyrene

Concen: 3.35 ng





RT: 23.465 min Scan# 2541

Delta R.T. 0.000 min

Lab File: BN014815.D

Acq: 01 Jun 2021 15:53

Tgt Ion:	252	Resp:	49666
Ion Ratio	Lower	Upper	
252	100		
253	22.6	25.1	37.7#
125	6.7	9.4	14.0#



#40

Dibenzo(a,h)anthracene

Concen: 3.60 ng




RT: 25.851 min Scan# 3238


Delta R.T. -0.010 min

Lab File: BN014815.D

Acq: 01 Jun 2021 15:53

Tgt Ion:	278	Resp:	56029
Ion Ratio	Lower	Upper	
278	100		
139	8.5	8.6	13.0#
279	23.6	22.8	34.2





#41
 Benzo(g,h,i)perylene
 Concen: 3.47 ng
 RT: 26.529 min Scan# 3436
 Delta R.T. -0.007 min
 Lab File: BN014815.D
 Acq: 01 Jun 2021 15:53

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	276	Resp:	58010
Ion Ratio	Lower	Upper	
276	100		
277	23.9	20.9	31.3
138	10.9	10.3	15.5

