




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
 Data File : BN014809.D
 Acq On : 01 Jun 2021 12:14
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 01 13:53:01 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 22 07:55:10 2021
 Response via : Initial Calibration






TIC: BN014809.D\data.ms

(77) Benzidine**19.362min (-0.018) 694789.99 ng****response 2548642**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.78
183.00	13.20	11.69
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
Data File : BN014809.D
Acq On : 01 Jun 2021 12:14
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 01 13:53:01 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 22 07:55:10 2021
Response via : Initial Calibration



TIC: BN014809.D\data.ms

(70) Pentachlorophenol (C)

16.757min (-0.018) 23476.90 ng

response 524902

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	77.62#
264.00	61.60	69.30
0.00	0.00	0.00