

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
 Data File : BN014817.D
 Acq On : 01 Jun 2021 17:30
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060121

Quant Time: Jun 01 18:09:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 17:40:45 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	0.00
2	1,4-Dioxane	0.400	0.343	14.2	114	0.00
3	n-Nitrosodimethylamine	0.400	0.292	27.0#	158	0.02
4 S	2-Fluorophenol	0.400	0.434	-8.5	111	0.00
5 S	Phenol-d6	0.400	0.399	0.3	111	0.00
6	bis(2-Chloroethyl)ether	0.400	0.387	3.3	120	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.369	7.8	107	0.00
9	Naphthalene	0.400	0.392	2.0	107	0.00
10	Hexachlorobutadiene	0.400	0.400	0.0	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.373	6.8	108	0.00
12	2-Methylnaphthalene	0.400	0.380	5.0	108	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	0.01
14 S	2,4,6-Tribromophenol	0.400	0.362	9.5	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.399	0.3	109	0.00
16	Acenaphthylene	0.400	0.369	7.8	106	0.00
17	Acenaphthene	0.400	0.392	2.0	107	0.01
18	Fluorene	0.400	0.385	3.8	107	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	108	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.409	-2.2	123	0.00
21	4-Bromophenyl-phenylether	0.400	0.366	8.5	107	0.01
22	Hexachlorobenzene	0.400	0.380	5.0	108	0.01
23	Atrazine	0.400	0.323	19.3	100	0.01
24	Pentachlorophenol	0.400	0.423	-5.7	95	0.00
25	Phenanthrene	0.400	0.376	6.0	109	0.00
26	Anthracene	0.400	0.346	13.5	106	0.00
27 SURR	Fluoranthene-d10	0.400	0.354	11.5	106	0.00
28	Fluoranthene	0.400	0.361	9.8	107	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	109	0.00
30	Pyrene	0.400	0.369	7.8	106	0.00
31 S	Terphenyl-d14	0.400	0.375	6.3	108	0.00
32	Benzo(a)anthracene	0.400	0.365	8.8	115	0.00
33	Chrysene	0.400	0.382	4.5	107	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.380	5.0	110	0.00
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.368	8.0	113	0.00
37	Benzo(b)fluoranthene	0.400	0.381	4.8	119	0.00
38	Benzo(k)fluoranthene	0.400	0.366	8.5	104	0.00
39 C	Benzo(a)pyrene	0.400	0.388	3.0	116	0.01
40	Dibenzo(a,h)anthracene	0.400	0.363	9.3	114	0.01
41	Benzo(g,h,i)perylene	0.400	0.366	8.5	109	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN060121