

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
 Data File : BN014812.D
 Acq On : 01 Jun 2021 14:06
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Manual Integrations
 APPROVED

mohammad
 6/2/2021 10:53:14 AM

Quant Time: Jun 01 14:46:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 14:44:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	552	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	2027	0.40	ng	0.00
13) Acenaphthene-d10	14.346	164	1449	0.40	ng	0.00
19) Phenanthrene-d10	17.104	188	3427	0.40	ng	0.00
29) Chrysene-d12	21.303	240	4496	0.40	ng	# 0.00
35) Perylene-d12	23.568	264	4497	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	628	0.43	ng	0.00
5) Phenol-d6	6.887	99	766	0.38	ng	0.00
8) Nitrobenzene-d5	8.857	82	629	0.34	ng	0.00
11) 2-Methylnaphthalene-d10	12.093	152	1391	0.34	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	277	0.33	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	1999	0.38	ng	0.00
27) Fluoranthene-d10	19.144	212	4553	0.34	ng	0.00
31) Terphenyl-d14	19.749	244	4038	0.38	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.271	88	308	0.48	ng	# 92
3) n-Nitrosodimethylamine	3.633	42	33	0.16	ng	# 88
6) bis(2-Chloroethyl)ether	7.144	93	565	0.33	ng	96
9) Naphthalene	10.543	128	2166	0.37	ng	95
10) Hexachlorobutadiene	10.834	225	581	0.38	ng	# 96
12) 2-Methylnaphthalene	12.164	142	1466	0.34	ng	96
16) Acenaphthylene	14.067	152	1988	0.34	ng	98
17) Acenaphthene	14.409	154	1533	0.36	ng	99
18) Fluorene	15.412	166	2052	0.34	ng	99
20) 4,6-Dinitro-2-methylph...	15.501	198	100	0.27	ng	90
21) 4-Bromophenyl-phenylether	16.301	248	840	0.36	ng	94
22) Hexachlorobenzene	16.410	284	1014	0.39	ng	98
23) Atrazine	16.568	200	491	0.32	ng	94
24) Pentachlorophenol	16.763	266	184	0.33	ng	# 86
25) Phenanthrene	17.153	178	3608	0.35	ng	99
26) Anthracene	17.238	178	2823	0.33	ng	97
28) Fluoranthene	19.173	202	4771	0.33	ng	99
30) Pyrene	19.541	202	4906	0.38	ng	99
32) Benzo(a)anthracene	21.292	228	4345	0.32	ng	99
33) Chrysene	21.345	228	5937	0.39	ng	99
34) Bis(2-ethylhexyl)phtha...	21.229	149	1341	0.35	ng	96
36) Indeno(1,2,3-cd)pyrene	25.844	276	6625	0.36	ng	100
37) Benzo(b)fluoranthene	22.887	252	5748	0.35	ng	99
38) Benzo(k)fluoranthene	22.931	252	6827m	0.39	ng	
39) Benzo(a)pyrene	23.465	252	5255	0.36	ng	98
40) Dibenzo(a,h)anthracene	25.861	278	5376	0.35	ng	99
41) Benzo(g,h,i)perylene	26.535	276	6269	0.38	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
Data File : BN014812.D
Acq On : 01 Jun 2021 14:06
Operator : CG/JU
Sample : SSTDICC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.4

Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:14 AM


Quant Time: Jun 01 14:46:07 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 01 14:44:32 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.708 min Scan# 612
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:152 Resp: 552
Ion Ratio Lower Upper
152 100
150 146.3 123.4 185.2
115 84.8 73.2 109.8

Manual Integrations
APPROVED



mohammad
6/2/2021 10:53:14 AM




#2
1,4-Dioxane
Concen: 0.48 ng
RT: 3.271 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Tgt Ion: 88 Resp: 308
Ion Ratio Lower Upper
88 100
58 52.9 44.7 67.1
43 18.2 21.0 31.6#









#3
n-Nitrosodimethylamine
Concen: 0.16 ng
RT: 3.633 min Scan# 106
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion: 42 Resp: 33
Ion Ratio Lower Upper
42 100
74 0.0 0.0 0.0
44 87.9 80.0 120.0

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:14 AM



#4
2-Fluorophenol
Concen: 0.43 ng
RT: 5.316 min Scan# 315
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion: 112 Resp: 628
Ion Ratio Lower Upper
112 100
64 55.6 43.6 65.4
63 30.6 25.0 37.6









#5
Phenol-d6
Concen: 0.38 ng
RT: 6.887 min Scan# 510
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:	99	Resp:	766
Ion	Ratio	Lower	Upper
99	100		
42	13.2	10.2	15.2
71	45.2	32.4	48.6

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:14 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.33 ng
RT: 7.144 min Scan# 542
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06


Tgt Ion:	93	Resp:	565
Ion	Ratio	Lower	Upper
93	100		
63	64.8	47.5	71.3
95	35.6	28.5	42.7





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:	136	Resp:	2027
Ion	Ratio	Lower	Upper
136	100		
137	14.8	13.2	19.8
54	9.1	9.0	13.4
68	9.1	8.8	13.2

Manual Integrations
APPROVED



mohammad
6/2/2021 10:53:14 AM




#8
Nitrobenzene-d5
Concen: 0.34 ng
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Tgt Ion:	82	Resp:	629
Ion	Ratio	Lower	Upper
82	100		
128	44.2	34.2	51.2
54	49.4	38.4	57.6









#9
Naphthalene
Concen: 0.37 ng
RT: 10.543 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:	128	Resp:	2166
Ion	Ratio	Lower	Upper
128	100		
129	13.6	13.0	19.4
127	16.9	15.0	22.6

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:14 AM



#10
Hexachlorobutadiene
Concen: 0.38 ng
RT: 10.834 min Scan# 886
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:	225	Resp:	581
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	66.3	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.34 ng
RT: 12.093 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:152 Resp: 1391
Ion Ratio Lower Upper
152 100
151 20.4 16.8 25.2

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:14 AM




#12
2-Methylnaphthalene
Concen: 0.34 ng
RT: 12.164 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:142 Resp: 1466
Ion Ratio Lower Upper
142 100
141 85.2 72.9 109.3
115 52.0 41.8 62.6






#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Tgt Ion:164 Resp: 1449


Ion	Ratio	Lower	Upper
164	100		
162	108.3	83.2	124.8
160	54.2	41.8	62.6




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:14 AM






#14
2,4,6-Tribromophenol
Concen: 0.33 ng
RT: 15.856 min Scan# 1428
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06




Tgt Ion:330 Resp: 277

Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	28.9	21.8	32.6









#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 12.976 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06





Tgt Ion:	172	Resp:	1999
Ion	Ratio	Lower	Upper
172	100		
171	38.6	30.9	46.3
170	25.7	21.4	32.0



#16
Acenaphthylene
Concen: 0.34 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06




Tgt Ion:	152	Resp:	1988
Ion	Ratio	Lower	Upper
152	100		
151	20.6	17.1	25.7
153	13.4	9.9	14.9



Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:14 AM



#17
Acenaphthene
Concen: 0.36 ng
RT: 14.409 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4





Tgt Ion:154 Resp: 1533


Ion	Ratio	Lower	Upper
154	100		
153	115.0	91.8	137.8
152	54.5	46.0	69.0

Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:14 AM






#18
Fluorene
Concen: 0.34 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Tgt Ion:166 Resp: 2052


Ion	Ratio	Lower	Upper
166	100		
165	100.8	79.9	119.9
167	13.5	10.5	15.7





#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:188 Resp: 3427
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.7 8.2 12.4

Manual Integrations
APPROVED



mohammad
6/2/2021 10:53:14 AM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.27 ng
RT: 15.501 min Scan# 1400
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06




Tgt Ion:198 Resp: 100
Ion Ratio Lower Upper
198 100
51 106.1 96.4 144.6
105 85.4 72.3 108.5









#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.301 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06





Tgt Ion:248 Resp: 840
Ion Ratio Lower Upper
248 100
250 99.1 77.5 116.3
141 72.9 50.2 75.4



#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.410 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06




Tgt Ion:284 Resp: 1014
Ion Ratio Lower Upper
284 100
142 26.5 20.3 30.5
249 32.0 26.4 39.6



Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:14 AM






#23
Atrazine
Concen: 0.32 ng
RT: 16.568 min Scan# 1486
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4





Tgt Ion:200 Resp: 491
Ion Ratio Lower Upper
200 100
173 42.2 29.1 43.7
215 59.2 45.1 67.7




#24
Pentachlorophenol
Concen: 0.33 ng
RT: 16.763 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06




Tgt Ion:266 Resp: 184
Ion Ratio Lower Upper
266 100
264 53.8 44.2 66.2
268 74.5 44.2 66.2#





#25
Phenanthrene
Concen: 0.35 ng
RT: 17.153 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:178 Resp: 3608
Ion Ratio Lower Upper
178 100
176 19.8 16.0 24.0
179 16.2 12.6 19.0

Manual Integrations
APPROVED



mohammad
6/2/2021 10:53:14 AM




#26
Anthracene
Concen: 0.33 ng
RT: 17.238 min Scan# 1541
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Tgt Ion:178 Resp: 2823
Ion Ratio Lower Upper
178 100
176 18.9 14.2 21.2
179 16.0 13.7 20.5









#27
Fluoranthene-d10
Concen: 0.34 ng
RT: 19.144 min Scan# 1832
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:212 Resp: 4553
Ion Ratio Lower Upper
212 100
106 9.0 7.4 11.2
104 5.4 4.1 6.1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:14 AM




#28
Fluoranthene
Concen: 0.33 ng
RT: 19.173 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:202 Resp: 4771
Ion Ratio Lower Upper
202 100
101 8.7 7.0 10.6
203 17.2 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06





Tgt Ion:240 Resp: 4496
Ion Ratio Lower Upper
240 100
120 9.8 5.2 7.8#
236 29.0 21.6 32.4



#30
Pyrene
Concen: 0.38 ng
RT: 19.541 min Scan# 1912
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06




Tgt Ion:202 Resp: 4906
Ion Ratio Lower Upper
202 100
200 21.4 16.6 24.8
203 17.8 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:14 AM



#31
Terphenyl-d14
Concen: 0.38 ng
RT: 19.749 min Scan# 1954
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:244 Resp: 4038
Ion Ratio Lower Upper
244 100
212 8.9 6.8 10.2
122 9.2 6.6 9.8

Manual Integrations
APPROVED



mohammad
6/2/2021 10:53:14 AM




#32
Benzo(a)anthracene
Concen: 0.32 ng
RT: 21.292 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Tgt Ion:228 Resp: 4345
Ion Ratio Lower Upper
228 100
226 28.2 22.9 34.3
229 20.5 16.2 24.2









#33
Chrysene
Concen: 0.39 ng
RT: 21.345 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:	228	Resp:	5937
Ion Ratio	Lower	Upper	
228	100		
226	30.3	24.3	36.5
229	21.3	16.4	24.6

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:14 AM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.35 ng
RT: 21.229 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:	149	Resp:	1341
Ion Ratio	Lower	Upper	
149	100		
167	38.0	28.3	42.5
279	7.0	5.5	8.3







#35
Perylene-d12
Concen: 0.40 ng
RT: 23.568 min Scan# 2571
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06




Tgt Ion:264 Resp: 4497
Ion Ratio Lower Upper
264 100
260 26.4 21.4 32.2
265 55.2 40.9 61.3




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Manual Integrations
APPROVED



mohammad
6/2/2021 10:53:14 AM




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.36 ng
RT: 25.844 min Scan# 3236
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06



Tgt Ion:276 Resp: 6625
Ion Ratio Lower Upper
276 100
138 12.6 10.2 15.4
277 25.6 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.35 ng
RT: 22.887 min Scan# 2372
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:252 Resp: 5748
Ion Ratio Lower Upper
252 100
253 27.3 21.5 32.3
125 9.4 6.6 10.0

Instrument :
BNA_N
ClientSampled :
SSTDICCC0.4




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:14 AM



#38
Benzo(k)fluoranthene
Concen: 0.39 ng m
RT: 22.931 min Scan# 2385
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:252 Resp: 6827
Ion Ratio Lower Upper
252 100
253 28.1 22.7 34.1
125 8.1 6.4 9.6









#39
Benzo(a)pyrene
Concen: 0.36 ng
RT: 23.465 min Scan# 2541
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:	252	Resp:	5255
Ion Ratio	Lower	Upper	
252	100		
253	32.3	25.1	37.7
125	13.1	9.4	14.0

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:14 AM



#40
Dibenzo(a,h)anthracene
Concen: 0.35 ng
RT: 25.861 min Scan# 3241
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06


Tgt Ion:	278	Resp:	5376
Ion Ratio	Lower	Upper	
278	100		
139	11.5	8.6	13.0
279	28.4	22.8	34.2





#41
Benzo(g,h,i)perylene
Concen: 0.38 ng
RT: 26.535 min Scan# 3438
Delta R.T. 0.000 min
Lab File: BN014812.D
Acq: 01 Jun 2021 14:06

Tgt Ion:	276	Resp:	6269
Ion Ratio	Lower	Upper	
276	100		
277	26.6	20.9	31.3
138	14.4	10.3	15.5



Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:14 AM