

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
 Data File : BN014816.D
 Acq On : 01 Jun 2021 16:28
 Operator : CG/JU
 Sample : SSTDIC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC5.0

Quant Time: Jun 01 16:59:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 14:44:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	655	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	2121	0.40	ng	# 0.00
13) Acenaphthene-d10	14.346	164	1535	0.40	ng	0.00
19) Phenanthrene-d10	17.104	188	3507	0.40	ng	0.00
29) Chrysene-d12	21.303	240	5130	0.40	ng	# 0.00
35) Perylene-d12	23.564	264	4466	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.308	112	7454	4.25	ng	0.00
5) Phenol-d6	6.879	99	10714	4.46	ng	0.00
8) Nitrobenzene-d5	8.857	82	9857	5.11	ng	0.00
11) 2-Methylnaphthalene-d10	12.084	152	19411	4.53	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	4886	5.51	ng	-0.01
15) 2-Fluorobiphenyl	12.963	172	27330	4.84	ng	-0.01
27) Fluoranthene-d10	19.143	212	63866	4.59	ng	0.00
31) Terphenyl-d14	19.749	244	58847	4.84	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.254	88	3758	4.92	ng	# 92
3) n-Nitrosodimethylamine	3.585	42	1448	6.03	ng	# 3
6) bis(2-Chloroethyl)ether	7.136	93	8848	4.38	ng	99
9) Naphthalene	10.543	128	28173	4.62	ng	# 88
10) Hexachlorobutadiene	10.834	225	7452	4.66	ng	# 99
12) 2-Methylnaphthalene	12.159	142	20616	4.61	ng	99
16) Acenaphthylene	14.067	152	31458	5.12	ng	98
17) Acenaphthene	14.409	154	21660	4.82	ng	95
18) Fluorene	15.399	166	28485	4.48	ng	99
20) 4,6-Dinitro-2-methylph...	15.488	198	3240	8.69	ng	# 18
21) 4-Bromophenyl-phenylether	16.301	248	12145	5.03	ng	93
22) Hexachlorobenzene	16.410	284	13314	5.05	ng	99
23) Atrazine	16.568	200	7995	5.14	ng	# 91
24) Pentachlorophenol	16.763	266	4274	7.54	ng	89
25) Phenanthrene	17.140	178	50004	4.74	ng	99
26) Anthracene	17.238	178	44654	5.03	ng	97
28) Fluoranthene	19.173	202	69304	4.72	ng	99
30) Pyrene	19.535	202	69955	4.75	ng	99
32) Benzo(a)anthracene	21.292	228	76906	4.92	ng	98
33) Chrysene	21.345	228	82860	4.82	ng	98
34) Bis(2-ethylhexyl)phtha...	21.228	149	20436	4.61	ng	95
36) Indeno(1,2,3-cd)pyrene	25.840	276	98920	5.37	ng	100
37) Benzo(b)fluoranthene	22.883	252	87033	5.31	ng	# 91
38) Benzo(k)fluoranthene	22.927	252	88603	5.14	ng	# 89
39) Benzo(a)pyrene	23.465	252	74648	5.12	ng	# 85
40) Dibenzo(a,h)anthracene	25.854	278	84559	5.52	ng	# 91
41) Benzo(g,h,i)perylene	26.532	276	85636	5.20	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
Data File : BN014816.D
Acq On : 01 Jun 2021 16:28
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Jun 01 16:59:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 01 14:44:32 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.708 min Scan# 612
Delta R.T. 0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 655
Ion Ratio Lower Upper
152 100
150 154.4 123.4 185.2
115 84.4 73.2 109.8



#2
1,4-Dioxane
Concen: 4.92 ng
RT: 3.254 min Scan# 59
Delta R.T. -0.016 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion: 88 Resp: 3758
Ion Ratio Lower Upper
88 100
58 54.3 44.7 67.1
43 16.6 21.0 31.6#









#3
n-Nitrosodimethylamine
Concen: 6.03 ng
RT: 3.585 min Scan# 100
Delta R.T. -0.049 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 1448
Ion Ratio Lower Upper
42 100
74 194.1 0.0 0.0#
44 2.7 80.0 120.0#



#4
2-Fluorophenol
Concen: 4.25 ng
RT: 5.308 min Scan# 314
Delta R.T. -0.008 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion: 112 Resp: 7454
Ion Ratio Lower Upper
112 100
64 57.2 43.6 65.4
63 30.4 25.0 37.6









#5
Phenol-d6
Concen: 4.46 ng
RT: 6.879 min Scan# 509
Delta R.T. -0.008 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	99	Resp:	10714
Ion	Ratio	Lower	Upper
99	100		
42	11.7	10.2	15.2
71	45.9	32.4	48.6



#6
bis(2-Chloroethyl)ether
Concen: 4.38 ng
RT: 7.136 min Scan# 541
Delta R.T. -0.008 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:	93	Resp:	8848
Ion	Ratio	Lower	Upper
93	100		
63	59.9	47.5	71.3
95	33.9	28.5	42.7









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	2121
Ion	Ratio	Lower	Upper
136	100		
137	14.3	13.2	19.8
54	8.7	9.0	13.4#
68	7.4	8.8	13.2#



#8
Nitrobenzene-d5
Concen: 5.11 ng
RT: 8.857 min Scan# 730
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:	82	Resp:	9857
Ion	Ratio	Lower	Upper
82	100		
128	35.6	34.2	51.2
54	33.0	38.4	57.6#









#9
Naphthalene
Concen: 4.62 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 28173
Ion Ratio Lower Upper
128 100
129 10.9 13.0 19.4#
127 13.7 15.0 22.6#



#10
Hexachlorobutadiene
Concen: 4.66 ng
RT: 10.834 min Scan# 886
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:225 Resp: 7452
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.8 76.2









#11
2-Methylnaphthalene-d10
Concen: 4.53 ng
RT: 12.084 min Scan# 1069
Delta R.T. -0.009 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 19411
Ion Ratio Lower Upper
152 100
151 21.1 16.8 25.2



#12
2-Methylnaphthalene
Concen: 4.61 ng
RT: 12.159 min Scan# 1086
Delta R.T. -0.005 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:142 Resp: 20616
Ion Ratio Lower Upper
142 100
141 90.5 72.9 109.3
115 50.3 41.8 62.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	164	Resp:	1535
Ion	Ratio	Lower	Upper
164	100		
162	107.6	83.2	124.8
160	48.8	41.8	62.6



#14
2,4,6-Tribromophenol
Concen: 5.51 ng
RT: 15.843 min Scan# 1427
Delta R.T. -0.013 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:	330	Resp:	4886
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	25.6	21.8	32.6









#15
2-Fluorobiphenyl
Concen: 4.84 ng
RT: 12.963 min Scan# 1200
Delta R.T. -0.013 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 27330
Ion Ratio Lower Upper
172 100
171 38.1 30.9 46.3
170 25.1 21.4 32.0



#16
Acenaphthylene
Concen: 5.12 ng
RT: 14.067 min Scan# 1287
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:152 Resp: 31458
Ion Ratio Lower Upper
152 100
151 20.8 17.1 25.7
153 13.2 9.9 14.9









#17
Acenaphthene
Concen: 4.82 ng
RT: 14.409 min Scan# 1314
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 21660
Ion Ratio Lower Upper
154 100
153 110.3 91.8 137.8
152 53.3 46.0 69.0



#18
Fluorene
Concen: 4.48 ng
RT: 15.399 min Scan# 1392
Delta R.T. -0.013 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:166 Resp: 28485
Ion Ratio Lower Upper
166 100
165 98.9 79.9 119.9
167 13.8 10.5 15.7









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.104 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 3507
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.1 8.2 12.4



#20
4,6-Dinitro-2-methylphenol
Concen: 8.69 ng
RT: 15.488 min Scan# 1399
Delta R.T. -0.013 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:198 Resp: 3240
Ion Ratio Lower Upper
198 100
51 17.7 96.4 144.6#
105 26.4 72.3 108.5#









#21
4-Bromophenyl-phenylether
Concen: 5.03 ng
RT: 16.301 min Scan# 1464
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	248	Resp:	12145
Ion Ratio	Lower	Upper	
248	100		
250	99.7	77.5	116.3
141	52.2	50.2	75.4



#22
Hexachlorobenzene
Concen: 5.05 ng
RT: 16.410 min Scan# 1473
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:	284	Resp:	13314
Ion Ratio	Lower	Upper	
284	100		
142	25.8	20.3	30.5
249	32.2	26.4	39.6









#23
Atrazine
Concen: 5.14 ng
RT: 16.568 min Scan# 1486
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:200 Resp: 7995
Ion Ratio Lower Upper
200 100
173 28.5 29.1 43.7#
215 51.4 45.1 67.7



#24
Pentachlorophenol
Concen: 7.54 ng
RT: 16.763 min Scan# 1502
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:266 Resp: 4274
Ion Ratio Lower Upper
266 100
264 63.3 44.2 66.2
268 62.7 44.2 66.2









#25
Phenanthrene
Concen: 4.74 ng
RT: 17.140 min Scan# 1533
Delta R.T. -0.012 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 50004
Ion Ratio Lower Upper
178 100
176 19.2 16.0 24.0
179 15.4 12.6 19.0



#26
Anthracene
Concen: 5.03 ng
RT: 17.238 min Scan# 1541
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:178 Resp: 44654
Ion Ratio Lower Upper
178 100
176 18.8 14.2 21.2
179 15.5 13.7 20.5









#27
Fluoranthene-d10
Concen: 4.59 ng
RT: 19.143 min Scan# 1832
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:212 Resp: 63866
Ion Ratio Lower Upper
212 100
106 9.4 7.4 11.2
104 5.5 4.1 6.1



#28
Fluoranthene
Concen: 4.72 ng
RT: 19.173 min Scan# 1838
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:202 Resp: 69304
Ion Ratio Lower Upper
202 100
101 9.0 7.0 10.6
203 17.5 13.5 20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.303 min Scan# 2128
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:240 Resp: 5130
Ion Ratio Lower Upper
240 100
120 8.0 5.2 7.8#
236 27.8 21.6 32.4



#30
Pyrene
Concen: 4.75 ng
RT: 19.535 min Scan# 1911
Delta R.T. -0.005 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:202 Resp: 69955
Ion Ratio Lower Upper
202 100
200 21.1 16.6 24.8
203 17.7 13.9 20.9









#31
Terphenyl-d14
Concen: 4.84 ng
RT: 19.749 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:244 Resp: 58847
Ion Ratio Lower Upper
244 100
212 7.1 6.8 10.2
122 6.7 6.6 9.8



#32
Benzo(a)anthracene
Concen: 4.92 ng
RT: 21.292 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:228 Resp: 76906
Ion Ratio Lower Upper
228 100
226 26.7 22.9 34.3
229 20.3 16.2 24.2









#33
Chrysene
Concen: 4.82 ng
RT: 21.345 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	228	Resp:	82860
Ion Ratio	Lower	Upper	
228	100		
226	28.6	24.3	36.5
229	20.1	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 4.61 ng
RT: 21.228 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:	149	Resp:	20436
Ion Ratio	Lower	Upper	
149	100		
167	38.5	28.3	42.5
279	7.1	5.5	8.3









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.564 min Scan# 2570
Delta R.T. -0.004 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampled :
SSTDICC5.0


Tgt Ion:264	Resp:	4466	
Ion	Ratio	Lower	Upper
264	100		
260	26.6	21.4	32.2
265	50.1	40.9	61.3



#36
Indeno(1,2,3-cd)pyrene
Concen: 5.37 ng
RT: 25.840 min Scan# 3235
Delta R.T. -0.004 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:276	Resp:	98920	
Ion	Ratio	Lower	Upper
276	100		
138	12.6	10.2	15.4
277	25.6	20.4	30.6









#37
Benzo(b)fluoranthene
Concen: 5.31 ng
RT: 22.883 min Scan# 2371
Delta R.T. -0.004 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 87033
Ion Ratio Lower Upper
252 100
253 22.3 21.5 32.3
125 5.6 6.6 10.0#



#38
Benzo(k)fluoranthene
Concen: 5.14 ng
RT: 22.927 min Scan# 2384
Delta R.T. -0.004 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:252 Resp: 88603
Ion Ratio Lower Upper
252 100
253 22.1 22.7 34.1#
125 5.6 6.4 9.6#









#39
Benzo(a)pyrene
Concen: 5.12 ng
RT: 23.465 min Scan# 2541
Delta R.T. -0.000 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 74648
Ion Ratio Lower Upper
252 100
253 22.5 25.1 37.7#
125 6.4 9.4 14.0#



#40
Dibenzo(a,h)anthracene
Concen: 5.52 ng
RT: 25.854 min Scan# 3239
Delta R.T. -0.007 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28


Tgt Ion:278 Resp: 84559
Ion Ratio Lower Upper
278 100
139 8.1 8.6 13.0#
279 23.7 22.8 34.2





#41
Benzo(g,h,i)perylene
Concen: 5.20 ng
RT: 26.532 min Scan# 3437
Delta R.T. -0.004 min
Lab File: BN014816.D
Acq: 01 Jun 2021 16:28

Tgt Ion:	276	Resp:	85636
Ion	Ratio	Lower	Upper
276	100		
277	23.8	20.9	31.3
138	10.7	10.3	15.5



Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0