

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
 Data File : BN014817.D
 Acq On : 01 Jun 2021 17:30
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060121

Manual Integrations
 APPROVED

mohammad
 6/2/2021 10:53:17 AM

Quant Time: Jun 01 18:09:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 17:40:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	618	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	2144	0.40	ng	0.00
13) Acenaphthene-d10	14.359	164	1499	0.40	ng	0.01
19) Phenanthrene-d10	17.104	188	3687	0.40	ng	# 0.00
29) Chrysene-d12	21.311	240	4893	0.40	ng	# 0.00
35) Perylene-d12	23.576	264	4944	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.317	112	697	0.43	ng	0.00
5) Phenol-d6	6.895	99	849	0.40	ng	0.00
8) Nitrobenzene-d5	8.857	82	674m	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.093	152	1506	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	285	0.36	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	2187	0.40	ng	0.00
27) Fluoranthene-d10	19.151	212	4806	0.35	ng	0.00
31) Terphenyl-d14	19.752	244	4379	0.38	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.271	88	352	0.34	ng	# 88
3) n-Nitrosodimethylamine	3.658	42	52	0.29	ng	# 1
6) bis(2-Chloroethyl)ether	7.145	93	678	0.39	ng	95
9) Naphthalene	10.543	128	2319	0.39	ng	96
10) Hexachlorobutadiene	10.835	225	648	0.40	ng	# 94
12) 2-Methylnaphthalene	12.169	142	1583	0.38	ng	98
16) Acenaphthylene	14.067	152	2100	0.37	ng	97
17) Acenaphthene	14.422	154	1644	0.39	ng	98
18) Fluorene	15.412	166	2196	0.38	ng	99
20) 4,6-Dinitro-2-methylph...	15.501	198	123	0.41	ng	# 81
21) 4-Bromophenyl-phenylether	16.313	248	896	0.37	ng	90
22) Hexachlorobenzene	16.422	284	1097	0.38	ng	98
23) Atrazine	16.580	200	491	0.32	ng	99
24) Pentachlorophenol	16.763	266	174	0.42	ng	# 87
25) Phenanthrene	17.152	178	3922	0.38	ng	99
26) Anthracene	17.238	178	2984	0.35	ng	97
28) Fluoranthene	19.181	202	5118	0.36	ng	99
30) Pyrene	19.543	202	5181	0.37	ng	99
32) Benzo(a)anthracene	21.300	228	4989	0.36	ng	99
33) Chrysene	21.343	228	6364	0.38	ng	97
34) Bis(2-ethylhexyl)phtha...	21.226	149	1471	0.38	ng	99
36) Indeno(1,2,3-cd)pyrene	25.852	276	7485	0.37	ng	98
37) Benzo(b)fluoranthene	22.894	252	6841m	0.38	ng	
38) Benzo(k)fluoranthene	22.939	252	7084	0.37	ng	98
39) Benzo(a)pyrene	23.476	252	6101	0.39	ng	# 90
40) Dibenzo(a,h)anthracene	25.872	278	6132	0.36	ng	99
41) Benzo(g,h,i)perylene	26.550	276	6860	0.37	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060121\
 Data File : BN014817.D
 Acq On : 01 Jun 2021 17:30
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampled :
 ICVBN060121

Manual Integrations
 APPROVED


mohammad
 6/2/2021 10:53:17 AM

Quant Time: Jun 01 18:09:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 17:40:45 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.717 min Scan# 613
Delta R.T. 0.009 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30





Tgt Ion: 152 Resp: 618
Ion Ratio Lower Upper
152 100
150 150.4 123.4 185.2
115 81.9 73.2 109.8



#2
1,4-Dioxane
Concen: 0.34 ng
RT: 3.271 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Tgt Ion: 88 Resp: 352
Ion Ratio Lower Upper
88 100
58 48.9 44.7 67.1
43 18.2 21.0 31.6



Instrument :
BNA_N
ClientSampleId :
ICVBN060121

Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:17 AM







#3
n-Nitrosodimethylamine
Concen: 0.29 ng
RT: 3.658 min Scan# 109
Delta R.T. 0.024 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion: 42 Resp: 52
Ion Ratio Lower Upper
42 100
74 196.2 0.0 0.0#
44 0.0 80.0 120.0#

Instrument :
BNA_N
ClientSampleId :
ICVBN060121




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM



#4
2-Fluorophenol
Concen: 0.43 ng
RT: 5.317 min Scan# 315
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion: 112 Resp: 697
Ion Ratio Lower Upper
112 100
64 55.4 43.6 65.4
63 27.4 25.0 37.6









#5
Phenol-d6
Concen: 0.40 ng
RT: 6.895 min Scan# 511
Delta R.T. 0.008 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion: 99	Resp: 849
Ion Ratio	Lower Upper
99 100	
42 11.1	10.2 15.2
71 45.0	32.4 48.6

Instrument :
BNA_N
ClientSampleId :
ICVBN060121




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.145 min Scan# 542
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30


Tgt Ion: 93	Resp: 678		
Ion Ratio	Lower	Upper	
93 100			
63 63.6	47.5	71.3	
95 33.5	28.5	42.7	





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30



Instrument :
BNA_N
ClientSampleId :
ICVBN060121




Tgt Ion:	136	Resp:	2144
Ion	Ratio	Lower	Upper
136	100		
137	14.2	13.2	19.8
54	10.9	9.0	13.4
68	9.8	8.8	13.2

Manual Integrations
APPROVED



mohammad
6/2/2021 10:53:17 AM




#8
Nitrobenzene-d5
Concen: 0.37 ng m
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30



Tgt Ion:	82	Resp:	674
Ion	Ratio	Lower	Upper
82	100		
128	34.0	34.2	51.2#
54	47.7	38.4	57.6









#9
Naphthalene
Concen: 0.39 ng
RT: 10.543 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:	128	Resp:	2319
Ion Ratio	Lower	Upper	
128	100		
129	13.7	13.0	19.4
127	17.5	15.0	22.6

Instrument :
BNA_N
ClientSampleId :
ICVBN060121




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM



#10
Hexachlorobutadiene
Concen: 0.40 ng
RT: 10.835 min Scan# 886
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:	225	Resp:	648
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	59.0	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.093 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:152 Resp: 1506
Ion Ratio Lower Upper
152 100
151 20.9 16.8 25.2

Instrument :
BNA_N
ClientSampleId :
ICVBN060121




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM




#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.169 min Scan# 1088
Delta R.T. 0.005 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:142 Resp: 1583
Ion Ratio Lower Upper
142 100
141 90.1 72.9 109.3
115 54.1 41.8 62.6








#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. 0.013 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Tgt Ion:164 Resp: 1499

Ion	Ratio	Lower	Upper
164	100		
162	106.4	83.2	124.8
160	48.2	41.8	62.6





#14
2,4,6-Tribromophenol
Concen: 0.36 ng
RT: 15.856 min Scan# 1428
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30



Tgt Ion:330 Resp: 285


Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	28.8	21.8	32.6




Instrument :
BNA_N
ClientSampleId :
ICVBN060121

Manual Integrations
APPROVED




mohammad
6/2/2021 10:53:17 AM




#15
2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.976 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30





Tgt Ion:	172	Resp:	2187
Ion	Ratio	Lower	Upper
172	100		
171	39.6	30.9	46.3
170	26.3	21.4	32.0



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Tgt Ion:	152	Resp:	2100
Ion	Ratio	Lower	Upper
152	100		
151	20.2	17.1	25.7
153	14.0	9.9	14.9



Instrument :
BNA_N
ClientSampleId :
ICVBN060121


Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:17 AM



#17
Acenaphthene
Concen: 0.39 ng
RT: 14.422 min Scan# 1315
Delta R.T. 0.013 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Instrument :
BNA_N
ClientSampleId :
ICVBN060121



Tgt Ion:154 Resp: 1644
Ion Ratio Lower Upper
154 100
153 112.9 91.8 137.8
152 56.1 46.0 69.0




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM




#18
Fluorene
Concen: 0.38 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:166 Resp: 2196
Ion Ratio Lower Upper
166 100
165 98.3 79.9 119.9
167 13.4 10.5 15.7









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.104 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30





Tgt Ion:188 Resp: 3687
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.4 8.2 12.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.41 ng
RT: 15.501 min Scan# 1400
Delta R.T. -0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Tgt Ion:198 Resp: 123
Ion Ratio Lower Upper
198 100
51 95.7 96.4 144.6#
105 77.4 72.3 108.5




Instrument :
BNA_N
ClientSampleId :
ICVBN060121

Manual Integrations
APPROVED




mohammad
6/2/2021 10:53:17 AM




#21
4-Bromophenyl-phenylether
Concen: 0.37 ng
RT: 16.313 min Scan# 1465
Delta R.T. 0.012 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30





Tgt Ion:248 Resp: 896
Ion Ratio Lower Upper
248 100
250 102.7 77.5 116.3
141 50.3 50.2 75.4



#22
Hexachlorobenzene
Concen: 0.38 ng
RT: 16.422 min Scan# 1474
Delta R.T. 0.012 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Tgt Ion:284 Resp: 1097
Ion Ratio Lower Upper
284 100
142 26.0 20.3 30.5
249 31.6 26.4 39.6




Instrument :
BNA_N
ClientSampleId :
ICVBN060121

Manual Integrations
APPROVED




mohammad
6/2/2021 10:53:17 AM




#23
Atrazine
Concen: 0.32 ng
RT: 16.580 min Scan# 1487
Delta R.T. 0.012 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30





Tgt Ion:200 Resp: 491
Ion Ratio Lower Upper
200 100
173 35.5 29.1 43.7
215 55.5 45.1 67.7



#24
Pentachlorophenol
Concen: 0.42 ng
RT: 16.763 min Scan# 1502
Delta R.T. -0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Tgt Ion:266 Resp: 174
Ion Ratio Lower Upper
266 100
264 60.9 44.2 66.2
268 69.0 44.2 66.2#



Instrument :
BNA_N
ClientSampleId :
ICVBN060121

Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:17 AM



#25
Phenanthrene
Concen: 0.38 ng
RT: 17.152 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30





Tgt Ion:178 Resp: 3922

Ion	Ratio	Lower	Upper
178	100		
176	19.1	16.0	24.0
179	15.7	12.6	19.0

Instrument :
BNA_N
ClientSampleId :
ICVBN060121

Manual Integrations
APPROVED




mohammad
6/2/2021 10:53:17 AM




#26
Anthracene
Concen: 0.35 ng
RT: 17.238 min Scan# 1541
Delta R.T. -0.000 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30


Tgt Ion:178 Resp: 2984

Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.2	21.2
179	16.0	13.7	20.5









#27
Fluoranthene-d10
Concen: 0.35 ng
RT: 19.151 min Scan# 1835
Delta R.T. 0.007 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30





Tgt Ion:212 Resp: 4806
Ion Ratio Lower Upper
212 100
106 9.4 7.4 11.2
104 5.4 4.1 6.1



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.181 min Scan# 1841
Delta R.T. 0.007 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Tgt Ion:202 Resp: 5118
Ion Ratio Lower Upper
202 100
101 8.7 7.0 10.6
203 17.3 13.5 20.3



Instrument :
BNA_N
ClientSampleId :
ICVBN060121

Manual Integrations
APPROVED

mohammad
6/2/2021 10:53:17 AM







#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.311 min Scan# 2130
Delta R.T. 0.008 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:240 Resp: 4893
Ion Ratio Lower Upper
240 100
120 8.3 5.2 7.8#
236 28.8 21.6 32.4

Instrument :
BNA_N
ClientSampleId :
ICVBN060121




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM




#30
Pyrene
Concen: 0.37 ng
RT: 19.543 min Scan# 1914
Delta R.T. 0.002 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:202 Resp: 5181
Ion Ratio Lower Upper
202 100
200 21.3 16.6 24.8
203 17.5 13.9 20.9







#31
Terphenyl-d14
Concen: 0.38 ng
RT: 19.752 min Scan# 1956
Delta R.T. 0.002 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Tgt Ion:244 Resp: 4379
Ion Ratio Lower Upper
244 100
212 8.2 6.8 10.2
122 9.5 6.6 9.8




Instrument :
BNA_N
ClientSampleId :
ICVBN060121

Manual Integrations
APPROVED



mohammad
6/2/2021 10:53:17 AM




#32
Benzo(a)anthracene
Concen: 0.36 ng
RT: 21.300 min Scan# 2129
Delta R.T. 0.008 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30



Tgt Ion:228 Resp: 4989
Ion Ratio Lower Upper
228 100
226 27.6 22.9 34.3
229 20.6 16.2 24.2









#33
Chrysene
Concen: 0.38 ng
RT: 21.343 min Scan# 2133
Delta R.T. -0.003 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:	228	Resp:	6364
Ion Ratio	Lower	Upper	
228	100		
226	32.4	24.3	36.5
229	19.3	16.4	24.6

Instrument :
BNA_N
ClientSampleId :
ICVBN060121




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.38 ng
RT: 21.226 min Scan# 2122
Delta R.T. -0.003 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30


Tgt Ion:	149	Resp:	1471
Ion Ratio	Lower	Upper	
149	100		
167	36.2	28.3	42.5
279	7.1	5.5	8.3





#35
Perylene-d12
Concen: 0.40 ng
RT: 23.576 min Scan# 2575
Delta R.T. 0.008 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30




Instrument :
BNA_N
ClientSampleId :
ICVBN060121



Tgt Ion:264 Resp: 4944
Ion Ratio Lower Upper
264 100
260 27.8 21.4 32.2
265 51.7 40.9 61.3




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.37 ng
RT: 25.852 min Scan# 3240
Delta R.T. 0.008 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:276 Resp: 7485
Ion Ratio Lower Upper
276 100
138 10.9 10.2 15.4
277 24.8 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.38 ng m
RT: 22.894 min Scan# 2376
Delta R.T. 0.008 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:252 Resp: 6841
Ion Ratio Lower Upper
252 100
253 27.5 21.5 32.3
125 11.6 6.6 10.0#

Instrument :
BNA_N
ClientSampled :
ICVBN060121




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM



#38
Benzo(k)fluoranthene
Concen: 0.37 ng
RT: 22.939 min Scan# 2389
Delta R.T. 0.008 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion:252 Resp: 7084
Ion Ratio Lower Upper
252 100
253 27.2 22.7 34.1
125 8.2 6.4 9.6









#39
Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.476 min Scan# 2546
Delta R.T. 0.011 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion	Ratio	Lower	Upper
252	100		
253	33.7	25.1	37.7
125	21.0	9.4	14.0

Instrument :
BNA_N
ClientSampleId :
ICVBN060121




Manual Integrations
APPROVED


mohammad
6/2/2021 10:53:17 AM




#40
Dibenzo(a,h)anthracene
Concen: 0.36 ng
RT: 25.872 min Scan# 3246
Delta R.T. 0.011 min
Lab File: BN014817.D
Acq: 01 Jun 2021 17:30

Tgt Ion	Ratio	Lower	Upper
278	100		
139	11.5	8.6	13.0
279	29.0	22.8	34.2







#41
 Benzo(g,h,i)perylene
 Concen: 0.37 ng
 RT: 26.550 min Scan# 3444
 Delta R.T. 0.015 min
 Lab File: BN014817.D
 Acq: 01 Jun 2021 17:30



Tgt Ion:	276	Resp:	6860
Ion Ratio	Lower	Upper	
276	100		
277	26.2	20.9	31.3
138	13.1	10.3	15.5



Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060121

Manual Integrations
 APPROVED

mohammad
 6/2/2021 10:53:17 AM