

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060122\
 Data File : BN020014.D
 Acq On : 01 Jun 2022 13:30
 Operator : CG/JU
 Sample : PB145099BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK099

Quant Time: Jun 01 13:58:07 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN053122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 31 16:55:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.846	152	3353	0.400	ng/ul	0.00
4) Naphthalene-d8	10.634	136	10483	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.469	164	6056	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.209	188	13028	0.400	ng/ul	0.00
17) Chrysene-d12	21.396	240	10675	0.400	ng/ul	0.00
23) Perylene-d12	23.731	264	10126	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.311	96	2922	0.765	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.223	152	6252	0.374	ng/ul	0.00
18) Fluoranthene-d10	19.234	212	13395	0.415	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060122\
Data File : BN020014.D
Acq On : 01 Jun 2022 13:30
Operator : CG/JU
Sample : PB145099BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK099

Quant Time: Jun 01 13:58:07 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN053122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 31 16:55:35 2022
Response via : Initial Calibration

