


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060122\
 Data File : BN020006.D
 Acq On : 01 Jun 2022 07:12
 Operator : CG/JU
 Sample : DFTPP253
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP253

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri May 27 01:51:59 2022



AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2469

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.7	181115	PASS
68	69	0.00	2	1.8	3105	PASS
69	198	0.00	100	41.6	168725	PASS
70	69	0.00	2	0.6	957	PASS
127	198	10	80	51.9	210368	PASS
197	198	0.00	2	0.3	1066	PASS
198	198	100	100	100.0	405120	PASS
199	198	5	9	6.8	27587	PASS
275	198	10	60	24.2	97925	PASS
365	198	1	100	3.0	12260	PASS
441	198	0.01	100	11.6	47088	PASS
442	442	50	100	100.0	310677	PASS
443	442	15	24	19.9	61691	PASS

m/z	Abundance
39.90	1327.0
41.00	393.0
42.90	345.0
44.00	1036.0
49.00	362.0
50.90	154.0
54.90	250.0
56.90	357.0
70.90	221.0
73.00	339.0
83.00	188.0
83.90	248.0
85.80	153.0
97.00	159.0
207.00	659.0
281.00	157.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

m/z	Abundance
35.90	160.0
37.00	724.0
38.00	2205.0
39.00	12910.0
40.00	1460.0
41.00	734.0
42.90	240.0
44.00	974.0
44.90	239.0
49.00	1415.0
50.00	44616.0
51.00	172864.0
52.00	8797.0
53.00	581.0
54.90	1119.0
56.00	5234.0
57.00	11508.0
57.90	537.0
61.00	2154.0
62.00	2449.0
63.00	6574.0
64.00	838.0
65.00	3043.0
65.90	224.0
66.90	237.0
68.10	2626.0
68.90	155584.0
70.00	952.0
70.90	216.0
73.00	1350.0
74.00	16028.0
75.00	24592.0
76.00	8783.0
77.00	174528.0
78.00	12694.0
79.00	12555.0
79.90	9461.0
81.00	12975.0
82.00	3626.0
83.00	2906.0
83.90	467.0
84.90	2335.0
86.00	4094.0
87.00	1657.0
88.00	593.0
89.00	264.0
90.90	2861.0
92.00	3308.0
93.00	20912.0
94.00	1510.0
95.00	487.0
96.00	982.0
97.10	471.0
98.00	15615.0
99.00	13749.0
99.90	1080.0
101.00	6786.0
101.90	418.0
103.00	2278.0
103.90	4532.0
104.90	4184.0
106.00	1605.0
107.00	54216.0
108.00	8325.0
109.00	1826.0
110.00	97384.0
111.00	15068.0
112.00	1777.0
112.90	578.0
114.80	222.0
115.90	3533.0
117.00	45976.0
117.90	3854.0
118.90	499.0
119.90	805.0
120.90	299.0
122.00	3496.0
123.00	5437.0
123.90	2817.0
125.00	2452.0
127.00	180672.0
128.00	14324.0
129.00	70792.0
130.00	6593.0
131.00	1074.0
132.00	734.0
132.80	280.0
133.90	2009.0
135.00	6136.0
136.00	2224.0
137.00	3354.0
137.80	911.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

138.80	390.0
140.00	902.0
141.00	8996.0
142.00	3029.0
142.90	1842.0
143.90	527.0
145.00	392.0
146.00	2022.0
146.90	4893.0
147.90	10692.0
149.00	2525.0
149.90	574.0
150.90	1140.0
151.90	593.0
153.00	2871.0
154.00	2369.0
155.00	5206.0
156.00	7286.0
157.00	1686.0
157.90	1961.0
158.90	1412.0
160.00	2954.0
161.00	3904.0
161.90	1297.0
162.80	356.0
163.90	537.0
164.90	3331.0
166.00	2660.0
167.00	18520.0
168.00	10942.0
168.90	1791.0
169.90	513.0
170.90	926.0
171.90	1772.0
173.10	1866.0
174.00	3980.0
175.00	6842.0
175.90	1886.0
177.00	3661.0
178.00	1219.0
178.90	13692.0
180.00	10293.0
181.00	4211.0
181.90	635.0
182.80	452.0
183.90	852.0
185.00	6095.0
186.00	48168.0
187.00	13914.0
188.00	1791.0
188.90	3019.0
190.00	613.0
190.90	1602.0
192.00	4090.0
193.00	4513.0
194.00	884.0
195.00	568.0
196.00	9791.0
197.90	334016.0
198.90	23880.0
199.90	1801.0
201.30	1366.0
202.90	2424.0
204.00	11495.0
205.00	21088.0
206.00	91280.0
207.00	12642.0
207.90	2864.0
209.00	1003.0
210.00	1543.0
211.00	3357.0
212.90	260.0
214.90	914.0
215.90	1636.0
216.90	22536.0
218.00	2963.0
219.00	284.0
221.00	11144.0
221.90	2747.0
223.00	5088.0
224.00	45480.0
225.00	12748.0
226.00	1756.0
227.00	19160.0
228.00	3096.0
228.90	4312.0
230.00	594.0
231.00	1812.0
232.00	227.0
232.90	402.0
234.00	1368.0
234.90	1309.0
235.90	929.0
237.00	2458.0
238.00	302.0
239.00	858.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

239.90	634.0
240.90	1020.0
242.00	2393.0
243.00	2518.0
244.00	35864.0
245.00	4469.0
246.00	7147.0
247.00	1548.0
247.90	384.0
249.00	1380.0
250.00	273.0
251.00	311.0
251.90	311.0
253.00	663.0
255.00	183232.0
256.00	27832.0
257.00	2382.0
258.00	9684.0
259.00	1860.0
259.90	379.0
260.90	345.0
263.90	239.0
265.00	4215.0
266.00	679.0
269.90	172.0
271.00	302.0
271.90	377.0
273.00	4940.0
274.00	14254.0
275.00	77120.0
276.10	9787.0
276.90	6133.0
277.90	1177.0
279.10	268.0
281.00	172.0
282.10	171.0
282.90	609.0
284.00	596.0
284.90	953.0
285.90	267.0
289.00	329.0
290.00	231.0
290.90	155.0
291.90	355.0
293.00	1689.0
293.90	531.0
296.00	21696.0
297.00	3045.0
297.80	218.0
301.10	182.0
301.90	452.0
303.00	2436.0
304.00	602.0
304.90	204.0
307.90	225.0
308.90	286.0
309.90	167.0
312.90	166.0
314.00	1043.0
315.00	2089.0
316.00	1182.0
317.00	382.0
321.00	881.0
321.80	395.0
323.00	5953.0
324.00	1164.0
324.90	212.0
325.80	153.0
326.90	1101.0
327.90	656.0
329.00	183.0
332.00	410.0
332.90	895.0
334.00	3960.0
334.90	1013.0
340.90	937.0
341.90	265.0
345.90	1251.0
346.80	276.0
351.10	195.0
351.90	1878.0
353.00	1447.0
354.00	1854.0
354.90	549.0
359.00	154.0
364.90	8917.0
365.90	1465.0
370.00	290.0
370.90	451.0
371.90	3033.0
373.00	690.0
383.00	1030.0
383.90	336.0
390.00	320.0
391.00	298.0
400.90	175.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

402.00	976.0
403.00	2039.0
404.00	669.0
404.80	185.0
420.90	1470.0
421.90	1517.0
423.00	9997.0
424.00	3542.0
424.90	423.0
441.00	29064.0
442.00	196480.0
443.00	37736.0
444.00	3534.0
445.00	242.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

m/z	Abundance
37.00	905.0
38.00	2483.0
39.00	15910.0
40.00	1504.0
41.00	688.0
43.00	426.0
44.00	935.0
45.10	402.0
46.90	153.0
49.00	1883.0
50.00	54936.0
51.00	207552.0
52.00	10417.0
53.00	396.0
55.00	1367.0
56.00	6199.0
57.00	13814.0
58.00	718.0
59.00	254.0
61.00	2757.0
62.00	3127.0
63.00	8161.0
64.00	1198.0
65.00	3833.0
66.00	251.0
66.90	296.0
68.00	3926.0
69.00	197376.0
70.00	1138.0
70.90	262.0
73.00	1662.0
74.00	20128.0
75.00	30592.0
76.00	10667.0
77.00	222592.0
78.00	14747.0
79.00	15460.0
80.00	11984.0
81.00	16035.0
82.00	4091.0
83.00	3841.0
84.00	427.0
84.90	3098.0
86.00	5137.0
87.00	2151.0
87.90	726.0
88.90	313.0
91.00	4169.0
91.90	4383.0
93.00	26520.0
93.90	2059.0
95.10	415.0
96.00	1217.0
97.10	616.0
98.00	21064.0
99.00	17040.0
99.90	1468.0
100.90	9792.0
101.90	442.0
102.90	2878.0
104.00	5550.0
105.00	5163.0
106.10	1985.0
107.00	71352.0
108.00	10861.0
109.00	2056.0
110.00	129352.0
111.00	19976.0
112.00	2589.0
112.90	823.0
114.90	353.0
116.00	4163.0
117.00	61896.0
118.00	4127.0
119.00	568.0
119.90	862.0
121.10	339.0
122.00	4869.0
123.00	7173.0
124.00	3573.0
125.00	3045.0
126.10	1403.0
127.00	246464.0
128.00	18688.0
129.00	94024.0
130.00	8165.0
130.90	1379.0
132.00	819.0
132.90	444.0
134.00	3033.0
135.00	7650.0
136.00	2949.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

137.00	4536.0
137.90	922.0
139.00	438.0
139.90	1136.0
140.90	12260.0
142.00	3753.0
143.00	2553.0
144.00	729.0
144.90	549.0
146.00	2148.0
147.00	6639.0
148.00	14747.0
149.00	3197.0
150.00	725.0
151.00	1537.0
151.80	1132.0
153.00	3891.0
154.00	3365.0
155.00	7344.0
156.00	10196.0
157.00	1824.0
158.00	2504.0
159.00	1593.0
160.00	3680.0
161.00	5652.0
162.00	1568.0
162.90	362.0
165.00	4521.0
166.00	4437.0
166.90	23944.0
168.00	13304.0
169.00	2175.0
170.00	896.0
170.90	1179.0
172.00	2460.0
173.00	2694.0
174.00	4666.0
175.00	9220.0
176.00	2460.0
176.90	4424.0
177.90	1602.0
178.90	17648.0
180.00	12765.0
181.00	6022.0
182.00	1053.0
183.00	442.0
183.90	1294.0
185.00	8292.0
186.00	66312.0
187.00	19896.0
188.00	2470.0
189.00	4118.0
189.90	812.0
191.00	1878.0
191.90	5242.0
193.00	6337.0
194.00	1431.0
195.00	782.0
195.90	12469.0
196.80	3198.0
197.90	455168.0
198.90	31480.0
200.00	2357.0
201.40	2047.0
203.00	3195.0
204.00	17232.0
205.00	29120.0
206.00	120720.0
207.00	17872.0
208.00	4114.0
209.00	1598.0
210.00	2343.0
211.00	4789.0
211.80	1072.0
212.90	334.0
213.80	203.0
215.00	1409.0
216.00	2786.0
216.90	32392.0
218.00	4068.0
219.00	571.0
221.00	14670.0
221.70	4294.0
223.00	6923.0
223.90	60488.0
225.00	16776.0
226.00	2129.0
227.00	27440.0
228.00	4041.0
229.00	5738.0
229.90	691.0
231.00	2592.0
232.00	517.0
232.90	536.0
233.90	1788.0
235.00	2320.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

236.00	1394.0
236.90	2311.0
237.90	286.0
238.90	1360.0
240.00	983.0
241.00	1577.0
242.00	3598.0
243.00	3460.0
244.00	48864.0
245.00	6530.0
246.00	10554.0
247.00	1861.0
248.10	289.0
249.00	2029.0
249.90	391.0
251.00	431.0
252.00	510.0
253.00	1345.0
255.00	257600.0
256.00	36128.0
257.00	3116.0
258.00	14450.0
258.90	2205.0
259.90	387.0
260.90	468.0
263.90	494.0
265.00	6403.0
265.90	905.0
270.00	348.0
270.90	676.0
272.00	632.0
273.00	7366.0
274.00	21680.0
275.00	114432.0
276.00	14080.0
277.00	10219.0
277.90	1655.0
279.00	402.0
280.90	204.0
282.00	343.0
282.90	955.0
284.00	644.0
285.10	1747.0
286.00	393.0
288.90	487.0
289.90	370.0
290.90	333.0
292.10	541.0
293.00	2219.0
294.00	588.0
294.90	549.0
296.00	32096.0
297.00	4323.0
297.90	331.0
300.90	361.0
302.00	827.0
303.00	3874.0
304.00	1215.0
307.80	464.0
308.90	359.0
310.00	513.0
313.00	332.0
314.00	1305.0
314.90	3404.0
316.00	2083.0
317.00	358.0
320.90	1014.0
321.90	545.0
323.00	9867.0
323.90	1792.0
325.00	247.0
326.10	231.0
326.90	2024.0
327.90	872.0
331.90	667.0
332.90	907.0
334.00	6391.0
334.90	1836.0
336.00	263.0
340.90	1169.0
341.90	309.0
346.00	2201.0
347.00	475.0
350.90	169.0
352.00	3072.0
353.00	1815.0
354.00	2888.0
354.90	594.0
358.90	283.0
364.90	14000.0
365.90	1972.0
370.00	280.0
371.00	762.0
372.00	4337.0
372.90	1248.0
383.00	1322.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

383.90	379.0
390.00	635.0
390.90	416.0
391.90	292.0
400.90	382.0
402.00	1827.0
402.90	2997.0
403.90	1102.0
404.90	160.0
420.90	2611.0
421.90	2574.0
423.00	17456.0
424.00	4554.0
425.00	563.0
439.20	169.0
440.10	528.0
441.00	51624.0
442.00	363840.0
443.00	71256.0
443.90	6385.0
445.00	313.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

m/z	Abundance
37.00	701.0
38.00	1677.0
39.00	12062.0
40.00	1451.0
41.00	601.0
43.00	299.0
44.00	1224.0
45.00	331.0
49.00	1383.0
50.00	41800.0
51.00	163392.0
52.00	8356.0
53.00	445.0
55.00	1093.0
56.00	4687.0
57.00	12073.0
58.00	598.0
58.90	221.0
59.90	204.0
60.90	2027.0
62.00	2156.0
63.00	6691.0
64.00	814.0
65.00	3337.0
66.00	252.0
67.00	317.0
68.00	2764.0
69.00	153216.0
70.00	780.0
70.90	179.0
73.00	1461.0
74.00	15431.0
75.00	25616.0
76.00	8302.0
77.00	171904.0
78.00	11508.0
79.00	12395.0
80.00	9161.0
81.00	13039.0
82.00	3045.0
83.00	3083.0
84.00	434.0
85.00	2963.0
86.00	3952.0
86.90	1669.0
87.90	604.0
88.90	375.0
91.00	3134.0
91.90	3289.0
93.00	21160.0
94.00	1181.0
95.10	323.0
96.00	1116.0
96.90	493.0
98.00	16792.0
99.00	13248.0
100.00	1264.0
100.90	8088.0
101.90	375.0
103.00	2170.0
104.00	5129.0
105.00	4298.0
106.00	1540.0
107.00	58160.0
108.00	8662.0
109.00	1720.0
110.00	102824.0
111.00	15410.0
112.00	1990.0
113.00	715.0
114.80	218.0
116.00	3275.0
117.00	46736.0
118.00	3531.0
118.90	475.0
120.00	653.0
121.00	189.0
122.00	4195.0
123.00	5917.0
123.90	2521.0
125.00	2331.0
127.00	203968.0
128.00	14492.0
129.00	78240.0
130.00	6459.0
131.00	1245.0
131.80	544.0
133.10	360.0
133.90	2391.0
135.00	6490.0
136.00	2363.0
137.00	3786.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

138.00	844.0
139.00	489.0
140.00	1092.0
141.00	10833.0
142.00	3335.0
142.90	2385.0
144.00	603.0
144.80	475.0
146.00	1781.0
147.00	5828.0
147.90	12472.0
149.00	2366.0
149.90	762.0
151.10	1132.0
151.90	1024.0
152.90	3358.0
153.90	2683.0
155.00	6113.0
156.00	8974.0
157.00	1510.0
157.90	2046.0
158.90	1442.0
160.00	3217.0
160.90	4478.0
161.90	1407.0
163.00	261.0
164.00	609.0
165.00	4213.0
166.00	2869.0
167.00	22104.0
168.00	10933.0
169.00	1764.0
170.10	602.0
171.00	1009.0
171.90	1879.0
173.00	2913.0
173.90	4698.0
175.00	7831.0
175.90	2343.0
177.00	3925.0
177.90	1302.0
178.90	15374.0
180.00	11022.0
181.00	5290.0
182.00	794.0
182.80	351.0
183.90	1289.0
185.00	7368.0
186.00	59384.0
187.00	16848.0
187.90	1921.0
188.90	3652.0
189.90	514.0
190.90	1804.0
191.90	5241.0
193.00	5886.0
194.00	1239.0
195.10	717.0
196.00	11967.0
197.90	426176.0
198.90	27400.0
199.90	1981.0
201.40	1443.0
203.00	3218.0
204.00	15036.0
205.00	25608.0
206.00	109104.0
207.00	14329.0
208.00	4095.0
208.90	1072.0
209.90	1723.0
211.00	4290.0
211.80	786.0
212.80	232.0
215.00	1259.0
216.00	2614.0
216.90	28320.0
217.90	3756.0
219.00	312.0
221.00	14468.0
221.70	3849.0
223.00	6545.0
224.00	59560.0
225.00	14340.0
226.00	1747.0
227.00	26688.0
228.00	4278.0
229.00	5011.0
229.90	720.0
230.90	2196.0
232.00	325.0
232.90	516.0
234.00	1711.0
234.90	1724.0
236.00	1246.0
236.90	2391.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

237.90	310.0
239.00	1074.0
240.00	860.0
240.90	1475.0
242.00	2799.0
243.00	3261.0
244.00	46088.0
245.00	5479.0
245.90	8802.0
247.00	2172.0
247.90	389.0
249.00	1717.0
250.00	333.0
251.00	568.0
251.90	330.0
252.90	1217.0
255.00	244288.0
256.00	34240.0
257.00	2796.0
257.90	13468.0
259.00	2246.0
259.80	472.0
260.90	381.0
262.80	213.0
264.00	412.0
265.00	6015.0
265.90	855.0
270.00	331.0
271.00	547.0
272.00	753.0
272.90	7232.0
274.00	18456.0
275.00	102224.0
276.00	14194.0
277.00	9494.0
278.00	1677.0
279.00	398.0
281.90	346.0
282.90	1032.0
284.10	811.0
285.00	1683.0
286.00	331.0
288.90	355.0
289.90	314.0
291.00	419.0
291.90	396.0
293.00	1716.0
293.90	456.0
296.00	30744.0
297.00	3994.0
297.90	360.0
300.90	427.0
301.90	742.0
303.00	3239.0
304.00	1203.0
304.90	158.0
307.80	469.0
308.90	258.0
309.90	384.0
313.00	416.0
314.00	1629.0
315.00	3460.0
316.00	1670.0
317.00	318.0
321.00	944.0
322.00	431.0
323.00	8665.0
324.00	1815.0
326.00	160.0
326.90	2022.0
327.90	1152.0
331.90	655.0
333.00	1067.0
334.00	6351.0
335.00	1346.0
335.90	219.0
338.80	204.0
339.70	231.0
341.00	1357.0
342.10	280.0
346.00	2298.0
346.90	314.0
350.90	191.0
351.90	3084.0
352.90	1649.0
353.90	3092.0
355.00	564.0
358.90	233.0
364.90	13864.0
366.00	1920.0
366.80	157.0
369.90	200.0
370.90	781.0
371.90	5172.0
372.90	1306.0
376.80	235.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

382.90	1501.0
383.90	649.0
384.90	197.0
390.00	797.0
391.00	466.0
391.90	292.0
400.90	331.0
401.90	2360.0
403.00	2648.0
403.90	998.0
404.90	300.0
420.90	2516.0
422.10	2641.0
423.00	18208.0
424.00	3753.0
424.90	554.0
441.00	60576.0
441.90	371712.0
443.00	76080.0
444.00	6666.0
444.90	377.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253