

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060221\
 Data File : BN014821.D
 Acq On : 02 Jun 2021 11:10
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad
 6/3/2021 11:07:37 AM

Quant Time: Jun 02 12:31:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 17:40:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	560	0.400	ng	0.00
7) Naphthalene-d8	10.492	136	1930	0.400	ng	0.00
13) Acenaphthene-d10	14.346	164	1418	0.400	ng	0.00
19) Phenanthrene-d10	17.104	188	3465	0.400	ng	0.00
29) Chrysene-d12	21.303	240	4507	0.400	ng	# 0.00
35) Perylene-d12	23.568	264	4379	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	570	0.392	ng	0.00
5) Phenol-d6	6.887	99	701	0.364	ng	0.00
8) Nitrobenzene-d5	8.857	82	599	0.365	ng	0.00
11) 2-Methylnaphthalene-d10	12.093	152	1356	0.373	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	246	0.330	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	1938	0.374	ng	0.00
27) Fluoranthene-d10	19.144	212	4531	0.356	ng	0.00
31) Terphenyl-d14	19.754	244	4093	0.381	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.271	88	319	0.343	ng	# 92
3) n-Nitrosodimethylamine	3.625	42	8	0.050	ng	# 1
6) bis(2-Chloroethyl)ether	7.145	93	552	0.348	ng	99
9) Naphthalene	10.543	128	2044	0.384	ng	98
10) Hexachlorobutadiene	10.834	225	593	0.407	ng	# 98
12) 2-Methylnaphthalene	12.164	142	1424	0.380	ng	98
16) Acenaphthylene	14.067	152	1906	0.354	ng	99
17) Acenaphthene	14.410	154	1503	0.378	ng	99
18) Fluorene	15.412	166	2023	0.375	ng	98
20) 4,6-Dinitro-2-methylph...	15.501	198	94	0.376	ng	92
21) 4-Bromophenyl-phenylether	16.301	248	829	0.360	ng	# 90
22) Hexachlorobenzene	16.410	284	1049	0.386	ng	97
23) Atrazine	16.568	200	436	0.305	ng	93
24) Pentachlorophenol	16.763	266	158	0.417	ng	90
25) Phenanthrene	17.153	178	3608	0.368	ng	99
26) Anthracene	17.238	178	2845	0.351	ng	97
28) Fluoranthene	19.173	202	4733	0.355	ng	99
30) Pyrene	19.541	202	4874	0.377	ng	99
32) Benzo(a)anthracene	21.292	228	4224	0.335	ng	99
33) Chrysene	21.345	228	5934	0.387	ng	99
34) Bis(2-ethylhexyl)phtha...	21.229	149	1130	0.317	ng	96
36) Indeno(1,2,3-cd)pyrene	25.841	276	6540	0.363	ng	99
37) Benzo(b)fluoranthene	22.887	252	5664	0.356	ng	98
38) Benzo(k)fluoranthene	22.931	252	6882m	0.402	ng	
39) Benzo(a)pyrene	23.465	252	5072	0.364	ng	97
40) Dibenzo(a,h)anthracene	25.861	278	5365	0.358	ng	99
41) Benzo(g,h,i)perylene	26.539	276	6287	0.378	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060221\
Data File : BN014821.D
Acq On : 02 Jun 2021 11:10
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/3/2021 11:07:37 AM


Quant Time: Jun 02 12:31:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 01 17:40:45 2021
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.708 min Scan# 612
 Delta R.T. 0.000 min
 Lab File: BN014821.D
 Acq: 02 Jun 2021 11:10




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4




Tgt Ion: 152 Resp: 560
 Ion Ratio Lower Upper
 152 100
 150 142.3 123.4 185.2
 115 84.8 73.2 109.8

Manual Integrations
 APPROVED



mohammad
 6/3/2021 11:07:37 AM




#2
 1,4-Dioxane
 Concen: 0.343 ng
 RT: 3.271 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN014821.D
 Acq: 02 Jun 2021 11:10



Tgt Ion: 88 Resp: 319
 Ion Ratio Lower Upper
 88 100
 58 53.9 44.7 67.1
 43 16.9 21.0 31.6#









#3
n-Nitrosodimethylamine
Concen: 0.050 ng
RT: 3.625 min Scan# 105
Delta R.T. -0.008 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion: 42 Resp: 8
Ion Ratio Lower Upper
42 100
74 0.0 0.0 0.0
44 0.0 80.0 120.0#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/3/2021 11:07:37 AM



#4
2-Fluorophenol
Concen: 0.392 ng
RT: 5.316 min Scan# 315
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion: 112 Resp: 570
Ion Ratio Lower Upper
112 100
64 58.8 43.6 65.4
63 31.8 25.0 37.6









#5
Phenol-d6
Concen: 0.364 ng
RT: 6.887 min Scan# 510
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion:	99	Resp:	701
Ion	Ratio	Lower	Upper
99	100		
42	12.7	10.2	15.2
71	47.4	32.4	48.6

Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/3/2021 11:07:37 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.348 ng
RT: 7.145 min Scan# 542
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10


Tgt Ion:	93	Resp:	552
Ion	Ratio	Lower	Upper
93	100		
63	58.7	47.5	71.3
95	36.1	28.5	42.7





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
136	100		
137	14.6	13.2	19.8
54	10.4	9.0	13.4
68	8.9	8.8	13.2

Manual Integrations
APPROVED



mohammad
6/3/2021 11:07:37 AM




#8
Nitrobenzene-d5
Concen: 0.365 ng
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Tgt Ion	Ratio	Lower	Upper
82	100		
128	37.1	34.2	51.2
54	46.4	38.4	57.6





#9
Naphthalene
Concen: 0.384 ng
RT: 10.543 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:	128	Resp:	2044
Ion Ratio	Lower	Upper	
128	100		
129	15.0	13.0	19.4
127	17.8	15.0	22.6

Manual Integrations
APPROVED



mohammad
6/3/2021 11:07:37 AM




#10
Hexachlorobutadiene
Concen: 0.407 ng
RT: 10.834 min Scan# 886
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10



Tgt Ion:	225	Resp:	593
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	61.9	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.373 ng
RT: 12.093 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion:152 Resp: 1356
Ion Ratio Lower Upper
152 100
151 21.0 16.8 25.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/3/2021 11:07:37 AM




#12
2-Methylnaphthalene
Concen: 0.380 ng
RT: 12.164 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion:142 Resp: 1424
Ion Ratio Lower Upper
142 100
141 88.2 72.9 109.3
115 51.1 41.8 62.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10





Tgt Ion:164 Resp: 1418
Ion Ratio Lower Upper
164 100
162 104.8 83.2 124.8
160 49.8 41.8 62.6



#14
2,4,6-Tribromophenol
Concen: 0.330 ng
RT: 15.856 min Scan# 1428
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Tgt Ion:330 Resp: 246
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 30.5 21.8 32.6



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad
6/3/2021 11:07:37 AM



#15
2-Fluorobiphenyl
Concen: 0.374 ng
RT: 12.976 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
172	100		
171	39.6	30.9	46.3
170	28.2	21.4	32.0

Manual Integrations
APPROVED



mohammad
6/3/2021 11:07:37 AM




#16
Acenaphthylene
Concen: 0.354 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Tgt Ion	Ratio	Lower	Upper
152	100		
151	21.1	17.1	25.7
153	12.7	9.9	14.9





#17
Acenaphthene
Concen: 0.378 ng
RT: 14.410 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:154 Resp: 1503
Ion Ratio Lower Upper
154 100
153 114.2 91.8 137.8
152 56.6 46.0 69.0

Manual Integrations
APPROVED



mohammad
6/3/2021 11:07:37 AM




#18
Fluorene
Concen: 0.375 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Tgt Ion:166 Resp: 2023
Ion Ratio Lower Upper
166 100
165 98.5 79.9 119.9
167 13.9 10.5 15.7





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.104 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:188 Resp: 3465
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.4 8.2 12.4

Manual Integrations
APPROVED



mohammad
6/3/2021 11:07:37 AM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.376 ng
RT: 15.501 min Scan# 1400
Delta R.T. -0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Tgt Ion:198 Resp: 94
Ion Ratio Lower Upper
198 100
51 110.5 96.4 144.6
105 84.2 72.3 108.5









#21
4-Bromophenyl-phenylether
Concen: 0.360 ng
RT: 16.301 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10





Tgt Ion:248 Resp: 829
Ion Ratio Lower Upper
248 100
250 98.7 77.5 116.3
141 80.3 50.2 75.4#



#22
Hexachlorobenzene
Concen: 0.386 ng
RT: 16.410 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Tgt Ion:284 Resp: 1049
Ion Ratio Lower Upper
284 100
142 25.2 20.3 30.5
249 30.6 26.4 39.6



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/3/2021 11:07:37 AM







#23
Atrazine
Concen: 0.305 ng
RT: 16.568 min Scan# 1486
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	200	Resp:	436
Ion Ratio	Lower	Upper	
200	100		
173	42.9	29.1	43.7
215	59.8	45.1	67.7




Manual Integrations
APPROVED


mohammad
6/3/2021 11:07:37 AM




#24
Pentachlorophenol
Concen: 0.417 ng
RT: 16.763 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion:	266	Resp:	158
Ion Ratio	Lower	Upper	
266	100		
264	50.6	44.2	66.2
268	65.2	44.2	66.2









#25
Phenanthrene
Concen: 0.368 ng
RT: 17.153 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10





Tgt Ion:178 Resp: 3608
Ion Ratio Lower Upper
178 100
176 19.5 16.0 24.0
179 15.4 12.6 19.0



#26
Anthracene
Concen: 0.351 ng
RT: 17.238 min Scan# 1541
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Tgt Ion:178 Resp: 2845
Ion Ratio Lower Upper
178 100
176 18.2 14.2 21.2
179 15.3 13.7 20.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/3/2021 11:07:37 AM







#27
Fluoranthene-d10
Concen: 0.356 ng
RT: 19.144 min Scan# 1832
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion: 212 Resp: 4531
Ion Ratio Lower Upper
212 100
106 9.5 7.4 11.2
104 6.2 4.1 6.1#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/3/2021 11:07:37 AM




#28
Fluoranthene
Concen: 0.355 ng
RT: 19.173 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion: 202 Resp: 4733
Ion Ratio Lower Upper
202 100
101 8.6 7.0 10.6
203 17.2 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10





Tgt Ion:240 Resp: 4507
Ion Ratio Lower Upper
240 100
120 9.4 5.2 7.8#
236 27.8 21.6 32.4



#30
Pyrene
Concen: 0.377 ng
RT: 19.541 min Scan# 1912
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Tgt Ion:202 Resp: 4874
Ion Ratio Lower Upper
202 100
200 20.8 16.6 24.8
203 18.2 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad
6/3/2021 11:07:37 AM



#31
Terphenyl-d14
Concen: 0.381 ng
RT: 19.754 min Scan# 1955
Delta R.T. 0.005 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:244 Resp: 4093
Ion Ratio Lower Upper
244 100
212 9.2 6.8 10.2
122 8.1 6.6 9.8




Manual Integrations
APPROVED


mohammad
6/3/2021 11:07:37 AM




#32
Benzo(a)anthracene
Concen: 0.335 ng
RT: 21.292 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion:228 Resp: 4224
Ion Ratio Lower Upper
228 100
226 28.2 22.9 34.3
229 20.8 16.2 24.2








#33
Chrysene
Concen: 0.387 ng
RT: 21.345 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10




Tgt Ion: 228 Resp: 5934

Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.3	36.5
229	20.6	16.4	24.6





#34
Bis(2-ethylhexyl)phthalate
Concen: 0.317 ng
RT: 21.229 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10



Tgt Ion: 149 Resp: 1130


Ion	Ratio	Lower	Upper
149	100		
167	38.2	28.3	42.5
279	7.1	5.5	8.3



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad
6/3/2021 11:07:37 AM



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.568 min Scan# 2571
 Delta R.T. 0.000 min
 Lab File: BN014821.D
 Acq: 02 Jun 2021 11:10

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4





Tgt Ion: 264 Resp: 4379


Ion	Ratio	Lower	Upper
264	100		
260	25.6	21.4	32.2
265	54.0	40.9	61.3

Manual Integrations
 APPROVED

mohammad
 6/3/2021 11:07:37 AM






#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.363 ng
 RT: 25.841 min Scan# 3235
 Delta R.T. -0.003 min
 Lab File: BN014821.D
 Acq: 02 Jun 2021 11:10



Tgt Ion: 276 Resp: 6540

Ion	Ratio	Lower	Upper
276	100		
138	11.7	10.2	15.4
277	25.8	20.4	30.6









#37
Benzo(b)fluoranthene
Concen: 0.356 ng
RT: 22.887 min Scan# 2372
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion:252 Resp: 5664
Ion Ratio Lower Upper
252 100
253 27.9 21.5 32.3
125 8.9 6.6 10.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/3/2021 11:07:37 AM



#38
Benzo(k)fluoranthene
Concen: 0.402 ng m
RT: 22.931 min Scan# 2385
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion:252 Resp: 6882
Ion Ratio Lower Upper
252 100
253 26.8 22.7 34.1
125 7.7 6.4 9.6









#39
Benzo(a)pyrene
Concen: 0.364 ng
RT: 23.465 min Scan# 2541
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion:	252	Resp:	5072
Ion Ratio	Lower	Upper	
252	100		
253	32.7	25.1	37.7
125	13.7	9.4	14.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/3/2021 11:07:37 AM




#40
Dibenzo(a,h)anthracene
Concen: 0.358 ng
RT: 25.861 min Scan# 3241
Delta R.T. 0.000 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10

Tgt Ion:	278	Resp:	5365
Ion Ratio	Lower	Upper	
278	100		
139	11.3	8.6	13.0
279	29.0	22.8	34.2







#41
Benzo(g,h,i)perylene
Concen: 0.378 ng
RT: 26.539 min Scan# 3439
Delta R.T. 0.003 min
Lab File: BN014821.D
Acq: 02 Jun 2021 11:10



Tgt Ion: 276 Resp: 6287
Ion Ratio Lower Upper
276 100
277 25.9 20.9 31.3
138 13.3 10.3 15.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/3/2021 11:07:37 AM