

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060221\
 Data File : BN014825.D
 Acq On : 02 Jun 2021 15:23
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

mohammad
 6/3/2021 5:06:28 PM

Quant Time: Jun 02 15:56:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 17:40:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	556	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	1993	0.40	ng	0.00
13) Acenaphthene-d10	14.359	164	1468	0.40	ng	0.01
19) Phenanthrene-d10	17.104	188	3547	0.40	ng	# 0.00
29) Chrysene-d12	21.314	240	4719	0.40	ng	0.01
35) Perylene-d12	23.571	264	4827	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.316	112	606	0.42	ng	0.00
5) Phenol-d6	6.895	99	749	0.39	ng	0.00
8) Nitrobenzene-d5	8.857	82	622m	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.093	152	1418	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	289	0.37	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	1986	0.37	ng	0.00
27) Fluoranthene-d10	19.149	212	4645	0.36	ng	0.00
31) Terphenyl-d14	19.754	244	4184	0.37	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.271	88	266	0.26	ng	# 88
3) n-Nitrosodimethylamine	3.665	42	17	0.11	ng	# 1
6) bis(2-Chloroethyl)ether	7.145	93	589	0.37	ng	98
9) Naphthalene	10.543	128	2067	0.38	ng	96
10) Hexachlorobutadiene	10.834	225	603	0.40	ng	# 99
12) 2-Methylnaphthalene	12.169	142	1448	0.37	ng	99
16) Acenaphthylene	14.067	152	1979	0.36	ng	97
17) Acenaphthene	14.422	154	1519	0.37	ng	99
18) Fluorene	15.412	166	2103	0.38	ng	98
20) 4,6-Dinitro-2-methylph...	15.501	198	93	0.37	ng	93
21) 4-Bromophenyl-phenylether	16.301	248	861	0.37	ng	# 88
22) Hexachlorobenzene	16.422	284	1058	0.38	ng	99
23) Atrazine	16.569	200	486	0.33	ng	95
24) Pentachlorophenol	16.763	266	189	0.45	ng	95
25) Phenanthrene	17.153	178	3744	0.37	ng	99
26) Anthracene	17.238	178	2945	0.35	ng	96
28) Fluoranthene	19.178	202	4965	0.36	ng	100
30) Pyrene	19.541	202	5085	0.38	ng	99
32) Benzo(a)anthracene	21.292	228	4957	0.38	ng	98
33) Chrysene	21.345	228	5861	0.37	ng	100
34) Bis(2-ethylhexyl)phtha...	21.229	149	1231	0.33	ng	96
36) Indeno(1,2,3-cd)pyrene	25.851	276	7160	0.36	ng	99
37) Benzo(b)fluoranthene	22.890	252	6116	0.35	ng	98
38) Benzo(k)fluoranthene	22.935	252	7304m	0.39	ng	
39) Benzo(a)pyrene	23.472	252	5528	0.36	ng	98
40) Dibenzo(a,h)anthracene	25.865	278	5879	0.36	ng	98
41) Benzo(g,h,i)perylene	26.542	276	6820	0.37	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060221\
Data File : BN014825.D
Acq On : 02 Jun 2021 15:23
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
6/3/2021 5:06:28 PM


Quant Time: Jun 02 15:56:37 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 01 17:40:45 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.716 min Scan# 613
Delta R.T. 0.008 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:152 Resp: 556
Ion Ratio Lower Upper
152 100
150 145.9 123.4 185.2
115 83.9 73.2 109.8

Manual Integrations
APPROVED



mohammad
6/3/2021 5:06:28 PM




#2
1,4-Dioxane
Concen: 0.26 ng
RT: 3.271 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23



Tgt Ion: 88 Resp: 266
Ion Ratio Lower Upper
88 100
58 62.0 44.7 67.1
43 16.9 21.0 31.6#









#3
n-Nitrosodimethylamine
Concen: 0.11 ng
RT: 3.665 min Scan# 110
Delta R.T. 0.032 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:	42	Resp:	17
Ion	Ratio	Lower	Upper
42	100		
74	0.0	0.0	0.0
44	0.0	80.0	120.0#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
6/3/2021 5:06:28 PM



#4
2-Fluorophenol
Concen: 0.42 ng
RT: 5.316 min Scan# 315
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:	112	Resp:	606
Ion	Ratio	Lower	Upper
112	100		
64	57.8	43.6	65.4
63	31.8	25.0	37.6









#5
Phenol-d6
Concen: 0.39 ng
RT: 6.895 min Scan# 511
Delta R.T. 0.008 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:	99	Resp:	749
Ion	Ratio	Lower	Upper
99	100		
42	12.6	10.2	15.2
71	46.9	32.4	48.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
6/3/2021 5:06:28 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.37 ng
RT: 7.145 min Scan# 542
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23


Tgt Ion:	93	Resp:	589
Ion	Ratio	Lower	Upper
93	100		
63	58.7	47.5	71.3
95	33.8	28.5	42.7





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:	136	Resp:	1993
Ion	Ratio	Lower	Upper
136	100		
137	15.9	13.2	19.8
54	10.2	9.0	13.4
68	9.2	8.8	13.2

Manual Integrations
APPROVED



mohammad
6/3/2021 5:06:28 PM




#8
Nitrobenzene-d5
Concen: 0.37 ng m
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:	82	Resp:	622
Ion	Ratio	Lower	Upper
82	100		
128	33.7	34.2	51.2#
54	49.2	38.4	57.6






#9
Naphthalene
Concen: 0.38 ng
RT: 10.543 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:	128	Resp:	2067
Ion	Ratio	Lower	Upper
128	100		
129	14.3	13.0	19.4
127	17.3	15.0	22.6




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED



mohammad
6/3/2021 5:06:28 PM




#10
Hexachlorobutadiene
Concen: 0.40 ng
RT: 10.834 min Scan# 886
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23



Tgt Ion:	225	Resp:	603
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.2	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 12.093 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:152 Resp: 1418
Ion Ratio Lower Upper
152 100
151 20.2 16.8 25.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
6/3/2021 5:06:28 PM




#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.169 min Scan# 1088
Delta R.T. 0.005 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:142 Resp: 1448
Ion Ratio Lower Upper
142 100
141 89.2 72.9 109.3
115 51.8 41.8 62.6








#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. 0.013 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:164 Resp: 1468

Ion	Ratio	Lower	Upper
164	100		
162	100.4	83.2	124.8
160	44.4	41.8	62.6





#14
2,4,6-Tribromophenol
Concen: 0.37 ng
RT: 15.856 min Scan# 1428
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23



Tgt Ion:330 Resp: 289


Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	28.7	21.8	32.6




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
6/3/2021 5:06:28 PM




#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 12.976 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23





Tgt Ion:	172	Resp:	1986
Ion	Ratio	Lower	Upper
172	100		
171	39.1	30.9	46.3
170	27.1	21.4	32.0



#16
Acenaphthylene
Concen: 0.36 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:	152	Resp:	1979
Ion	Ratio	Lower	Upper
152	100		
151	20.1	17.1	25.7
153	13.7	9.9	14.9



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
6/3/2021 5:06:28 PM



#17
Acenaphthene
Concen: 0.37 ng
RT: 14.422 min Scan# 1315
Delta R.T. 0.013 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:154 Resp: 1519
Ion Ratio Lower Upper
154 100
153 114.0 91.8 137.8
152 57.0 46.0 69.0




Manual Integrations
APPROVED


mohammad
6/3/2021 5:06:28 PM



#18
Fluorene
Concen: 0.38 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23


Tgt Ion:166 Resp: 2103
Ion Ratio Lower Upper
166 100
165 97.4 79.9 119.9
167 14.0 10.5 15.7





#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.104 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:188 Resp: 3547
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.8 8.2 12.4#

Manual Integrations
APPROVED



mohammad
6/3/2021 5:06:28 PM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.37 ng
RT: 15.501 min Scan# 1400
Delta R.T. -0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:198 Resp: 93
Ion Ratio Lower Upper
198 100
51 111.4 96.4 144.6
105 84.8 72.3 108.5





#21
4-Bromophenyl-phenylether
Concen: 0.37 ng
RT: 16.301 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:248 Resp: 861
Ion Ratio Lower Upper
248 100
250 92.1 77.5 116.3
141 80.2 50.2 75.4#

Manual Integrations
APPROVED



mohammad
6/3/2021 5:06:28 PM




#22
Hexachlorobenzene
Concen: 0.38 ng
RT: 16.422 min Scan# 1474
Delta R.T. 0.012 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:284 Resp: 1058
Ion Ratio Lower Upper
284 100
142 25.7 20.3 30.5
249 32.4 26.4 39.6





#23
Atrazine
Concen: 0.33 ng
RT: 16.569 min Scan# 1486
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:200 Resp: 486
Ion Ratio Lower Upper
200 100
173 43.3 29.1 43.7
215 57.4 45.1 67.7

Manual Integrations
APPROVED



mohammad
6/3/2021 5:06:28 PM




#24
Pentachlorophenol
Concen: 0.45 ng
RT: 16.763 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:266 Resp: 189
Ion Ratio Lower Upper
266 100
264 60.3 44.2 66.2
268 52.9 44.2 66.2





#25
Phenanthrene
Concen: 0.37 ng
RT: 17.153 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC





Tgt Ion:178 Resp: 3744


Ion	Ratio	Lower	Upper
178	100		
176	19.6	16.0	24.0
179	15.7	12.6	19.0

Manual Integrations
APPROVED

mohammad
6/3/2021 5:06:28 PM






#26
Anthracene
Concen: 0.35 ng
RT: 17.238 min Scan# 1541
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:178 Resp: 2945

Ion	Ratio	Lower	Upper
178	100		
176	19.6	14.2	21.2
179	15.8	13.7	20.5









#27
Fluoranthene-d10
Concen: 0.36 ng
RT: 19.149 min Scan# 1833
Delta R.T. 0.005 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23





Tgt Ion:212 Resp: 4645
Ion Ratio Lower Upper
212 100
106 9.4 7.4 11.2
104 5.5 4.1 6.1



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.178 min Scan# 1839
Delta R.T. 0.005 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:202 Resp: 4965
Ion Ratio Lower Upper
202 100
101 8.8 7.0 10.6
203 16.8 13.5 20.3




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
6/3/2021 5:06:28 PM




#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.314 min Scan# 2129
Delta R.T. 0.011 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23





Tgt Ion:240 Resp: 4719
Ion Ratio Lower Upper
240 100
120 6.9 5.2 7.8
236 27.0 21.6 32.4



#30
Pyrene
Concen: 0.38 ng
RT: 19.541 min Scan# 1912
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:202 Resp: 5085
Ion Ratio Lower Upper
202 100
200 21.1 16.6 24.8
203 17.8 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
6/3/2021 5:06:28 PM



#31
Terphenyl-d14
Concen: 0.37 ng
RT: 19.754 min Scan# 1955
Delta R.T. 0.005 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:244 Resp: 4184
Ion Ratio Lower Upper
244 100
212 8.3 6.8 10.2
122 8.7 6.6 9.8




Manual Integrations
APPROVED


mohammad
6/3/2021 5:06:28 PM




#32
Benzo(a)anthracene
Concen: 0.38 ng
RT: 21.292 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:228 Resp: 4957
Ion Ratio Lower Upper
228 100
226 27.7 22.9 34.3
229 21.0 16.2 24.2









#33
Chrysene
Concen: 0.37 ng
RT: 21.345 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23





Tgt Ion:228 Resp: 5861
Ion Ratio Lower Upper
228 100
226 30.7 24.3 36.5
229 20.3 16.4 24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.33 ng
RT: 21.229 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Tgt Ion:149 Resp: 1231
Ion Ratio Lower Upper
149 100
167 37.7 28.3 42.5
279 7.6 5.5 8.3



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
6/3/2021 5:06:28 PM



#35
Perylene-d12
Concen: 0.40 ng
RT: 23.571 min Scan# 2572
Delta R.T. 0.003 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:264 Resp: 4827
Ion Ratio Lower Upper
264 100
260 26.6 21.4 32.2
265 48.9 40.9 61.3




Manual Integrations
APPROVED


mohammad
6/3/2021 5:06:28 PM



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.36 ng
RT: 25.851 min Scan# 3238
Delta R.T. 0.007 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:276 Resp: 7160
Ion Ratio Lower Upper
276 100
138 11.9 10.2 15.4
277 25.8 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.35 ng
RT: 22.890 min Scan# 2373
Delta R.T. 0.003 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.6	21.5	32.3
125	9.9	6.6	10.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
6/3/2021 5:06:28 PM



#38
Benzo(k)fluoranthene
Concen: 0.39 ng m
RT: 22.935 min Scan# 2386
Delta R.T. 0.003 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.4	22.7	34.1
125	8.3	6.4	9.6









#39
Benzo(a)pyrene
Concen: 0.36 ng
RT: 23.472 min Scan# 2543
Delta R.T. 0.007 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:	252	Resp:	5528
Ion Ratio	Lower	Upper	
252	100		
253	30.8	25.1	37.7
125	13.9	9.4	14.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
6/3/2021 5:06:28 PM



#40
Dibenzo(a,h)anthracene
Concen: 0.36 ng
RT: 25.865 min Scan# 3242
Delta R.T. 0.003 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:	278	Resp:	5879
Ion Ratio	Lower	Upper	
278	100		
139	11.8	8.6	13.0
279	27.7	22.8	34.2





#41
Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 26.542 min Scan# 3440
Delta R.T. 0.007 min
Lab File: BN014825.D
Acq: 02 Jun 2021 15:23

Tgt Ion:	276	Resp:	6820
Ion Ratio	Lower	Upper	
276	100		
277	25.9	20.9	31.3
138	13.1	10.3	15.5

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
6/3/2021 5:06:28 PM

