

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060321\
 Data File : BN014830.D
 Acq On : 03 Jun 2021 17:43
 Operator : CG/JU
 Sample : SSTDIC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.8

Quant Time: Jun 03 19:10:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 03 17:56:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.709	152	334	0.40	ng	0.00
7) Naphthalene-d8	10.482	136	1192	0.40	ng	0.00
13) Acenaphthene-d10	14.345	164	811	0.40	ng	0.00
19) Phenanthrene-d10	17.103	188	1891	0.40	ng	0.00
29) Chrysene-d12	21.302	240	2449	0.40	ng	# 0.00
35) Perylene-d12	23.556	264	2245	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.308	112	509	0.59	ng	0.00
5) Phenol-d6	6.889	99	702	0.61	ng	0.00
8) Nitrobenzene-d5	8.860	82	766	0.76	ng	0.00
11) 2-Methylnaphthalene-d10	12.083	152	1680	0.75	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	325	0.76	ng	-0.01
15) 2-Fluorobiphenyl	12.962	172	2466	0.83	ng	-0.01
27) Fluoranthene-d10	19.137	212	4978	0.72	ng	0.00
31) Terphenyl-d14	19.743	244	5204	0.89	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.251	88	348	0.74	ng	# 93
3) n-Nitrosodimethylamine	3.679	42	63	0.65	ng	# 29
6) bis(2-Chloroethyl)ether	7.138	93	682	0.72	ng	96
9) Naphthalene	10.533	128	2633	0.80	ng	96
10) Hexachlorobutadiene	10.825	225	758	0.84	ng	# 100
12) 2-Methylnaphthalene	12.158	142	1832	0.79	ng	96
16) Acenaphthylene	14.066	152	2290	0.74	ng	97
17) Acenaphthene	14.408	154	1792	0.79	ng	99
18) Fluorene	15.398	166	2412	0.78	ng	100
20) 4,6-Dinitro-2-methylph...	15.499	198	85	0.37	ng	90
21) 4-Bromophenyl-phenylether	16.299	248	989	0.79	ng	95
22) Hexachlorobenzene	16.409	284	1241	0.84	ng	99
23) Atrazine	16.567	200	495	0.64	ng	96
24) Pentachlorophenol	16.762	266	166	0.48	ng	# 83
25) Phenanthrene	17.139	178	4118	0.77	ng	99
26) Anthracene	17.236	178	3239	0.73	ng	97
28) Fluoranthene	19.167	202	5435	0.75	ng	99
30) Pyrene	19.534	202	5399	0.77	ng	100
32) Benzo(a)anthracene	21.280	228	4633	0.68	ng	98
33) Chrysene	21.333	228	6353	0.76	ng	99
34) Bis(2-ethylhexyl)phtha...	21.217	149	1248	0.64	ng	97
36) Indeno(1,2,3-cd)pyrene	25.832	276	7085	0.77	ng	99
37) Benzo(b)fluoranthene	22.875	252	6334	0.78	ng	99
38) Benzo(k)fluoranthene	22.923	252	7211	0.82	ng	98
39) Benzo(a)pyrene	23.457	252	5626	0.79	ng	# 97
40) Dibenzo(a,h)anthracene	25.846	278	5829	0.76	ng	98
41) Benzo(g,h,i)perylene	26.517	276	6681	0.78	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060321\
Data File : BN014830.D
Acq On : 03 Jun 2021 17:43
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Jun 03 19:10:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 03 17:56:36 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.709 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: BN014830.D
 Acq: 03 Jun 2021 17:43




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:152 Resp: 334
 Ion Ratio Lower Upper
 152 100
 150 138.4 123.4 185.2
 115 83.8 73.2 109.8



#2
 1,4-Dioxane
 Concen: 0.74 ng
 RT: 3.251 min Scan# 40
 Delta R.T. -0.012 min
 Lab File: BN014830.D
 Acq: 03 Jun 2021 17:43

Tgt Ion: 88 Resp: 348
 Ion Ratio Lower Upper
 88 100
 58 58.3 44.7 67.1
 43 19.0 21.0 31.6#









#3
n-Nitrosodimethylamine
Concen: 0.65 ng
RT: 3.679 min Scan# 76
Delta R.T. 0.024 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 63
Ion Ratio Lower Upper
42 100
74 134.9 0.0 0.0#
44 28.6 80.0 120.0#



#4
2-Fluorophenol
Concen: 0.59 ng
RT: 5.308 min Scan# 213
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion: 112 Resp: 509
Ion Ratio Lower Upper
112 100
64 56.4 43.6 65.4
63 28.7 25.0 37.6









#5
Phenol-d6
Concen: 0.61 ng
RT: 6.889 min Scan# 346
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampled :
SSTDIC0.8


Tgt Ion:	99	Resp:	702
Ion	Ratio	Lower	Upper
99	100		
42	11.1	10.2	15.2
71	44.4	32.4	48.6



#6
bis(2-Chloroethyl)ether
Concen: 0.72 ng
RT: 7.138 min Scan# 367
Delta R.T. -0.012 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	93	Resp:	682
Ion	Ratio	Lower	Upper
93	100		
63	58.2	47.5	71.3
95	30.8	28.5	42.7









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.482 min Scan# 637
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	136	Resp:	1192
Ion	Ratio	Lower	Upper
136	100		
137	16.6	13.2	19.8
54	13.3	9.0	13.4
68	11.9	8.8	13.2



#8
Nitrobenzene-d5
Concen: 0.76 ng
RT: 8.860 min Scan# 509
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	82	Resp:	766
Ion	Ratio	Lower	Upper
82	100		
128	43.8	34.2	51.2
54	44.1	38.4	57.6









#9
Naphthalene
Concen: 0.80 ng
RT: 10.533 min Scan# 641
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	128	Resp:	2633
Ion Ratio	Lower	Upper	
128	100		
129	14.0	13.0	19.4
127	17.1	15.0	22.6



#10
Hexachlorobutadiene
Concen: 0.84 ng
RT: 10.825 min Scan# 664
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	225	Resp:	758
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.3	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.75 ng
RT: 12.083 min Scan# 847
Delta R.T. -0.005 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 1680
Ion Ratio Lower Upper
152 100
151 21.0 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.79 ng
RT: 12.158 min Scan# 864
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:142 Resp: 1832
Ion Ratio Lower Upper
142 100
141 87.7 72.9 109.3
115 49.7 41.8 62.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.345 min Scan# 1087
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164	Resp:	811	
Ion	Ratio	Lower	Upper
164	100		
162	102.9	83.2	124.8
160	49.6	41.8	62.6



#14
2,4,6-Tribromophenol
Concen: 0.76 ng
RT: 15.842 min Scan# 1205
Delta R.T. -0.013 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:330	Resp:	325	
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	24.9	21.8	32.6









#15
2-Fluorobiphenyl
Concen: 0.83 ng
RT: 12.962 min Scan# 978
Delta R.T. -0.013 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:172	Resp:	2466
Ion Ratio	Lower	Upper
172	100	
171	38.9	30.9 46.3
170	27.4	21.4 32.0



#16
Acenaphthylene
Concen: 0.74 ng
RT: 14.066 min Scan# 1065
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:152	Resp:	2290
Ion Ratio	Lower	Upper
152	100	
151	20.2	17.1 25.7
153	13.6	9.9 14.9









#17
Acenaphthene
Concen: 0.79 ng
RT: 14.408 min Scan# 1092
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 1792
Ion Ratio Lower Upper
154 100
153 114.2 91.8 137.8
152 58.6 46.0 69.0



#18
Fluorene
Concen: 0.78 ng
RT: 15.398 min Scan# 1170
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:166 Resp: 2412
Ion Ratio Lower Upper
166 100
165 99.6 79.9 119.9
167 13.5 10.5 15.7









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.103 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	188	Resp:	1891
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.9	8.2	12.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.37 ng
RT: 15.499 min Scan# 1178
Delta R.T. -0.038 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	198	Resp:	85
Ion Ratio	Lower	Upper	
198	100		
51	101.4	96.4	144.6
105	91.7	72.3	108.5









#21
4-Bromophenyl-phenylether
Concen: 0.79 ng
RT: 16.299 min Scan# 1242
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampled :
SSTDICC0.8


Tgt Ion:	248	Resp:	989
Ion	Ratio	Lower	Upper
248	100		
250	101.0	77.5	116.3
141	58.0	50.2	75.4



#22
Hexachlorobenzene
Concen: 0.84 ng
RT: 16.409 min Scan# 1251
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	284	Resp:	1241
Ion	Ratio	Lower	Upper
284	100		
142	25.9	20.3	30.5
249	32.6	26.4	39.6









#23
Atrazine
Concen: 0.64 ng
RT: 16.567 min Scan# 1264
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:200 Resp: 495
Ion Ratio Lower Upper
200 100
173 38.2 29.1 43.7
215 59.7 45.1 67.7



#24
Pentachlorophenol
Concen: 0.48 ng
RT: 16.762 min Scan# 1280
Delta R.T. -0.012 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:266 Resp: 166
Ion Ratio Lower Upper
266 100
264 61.4 44.2 66.2
268 72.9 44.2 66.2#









#25
Phenanthrene
Concen: 0.77 ng
RT: 17.139 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 4118
Ion Ratio Lower Upper
178 100
176 19.5 16.0 24.0
179 15.2 12.6 19.0



#26
Anthracene
Concen: 0.73 ng
RT: 17.236 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:178 Resp: 3239
Ion Ratio Lower Upper
178 100
176 19.0 14.2 21.2
179 16.1 13.7 20.5









#27
Fluoranthene-d10
Concen: 0.72 ng
RT: 19.137 min Scan# 1609
Delta R.T. -0.005 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:212 Resp: 4978
Ion Ratio Lower Upper
212 100
106 9.6 7.4 11.2
104 5.5 4.1 6.1



#28
Fluoranthene
Concen: 0.75 ng
RT: 19.167 min Scan# 1615
Delta R.T. -0.005 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:202 Resp: 5435
Ion Ratio Lower Upper
202 100
101 8.3 7.0 10.6
203 17.1 13.5 20.3





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.302 min Scan# 1906

Delta R.T. -0.000 min

Lab File: BN014830.D

Acq: 03 Jun 2021 17:43





Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8

Tgt Ion:240 Resp: 2449
 Ion Ratio Lower Upper
 240 100
 120 9.5 5.2 7.8#
 236 28.8 21.6 32.4



#30

Pyrene

Concen: 0.77 ng




RT: 19.534 min Scan# 1689


Delta R.T. -0.000 min

Lab File: BN014830.D

Acq: 03 Jun 2021 17:43

Tgt Ion:202 Resp: 5399
 Ion Ratio Lower Upper
 202 100
 200 20.7 16.6 24.8
 203 17.6 13.9 20.9









#31
Terphenyl-d14
Concen: 0.89 ng
RT: 19.743 min Scan# 1731
Delta R.T. -0.005 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:244 Resp: 5204
Ion Ratio Lower Upper
244 100
212 8.8 6.8 10.2
122 9.4 6.6 9.8



#32
Benzo(a)anthracene
Concen: 0.68 ng
RT: 21.280 min Scan# 1904
Delta R.T. -0.011 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:228 Resp: 4633
Ion Ratio Lower Upper
228 100
226 29.7 22.9 34.3
229 20.5 16.2 24.2









#33
Chrysene
Concen: 0.76 ng
RT: 21.333 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	228	Resp:	6353
Ion Ratio	Lower	Upper	
228	100		
226	30.9	24.3	36.5
229	20.5	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.64 ng
RT: 21.217 min Scan# 1898
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43


Tgt Ion:	149	Resp:	1248
Ion Ratio	Lower	Upper	
149	100		
167	37.5	28.3	42.5
279	7.9	5.5	8.3








#35
Perylene-d12
Concen: 0.40 ng
RT: 23.556 min Scan# 2346
Delta R.T. -0.004 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8





Tgt Ion:264 Resp: 2245
Ion Ratio Lower Upper
264 100
260 28.4 21.4 32.2
265 83.1 40.9 61.3#




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.77 ng
RT: 25.832 min Scan# 3011
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Tgt Ion:276 Resp: 7085
Ion Ratio Lower Upper
276 100
138 11.9 10.2 15.4
277 25.2 20.4 30.6







#37
Benzo(b)fluoranthene
Concen: 0.78 ng
RT: 22.875 min Scan# 2147
Delta R.T. -0.003 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:252 Resp: 6334
Ion Ratio Lower Upper
252 100
253 27.0 21.5 32.3
125 8.8 6.6 10.0



#38
Benzo(k)fluoranthene
Concen: 0.82 ng
RT: 22.923 min Scan# 2161
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:252 Resp: 7211
Ion Ratio Lower Upper
252 100
253 27.3 22.7 34.1
125 8.1 6.4 9.6









#39
Benzo(a)pyrene
Concen: 0.79 ng
RT: 23.457 min Scan# 2317
Delta R.T. -0.004 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 5626
Ion Ratio Lower Upper
252 100
253 31.8 25.1 37.7
125 14.6 9.4 14.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.76 ng
RT: 25.846 min Scan# 3015
Delta R.T. -0.004 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:278 Resp: 5829
Ion Ratio Lower Upper
278 100
139 11.1 8.6 13.0
279 27.0 22.8 34.2





#41
 Benzo(g,h,i)perylene
 Concen: 0.78 ng
 RT: 26.517 min Scan# 3211
 Delta R.T. -0.004 min
 Lab File: BN014830.D
 Acq: 03 Jun 2021 17:43

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	276	Resp:	6681
Ion	Ratio	Lower	Upper
276	100		
277	25.7	20.9	31.3
138	12.4	10.3	15.5

