

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060321\
 Data File : BN014834.D
 Acq On : 03 Jun 2021 20:06
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060321

Quant Time: Jun 04 05:28:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 04 05:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.709	152	310	0.40	ng	0.00
7) Naphthalene-d8	10.483	136	1131	0.40	ng	# 0.00
13) Acenaphthene-d10	14.345	164	795	0.40	ng	0.00
19) Phenanthrene-d10	17.103	188	1775	0.40	ng	# 0.00
29) Chrysene-d12	21.302	240	2284	0.40	ng	# 0.00
35) Perylene-d12	23.556	264	2315	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.308	112	226	0.36	ng	0.00
5) Phenol-d6	6.889	99	293	0.36	ng	0.00
8) Nitrobenzene-d5	8.860	82	326	0.36	ng	0.00
11) 2-Methylnaphthalene-d10	12.087	152	747	0.36	ng	0.00
14) 2,4,6-Tribromophenol	15.854	330	126	0.35	ng	0.00
15) 2-Fluorobiphenyl	12.974	172	1070	0.38	ng	0.00
27) Fluoranthene-d10	19.142	212	2245	0.36	ng	0.00
31) Terphenyl-d14	19.748	244	2080	0.38	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.263	88	159	0.38	ng	# 85
6) bis(2-Chloroethyl)ether	7.150	93	219	0.30	ng	# 70
9) Naphthalene	10.533	128	1184	0.38	ng	96
10) Hexachlorobutadiene	10.825	225	347	0.39	ng	# 100
12) 2-Methylnaphthalene	12.163	142	778	0.37	ng	97
16) Acenaphthylene	14.066	152	1001	0.35	ng	97
17) Acenaphthene	14.408	154	843	0.38	ng	97
18) Fluorene	15.410	166	1073	0.37	ng	99
20) 4,6-Dinitro-2-methylph...	15.512	198	19	0.39	ng	89
21) 4-Bromophenyl-phenylether	16.299	248	431	0.38	ng	# 89
22) Hexachlorobenzene	16.409	284	575	0.39	ng	99
23) Atrazine	16.567	200	238	0.36	ng	# 82
24) Pentachlorophenol	16.774	266	50	0.23	ng	# 89
25) Phenanthrene	17.139	178	1856	0.38	ng	98
26) Anthracene	17.236	178	1355	0.35	ng	99
28) Fluoranthene	19.172	202	2430	0.37	ng	99
30) Pyrene	19.534	202	2462	0.37	ng	97
32) Benzo(a)anthracene	21.291	228	2057	0.37	ng	95
33) Chrysene	21.333	228	2916	0.38	ng	96
34) Bis(2-ethylhexyl)phtha...	21.217	149	632	0.40	ng	90
36) Indeno(1,2,3-cd)pyrene	25.829	276	3666	0.39	ng	# 97
37) Benzo(b)fluoranthene	22.878	252	3127	0.38	ng	# 86
38) Benzo(k)fluoranthene	22.923	252	3307	0.35	ng	# 92
39) Benzo(a)pyrene	23.460	252	2832	0.39	ng	# 77
40) Dibenzo(a,h)anthracene	25.846	278	2944	0.38	ng	# 91
41) Benzo(g,h,i)perylene	26.517	276	3484	0.39	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060321\
 Data File : BN014834.D
 Acq On : 03 Jun 2021 20:06
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060321

Quant Time: Jun 04 05:28:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 04 05:24:34 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.709 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: BN014834.D
 Acq: 03 Jun 2021 20:06




Instrument :
 BNA_N
 ClientSampled :
 ICVBN060321


Tgt Ion: 152 Resp: 310
 Ion Ratio Lower Upper
 152 100
 150 140.7 123.4 185.2
 115 85.9 73.2 109.8



#2
 1,4-Dioxane
 Concen: 0.38 ng
 RT: 3.263 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: BN014834.D
 Acq: 03 Jun 2021 20:06

Tgt Ion: 88 Resp: 159
 Ion Ratio Lower Upper
 88 100
 58 47.8 44.7 67.1
 43 15.1 21.0 31.6#









#4
2-Fluorophenol
Concen: 0.36 ng
RT: 5.308 min Scan# 213
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:	112	Resp:	226
Ion	Ratio	Lower	Upper
112	100		
64	65.5	43.6	65.4#
63	30.1	25.0	37.6



#5
Phenol-d6
Concen: 0.36 ng
RT: 6.889 min Scan# 346
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:	99	Resp:	293
Ion	Ratio	Lower	Upper
99	100		
42	14.7	10.2	15.2
71	47.1	32.4	48.6









#6
bis(2-Chloroethyl)ether
Concen: 0.30 ng
RT: 7.150 min Scan# 368
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion: 93	Resp: 219
Ion Ratio	Lower Upper
93 100	
63 75.8	47.5 71.3#
95 61.2	28.5 42.7#



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.483 min Scan# 637
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt	Ion:136	Resp:	1131
Ion	Ratio	Lower	Upper
136	100		
137	18.2	13.2	19.8
54	15.8	9.0	13.4#
68	15.1	8.8	13.2#









#8
Nitrobenzene-d5
Concen: 0.36 ng
RT: 8.860 min Scan# 509
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:	82	Resp:	326
Ion	Ratio	Lower	Upper
82	100		
128	49.0	34.2	51.2
54	59.1	38.4	57.6



#9
Naphthalene
Concen: 0.38 ng
RT: 10.533 min Scan# 641
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:	128	Resp:	1184
Ion	Ratio	Lower	Upper
128	100		
129	17.5	13.0	19.4
127	20.9	15.0	22.6









#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.825 min Scan# 664
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:	225	Resp:	347
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.4	50.8	76.2



#11
2-Methylnaphthalene-d10
Concen: 0.36 ng
RT: 12.087 min Scan# 848
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:	152	Resp:	747
Ion Ratio	Lower	Upper	
152	100		
151	21.3	16.8	25.2









#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.163 min Scan# 865
Delta R.T. 0.004 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:142 Resp: 778
Ion Ratio Lower Upper
142 100
141 92.5 72.9 109.3
115 57.1 41.8 62.6



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.345 min Scan# 1087
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:164 Resp: 795
Ion Ratio Lower Upper
164 100
162 104.8 83.2 124.8
160 49.0 41.8 62.6









#14
2,4,6-Tribromophenol
Concen: 0.35 ng
RT: 15.854 min Scan# 1206
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:	330	Resp:	126
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	31.7	21.8	32.6



#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 12.974 min Scan# 979
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:	172	Resp:	1070
Ion Ratio	Lower	Upper	
172	100		
171	40.2	30.9	46.3
170	30.2	21.4	32.0









#16
Acenaphthylene
Concen: 0.35 ng
RT: 14.066 min Scan# 1065
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:	152	Resp:	1001
Ion	Ratio	Lower	Upper
152	100		
151	20.5	17.1	25.7
153	13.9	9.9	14.9



#17
Acenaphthene
Concen: 0.38 ng
RT: 14.408 min Scan# 1092
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:	154	Resp:	843
Ion	Ratio	Lower	Upper
154	100		
153	111.4	91.8	137.8
152	56.1	46.0	69.0









#18
Fluorene
Concen: 0.37 ng
RT: 15.410 min Scan# 1171
Delta R.T. 0.013 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:166 Resp: 1073
Ion Ratio Lower Upper
166 100
165 100.8 79.9 119.9
167 14.5 10.5 15.7



#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.103 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:188 Resp: 1775
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.3 8.2 12.4#









#20
4,6-Dinitro-2-methylphenol
Concen: 0.39 ng
RT: 15.512 min Scan# 1179
Delta R.T. -0.025 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:198 Resp: 19
Ion Ratio Lower Upper
198 100
51 127.5 96.4 144.6
105 105.9 72.3 108.5



#21
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.299 min Scan# 1242
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:248 Resp: 431
Ion Ratio Lower Upper
248 100
250 101.5 77.5 116.3
141 78.9 50.2 75.4#









#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.409 min Scan# 1251
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:284	Resp:	575
Ion	Ratio	Lower Upper
284	100	
142	25.2	20.3 30.5
249	33.7	26.4 39.6



#23
Atrazine
Concen: 0.36 ng
RT: 16.567 min Scan# 1264
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:200	Resp:	238
Ion	Ratio	Lower Upper
200	100	
173	50.0	29.1 43.7#
215	67.8	45.1 67.7#









#24
Pentachlorophenol
Concen: 0.23 ng
RT: 16.774 min Scan# 1281
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:266 Resp: 50
Ion Ratio Lower Upper
266 100
264 58.0 44.2 66.2
268 68.0 44.2 66.2#



#25
Phenanthrene
Concen: 0.38 ng
RT: 17.139 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:178 Resp: 1856
Ion Ratio Lower Upper
178 100
176 19.0 16.0 24.0
179 15.2 12.6 19.0









#26
 Anthracene
 Concen: 0.35 ng
 RT: 17.236 min Scan# 1319
 Delta R.T. -0.000 min
 Lab File: BN014834.D
 Acq: 03 Jun 2021 20:06




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060321


Tgt Ion:178	Resp:	1355
Ion Ratio	Lower	Upper
178	100	
176	18.2	14.2 21.2
179	16.8	13.7 20.5



#27
 Fluoranthene-d10
 Concen: 0.36 ng
 RT: 19.142 min Scan# 1610
 Delta R.T. -0.000 min
 Lab File: BN014834.D
 Acq: 03 Jun 2021 20:06

Tgt Ion:212	Resp:	2245
Ion Ratio	Lower	Upper
212	100	
106	9.7	7.4 11.2
104	5.2	4.1 6.1









#28
Fluoranthene
Concen: 0.37 ng
RT: 19.172 min Scan# 1616
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:202 Resp: 2430
Ion Ratio Lower Upper
202 100
101 8.1 7.0 10.6
203 16.9 13.5 20.3



#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.302 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:240 Resp: 2284
Ion Ratio Lower Upper
240 100
120 10.9 5.2 7.8#
236 30.2 21.6 32.4









#30
Pyrene
Concen: 0.37 ng
RT: 19.534 min Scan# 1689
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:	202	Resp:	2462
Ion Ratio	Lower	Upper	
202	100		
200	21.6	16.6	24.8
203	18.9	13.9	20.9



#31
Terphenyl-d14
Concen: 0.38 ng
RT: 19.748 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:	244	Resp:	2080
Ion Ratio	Lower	Upper	
244	100		
212	10.7	6.8	10.2#
122	11.8	6.6	9.8#









#32
Benzo(a)anthracene
Concen: 0.37 ng
RT: 21.291 min Scan# 1905
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:228 Resp: 2057
Ion Ratio Lower Upper
228 100
226 30.2 22.9 34.3
229 23.6 16.2 24.2



#33
Chrysene
Concen: 0.38 ng
RT: 21.333 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:228 Resp: 2916
Ion Ratio Lower Upper
228 100
226 33.5 24.3 36.5
229 21.6 16.4 24.6









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.40 ng
RT: 21.217 min Scan# 1898
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:149 Resp: 632
Ion Ratio Lower Upper
149 100
167 42.1 28.3 42.5
279 7.8 5.5 8.3



#35
Perylene-d12
Concen: 0.40 ng
RT: 23.556 min Scan# 2346
Delta R.T. -0.004 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:264 Resp: 2315
Ion Ratio Lower Upper
264 100
260 27.9 21.4 32.2
265 80.1 40.9 61.3#









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.39 ng
RT: 25.829 min Scan# 3010
Delta R.T. -0.004 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:276 Resp: 3666
Ion Ratio Lower Upper
276 100
138 9.8 10.2 15.4#
277 24.9 20.4 30.6



#37
Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 22.878 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:252 Resp: 3127
Ion Ratio Lower Upper
252 100
253 33.0 21.5 32.3#
125 15.6 6.6 10.0#









#38
Benzo(k)fluoranthene
Concen: 0.35 ng
RT: 22.923 min Scan# 2161
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06




Instrument :
BNA_N
ClientSampleId :
ICVBN060321


Tgt Ion:252 Resp: 3307
Ion Ratio Lower Upper
252 100
253 32.6 22.7 34.1
125 10.1 6.4 9.6#



#39
Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.460 min Scan# 2318
Delta R.T. -0.000 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:252 Resp: 2832
Ion Ratio Lower Upper
252 100
253 41.5 25.1 37.7#
125 25.6 9.4 14.0#









#40
Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 25.846 min Scan# 3015
Delta R.T. -0.003 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Instrument :
BNA_N
ClientSampleId :
ICVBN060321

Tgt Ion:278 Resp: 2944
Ion Ratio Lower Upper
278 100
139 13.7 8.6 13.0#
279 33.6 22.8 34.2



#41
Benzo(g,h,i)perylene
Concen: 0.39 ng
RT: 26.517 min Scan# 3211
Delta R.T. -0.003 min
Lab File: BN014834.D
Acq: 03 Jun 2021 20:06

Tgt Ion:276 Resp: 3484
Ion Ratio Lower Upper
276 100
277 27.9 20.9 31.3
138 15.0 10.3 15.5

