

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060321\
 Data File : BN014830.D
 Acq On : 03 Jun 2021 17:43
 Operator : CG/JU
 Sample : SSTDIC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.8

Quant Time: Jun 03 19:10:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 03 17:56:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.709	152	334	0.400	ng	0.00
7) Naphthalene-d8	10.482	136	1192	0.400	ng	0.00
13) Acenaphthene-d10	14.345	164	811	0.400	ng	0.00
19) Phenanthrene-d10	17.103	188	1891	0.400	ng	0.00
29) Chrysene-d12	21.302	240	2449	0.400	ng	# 0.00
35) Perylene-d12	23.556	264	2245	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.308	112	509	0.587	ng	0.00
5) Phenol-d6	6.889	99	702	0.611	ng	0.00
8) Nitrobenzene-d5	8.860	82	766	0.755	ng	0.00
11) 2-Methylnaphthalene-d10	12.083	152	1680	0.749	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	325	0.762	ng	-0.01
15) 2-Fluorobiphenyl	12.962	172	2466	0.832	ng	-0.01
27) Fluoranthene-d10	19.137	212	4978	0.716	ng	0.00
31) Terphenyl-d14	19.743	244	5204	0.891	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.251	88	348	0.736	ng	# 93
3) n-Nitrosodimethylamine	3.679	42	63	0.655	ng	# 29
6) bis(2-Chloroethyl)ether	7.138	93	682	0.720	ng	96
9) Naphthalene	10.533	128	2633	0.800	ng	96
10) Hexachlorobutadiene	10.825	225	758	0.842	ng	# 100
12) 2-Methylnaphthalene	12.158	142	1832	0.791	ng	96
16) Acenaphthylene	14.066	152	2290	0.744	ng	97
17) Acenaphthene	14.408	154	1792	0.789	ng	99
18) Fluorene	15.398	166	2412	0.781	ng	100
20) 4,6-Dinitro-2-methylph...	15.499	198	85	0.368	ng	90
21) 4-Bromophenyl-phenylether	16.299	248	989	0.787	ng	95
22) Hexachlorobenzene	16.409	284	1241	0.837	ng	99
23) Atrazine	16.567	200	495	0.635	ng	96
24) Pentachlorophenol	16.762	266	166	0.481	ng	# 83
25) Phenanthrene	17.139	178	4118	0.770	ng	99
26) Anthracene	17.236	178	3239	0.731	ng	97
28) Fluoranthene	19.167	202	5435	0.747	ng	99
30) Pyrene	19.534	202	5399	0.768	ng	100
32) Benzo(a)anthracene	21.280	228	4633	0.677	ng	98
33) Chrysene	21.333	228	6353	0.763	ng	99
34) Bis(2-ethylhexyl)phtha...	21.217	149	1248	0.643	ng	97
36) Indeno(1,2,3-cd)pyrene	25.832	276	7085	0.767	ng	99
37) Benzo(b)fluoranthene	22.875	252	6334	0.777	ng	99
38) Benzo(k)fluoranthene	22.923	252	7211	0.821	ng	98
39) Benzo(a)pyrene	23.457	252	5626	0.788	ng	# 97
40) Dibenzo(a,h)anthracene	25.846	278	5829	0.759	ng	98
41) Benzo(g,h,i)perylene	26.517	276	6681	0.784	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060321\
Data File : BN014830.D
Acq On : 03 Jun 2021 17:43
Operator : CG/JU
Sample : SSTDIC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Quant Time: Jun 03 19:10:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 03 17:56:36 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.709 min Scan# 415
Delta R.T. 0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 334
Ion Ratio Lower Upper
152 100
150 138.4 123.4 185.2
115 83.8 73.2 109.8



#2
1,4-Dioxane
Concen: 0.736 ng
RT: 3.251 min Scan# 40
Delta R.T. -0.012 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion: 88 Resp: 348
Ion Ratio Lower Upper
88 100
58 58.3 44.7 67.1
43 19.0 21.0 31.6#









#3
n-Nitrosodimethylamine
Concen: 0.655 ng
RT: 3.679 min Scan# 76
Delta R.T. 0.024 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion: 42 Resp: 63
Ion Ratio Lower Upper
42 100
74 134.9 0.0 0.0#
44 28.6 80.0 120.0#



#4
2-Fluorophenol
Concen: 0.587 ng
RT: 5.308 min Scan# 213
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion: 112 Resp: 509
Ion Ratio Lower Upper
112 100
64 56.4 43.6 65.4
63 28.7 25.0 37.6









#5
Phenol-d6
Concen: 0.611 ng
RT: 6.889 min Scan# 346
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Instrument :
BNA_N
ClientSampled :
SSTDIC0.8

Tgt Ion: 99 Resp: 702




Ion	Ratio	Lower	Upper
99	100		
42	11.1	10.2	15.2
71	44.4	32.4	48.6




#6
bis(2-Chloroethyl)ether
Concen: 0.720 ng
RT: 7.138 min Scan# 367
Delta R.T. -0.012 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion: 93 Resp: 682

Ion	Ratio	Lower	Upper
93	100		
63	58.2	47.5	71.3
95	30.8	28.5	42.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.482 min Scan# 637
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	1192
Ion	Ratio	Lower	Upper
136	100		
137	16.6	13.2	19.8
54	13.3	9.0	13.4
68	11.9	8.8	13.2



#8
Nitrobenzene-d5
Concen: 0.755 ng
RT: 8.860 min Scan# 509
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	82	Resp:	766
Ion	Ratio	Lower	Upper
82	100		
128	43.8	34.2	51.2
54	44.1	38.4	57.6









#9
Naphthalene
Concen: 0.800 ng
RT: 10.533 min Scan# 641
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	128	Resp:	2633
Ion	Ratio	Lower	Upper
128	100		
129	14.0	13.0	19.4
127	17.1	15.0	22.6



#10
Hexachlorobutadiene
Concen: 0.842 ng
RT: 10.825 min Scan# 664
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	225	Resp:	758
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.3	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.749 ng
RT: 12.083 min Scan# 847
Delta R.T. -0.005 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampled :
SSTDIC0.8


Tgt Ion:152 Resp: 1680
Ion Ratio Lower Upper
152 100
151 21.0 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.791 ng
RT: 12.158 min Scan# 864
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:142 Resp: 1832
Ion Ratio Lower Upper
142 100
141 87.7 72.9 109.3
115 49.7 41.8 62.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.345 min Scan# 1087
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:164 Resp: 811
Ion Ratio Lower Upper
164 100
162 102.9 83.2 124.8
160 49.6 41.8 62.6



#14
2,4,6-Tribromophenol
Concen: 0.762 ng
RT: 15.842 min Scan# 1205
Delta R.T. -0.013 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:330 Resp: 325
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 24.9 21.8 32.6









#15
2-Fluorobiphenyl
Concen: 0.832 ng
RT: 12.962 min Scan# 978
Delta R.T. -0.013 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	172	Resp:	2466
Ion	Ratio	Lower	Upper
172	100		
171	38.9	30.9	46.3
170	27.4	21.4	32.0



#16
Acenaphthylene
Concen: 0.744 ng
RT: 14.066 min Scan# 1065
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	152	Resp:	2290
Ion	Ratio	Lower	Upper
152	100		
151	20.2	17.1	25.7
153	13.6	9.9	14.9









#17
Acenaphthene
Concen: 0.789 ng
RT: 14.408 min Scan# 1092
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	154	Resp:	1792
Ion	Ratio	Lower	Upper
154	100		
153	114.2	91.8	137.8
152	58.6	46.0	69.0



#18
Fluorene
Concen: 0.781 ng
RT: 15.398 min Scan# 1170
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	166	Resp:	2412
Ion	Ratio	Lower	Upper
166	100		
165	99.6	79.9	119.9
167	13.5	10.5	15.7









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.103 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	188	Resp:	1891
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	10.9	8.2	12.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.368 ng
RT: 15.499 min Scan# 1178
Delta R.T. -0.038 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	198	Resp:	85
Ion	Ratio	Lower	Upper
198	100		
51	101.4	96.4	144.6
105	91.7	72.3	108.5









#21
4-Bromophenyl-phenylether
Concen: 0.787 ng
RT: 16.299 min Scan# 1242
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampled :
SSTDICC0.8


Tgt Ion:	248	Resp:	989
Ion	Ratio	Lower	Upper
248	100		
250	101.0	77.5	116.3
141	58.0	50.2	75.4



#22
Hexachlorobenzene
Concen: 0.837 ng
RT: 16.409 min Scan# 1251
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	284	Resp:	1241
Ion	Ratio	Lower	Upper
284	100		
142	25.9	20.3	30.5
249	32.6	26.4	39.6









#23
Atrazine
Concen: 0.635 ng
RT: 16.567 min Scan# 1264
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	200	Resp:	495
Ion	Ratio	Lower	Upper
200	100		
173	38.2	29.1	43.7
215	59.7	45.1	67.7



#24
Pentachlorophenol
Concen: 0.481 ng
RT: 16.762 min Scan# 1280
Delta R.T. -0.012 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	266	Resp:	166
Ion	Ratio	Lower	Upper
266	100		
264	61.4	44.2	66.2
268	72.9	44.2	66.2









#25
Phenanthrene
Concen: 0.770 ng
RT: 17.139 min Scan# 1311
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	178	Resp:	4118
Ion	Ratio	Lower	Upper
178	100		
176	19.5	16.0	24.0
179	15.2	12.6	19.0



#26
Anthracene
Concen: 0.731 ng
RT: 17.236 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	178	Resp:	3239
Ion	Ratio	Lower	Upper
178	100		
176	19.0	14.2	21.2
179	16.1	13.7	20.5









#27
Fluoranthene-d10
Concen: 0.716 ng
RT: 19.137 min Scan# 1609
Delta R.T. -0.005 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	212	Resp:	4978
Ion	Ratio	Lower	Upper
212	100		
106	9.6	7.4	11.2
104	5.5	4.1	6.1



#28
Fluoranthene
Concen: 0.747 ng
RT: 19.167 min Scan# 1615
Delta R.T. -0.005 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:	202	Resp:	5435
Ion	Ratio	Lower	Upper
202	100		
101	8.3	7.0	10.6
203	17.1	13.5	20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.302 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:240 Resp: 2449
Ion Ratio Lower Upper
240 100
120 9.5 5.2 7.8#
236 28.8 21.6 32.4



#30
Pyrene
Concen: 0.768 ng
RT: 19.534 min Scan# 1689
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:202 Resp: 5399
Ion Ratio Lower Upper
202 100
200 20.7 16.6 24.8
203 17.6 13.9 20.9









#31
Terphenyl-d14
Concen: 0.891 ng
RT: 19.743 min Scan# 1731
Delta R.T. -0.005 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:244 Resp: 5204
Ion Ratio Lower Upper
244 100
212 8.8 6.8 10.2
122 9.4 6.6 9.8



#32
Benzo(a)anthracene
Concen: 0.677 ng
RT: 21.280 min Scan# 1904
Delta R.T. -0.011 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:228 Resp: 4633
Ion Ratio Lower Upper
228 100
226 29.7 22.9 34.3
229 20.5 16.2 24.2









#33
Chrysene
Concen: 0.763 ng
RT: 21.333 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	228	Resp:	6353
Ion	Ratio	Lower	Upper
228	100		
226	30.9	24.3	36.5
229	20.5	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.643 ng
RT: 21.217 min Scan# 1898
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43


Tgt Ion:	149	Resp:	1248
Ion	Ratio	Lower	Upper
149	100		
167	37.5	28.3	42.5
279	7.9	5.5	8.3








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.556 min Scan# 2346
Delta R.T. -0.004 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8







Tgt Ion:264 Resp: 2245
Ion Ratio Lower Upper
264 100
260 28.4 21.4 32.2
265 83.1 40.9 61.3#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.767 ng
RT: 25.832 min Scan# 3011
Delta R.T. -0.000 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:276 Resp: 7085
Ion Ratio Lower Upper
276 100
138 11.9 10.2 15.4
277 25.2 20.4 30.6





#37

Benzo(b)fluoranthene

Concen: 0.777 ng

RT: 22.875 min Scan# 2147

Delta R.T. -0.003 min

Lab File: BN014830.D

Acq: 03 Jun 2021 17:43





Tgt Ion:	252	Resp:	6334
Ion	Ratio	Lower	Upper
252	100		
253	27.0	21.5	32.3
125	8.8	6.6	10.0

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8



#38

Benzo(k)fluoranthene

Concen: 0.821 ng




RT: 22.923 min Scan# 2161


Delta R.T. -0.000 min

Lab File: BN014830.D

Acq: 03 Jun 2021 17:43

Tgt Ion:	252	Resp:	7211
Ion	Ratio	Lower	Upper
252	100		
253	27.3	22.7	34.1
125	8.1	6.4	9.6









#39
Benzo(a)pyrene
Concen: 0.788 ng
RT: 23.457 min Scan# 2317
Delta R.T. -0.004 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 5626
Ion Ratio Lower Upper
252 100
253 31.8 25.1 37.7
125 14.6 9.4 14.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.759 ng
RT: 25.846 min Scan# 3015
Delta R.T. -0.004 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Tgt Ion:278 Resp: 5829
Ion Ratio Lower Upper
278 100
139 11.1 8.6 13.0
279 27.0 22.8 34.2





#41
Benzo(g,h,i)perylene
Concen: 0.784 ng
RT: 26.517 min Scan# 3211
Delta R.T. -0.004 min
Lab File: BN014830.D
Acq: 03 Jun 2021 17:43

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:	276	Resp:	6681
Ion	Ratio	Lower	Upper
276	100		
277	25.7	20.9	31.3
138	12.4	10.3	15.5

