

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060324\
 Data File : BN031766.D
 Acq On : 03 Jun 2024 19:55
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020271

Quant Time: Jun 04 01:20:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 29 15:13:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.681	152	682454	20.000	ng/u1	0.00
20) Naphthalene-d8	10.463	136	2981827	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.322	164	1740624	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.069	188	2862342	20.000	ng/u1	0.00
79) Chrysene-d12	21.310	240	2057716	20.000	ng/u1	0.00
88) Perylene-d12	24.251	264	2381156	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	111357	6.750	ng/uL	0.00
4) Pyridine-d5	3.593	84	844168	18.179	ng/u1	0.00
7) Phenol-d5	6.852	99	1141746	19.515	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.022	67	633165	18.502	ng/u1	0.00
11) 2-Chlorophenol-d4	7.216	132	873827	19.575	ng/u1	0.00
15) 4-Methylphenol-d8	8.393	113	902866	18.932	ng/u1	0.00
21) Nitrobenzene-d5	8.840	128	448451	19.670	ng/u1	0.00
24) 2-Nitrophenol-d4	9.557	143	482352	20.481	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.087	165	861799	19.601	ng/u1	0.00
31) 4-Chloroaniline-d4	10.604	131	1149253	17.605	ng/u1	0.00
46) Dimethylphthalate-d6	13.740	166	2222251	18.326	ng/u1	0.00
49) Acenaphthylene-d8	14.016	160	2626148	18.834	ng/u1	0.00
54) 4-Nitrophenol-d4	14.534	143	370051	15.241	ng/u1	0.00
60) Fluorene-d10	15.322	176	1807772	17.619	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.445	200	210402	14.624	ng/u1	0.00
73) Anthracene-d10	17.169	188	2353703	19.207	ng/u1	0.00
81) Pyrene-d10	19.469	212	2288196	20.780	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.045	264	2138290	18.806	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.228	88	119297	6.759	ng/uL	97
5) Pyridine	3.617	79	840866	18.110	ng/u1	95
6) Benzaldehyde	6.834	77	426343	15.537	ng/u1	97
8) Phenol	6.881	94	1158871	19.463	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.116	93	928766	18.864	ng/u1	97
12) 2-Chlorophenol	7.246	128	913802	19.561	ng/u1	99
13) 2-Methylphenol	8.122	108	879957	19.135	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.210	45	1008984	17.803	ng/u1	98
16) Acetophenone	8.505	105	1381561	19.115	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.493	70	669540	17.877	ng/u1	97
18) 4-Methylphenol	8.452	108	941752	18.891	ng/u1	96
19) Hexachloroethane	8.752	117	350054	17.695	ng/u1	91
22) Nitrobenzene	8.881	77	1036403	19.080	ng/u1	96
23) Isophorone	9.405	82	1933205	17.349	ng/u1	99
25) 2-Nitrophenol	9.587	139	521254	20.258	ng/u1	98
26) 2,4-Dimethylphenol	9.646	107	924445	18.469	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.887	93	1261578	18.166	ng/u1	99
29) 2,4-Dichlorophenol	10.116	162	848410	19.543	ng/u1	95
30) Naphthalene	10.516	128	2934640	19.016	ng/u1	100
32) 4-Chloroaniline	10.634	127	1129029	17.814	ng/u1	100
33) Hexachlorobutadiene	10.793	225	491172	18.838	ng/u1	98
34) Caprolactam	11.410	113	249740	14.738	ng/u1	96
35) 4-Chloro-3-methylphenol	11.757	107	907022	17.874	ng/u1	98
36) 2-Methylnaphthalene	12.134	142	1956016	18.543	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060324\
 Data File : BN031766.D
 Acq On : 03 Jun 2024 19:55
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020271


Quant Time: Jun 04 01:20:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 29 15:13:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.351	142	1928577	18.222	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.504	216	970794	21.089	ng/ul	98
40) Hexachlorocyclopentadiene	12.475	237	591016	22.276	ng/ul	96
41) 2,4,6-Trichlorophenol	12.751	196	630218	20.170	ng/ul	97
42) 2,4,5-Trichlorophenol	12.822	196	646000	19.026	ng/ul	99
43) 1,1'-Biphenyl	13.157	154	2448294	19.997	ng/ul	98
44) 2-Chloronaphthalene	13.193	162	1888246	19.691	ng/ul	98
45) 2-Nitroaniline	13.404	65	515505	18.545	ng/ul	92
47) Dimethylphthalate	13.787	163	2275558	18.349	ng/ul	99
48) 2,6-Dinitrotoluene	13.910	165	457925	18.576	ng/ul	98
50) Acenaphthylene	14.045	152	3015993	19.139	ng/ul	100
51) 3-Nitroaniline	14.234	138	483319	18.338	ng/ul	99
52) Acenaphthene	14.387	153	1956706	18.767	ng/ul	97
53) 2,4-Dinitrophenol	14.445	184	148556	11.937	ng/ul	99
55) 4-Nitrophenol	14.545	109	268198	14.891	ng/ul	98
56) Dibenzofuran	14.728	168	2593033	18.305	ng/ul	98
57) 2,4-Dinitrotoluene	14.698	165	603402	16.682	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.951	232	487117	17.151	ng/ul	97
59) Diethylphthalate	15.157	149	2235964	17.473	ng/ul	99
61) Fluorene	15.375	166	2033524	17.827	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.375	204	1013251	18.339	ng/ul	99
63) 4-Nitroaniline	15.398	138	457764	18.245	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.463	198	229992	15.037	ng/ul	97
67) N-Nitrosodiphenylamine	15.587	169	1718350	21.779	ng/ul	98
68) 4-Bromophenyl-phenylether	16.269	248	592398	21.360	ng/ul	94
69) Hexachlorobenzene	16.375	284	624454	19.866	ng/ul	97
70) Atrazine	16.545	200	479780	19.260	ng/ul	98
71) Pentachlorophenol	16.722	266	360345	17.918	ng/ul	86
72) Phenanthrene	17.116	178	2808946	19.264	ng/ul	99
74) Anthracene	17.204	178	2872499	19.442	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.116	216	940179	25.519	ng/uL	97
76) Pentachlorobenzene	14.640	250	828751	23.095	ng/uL	98
77) Carbazole	17.475	167	2463362	19.553	ng/ul	99
78) Di-n-butylphthalate	18.045	149	3353273	19.347	ng/ul	100
80) Fluoranthene	19.133	202	2796197	21.248	ng/ul	99
82) Pyrene	19.498	202	2701624	20.010	ng/ul	99
83) Butylbenzylphthalate	20.404	149	1288549	20.206	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.216	252	594216	15.059	ng/ul	93
85) Benzo(a)anthracene	21.286	228	2611646	18.962	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.216	149	1860373	20.013	ng/ul	98
87) Chrysene	21.351	228	2422380	18.829	ng/ul	98
89) Di-n-octyl phthalate	22.333	149	3045147	19.102	ng/ul	100
90) Benzo(b)fluoranthene	23.316	252	2589938	18.197	ng/ul	99
91) Benzo(k)fluoranthene	23.374	252	2634205	18.861	ng/ul	99
93) Benzo(a)pyrene	24.110	252	2466314	18.502	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.557	276	3068178	19.974	ng/ul	98
95) Dibenzo(a,h)anthracene	27.615	278	2526072	20.143	ng/ul	98
96) Benzo(g,h,i)perylene	28.592	276	2455898	20.113	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 04 01:20:05 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 29 15:13:07 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 6.759 ng/uL
 RT: 3.228 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55




Instrument :
 BNA_N
 ClientSampleId :
 SST020271


Tgt Ion: 88 Resp: 119297
 Ion Ratio Lower Upper
 88 100
 43 22.5 18.5 27.7
 58 63.0 52.6 78.8



#5
 Pyridine
 Concen: 18.110 ng/uL
 RT: 3.617 min Scan# 107
 Delta R.T. -0.000 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55

Tgt Ion: 79 Resp: 840866
 Ion Ratio Lower Upper
 79 100
 52 61.0 53.0 79.6
 51 28.2 23.3 34.9









#6
Benzaldehyde
Concen: 15.537 ng/ul
RT: 6.834 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion: 77 Resp: 426343
Ion Ratio Lower Upper
77 100
105 104.2 80.2 120.4
106 98.8 76.8 115.2



#8
Phenol
Concen: 19.463 ng/ul
RT: 6.881 min Scan# 662
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion: 94 Resp: 1158871
Ion Ratio Lower Upper
94 100
65 24.6 20.3 30.5
66 33.2 25.8 38.8









#10
Bis(2-Chloroethyl)ether
Concen: 18.864 ng/ul
RT: 7.116 min Scan# 702
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion: 93 Resp: 928766
Ion Ratio Lower Upper
93 100
63 67.7 56.8 85.2
95 32.2 26.0 39.0



#12
2-Chlorophenol
Concen: 19.561 ng/ul
RT: 7.246 min Scan# 724
Delta R.T. -0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:128 Resp: 913802
Ion Ratio Lower Upper
128 100
64 42.1 33.9 50.9
130 32.6 26.3 39.5









#13
2-Methylphenol
Concen: 19.135 ng/ul
RT: 8.122 min Scan# 873
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:108 Resp: 879957
Ion Ratio Lower Upper
108 100
107 90.4 71.6 107.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.803 ng/ul
RT: 8.210 min Scan# 888
Delta R.T. -0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion: 45 Resp: 1008984
Ion Ratio Lower Upper
45 100
77 18.3 13.6 20.4
79 13.4 10.6 15.8









#16
Acetophenone
Concen: 19.115 ng/ul
RT: 8.505 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:105 Resp: 1381561
Ion Ratio Lower Upper
105 100
77 74.1 58.6 87.8
51 21.2 17.2 25.8



#17
N-Nitroso-di-n-propylamine
Concen: 17.877 ng/ul
RT: 8.493 min Scan# 936
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion: 70 Resp: 669540
Ion Ratio Lower Upper
70 100
42 45.3 37.3 55.9
101 10.3 7.8 11.6
130 22.7 16.4 24.6









#18
4-Methylphenol
Concen: 18.891 ng/ul
RT: 8.452 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:108 Resp: 941752
Ion Ratio Lower Upper
108 100
107 116.7 89.6 134.4



#19
Hexachloroethane
Concen: 17.695 ng/ul
RT: 8.752 min Scan# 980
Delta R.T. -0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:117 Resp: 350054
Ion Ratio Lower Upper
117 100
201 98.9 71.6 107.4
199 61.9 45.6 68.4









#22
Nitrobenzene
Concen: 19.080 ng/ul
RT: 8.881 min Scan# 1002
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion: 77 Resp: 1036403
Ion Ratio Lower Upper
77 100
123 48.7 36.6 55.0
65 13.0 10.2 15.4



#23
Isophorone
Concen: 17.349 ng/ul
RT: 9.405 min Scan# 1091
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion: 82 Resp: 1933205
Ion Ratio Lower Upper
82 100
95 6.7 5.5 8.3
138 17.9 14.1 21.1









#25
2-Nitrophenol
Concen: 20.258 ng/ul
RT: 9.587 min Scan# 1122
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:139 Resp: 521254
Ion Ratio Lower Upper
139 100
65 45.9 38.3 57.5
109 27.7 22.1 33.1



#26
2,4-Dimethylphenol
Concen: 18.469 ng/ul
RT: 9.646 min Scan# 1132
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:107 Resp: 924445
Ion Ratio Lower Upper
107 100
121 50.2 40.8 61.2
122 84.3 70.4 105.6





#27

Bis(2-Chloroethoxy)methane

Concen: 18.166 ng/ul

RT: 9.887 min Scan# 1173

Delta R.T. -0.006 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Instrument :

BNA_N

ClientSampleId :

SSTD020271





Tgt Ion: 93 Resp: 1261578

Ion Ratio Lower Upper

93 100

95 32.1 26.1 39.1

123 11.1 9.0 13.6



#29

2,4-Dichlorophenol

Concen: 19.543 ng/ul

RT: 10.116 min Scan# 1212

Delta R.T. -0.000 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55




Tgt Ion: 162 Resp: 848410


Ion Ratio Lower Upper

162 100

164 61.8 52.7 79.1

98 34.9 30.5 45.7





#30

Naphthalene

Concen: 19.016 ng/ul

RT: 10.516 min Scan# 1300

Delta R.T. -0.000 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Instrument :

BNA_N

ClientSampleId :

SSTD020271





Tgt Ion:128 Resp: 2934640

Ion Ratio Lower Upper

128 100

129 10.7 8.5 12.7

127 13.0 10.3 15.5



#32

4-Chloroaniline

Concen: 17.814 ng/ul

RT: 10.634 min Scan# 1300

Delta R.T. 0.006 min

Lab File: BN031766.D




Acq: 03 Jun 2024 19:55


Tgt Ion:127 Resp: 1129029

Ion Ratio Lower Upper

127 100


129 31.1 25.0 37.4








#33
Hexachlorobutadiene
Concen: 18.838 ng/ul
RT: 10.793 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Instrument :
BNA_N
ClientSampleId :
SSTD020271







Tgt Ion:225 Resp: 491172
Ion Ratio Lower Upper
225 100
223 64.1 52.3 78.5
227 64.2 52.5 78.7



#34
Caprolactam
Concen: 14.738 ng/ul
RT: 11.410 min Scan# 1432
Delta R.T. 0.012 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:113 Resp: 249740
Ion Ratio Lower Upper
113 100
55 141.4 117.2 175.8
56 106.4 89.4 134.0









#35
4-Chloro-3-methylphenol
Concen: 17.874 ng/ul
RT: 11.757 min Scan# 1491
Delta R.T. 0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:107 Resp: 907022
Ion Ratio Lower Upper
107 100
144 26.0 20.5 30.7
142 78.1 64.0 96.0



#36
2-Methylnaphthalene
Concen: 18.543 ng/ul
RT: 12.134 min Scan# 1555
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:142 Resp: 1956016
Ion Ratio Lower Upper
142 100
141 88.8 70.1 105.1









#37
1-Methylnaphthalene
Concen: 18.222 ng/ul
RT: 12.351 min Scan# 1592
Delta R.T. -0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:142 Resp: 1928577
Ion Ratio Lower Upper
142 100
141 91.0 71.7 107.5
116 3.9 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.089 ng/ul
RT: 12.504 min Scan# 1618
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:216 Resp: 970794
Ion Ratio Lower Upper
216 100
214 79.1 61.7 92.5
179 20.2 16.7 25.1
108 17.1 14.5 21.7









#40
Hexachlorocyclopentadiene
Concen: 22.276 ng/ul
RT: 12.475 min Scan# 1613
Delta R.T. -0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:237 Resp: 591016
Ion Ratio Lower Upper
237 100
235 63.0 47.8 71.6
272 13.4 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 20.170 ng/ul
RT: 12.751 min Scan# 1660
Delta R.T. 0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:196 Resp: 630218
Ion Ratio Lower Upper
196 100
198 98.7 76.4 114.6
200 29.9 24.2 36.4









#42
2,4,5-Trichlorophenol
Concen: 19.026 ng/ul
RT: 12.822 min Scan# 1672
Delta R.T. 0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:196 Resp: 646000
Ion Ratio Lower Upper
196 100
198 91.4 73.8 110.6
200 30.9 25.6 38.4



#43
1,1'-Biphenyl
Concen: 19.997 ng/ul
RT: 13.157 min Scan# 1729
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:154 Resp: 2448294
Ion Ratio Lower Upper
154 100
153 39.4 30.6 46.0
76 13.1 11.1 16.7





#44

2-Chloronaphthalene

Concen: 19.691 ng/ul

RT: 13.193 min Scan# 1735

Delta R.T. -0.006 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Instrument :

BNA_N

ClientSampleId :

SSTD020271





Tgt Ion:162 Resp: 1888246

Ion Ratio Lower Upper

162 100

164 32.0 27.0 40.4

127 40.3 31.8 47.6



#45

2-Nitroaniline

Concen: 18.545 ng/ul

RT: 13.404 min Scan# 1771

Delta R.T. -0.000 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55




Tgt Ion: 65 Resp: 515505


Ion Ratio Lower Upper

65 100

92 77.8 58.8 88.2

138 118.4 86.2 129.4









#47
Dimethylphthalate
Concen: 18.349 ng/ul
RT: 13.787 min Scan# 1836
Delta R.T. -0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:163 Resp: 2275558
Ion Ratio Lower Upper
163 100
194 4.2 3.2 4.8
164 9.9 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 18.576 ng/ul
RT: 13.910 min Scan# 1857
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:165 Resp: 457925
Ion Ratio Lower Upper
165 100
89 53.7 44.6 67.0
121 20.9 17.3 25.9









#50
Acenaphthylene
Concen: 19.139 ng/ul
RT: 14.045 min Scan# 1880
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:152 Resp: 3015993
Ion Ratio Lower Upper
152 100
151 20.0 15.9 23.9
153 13.2 10.4 15.6



#51
3-Nitroaniline
Concen: 18.338 ng/ul
RT: 14.234 min Scan# 1912
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:138 Resp: 483319
Ion Ratio Lower Upper
138 100
108 11.3 9.2 13.8
92 115.8 94.0 141.0









#52
Acenaphthene
Concen: 18.767 ng/ul
RT: 14.387 min Scan# 1938
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:153 Resp: 1956706
Ion Ratio Lower Upper
153 100
152 49.0 37.8 56.6
154 90.7 69.8 104.8



#53
2,4-Dinitrophenol
Concen: 11.937 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. 0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:184 Resp: 148556
Ion Ratio Lower Upper
184 100
63 62.4 50.2 75.4
154 63.2 49.1 73.7









#55
4-Nitrophenol
Concen: 14.891 ng/ul
RT: 14.545 min Scan# 1965
Delta R.T. 0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:109 Resp: 268198
Ion Ratio Lower Upper
109 100
139 141.8 113.4 170.2
65 127.5 106.2 159.2



#56
Dibenzofuran
Concen: 18.305 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:168 Resp: 2593033
Ion Ratio Lower Upper
168 100
139 37.7 29.4 44.0









#57
 2,4-Dinitrotoluene
 Concen: 16.682 ng/ul
 RT: 14.698 min Scan# 1991
 Delta R.T. 0.006 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55

Instrument :
 BNA_N
 ClientSampleId :
 SST020271

Tgt Ion:165 Resp: 603402




Ion	Ratio	Lower	Upper
165	100		
63	39.3	32.2	48.2
182	2.8	2.3	3.5




#58
 2,3,4,6-Tetrachlorophenol
 Concen: 17.151 ng/ul
 RT: 14.951 min Scan# 2034
 Delta R.T. -0.000 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55

Tgt Ion:232 Resp: 487117

Ion	Ratio	Lower	Upper
232	100		
131	48.4	36.6	55.0
130	2.2	1.8	2.8
166	32.7	25.0	37.4









#59
 Diethylphthalate
 Concen: 17.473 ng/ul
 RT: 15.157 min Scan# 2069
 Delta R.T. -0.000 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55

Instrument : BNA_N
 ClientSampleId : SSTD020271

Tgt Ion:149 Resp: 2235964




Ion	Ratio	Lower	Upper
149	100		
177	21.9	17.1	25.7
150	12.3	10.0	15.0




#61
 Fluorene
 Concen: 17.827 ng/ul
 RT: 15.375 min Scan# 2106
 Delta R.T. -0.000 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55

Tgt Ion:166 Resp: 2033524

Ion	Ratio	Lower	Upper
166	100		
165	95.5	80.3	120.5
167	13.5	11.3	16.9









#62
4-Chlorophenyl-phenylether
Concen: 18.339 ng/ul
RT: 15.375 min Scan# 2106
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:204 Resp: 1013251
Ion Ratio Lower Upper
204 100
206 32.6 25.7 38.5
141 60.9 47.7 71.5



#63
4-Nitroaniline
Concen: 18.245 ng/ul
RT: 15.398 min Scan# 2110
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:138 Resp: 457764
Ion Ratio Lower Upper
138 100
92 53.1 41.0 61.4
108 73.8 60.1 90.1









#66
 4,6-Dinitro-2-methylphenol
 Concen: 15.037 ng/ul
 RT: 15.463 min Scan# 2121
 Delta R.T. 0.006 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55




Instrument :
 BNA_N
 ClientSampleId :
 SST020271


Tgt Ion:198 Resp: 229992
 Ion Ratio Lower Upper
 198 100
 182 4.4 3.0 4.6
 77 25.6 19.4 29.2



#67
 N-Nitrosodiphenylamine
 Concen: 21.779 ng/ul
 RT: 15.587 min Scan# 2142
 Delta R.T. -0.000 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55

Tgt Ion:169 Resp: 1718350
 Ion Ratio Lower Upper
 169 100
 168 67.8 52.7 79.1
 167 36.1 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 21.360 ng/ul
RT: 16.269 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:248 Resp: 592398
Ion Ratio Lower Upper
248 100
250 99.6 76.5 114.7
141 72.4 63.9 95.9



#69
Hexachlorobenzene
Concen: 19.866 ng/ul
RT: 16.375 min Scan# 2276
Delta R.T. 0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:284 Resp: 624454
Ion Ratio Lower Upper
284 100
142 36.2 31.1 46.7
249 31.6 25.7 38.5





#70

Atrazine

Concen: 19.260 ng/ul

RT: 16.545 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Instrument :

BNA_N

ClientSampleId :

SSTD020271





Tgt Ion:200 Resp: 479780

Ion Ratio Lower Upper

200 100

173 26.8 21.2 31.8

215 50.3 38.9 58.3



#71

Pentachlorophenol

Concen: 17.918 ng/ul

RT: 16.722 min Scan# 2335

Delta R.T. 0.006 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55




Tgt Ion:266 Resp: 360345


Ion Ratio Lower Upper

266 100

264 55.0 54.7 82.1

268 58.4 53.4 80.2









#72
Phenanthrene
Concen: 19.264 ng/ul
RT: 17.116 min Scan# 2402
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:178 Resp: 2808946
Ion Ratio Lower Upper
178 100
179 15.3 12.6 19.0
176 19.9 15.8 23.6



#74
Anthracene
Concen: 19.442 ng/ul
RT: 17.204 min Scan# 2417
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:178 Resp: 2872499
Ion Ratio Lower Upper
178 100
179 15.5 11.8 17.6
176 18.6 14.8 22.2









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 25.519 ng/uL
 RT: 13.116 min Scan# 1722
 Delta R.T. -0.000 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55

Instrument :
 BNA_N
 ClientSampleId :
 SST020271

Tgt Ion: 216 Resp: 940179




Ion	Ratio	Lower	Upper
216	100		
214	75.1	62.1	93.1
218	48.0	40.2	60.4




#76
 Pentachlorobenzene
 Concen: 23.095 ng/uL
 RT: 14.640 min Scan# 1981
 Delta R.T. -0.000 min
 Lab File: BN031766.D
 Acq: 03 Jun 2024 19:55

Tgt Ion: 250 Resp: 828751

Ion	Ratio	Lower	Upper
250	100		
248	63.5	49.8	74.6
252	63.4	52.3	78.5





#77

Carbazole

Concen: 19.553 ng/ul

RT: 17.475 min Scan# 2463

Delta R.T. -0.000 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Instrument :

BNA_N

ClientSampleId :

SSTD020271





Tgt Ion:167 Resp: 2463362

Ion Ratio Lower Upper

167 100

166 21.9 17.9 26.9

139 13.0 10.5 15.7



#78

Di-n-butylphthalate

Concen: 19.347 ng/ul

RT: 18.045 min Scan# 2560

Delta R.T. -0.000 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55




Tgt Ion:149 Resp: 3353273


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.0

104 5.0 4.1 6.1









#80
Fluoranthene
Concen: 21.248 ng/u1
RT: 19.133 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:202 Resp: 2796197
Ion Ratio Lower Upper
202 100
101 12.3 10.3 15.5
100 9.6 8.0 12.0



#82
Pyrene
Concen: 20.010 ng/u1
RT: 19.498 min Scan# 2807
Delta R.T. 0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:202 Resp: 2701624
Ion Ratio Lower Upper
202 100
101 14.8 12.6 18.8
100 12.5 10.2 15.2









#83
Butylbenzylphthalate
Concen: 20.206 ng/ul
RT: 20.404 min Scan# 2961
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:149 Resp: 1288549
Ion Ratio Lower Upper
149 100
91 72.5 61.0 91.4
206 20.6 16.5 24.7



#84
3,3'-Dichlorobenzidine
Concen: 15.059 ng/ul
RT: 21.216 min Scan# 3099
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:252 Resp: 594216
Ion Ratio Lower Upper
252 100
254 60.3 53.6 80.4
126 12.9 10.2 15.4





#85

Benzo(a)anthracene

Concen: 18.962 ng/ul

RT: 21.286 min Scan# 3112

Delta R.T. -0.000 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Instrument :

BNA_N

ClientSampleId :

SSTD020271





Tgt Ion:228 Resp: 2611646

Ion Ratio Lower Upper

228 100

229 20.7 16.7 25.1

226 27.7 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.013 ng/ul

RT: 21.216 min Scan# 3099

Delta R.T. -0.006 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55




Tgt Ion:149 Resp: 1860373


Ion Ratio Lower Upper

149 100

167 27.8 21.4 32.0

279 4.1 3.3 4.9





#87

Chrysene

Concen: 18.829 ng/ul

RT: 21.351 min Scan# 31

Delta R.T. -0.000 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Instrument :

BNA_N

ClientSampleId :

SSTD020271





Tgt Ion:228 Resp: 2422380

Ion Ratio Lower Upper

228 100

226 30.6 23.7 35.5

229 20.3 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 19.102 ng/ul




RT: 22.333 min Scan# 3289


Delta R.T. -0.006 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Tgt Ion:149 Resp: 3045147









#90
Benzo(b)fluoranthene
Concen: 18.197 ng/ul
RT: 23.316 min Scan# 3457
Delta R.T. 0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:252 Resp: 2589938
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.6
125 11.9 9.2 13.8



#91
Benzo(k)fluoranthene
Concen: 18.861 ng/ul
RT: 23.374 min Scan# 3466
Delta R.T. -0.000 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:252 Resp: 2634205
Ion Ratio Lower Upper
252 100
253 21.5 17.6 26.4
125 12.0 9.4 14.2









#93
Benzo(a)pyrene
Concen: 18.502 ng/ul
RT: 24.110 min Scan# 3591
Delta R.T. 0.006 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55




Instrument :
BNA_N
ClientSampleId :
SSTD020271


Tgt Ion:252 Resp: 2466314
Ion Ratio Lower Upper
252 100
253 22.4 16.6 24.8
125 13.2 9.8 14.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.974 ng/ul
RT: 27.557 min Scan# 4177
Delta R.T. 0.023 min
Lab File: BN031766.D
Acq: 03 Jun 2024 19:55

Tgt Ion:276 Resp: 3068178
Ion Ratio Lower Upper
276 100
138 24.0 18.1 27.1
277 23.2 18.2 27.2





#95

Dibenzo(a,h)anthracene

Concen: 20.143 ng/ul

RT: 27.615 min Scan# 4187

Delta R.T. 0.012 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Instrument :

BNA_N

ClientSampleId :

SSTD020271





Tgt Ion:278 Resp: 2526072

Ion Ratio Lower Upper

278 100

139 19.1 15.0 22.6

279 23.3 17.4 26.0



#96

Benzo(g,h,i)perylene

Concen: 20.113 ng/ul

RT: 28.592 min Scan# 4353

Delta R.T. 0.023 min

Lab File: BN031766.D

Acq: 03 Jun 2024 19:55

Tgt Ion:276 Resp: 2455898

Ion Ratio Lower Upper

276 100

138 25.2 20.9 31.3

277 22.8 18.0 27.0

