




Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN060325\  
Data File : BN037155.D  
Acq On : 03 Jun 2025 20:10  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 04 02:36:15 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jun 04 02:35:39 2025  
Response via : Initial Calibration



TIC: BN037155.D\data.ms

**(70) Pentachlorophenol (C)**




16.645min ( 0.000) 277.78 ng

response 233660

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 66.51  |
| 264.00 | 61.60  | 62.18  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN060325\  
Data File : BN037155.D  
Acq On : 03 Jun 2025 20:10  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 04 02:36:15 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jun 04 02:35:39 2025  
Response via : Initial Calibration



TIC: BN037155.D\data.ms

**(77) Benzidine**

19.251min ( 0.000) 0.00 ng

response 1091920

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 13.66  |
| 183.00 | 13.20  | 11.92  |
| 0.00   | 0.00   | 0.00   |