

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN060420\
 Data File : BN010757.D
 Acq On : 04 Jun 2020 10:05
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.411

Quant Time: Jun 04 12:06:01 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN060320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 03 16:40:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	1930	0.40	ng/ul	0.00
2) Naphthalene-d8	10.32	136	6158	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.19	164	3212	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.94	188	6515	0.40	ng/ul	0.00
16) Chrysene-d12	21.15	240	5465	0.40	ng/ul	0.00
20) Perylene-d12	23.30	264	5059	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.91	152	6241	0.43	ng/ul	0.00
14) Fluoranthene-d10	18.97	212	11804	0.42	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.37	128	7738	0.426	ng/ul	98
5) 2-Methylnaphthalene	11.99	142	4896	0.430	ng/ul	99
7) Acenaphthylene	13.90	152	6047	0.422	ng/ul	99
8) Acenaphthene	14.25	153	5120	0.427	ng/ul	99
9) Fluorene	15.25	166	5659	0.429	ng/ul	99
11) Pentachlorophenol	16.60	266	642	0.421	ng/ul	97
12) Phenanthrene	16.98	178	8932	0.421	ng/ul	99
13) Anthracene	17.07	178	7329	0.418	ng/ul	99
15) Fluoranthene	19.00	202	9942	0.418	ng/ul	97
17) Pyrene	19.37	202	9853	0.415	ng/ul	99
18) Benzo(a)anthracene	21.13	228	8149	0.420	ng/ul	98
19) Chrysene	21.18	228	10406	0.422	ng/ul	99
21) Benzo(b)fluoranthene	22.66	252	9232	0.410	ng/ul	97
22) Benzo(k)fluoranthene	22.71	252	10210	0.428	ng/ul	96
23) Benzo(a)pyrene	23.21	252	7790	0.409	ng/ul	95
24) Indeno(1,2,3-cd)pyrene	25.42	276	10769	0.413	ng/ul#	98
25) Dibenzo(a,h)anthracene	25.43	278	8838	0.415	ng/ul	98
26) Benzo(g,h,i)perylene	26.06	276	9743	0.404	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN060420\
Data File : BN010757.D
Acq On : 04 Jun 2020 10:05
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.411

Quant Time: Jun 04 12:06:01 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN060320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 03 16:40:15 2020
Response via : Initial Calibration

