


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060420\
 Data File : BN010756.D
 Acq On : 04 Jun 2020 09:29
 Operator : CG/JU
 Sample : DFTPP64
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP64

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN060320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 05 10:29:34 2020



AutoFind: Scans 2485, 2486, 2487; Background Corrected with Scan 2479

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.2	126893	PASS
68	69	0.00	2	1.5	3208	PASS
69	198	0.00	100	39.7	217514	PASS
70	69	0.00	2	0.5	1127	PASS
127	198	10	80	48.4	264938	PASS
197	198	0.00	2	0.2	1258	PASS
198	198	100	100	100.0	547541	PASS
199	198	5	9	6.8	37485	PASS
275	198	10	60	28.3	154936	PASS
365	198	1	100	4.2	23032	PASS
441	198	0.01	100	11.0	60002	PASS
442	198	50	100	72.4	396394	PASS
443	442	15	24	20.3	80562	PASS

Scan 2479 (17.369 min): BN010756.D
DFTPP64

m/z	Abundance
40.00	583.0
44.00	1098.0
73.00	875.0
96.00	334.0
192.80	500.0
207.00	2261.0
208.00	379.0
208.90	490.0
221.00	321.0
281.10	704.0
282.00	369.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
37.00	1219.0
38.00	2432.0
39.00	10862.0
40.00	1347.0
43.90	1401.0
49.10	904.0
50.00	35176.0
51.00	127888.0
52.00	6315.0
52.90	544.0
54.90	800.0
55.90	3364.0
57.00	9069.0
58.00	385.0
61.00	2896.0
62.00	2920.0
63.00	8597.0
64.00	1293.0
65.00	3691.0
66.10	554.0
68.00	4315.0
68.90	214016.0
70.00	1141.0
73.00	3913.0
74.00	25464.0
75.00	34568.0
76.00	13000.0
77.00	245632.0
78.00	18248.0
79.00	17752.0
80.00	14347.0
81.00	18128.0
82.00	4265.0
83.00	3736.0
83.80	560.0
85.00	3178.0
86.00	5993.0
86.90	2800.0
88.10	894.0
89.00	329.0
91.00	3317.0
91.90	5620.0
93.00	34240.0
93.90	2175.0
94.80	472.0
96.00	1421.0
97.00	710.0
98.00	27576.0
99.00	25632.0
100.00	1790.0
101.00	9016.0
102.10	784.0
103.00	3303.0
103.90	6178.0
105.00	5448.0
105.90	2241.0
107.00	74680.0
108.00	11430.0
110.00	128440.0
111.00	22032.0
111.90	2948.0
113.00	982.0
114.70	377.0
115.90	4621.0
117.00	84544.0
118.00	7120.0
118.90	838.0
120.00	669.0
122.00	6700.0
123.00	7536.0
124.00	4116.0
125.00	4045.0
126.10	1734.0
127.00	239936.0
128.00	20048.0
129.00	94840.0
130.00	9994.0
131.00	1399.0
132.00	622.0
133.10	565.0
134.00	2861.0
135.00	9071.0
136.00	3170.0
137.00	5455.0
138.80	518.0
139.90	1806.0
140.90	13810.0
142.00	4820.0
143.00	2574.0
143.90	1224.0
145.00	685.0
146.00	2643.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

147.00	7526.0
147.90	17840.0
149.00	4202.0
150.00	1471.0
151.10	1679.0
151.70	767.0
153.00	5144.0
154.00	3699.0
155.00	8384.0
156.00	11286.0
157.00	1830.0
158.00	3201.0
158.90	2172.0
159.90	4147.0
161.00	6493.0
161.90	1995.0
162.90	763.0
163.80	650.0
164.90	5188.0
166.00	3427.0
167.00	31216.0
168.00	24040.0
168.90	3207.0
170.00	1427.0
170.90	1476.0
172.00	2762.0
172.80	2828.0
174.00	6255.0
175.00	10642.0
176.10	3201.0
177.00	4714.0
178.00	1592.0
179.00	21888.0
180.00	17136.0
181.00	6992.0
181.90	937.0
182.80	880.0
183.90	2285.0
184.90	10006.0
186.00	75768.0
187.00	22664.0
187.90	3274.0
189.00	4622.0
189.90	1245.0
191.00	2392.0
192.00	6571.0
193.00	7155.0
193.70	893.0
194.20	1253.0
195.00	747.0
196.10	11460.0
197.00	3774.0
197.90	449152.0
199.00	30968.0
199.90	2188.0
201.70	1662.0
203.10	4435.0
204.00	20968.0
205.00	32408.0
206.00	139264.0
207.00	20336.0
208.00	6664.0
209.00	2478.0
210.00	2926.0
210.90	5070.0
211.90	1598.0
213.90	325.0
214.90	1672.0
216.00	2784.0
217.00	35592.0
218.00	5083.0
218.90	598.0
219.90	574.0
221.00	16208.0
221.90	4661.0
223.00	7623.0
224.00	74736.0
225.00	19544.0
226.00	2451.0
226.90	35120.0
228.00	6538.0
229.00	6853.0
230.00	1374.0
231.00	2501.0
232.00	350.0
233.00	613.0
234.00	2013.0
235.00	2753.0
236.00	1556.0
237.00	5065.0
238.00	613.0
239.10	1579.0
240.00	1651.0
240.90	2104.0
242.00	4076.0
243.00	3129.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

244.00	52848.0
245.00	8777.0
246.00	12302.0
247.00	2912.0
248.00	761.0
249.00	2348.0
250.00	366.0
250.90	645.0
251.90	671.0
253.00	1479.0
255.00	292672.0
256.00	43088.0
257.00	4553.0
258.00	17160.0
259.00	3348.0
260.00	633.0
260.90	389.0
263.90	556.0
265.00	7792.0
266.00	1101.0
270.00	356.0
271.00	1035.0
272.10	518.0
272.90	8127.0
273.90	20968.0
275.00	116328.0
276.00	17096.0
276.90	11563.0
278.00	2445.0
279.00	741.0
281.10	812.0
281.80	384.0
283.00	1909.0
284.10	947.0
285.00	1625.0
285.90	761.0
288.80	522.0
289.90	312.0
291.00	439.0
292.00	568.0
293.00	2538.0
293.80	800.0
295.20	1034.0
295.90	41664.0
297.00	5244.0
298.00	422.0
300.90	381.0
301.80	745.0
302.90	4369.0
304.00	1198.0
305.00	502.0
307.90	415.0
308.70	484.0
310.00	493.0
312.90	468.0
313.90	2464.0
315.00	5059.0
316.00	2055.0
320.90	861.0
321.90	586.0
323.00	10469.0
323.90	1568.0
325.80	484.0
326.90	1878.0
328.00	1332.0
331.80	698.0
333.00	867.0
334.00	6743.0
335.00	1527.0
340.10	473.0
341.00	1307.0
341.90	450.0
346.00	2161.0
346.90	876.0
352.00	2471.0
353.00	2845.0
354.10	3070.0
355.10	952.0
358.90	323.0
365.00	17016.0
365.90	2699.0
369.80	363.0
371.00	957.0
372.00	4045.0
373.10	1267.0
373.90	399.0
382.90	1475.0
389.90	803.0
390.90	717.0
391.90	375.0
402.10	1948.0
402.90	2740.0
403.80	1183.0
405.10	425.0
421.00	2375.0
422.00	1751.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

423.00	16920.0
424.00	8199.0
425.00	551.0
439.80	329.0
441.00	33184.0
442.00	247488.0
443.00	47168.0
444.00	3645.0
444.90	447.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
37.10	1572.0
38.00	3211.0
39.00	13942.0
40.00	1060.0
44.00	1730.0
44.80	476.0
49.00	1174.0
50.00	41200.0
51.00	157760.0
52.00	7754.0
52.90	317.0
54.90	1159.0
56.00	4725.0
57.00	11795.0
57.80	557.0
60.00	363.0
61.00	3056.0
62.00	3504.0
63.00	9108.0
63.90	1030.0
65.00	5194.0
67.00	512.0
68.00	5310.0
68.90	272768.0
70.10	1392.0
73.00	3354.0
74.00	33280.0
75.00	50408.0
76.10	17992.0
77.00	313088.0
78.00	21024.0
79.00	23920.0
80.00	17592.0
81.00	23856.0
81.90	5724.0
82.90	4099.0
83.90	746.0
85.00	3742.0
86.00	7882.0
86.90	2934.0
88.00	1058.0
89.10	715.0
91.00	5271.0
92.00	6612.0
93.00	41968.0
94.00	3288.0
95.00	796.0
96.00	1980.0
97.00	687.0
98.00	34128.0
99.00	28784.0
100.00	2766.0
101.00	13459.0
102.00	593.0
103.00	4347.0
104.00	9455.0
105.00	8892.0
106.20	2794.0
107.00	98480.0
108.00	15225.0
109.00	3433.0
110.00	174208.0
111.00	26400.0
111.90	2797.0
113.00	1122.0
114.80	574.0
116.00	6622.0
117.00	107912.0
118.00	6518.0
118.90	1063.0
119.90	1373.0
121.00	312.0
122.00	7281.0
123.00	10202.0
124.00	5040.0
124.90	4249.0
127.00	336832.0
128.00	26336.0
129.00	131968.0
129.90	11387.0
130.90	2730.0
132.00	738.0
132.80	587.0
134.00	4474.0
135.00	12017.0
136.00	5851.0
137.00	6618.0
138.00	1706.0
139.00	945.0
140.00	1929.0
141.00	21744.0
142.00	6227.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

142.90	4108.0
144.00	1009.0
144.90	710.0
146.00	4295.0
147.00	10304.0
148.00	28272.0
149.00	7079.0
150.00	1544.0
151.00	1767.0
151.60	1531.0
153.00	6122.0
154.00	3838.0
155.00	11173.0
156.00	15503.0
157.10	2636.0
158.00	3805.0
159.00	2775.0
159.90	6351.0
160.90	8454.0
162.00	3321.0
163.00	963.0
164.10	1782.0
165.00	7874.0
166.00	5490.0
167.00	41392.0
168.00	29408.0
168.90	3946.0
169.90	1296.0
171.00	1588.0
171.90	4264.0
173.00	4830.0
174.00	8529.0
175.00	14152.0
176.00	4423.0
177.00	6689.0
177.90	2777.0
178.90	31584.0
180.00	23760.0
181.00	9549.0
182.00	1548.0
182.70	834.0
184.00	3436.0
185.00	15162.0
186.00	106776.0
187.00	31360.0
188.00	4451.0
189.00	8681.0
190.00	1107.0
190.90	3248.0
191.90	10535.0
193.00	11655.0
194.00	2769.0
195.00	1225.0
196.00	20544.0
198.00	701440.0
198.90	49032.0
199.90	3163.0
201.30	2486.0
203.00	6800.0
204.00	29336.0
205.00	49976.0
206.00	211648.0
207.00	29016.0
208.00	8056.0
208.90	3528.0
210.30	3478.0
211.00	8878.0
211.80	1439.0
212.90	536.0
214.20	345.0
214.90	2208.0
216.00	5582.0
216.90	60632.0
218.00	7649.0
218.80	646.0
219.20	569.0
220.10	1237.0
221.00	24856.0
221.90	7266.0
222.90	11849.0
224.00	108952.0
225.00	28800.0
226.10	4130.0
227.00	53520.0
228.00	7050.0
229.00	11110.0
229.90	1240.0
231.00	4192.0
232.00	598.0
233.00	1240.0
234.10	2746.0
235.00	4146.0
236.00	2624.0
237.00	6424.0
237.80	812.0
239.00	1810.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

240.00	2016.0
241.00	3047.0
242.00	6979.0
243.00	5616.0
244.00	81352.0
245.00	11507.0
246.00	18880.0
247.00	4707.0
248.00	1130.0
248.90	2956.0
250.00	800.0
251.10	692.0
251.90	840.0
253.00	1933.0
255.00	455424.0
256.00	70896.0
256.90	5553.0
258.00	29808.0
259.00	5430.0
260.00	545.0
260.90	654.0
264.10	803.0
264.90	12075.0
266.10	1478.0
266.90	531.0
267.90	315.0
270.00	886.0
270.80	1062.0
271.90	1234.0
273.00	14335.0
273.90	33472.0
275.00	201728.0
276.00	23088.0
277.00	18128.0
278.00	3790.0
278.90	944.0
281.00	1165.0
282.00	497.0
283.10	2509.0
284.00	1210.0
285.10	3531.0
285.90	459.0
288.80	623.0
289.80	433.0
291.00	408.0
292.00	1099.0
293.00	4930.0
294.00	1691.0
295.00	1490.0
296.00	62824.0
297.00	10078.0
298.00	640.0
301.10	622.0
301.90	1401.0
303.00	7665.0
303.80	1524.0
304.90	563.0
307.80	992.0
309.10	347.0
309.90	884.0
311.10	325.0
311.80	326.0
312.90	433.0
313.90	2661.0
314.90	7940.0
316.00	5237.0
317.10	479.0
321.00	2239.0
322.00	780.0
323.00	16776.0
324.10	3046.0
324.90	801.0
325.80	561.0
326.90	3563.0
328.00	2397.0
329.20	310.0
331.90	1501.0
333.00	1966.0
334.00	11214.0
335.00	2755.0
335.80	356.0
336.10	391.0
339.10	439.0
341.00	2157.0
341.90	447.0
346.00	3507.0
346.80	833.0
352.00	4829.0
353.00	4051.0
354.00	5775.0
355.00	995.0
359.00	507.0
364.90	24928.0
365.90	3891.0
370.00	831.0
371.00	1560.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

372.10	9228.0
373.00	2198.0
382.90	3190.0
384.00	1232.0
390.10	1338.0
391.00	914.0
392.00	598.0
401.00	762.0
401.90	3924.0
403.00	5882.0
403.90	1560.0
404.90	611.0
409.90	353.0
415.10	419.0
420.90	4198.0
422.00	5544.0
423.00	31488.0
424.00	11634.0
424.90	1508.0
438.10	326.0
439.50	324.0
441.00	69448.0
442.00	480192.0
443.00	101080.0
444.00	7813.0
444.90	699.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
37.20	632.0
38.00	2027.0
39.00	7451.0
39.90	975.0
40.90	390.0
43.90	1705.0
48.90	797.0
50.00	24512.0
51.00	95032.0
52.10	5629.0
55.00	588.0
56.00	3224.0
57.00	7506.0
61.00	1685.0
62.00	2614.0
63.00	6589.0
63.90	539.0
65.00	2791.0
68.90	165760.0
70.00	848.0
73.10	2256.0
74.00	20408.0
75.00	29568.0
76.00	10661.0
77.00	200192.0
78.00	13283.0
79.00	15171.0
80.00	10621.0
81.00	16094.0
82.00	3239.0
83.00	2922.0
83.80	787.0
85.00	2042.0
85.90	5019.0
86.90	1797.0
87.90	862.0
89.00	547.0
91.00	3697.0
92.00	4322.0
93.00	28472.0
94.00	2224.0
95.00	324.0
96.10	835.0
96.90	763.0
98.00	22352.0
99.00	20112.0
100.00	1622.0
100.90	7237.0
102.00	359.0
102.80	3151.0
104.00	5279.0
105.00	5709.0
106.10	1824.0
107.00	64504.0
108.00	9429.0
109.00	1601.0
110.00	105240.0
111.00	18856.0
112.00	1985.0
112.90	538.0
114.70	342.0
116.00	3645.0
117.00	71632.0
118.00	4371.0
118.90	536.0
119.80	1024.0
121.90	5043.0
123.00	6847.0
123.90	4156.0
124.90	3406.0
127.00	218048.0
128.00	14978.0
129.00	88528.0
130.00	8895.0
130.80	1402.0
132.10	534.0
132.90	630.0
133.90	2487.0
135.00	7741.0
136.00	3329.0
137.00	5988.0
138.00	951.0
138.60	310.0
139.90	1279.0
141.00	12694.0
142.00	4377.0
142.90	3052.0
143.90	885.0
145.00	785.0
146.00	3001.0
147.00	6965.0
148.00	17552.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

149.00	3236.0
149.90	1124.0
150.90	1501.0
151.90	541.0
153.00	4464.0
154.00	3349.0
155.00	7235.0
156.10	10474.0
157.00	2509.0
158.00	2963.0
159.00	2281.0
159.90	4016.0
160.90	6323.0
162.10	2082.0
163.00	567.0
163.80	518.0
165.00	6688.0
166.00	4357.0
167.00	28160.0
168.00	17424.0
169.00	2234.0
170.00	1063.0
171.00	1479.0
172.00	2436.0
173.00	3080.0
174.00	5987.0
175.00	10384.0
176.00	2594.0
177.00	5031.0
178.00	1856.0
179.00	21304.0
180.00	15913.0
181.00	7361.0
182.00	1280.0
183.00	911.0
184.10	1318.0
185.00	10058.0
186.00	71664.0
187.00	20704.0
188.00	4072.0
189.00	5067.0
190.00	775.0
191.00	2367.0
191.90	7511.0
193.00	8400.0
193.90	1657.0
195.10	988.0
196.00	14529.0
197.90	492032.0
199.00	32456.0
200.00	2896.0
201.30	2048.0
201.60	1973.0
203.00	4176.0
204.00	22440.0
205.00	36016.0
206.00	153472.0
207.00	23008.0
208.00	6411.0
208.90	2479.0
210.00	2975.0
211.00	5663.0
211.80	1258.0
212.90	626.0
215.20	1552.0
216.00	3268.0
216.90	43624.0
217.90	5023.0
219.00	860.0
220.00	937.0
221.00	16136.0
223.00	10642.0
224.00	89544.0
225.00	20400.0
226.00	2749.0
227.00	39712.0
228.00	5940.0
229.00	7535.0
230.00	1203.0
231.10	2363.0
232.10	617.0
232.90	621.0
233.90	2291.0
235.00	3253.0
235.90	2333.0
237.00	5242.0
237.90	749.0
238.90	1698.0
240.00	1247.0
241.00	1952.0
242.00	4486.0
243.00	5021.0
244.00	67592.0
245.00	8889.0
245.90	13879.0
247.00	3222.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

247.90	621.0
249.00	2368.0
249.90	361.0
251.10	780.0
252.00	952.0
252.90	1570.0
255.00	346368.0
256.00	42992.0
257.00	4322.0
258.00	22608.0
259.00	3558.0
260.90	520.0
263.90	1038.0
265.00	9353.0
266.00	1748.0
270.00	419.0
270.80	881.0
272.00	995.0
273.00	9816.0
274.00	25752.0
275.00	146752.0
276.00	19272.0
277.00	15535.0
277.90	3838.0
279.00	753.0
281.00	1023.0
281.90	609.0
283.00	1421.0
283.90	1468.0
284.90	2036.0
286.10	357.0
287.80	367.0
288.90	659.0
290.00	425.0
292.00	1134.0
292.90	3495.0
294.00	1744.0
295.00	1209.0
296.00	50592.0
296.90	7624.0
298.10	438.0
300.90	558.0
301.90	1236.0
303.00	5702.0
304.00	1399.0
305.10	438.0
308.00	852.0
309.10	564.0
310.00	779.0
311.10	304.0
312.90	683.0
313.90	2880.0
314.90	5942.0
316.00	3411.0
317.00	749.0
320.90	1491.0
321.90	398.0
323.00	14733.0
324.00	2558.0
325.00	321.0
325.90	453.0
326.90	3121.0
328.00	2013.0
332.10	1076.0
333.10	1442.0
334.00	11528.0
335.10	1996.0
341.00	1921.0
346.00	3996.0
347.00	618.0
350.80	319.0
352.00	4967.0
353.00	2810.0
354.00	6052.0
354.80	945.0
355.10	700.0
359.00	395.0
365.00	27152.0
366.00	3418.0
370.10	493.0
370.90	1270.0
372.00	8456.0
373.00	1575.0
374.00	511.0
383.00	3152.0
384.00	878.0
389.90	1140.0
391.00	632.0
392.00	635.0
400.90	534.0
401.90	3979.0
402.90	4207.0
404.00	1896.0
404.90	671.0
420.90	3949.0
422.00	4385.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

423.00	24984.0
424.00	10164.0
424.90	971.0
441.00	77376.0
442.00	461504.0
443.00	93440.0
444.00	7716.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64