

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN060519\
 Data File : BN006104.D
 Acq On : 05 Jun 2019 17:54
 Operator : JU/SJ
 Sample : PB120074BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK74

Quant Time: Jun 05 18:31:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN060419MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 04 17:47:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	278701	20.00	ng/ul	0.00
18) Naphthalene-d8	10.43	136	1260718	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.30	164	694293	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.05	188	1354188	20.00	ng/ul	0.00
77) Chrysene-d12	21.28	240	1152832	20.00	ng/ul	0.00
85) Perylene-d12	23.53	264	1314503	20.00	ng/ul	0.02

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.15	96	61437	9.55	ng/uL	0.00
5) Phenol-d5	6.82	99	1033991	46.43	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.98	67	623699	48.98	ng/ul	0.00
9) 2-Chlorophenol-d4	7.18	132	799601	44.68	ng/ul	0.00
13) 4-Methylphenol-d8	8.35	113	809518	46.92	ng/ul	0.00
19) Nitrobenzene-d5	8.80	128	399973	44.45	ng/ul	0.00
22) 2-Nitrophenol-d4	9.52	143	441477	44.21	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.05	165	748752	41.48	ng/ul	0.00
29) 4-Chloroaniline-d4	10.57	131	1143678	53.06	ng/ul	0.00
43) Dimethylphthalate-d6	13.71	166	2242592	45.53	ng/ul	0.00
46) Acenaphthylene-d8	13.99	160	2927150	45.92	ng/ul	0.00
51) 4-Nitrophenol-d4	14.51	143	328817	37.80	ng/ul	0.00
57) Fluorene-d10	15.30	176	1892800	45.48	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.43	200	235636	31.10	ng/ul	0.00
70) Anthracene-d10	17.15	188	2703146	46.78	ng/ul	0.00
78) Pyrene-d10	19.46	212	2745431	49.28	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.39	264	2699414	46.12	ng/ul	0.02

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN060519\
Data File : BN006104.D
Acq On : 05 Jun 2019 17:54
Operator : JU/SJ
Sample : PB120074BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK74

Quant Time: Jun 05 18:31:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN060419MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jun 04 17:47:25 2019
Response via : Initial Calibration

