

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN060620\
 Data File : BN010792.D
 Acq On : 05 Jun 2020 21:42
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060620

Quant Time: Jun 05 22:51:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN060620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 05 22:48:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	123	0.00
2	1,4-Dioxane	0.396	0.400	-1.0	123	0.00
3	n-Nitrosodimethylamine	0.230	0.221	3.9	135	0.00
4 S	2-Fluorophenol	0.798	0.792	0.8	127	0.00
5 S	Phenol-d6	0.943	0.937	0.6	128	0.00
6	bis(2-Chloroethyl)ether	0.899	0.883	1.8	119	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
8 S	Nitrobenzene-d5	0.293	0.286	2.4	126	0.00
9	Naphthalene	0.996	0.994	0.2	127	0.00
10	Hexachlorobutadiene	0.243	0.235	3.3	120	0.00
11 SURR	2-Methylnaphthalene-d10	0.607	0.587	3.3	122	0.00
12	2-Methylnaphthalene	0.660	0.643	2.6	124	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
14 S	2,4,6-Tribromophenol	0.140	0.123	12.1	115	0.00
15 S	2-Fluorobiphenyl	1.558	1.560	-0.1	124	-0.01
16	Acenaphthylene	1.621	1.544	4.8	124	0.00
17	Acenaphthene	1.099	1.087	1.1	122	0.00
18	Fluorene	1.426	1.375	3.6	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.233	0.230	1.3	119	0.00
21	Hexachlorobenzene	0.246	0.251	-2.0	120	0.00
22	Pentachlorophenol	0.084	0.074	11.9	115	0.00
23	Phenanthrene	1.081	1.066	1.4	119	0.00
24	Anthracene	0.916	0.866	5.5	119	0.00
25 SURR	Fluoranthene-d10	1.110	1.056	4.9	117	0.00
26	Fluoranthene	1.252	1.186	5.3	117	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
28	Pvrene	1.268	1.290	-1.7	117	0.00
29 S	Terphenyl-d14	0.892	0.904	-1.3	116	0.00
30	Benzo(a)anthracene	1.174	1.146	2.4	117	0.00
31	Chrysene	1.334	1.299	2.6	114	0.00
32	Bis(2-ethylhexyl)phthalate	0.388	0.371	4.4	116	0.00
33	Indeno(1,2,3-cd)pyrene	1.602	1.663	-3.8	114	0.00
34 I	Perylene-d12	1.000	1.000	0.0	114	0.00
35	Benzo(b)fluoranthene	1.303	1.217	6.6	115	0.00
36	Benzo(k)fluoranthene	1.433	1.448	-1.0	115	0.00
37 C	Benzo(a)pyrene	1.163	1.117	4.0	115	0.00
38	Dibenzo(a,h)anthracene	1.310	1.271	3.0	114	0.00
39	Benzo(a,h,i)perylene	1.354	1.266	6.5	111	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN060620\
Data File : BN010792.D
Acq On : 05 Jun 2020 21:42
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN060620

Quant Time: Jun 05 22:51:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN060620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 05 22:48:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			