

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060622\
 Data File : BN020017.D
 Acq On : 06 Jun 2022 11:33
 Operator : CG/JU
 Sample : SST00517
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST005217

Quant Time: Jun 06 14:36:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 14:35:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	144845	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	652137	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	428535	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	893912	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	877976	20.000	ng/u1	0.00
88) Perylene-d12	23.710	264	891211	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	6562	1.710	ng/uL	0.00
4) Pyridine-d5	3.693	84	39626	4.077	ng/u1	0.00
7) Phenol-d5	6.987	99	49241	3.923	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	34965	4.802	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	43347	4.354	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	43308	4.504	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	21567	4.536	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	21259	4.537	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	40100	4.275	ng/u1	0.00
31) 4-Chloroaniline-d4	10.746	131	67666	4.752	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	149948	5.094	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	171046	4.298	ng/u1	0.00
54) 4-Nitrophenol-d4	14.639	143	24012	4.195	ng/u1	0.00
60) Fluorene-d10	15.439	176	129575	4.987	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	18380	4.028	ng/u1	0.00
73) Anthracene-d10	17.286	188	197555	4.693	ng/u1	0.00
81) Pyrene-d10	19.586	212	220139	4.361	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.563	264	204191	4.482	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	7302	1.865	ng/uL	94
5) Pyridine	3.711	79	43542	4.251	ng/u1	91
6) Benzaldehyde	6.958	77	32239	5.011	ng/u1	87
8) Phenol	7.010	94	55403	4.386	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.246	93	47350	4.831	ng/u1	97
12) 2-Chlorophenol	7.381	128	46842	4.633	ng/u1	98
13) 2-Methylphenol	8.263	108	41818	4.471	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.346	45	76994	5.466	ng/u1	97
16) Acetophenone	8.634	105	74364	5.154	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.622	70	37169	5.535	ng/u1	93
18) 4-Methylphenol	8.587	108	47807	4.774	ng/u1	100
19) Hexachloroethane	8.899	117	19580	4.777	ng/u1	94
22) Nitrobenzene	9.016	77	52550	4.688	ng/u1	98
23) Isophorone	9.534	82	100891	5.046	ng/u1	100
25) 2-Nitrophenol	9.728	139	24720	4.672	ng/u1	94
26) 2,4-Dimethylphenol	9.787	107	54855	4.654	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.022	93	66181	5.034	ng/u1	98
29) 2,4-Dichlorophenol	10.257	162	43286	4.603	ng/u1	98
30) Naphthalene	10.663	128	166594	4.770	ng/u1	100
32) 4-Chloroaniline	10.769	127	71240	4.978	ng/u1	100
33) Hexachlorobutadiene	10.957	225	26351	4.671	ng/u1	97
34) Caprolactam	11.504	113	13280	4.993	ng/u1	99
35) 4-Chloro-3-methylphenol	11.893	107	48714	4.894	ng/u1	98
36) 2-Methylnaphthalene	12.275	142	110543	4.918	ng/u1	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060622\
 Data File : BN020017.D
 Acq On : 06 Jun 2022 11:33
 Operator : CG/JU
 Sample : SST00517
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST005217

Quant Time: Jun 06 14:36:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 14:35:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.493	142	115550	5.135	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.645	216	53100	4.466	ng/ul	99
40) Hexachlorocyclopentadiene	12.634	237	26987	3.689	ng/ul	95
41) 2,4,6-Trichlorophenol	12.881	196	32604	4.396	ng/ul	91
42) 2,4,5-Trichlorophenol	12.951	196	36586	4.543	ng/ul	98
43) 1,1'-Biphenyl	13.287	154	156591	4.622	ng/ul	96
44) 2-Chloronaphthalene	13.322	162	117659	4.545	ng/ul	95
45) 2-Nitroaniline	13.522	65	29603	4.766	ng/ul	93
47) Dimethylphthalate	13.904	163	153960	5.267	ng/ul	99
48) 2,6-Dinitrotoluene	14.022	165	25035	4.965	ng/ul	98
50) Acenaphthylene	14.169	152	195624	4.785	ng/ul	99
51) 3-Nitroaniline	14.345	138	26259	5.081	ng/ul	94
52) Acenaphthene	14.516	153	132525	4.953	ng/ul	98
53) 2,4-Dinitrophenol	14.557	184	10554	3.773	ng/ul	96
55) 4-Nitrophenol	14.651	109	21193	4.659	ng/ul	95
56) Dibenzofuran	14.851	168	188664	5.047	ng/ul	98
57) 2,4-Dinitrotoluene	14.810	165	37630	5.132	ng/ul	88
58) 2,3,4,6-Tetrachlorophenol	15.075	232	29157	4.963	ng/ul	91
59) Diethylphthalate	15.269	149	153777	5.349	ng/ul	99
61) Fluorene	15.498	166	152441	5.253	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.492	204	70964	5.409	ng/ul	92
63) 4-Nitroaniline	15.504	138	25316	5.715	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.569	198	19992	4.307	ng/ul#	88
67) N-Nitrosodiphenylamine	15.704	169	128106	4.671	ng/ul	96
68) 4-Bromophenyl-phenylether	16.386	248	37009	4.407	ng/ul#	76
69) Hexachlorobenzene	16.504	284	45659	4.834	ng/ul	94
70) Atrazine	16.651	200	40980	4.684	ng/ul#	95
71) Pentachlorophenol	16.845	266	21083	3.678	ng/ul	97
72) Phenanthrene	17.233	178	245932	4.923	ng/ul	99
74) Anthracene	17.322	178	245017	4.910	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.251	216	64123	4.653	ng/ul	94
76) Pentachlorobenzene	14.775	250	52140	4.343	ng/ul	98
77) Carbazole	17.592	167	218025	4.806	ng/ul	97
78) Di-n-butylphthalate	18.163	149	242693	5.001	ng/ul	99
80) Fluoranthene	19.251	202	269206	4.524	ng/ul	97
82) Pyrene	19.610	202	287635	4.641	ng/ul#	94
83) Butylbenzylphthalate	20.516	149	103523	4.516	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.298	252	77102	4.373	ng/ul#	92
85) Benzo(a)anthracene	21.363	228	267535	4.758	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.304	149	162982	4.633	ng/ul	98
87) Chrysene	21.416	228	276124	4.999	ng/ul	98
89) Di-n-octyl phthalate	22.204	149	272469	4.466	ng/ul	100
90) Benzo(b)fluoranthene	23.004	252	271244	4.705	ng/ul#	96
91) Benzo(k)fluoranthene	23.051	252	248068	4.744	ng/ul	98
93) Benzo(a)pyrene	23.604	252	265771	4.854	ng/ul#	94
94) Indeno(1,2,3-cd)pyrene	26.098	276	277107	4.417	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.115	278	234648	4.374	ng/ul	97
96) Benzo(g,h,i)perylene	26.827	276	228209	4.135	ng/ul#	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA N

BNA N

ClientSampleId :

SSTD005217


Quant Time: Jun 06 14:36:58 2022


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M

Quant Title : SVOA CALIBRATION

QLast Update : Mon Jun 06 14:35:54 2022


Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 1.865 ng/uL
RT: 3.317 min Scan# 39
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Instrument :
BNA_N
ClientSampleId :
SSTD005217




Tgt Ion: 88 Resp: 7302

Ion	Ratio	Lower	Upper
88	100		
43	28.3	22.4	33.6
58	83.6	61.0	91.4






#5
Pyridine
Concen: 4.251 ng/uL
RT: 3.711 min Scan# 106
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33



Tgt Ion: 79 Resp: 43542

Ion	Ratio	Lower	Upper
79	100		
52	89.4	65.0	97.4
51	44.8	31.6	47.4









#6
Benzaldehyde
Concen: 5.011 ng/ul
RT: 6.958 min Scan# 61
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion: 77 Resp: 32239
Ion Ratio Lower Upper
77 100
105 105.8 74.7 112.1
106 99.8 70.2 105.4



#8
Phenol
Concen: 4.386 ng/ul
RT: 7.010 min Scan# 667
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion: 94 Resp: 55403
Ion Ratio Lower Upper
94 100
65 30.3 23.6 35.4
66 43.3 31.2 46.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 4.831 ng/ul
 RT: 7.246 min Scan# 707
 Delta R.T. 0.000 min
 Lab File: BN020017.D
 Acq: 06 Jun 2022 11:33




Instrument :
 BNA_N
 ClientSampleId :
 SST005217


Tgt Ion: 93	Resp: 47350
Ion Ratio	Lower Upper
93 100	
63 75.5	63.4 95.0
95 32.9	26.2 39.2



#12
 2-Chlorophenol
 Concen: 4.633 ng/ul
 RT: 7.381 min Scan# 730
 Delta R.T. -0.006 min
 Lab File: BN020017.D
 Acq: 06 Jun 2022 11:33

Tgt Ion:128	Resp: 46842
Ion Ratio	Lower Upper
128 100	
64 47.2	38.6 57.8
130 30.1	25.0 37.4









#13
2-Methylphenol
Concen: 4.471 ng/ul
RT: 8.263 min Scan# 880
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:108 Resp: 41818
Ion Ratio Lower Upper
108 100
107 90.7 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 5.466 ng/ul
RT: 8.346 min Scan# 894
Delta R.T. -0.012 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion: 45 Resp: 76994
Ion Ratio Lower Upper
45 100
77 12.4 10.9 16.3
79 9.9 8.6 13.0









#16
Acetophenone
Concen: 5.154 ng/ul
RT: 8.634 min Scan# 944
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:105 Resp: 74364
Ion Ratio Lower Upper
105 100
77 79.3 63.5 95.3
51 29.6 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 5.535 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion: 70 Resp: 37169
Ion Ratio Lower Upper
70 100
42 66.5 47.8 71.6
101 9.0 7.8 11.6
130 21.0 15.8 23.8









#18
4-Methylphenol
Concen: 4.774 ng/ul
RT: 8.587 min Scan# 911
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:108 Resp: 47807
Ion Ratio Lower Upper
108 100
107 110.5 88.0 132.0



#19
Hexachloroethane
Concen: 4.777 ng/ul
RT: 8.899 min Scan# 988
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:117 Resp: 19580
Ion Ratio Lower Upper
117 100
201 98.5 83.2 124.8
199 57.9 50.2 75.4









#22
Nitrobenzene
Concen: 4.688 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion: 77 Resp: 52550
Ion Ratio Lower Upper
77 100
123 44.9 34.5 51.7
65 14.0 11.1 16.7



#23
Isophorone
Concen: 5.046 ng/ul
RT: 9.534 min Scan# 1096
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion: 82 Resp: 100891
Ion Ratio Lower Upper
82 100
95 7.6 5.8 8.8
138 17.3 13.8 20.8





#25

2-Nitrophenol

Concen: 4.672 ng/ul

RT: 9.728 min Scan# 1129

Delta R.T. 0.000 min

Lab File: BN020017.D

Acq: 06 Jun 2022 11:33

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:139 Resp: 24720

Ion Ratio Lower Upper

139 100

65 50.4 43.7 65.5

109 28.0 25.8 38.6



#26

2,4-Dimethylphenol

Concen: 4.654 ng/ul

RT: 9.787 min Scan# 1139

Delta R.T. -0.006 min

Lab File: BN020017.D

Acq: 06 Jun 2022 11:33




Tgt Ion:107 Resp: 54855


Ion Ratio Lower Upper

107 100

121 48.5 38.3 57.5

122 82.7 66.3 99.5









#27
Bis(2-Chloroethoxy)methane
Concen: 5.034 ng/ul
RT: 10.022 min Scan# 1180
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion: 93 Resp: 66181
Ion Ratio Lower Upper
93 100
95 32.4 25.4 38.0
123 11.6 8.3 12.5



#29
2,4-Dichlorophenol
Concen: 4.603 ng/ul
RT: 10.257 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion: 162 Resp: 43286
Ion Ratio Lower Upper
162 100
164 65.1 52.9 79.3
98 39.7 29.8 44.8









#30
Naphthalene
Concen: 4.770 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:128 Resp: 166594
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.4 10.6 16.0



#32
4-Chloroaniline
Concen: 4.978 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:127 Resp: 71240
Ion Ratio Lower Upper
127 100
129 32.7 26.1 39.1









#33
Hexachlorobutadiene
Concen: 4.671 ng/ul
RT: 10.957 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:225 Resp: 26351
Ion Ratio Lower Upper
225 100
223 66.4 51.8 77.6
227 66.1 50.6 76.0



#34
Caprolactam
Concen: 4.993 ng/ul
RT: 11.504 min Scan# 1431
Delta R.T. -0.012 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:113 Resp: 13280
Ion Ratio Lower Upper
113 100
55 181.5 145.5 218.3
56 131.1 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 4.894 ng/ul
RT: 11.893 min Scan# 1497
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:107 Resp: 48714
Ion Ratio Lower Upper
107 100
144 26.9 20.6 31.0
142 79.6 62.2 93.4



#36
2-Methylnaphthalene
Concen: 4.918 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:142 Resp: 110543
Ion Ratio Lower Upper
142 100
141 96.4 70.6 105.8









#37
1-Methylnaphthalene
Concen: 5.135 ng/ul
RT: 12.493 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:142 Resp: 115550
Ion Ratio Lower Upper
142 100
141 94.3 70.9 106.3
116 4.0 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.466 ng/ul
RT: 12.645 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:216 Resp: 53100
Ion Ratio Lower Upper
216 100
214 79.6 63.4 95.0
179 22.8 17.8 26.8
108 20.8 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 3.689 ng/ul
RT: 12.634 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:237	Resp:	26987
Ion Ratio	Lower	Upper
237	100	
235	56.6	48.8 73.2
272	11.7	9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 4.396 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:196	Resp:	32604
Ion Ratio	Lower	Upper
196	100	
198	88.8	79.8 119.8
200	28.8	25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 4.543 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:196 Resp: 36586
Ion Ratio Lower Upper
196 100
198 94.7 77.6 116.4
200 30.6 24.4 36.6



#43
1,1'-Biphenyl
Concen: 4.622 ng/ul
RT: 13.287 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:154 Resp: 156591
Ion Ratio Lower Upper
154 100
153 42.3 31.5 47.3
76 12.5 10.5 15.7





#44

2-Chloronaphthalene

Concen: 4.545 ng/ul

RT: 13.322 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN020017.D

Acq: 06 Jun 2022 11:33

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:162 Resp: 117659

Ion Ratio Lower Upper

162 100

164 33.8 25.8 38.6

127 43.4 31.4 47.2



#45

2-Nitroaniline

Concen: 4.766 ng/ul

RT: 13.522 min Scan# 1774

Delta R.T. -0.006 min

Lab File: BN020017.D

Acq: 06 Jun 2022 11:33




Tgt Ion: 65 Resp: 29603


Ion Ratio Lower Upper

65 100

92 69.6 54.1 81.1

138 105.5 77.0 115.4









#47
Dimethylphthalate
Concen: 5.267 ng/ul
RT: 13.904 min Scan# 1839
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:163 Resp: 153960
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.0
164 9.6 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 4.965 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:165 Resp: 25035
Ion Ratio Lower Upper
165 100
89 63.9 50.6 75.8
121 24.8 17.9 26.9









#50
Acenaphthylene
Concen: 4.785 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:152 Resp: 195624
Ion Ratio Lower Upper
152 100
151 20.2 16.0 24.0
153 13.9 10.7 16.1



#51
3-Nitroaniline
Concen: 5.081 ng/ul
RT: 14.345 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:138 Resp: 26259
Ion Ratio Lower Upper
138 100
108 12.3 8.6 12.8
92 122.5 92.9 139.3









#52
Acenaphthene
Concen: 4.953 ng/ul
RT: 14.516 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:153 Resp: 132525
Ion Ratio Lower Upper
153 100
152 47.1 38.4 57.6
154 88.7 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 3.773 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:184 Resp: 10554
Ion Ratio Lower Upper
184 100
63 73.6 55.9 83.9
154 68.7 53.1 79.7









#55
4-Nitrophenol
Concen: 4.659 ng/ul
RT: 14.651 min Scan# 1967
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:109 Resp: 21193
Ion Ratio Lower Upper
109 100
139 128.2 100.1 150.1
65 139.4 104.6 157.0



#56
Dibenzofuran
Concen: 5.047 ng/ul
RT: 14.851 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:168 Resp: 188664
Ion Ratio Lower Upper
168 100
139 37.7 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 5.132 ng/ul
RT: 14.810 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:165 Resp: 37630
Ion Ratio Lower Upper
165 100
63 53.9 36.3 54.5
182 2.6 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.963 ng/ul
RT: 15.075 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:232 Resp: 29157
Ion Ratio Lower Upper
232 100
131 54.8 38.5 57.7
130 3.1 2.2 3.2
166 37.4 25.8 38.8









#59
Diethylphthalate
Concen: 5.349 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:149 Resp: 153777
Ion Ratio Lower Upper
149 100
177 21.9 17.8 26.8
150 13.0 10.5 15.7



#61
Fluorene
Concen: 5.253 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:166 Resp: 152441
Ion Ratio Lower Upper
166 100
165 99.8 77.7 116.5
167 13.2 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 5.409 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:204 Resp: 70964
Ion Ratio Lower Upper
204 100
206 32.4 26.7 40.1
141 71.5 50.2 75.2



#63
4-Nitroaniline
Concen: 5.715 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:138 Resp: 25316
Ion Ratio Lower Upper
138 100
92 55.0 41.0 61.6
108 79.9 65.3 97.9









#66
4,6-Dinitro-2-methylphenol
Concen: 4.307 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:198 Resp: 19992
Ion Ratio Lower Upper
198 100
182 6.6 3.0 4.4#
77 31.2 20.1 30.1#



#67
N-Nitrosodiphenylamine
Concen: 4.671 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:169 Resp: 128106
Ion Ratio Lower Upper
169 100
168 68.6 52.4 78.6
167 36.5 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 4.407 ng/ul
RT: 16.386 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:248 Resp: 37009
Ion Ratio Lower Upper
248 100
250 118.4 77.8 116.8#
141 101.4 62.6 93.8#



#69
Hexachlorobenzene
Concen: 4.834 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:284 Resp: 45659
Ion Ratio Lower Upper
284 100
142 41.2 27.8 41.8
249 31.4 25.7 38.5





#70

Atrazine

Concen: 4.684 ng/ul

RT: 16.651 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN020017.D

Acq: 06 Jun 2022 11:33

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:200 Resp: 40980

Ion Ratio Lower Upper

200 100

173 31.9 20.5 30.7#

215 48.1 38.9 58.3



#71

Pentachlorophenol

Concen: 3.678 ng/ul

RT: 16.845 min Scan# 2339

Delta R.T. 0.000 min

Lab File: BN020017.D

Acq: 06 Jun 2022 11:33




Tgt Ion:266 Resp: 21083


Ion Ratio Lower Upper

266 100

264 62.4 50.2 75.2

268 65.7 49.3 73.9









#72
Phenanthrene
Concen: 4.923 ng/ul
RT: 17.233 min Scan# 2405
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:178 Resp: 245932
Ion Ratio Lower Upper
178 100
179 16.3 12.2 18.4
176 19.3 15.2 22.8



#74
Anthracene
Concen: 4.910 ng/ul
RT: 17.322 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:178 Resp: 245017
Ion Ratio Lower Upper
178 100
179 14.9 13.0 19.4
176 18.3 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.653 ng/uL
RT: 13.251 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:216 Resp: 64123
Ion Ratio Lower Upper
216 100
214 83.5 63.2 94.8
218 51.8 37.9 56.9



#76
Pentachlorobenzene
Concen: 4.343 ng/uL
RT: 14.775 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:250 Resp: 52140
Ion Ratio Lower Upper
250 100
248 67.8 53.0 79.4
252 64.7 53.4 80.2









#77
 Carbazole
 Concen: 4.806 ng/ul
 RT: 17.592 min Scan# 2466
 Delta R.T. 0.000 min
 Lab File: BN020017.D
 Acq: 06 Jun 2022 11:33




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005217


Tgt Ion:167	Resp:	218025
Ion	Ratio	Lower Upper
167	100	
166	21.8	16.1 24.1
139	13.5	10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 5.001 ng/ul
 RT: 18.163 min Scan# 2563
 Delta R.T. 0.000 min
 Lab File: BN020017.D
 Acq: 06 Jun 2022 11:33

Tgt Ion:149	Resp:	242693
Ion	Ratio	Lower Upper
149	100	
150	8.7	7.1 10.7
104	4.7	4.1 6.1









#80
Fluoranthene
Concen: 4.524 ng/u1
RT: 19.251 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:202 Resp: 269206
Ion Ratio Lower Upper
202 100
101 12.2 8.9 13.3
100 9.3 6.7 10.1



#82
Pyrene
Concen: 4.641 ng/u1
RT: 19.610 min Scan# 2809
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:202 Resp: 287635
Ion Ratio Lower Upper
202 100
101 14.8 9.7 14.5#
100 12.0 8.0 12.0#









#83
 Butylbenzylphthalate
 Concen: 4.516 ng/ul
 RT: 20.516 min Scan# 2963
 Delta R.T. -0.006 min
 Lab File: BN020017.D
 Acq: 06 Jun 2022 11:33




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005217


Tgt Ion:149 Resp: 103523
 Ion Ratio Lower Upper
 149 100
 91 73.5 58.8 88.2
 206 19.7 18.1 27.1



#84
 3,3'-Dichlorobenzidine
 Concen: 4.373 ng/ul
 RT: 21.298 min Scan# 3096
 Delta R.T. 0.000 min
 Lab File: BN020017.D
 Acq: 06 Jun 2022 11:33

Tgt Ion:252 Resp: 77102
 Ion Ratio Lower Upper
 252 100
 254 69.3 50.8 76.2
 126 14.9 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 4.758 ng/ul
RT: 21.363 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:228 Resp: 267535
Ion Ratio Lower Upper
228 100
229 20.0 16.3 24.5
226 26.4 23.2 34.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 4.633 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:149 Resp: 162982
Ion Ratio Lower Upper
149 100
167 27.0 22.5 33.7
279 3.6 3.5 5.3









#87
Chrysene
Concen: 4.999 ng/ul
RT: 21.416 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:228 Resp: 276124
Ion Ratio Lower Upper
228 100
226 29.2 24.2 36.2
229 18.8 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 4.466 ng/ul
RT: 22.204 min Scan# 3250
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:149 Resp: 272469





#90

Benzo(b)fluoranthene

Concen: 4.705 ng/ul

RT: 23.004 min Scan# 3386

Delta R.T. 0.000 min

Lab File: BN020017.D

Acq: 06 Jun 2022 11:33

Instrument :

BNA_N

ClientSampleId :

SSTD005217





Tgt Ion:252 Resp: 271244

Ion Ratio Lower Upper

252 100

253 21.4 16.5 24.7

125 13.0 7.8 11.8#



#91

Benzo(k)fluoranthene

Concen: 4.744 ng/ul

RT: 23.051 min Scan# 3394

Delta R.T. 0.000 min

Lab File: BN020017.D

Acq: 06 Jun 2022 11:33




Tgt Ion:252 Resp: 248068


Ion Ratio Lower Upper

252 100

253 23.1 18.2 27.4

125 12.0 8.2 12.4









#93
Benzo(a)pyrene
Concen: 4.854 ng/ul
RT: 23.604 min Scan# 3489
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33




Instrument :
BNA_N
ClientSampleId :
SSTD005217


Tgt Ion:252 Resp: 265771
Ion Ratio Lower Upper
252 100
253 20.5 18.7 28.1
125 13.9 9.2 13.8#



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.417 ng/ul
RT: 26.098 min Scan# 3912
Delta R.T. 0.000 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:276 Resp: 277107
Ion Ratio Lower Upper
276 100
138 31.5 19.5 29.3#
277 25.6 21.4 32.0









#95
Dibenzo(a,h)anthracene
Concen: 4.374 ng/ul
RT: 26.115 min Scan# 3915
Delta R.T. -0.006 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Instrument :
BNA_N
ClientSampleId :
SSTD005217

Tgt Ion:278 Resp: 234648
Ion Ratio Lower Upper
278 100
139 19.7 13.8 20.8
279 22.2 18.3 27.5



#96
Benzo(g,h,i)perylene
Concen: 4.135 ng/ul
RT: 26.827 min Scan# 4036
Delta R.T. -0.018 min
Lab File: BN020017.D
Acq: 06 Jun 2022 11:33

Tgt Ion:276 Resp: 228209
Ion Ratio Lower Upper
276 100
138 26.4 17.0 25.6#
277 23.6 17.8 26.6

