

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060622\
 Data File : BN020018.D
 Acq On : 06 Jun 2022 12:09
 Operator : CG/JU
 Sample : SSTD01018
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010218

Quant Time: Jun 06 14:37:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 14:35:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	133117	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	604621	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	400554	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	831934	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	817161	20.000	ng/u1	# 0.00
88) Perylene-d12	23.710	264	890644	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	12018	3.407	ng/uL	0.00
4) Pyridine-d5	3.693	84	73472	8.225	ng/u1	0.00
7) Phenol-d5	6.987	99	94918	8.228	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	64728	9.674	ng/u1	0.00
11) 2-Chlorophenol-d4	7.358	132	82143	8.977	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	83540	9.453	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	40521	9.192	ng/u1	0.00
24) 2-Nitrophenol-d4	9.699	143	42057	9.681	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	77899	8.958	ng/u1	0.00
31) 4-Chloroaniline-d4	10.746	131	128039	9.699	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	274515	9.978	ng/u1	0.00
49) Acenaphthylene-d8	14.140	160	314804	8.463	ng/u1	0.00
54) 4-Nitrophenol-d4	14.640	143	49709	9.290	ng/u1	0.00
60) Fluorene-d10	15.440	176	227249	9.357	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	39705	9.349	ng/u1	0.00
73) Anthracene-d10	17.287	188	354934	9.059	ng/u1	0.00
81) Pyrene-d10	19.586	212	413396	8.799	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.563	264	407352	8.947	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	12410	3.450	ng/uL	98
5) Pyridine	3.711	79	80631	8.566	ng/u1	89
6) Benzaldehyde	6.964	77	57229	9.679	ng/u1	96
8) Phenol	7.017	94	104359	8.990	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.246	93	85682	9.513	ng/u1	99
12) 2-Chlorophenol	7.387	128	87789	9.447	ng/u1	99
13) 2-Methylphenol	8.264	108	79422	9.240	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.352	45	136618	10.554	ng/u1	96
16) Acetophenone	8.640	105	140281	10.578	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.622	70	70106	11.359	ng/u1	96
18) 4-Methylphenol	8.587	108	89369	9.711	ng/u1	91
19) Hexachloroethane	8.905	117	34730	9.220	ng/u1	91
22) Nitrobenzene	9.016	77	98887	9.514	ng/u1	92
23) Isophorone	9.540	82	192164	10.366	ng/u1	98
25) 2-Nitrophenol	9.728	139	48578	9.902	ng/u1	99
26) 2,4-Dimethylphenol	9.793	107	103601	9.480	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.028	93	121734	9.987	ng/u1	99
29) 2,4-Dichlorophenol	10.258	162	81298	9.324	ng/u1	98
30) Naphthalene	10.663	128	302658	9.347	ng/u1	99
32) 4-Chloroaniline	10.769	127	132781	10.007	ng/u1	99
33) Hexachlorobutadiene	10.958	225	47845	9.147	ng/u1	92
34) Caprolactam	11.505	113	28133	11.409	ng/u1	96
35) 4-Chloro-3-methylphenol	11.899	107	94420	10.231	ng/u1	96
36) 2-Methylnaphthalene	12.275	142	203010	9.741	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060622\
 Data File : BN020018.D
 Acq On : 06 Jun 2022 12:09
 Operator : CG/JU
 Sample : SST01018
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST010218

Quant Time: Jun 06 14:37:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 14:35:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.493	142	209451	10.039	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.646	216	96214	8.657	ng/ul	99
40) Hexachlorocyclopentadiene	12.634	237	51508	7.532	ng/ul	96
41) 2,4,6-Trichlorophenol	12.881	196	61467	8.866	ng/ul	96
42) 2,4,5-Trichlorophenol	12.951	196	70594	9.378	ng/ul	95
43) 1,1'-Biphenyl	13.287	154	280670	8.863	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	210152	8.685	ng/ul	98
45) 2-Nitroaniline	13.522	65	59442	10.237	ng/ul	99
47) Dimethylphthalate	13.904	163	277058	10.141	ng/ul	98
48) 2,6-Dinitrotoluene	14.022	165	49202	10.439	ng/ul	97
50) Acenaphthylene	14.169	152	360057	9.421	ng/ul	99
51) 3-Nitroaniline	14.346	138	53639	11.103	ng/ul	97
52) Acenaphthene	14.516	153	233449	9.334	ng/ul	98
53) 2,4-Dinitrophenol	14.557	184	24048	9.198	ng/ul	92
55) 4-Nitrophenol	14.651	109	43087	10.134	ng/ul	98
56) Dibenzofuran	14.846	168	332491	9.516	ng/ul	98
57) 2,4-Dinitrotoluene	14.810	165	77818	11.354	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	15.075	232	56103	10.217	ng/ul	94
59) Diethylphthalate	15.269	149	286663	10.668	ng/ul	97
61) Fluorene	15.498	166	267350	9.857	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.493	204	123973	10.109	ng/ul	95
63) 4-Nitroaniline	15.504	138	53240	12.859	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.575	198	42882	9.927	ng/ul#	98
67) N-Nitrosodiphenylamine	15.704	169	230223	9.019	ng/ul	94
68) 4-Bromophenyl-phenylether	16.387	248	69558	8.900	ng/ul	91
69) Hexachlorobenzene	16.504	284	81885	9.315	ng/ul#	93
70) Atrazine	16.651	200	76559	9.402	ng/ul	98
71) Pentachlorophenol	16.845	266	42756	8.014	ng/ul	95
72) Phenanthrene	17.234	178	431092	9.272	ng/ul	100
74) Anthracene	17.322	178	443801	9.556	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.252	216	104848	8.176	ng/uL	100
76) Pentachlorobenzene	14.769	250	95504	8.548	ng/uL	93
77) Carbazole	17.592	167	406126	9.619	ng/ul	96
78) Di-n-butylphthalate	18.163	149	460120	10.188	ng/ul	100
80) Fluoranthene	19.251	202	497687	8.987	ng/ul	98
82) Pyrene	19.610	202	527406	9.143	ng/ul#	93
83) Butylbenzylphthalate	20.516	149	210877	9.883	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.298	252	158227	9.641	ng/ul#	99
85) Benzo(a)anthracene	21.363	228	507957	9.705	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.304	149	317812	9.707	ng/ul#	97
87) Chrysene	21.416	228	499163	9.709	ng/ul	98
89) Di-n-octyl phthalate	22.204	149	555710	9.114	ng/ul	100
90) Benzo(b)fluoranthene	23.004	252	530547	9.208	ng/ul#	97
91) Benzo(k)fluoranthene	23.051	252	504377	9.651	ng/ul	96
93) Benzo(a)pyrene	23.604	252	518667	9.479	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.104	276	550856	8.785	ng/ul#	91
95) Dibenzo(a,h)anthracene	26.116	278	478221	8.920	ng/ul	97
96) Benzo(g,h,i)perylene	26.833	276	456838	8.283	ng/ul#	92





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 3.450 ng/uL
 RT: 3.317 min Scan# 39
 Delta R.T. 0.000 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09




Instrument :
 BNA_N
 ClientSampleId :
 SST010218


Tgt Ion: 88 Resp: 12410
 Ion Ratio Lower Upper
 88 100
 43 31.3 22.4 33.6
 58 75.9 61.0 91.4



#5
 Pyridine
 Concen: 8.566 ng/uL
 RT: 3.711 min Scan# 106
 Delta R.T. 0.000 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09

Tgt Ion: 79 Resp: 80631
 Ion Ratio Lower Upper
 79 100
 52 92.8 65.0 97.4
 51 43.8 31.6 47.4









#6
Benzaldehyde
Concen: 9.679 ng/ul
RT: 6.964 min Scan# 659
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion: 77 Resp: 57229
Ion Ratio Lower Upper
77 100
105 97.1 74.7 112.1
106 91.5 70.2 105.4



#8
Phenol
Concen: 8.990 ng/ul
RT: 7.017 min Scan# 668
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion: 94 Resp: 104359
Ion Ratio Lower Upper
94 100
65 29.6 23.6 35.4
66 39.6 31.2 46.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 9.513 ng/ul
 RT: 7.246 min Scan# 707
 Delta R.T. 0.000 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09

Instrument :
 BNA_N
 ClientSampleId :
 SST010218

Tgt Ion: 93 Resp: 85682




Ion	Ratio	Lower	Upper
93	100		
63	78.6	63.4	95.0
95	30.8	26.2	39.2




#12
 2-Chlorophenol
 Concen: 9.447 ng/ul
 RT: 7.387 min Scan# 731
 Delta R.T. 0.000 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09

Tgt Ion: 128 Resp: 87789

Ion	Ratio	Lower	Upper
128	100		
64	48.7	38.6	57.8
130	31.9	25.0	37.4









#13
2-Methylphenol
Concen: 9.240 ng/ul
RT: 8.264 min Scan# 81
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:108 Resp: 79422
Ion Ratio Lower Upper
108 100
107 88.6 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 10.554 ng/ul
RT: 8.352 min Scan# 895
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion: 45 Resp: 136618
Ion Ratio Lower Upper
45 100
77 12.3 10.9 16.3
79 9.1 8.6 13.0









#16
Acetophenone
Concen: 10.578 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:105 Resp: 140281
Ion Ratio Lower Upper
105 100
77 76.0 63.5 95.3
51 29.4 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 11.359 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion: 70 Resp: 70106
Ion Ratio Lower Upper
70 100
42 62.9 47.8 71.6
101 9.8 7.8 11.6
130 21.4 15.8 23.8









#18
4-Methylphenol
Concen: 9.711 ng/ul
RT: 8.587 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:108 Resp: 89369
Ion Ratio Lower Upper
108 100
107 119.5 88.0 132.0



#19
Hexachloroethane
Concen: 9.220 ng/ul
RT: 8.905 min Scan# 989
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:117 Resp: 34730
Ion Ratio Lower Upper
117 100
201 92.9 83.2 124.8
199 58.7 50.2 75.4









#22
Nitrobenzene
Concen: 9.514 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion: 77 Resp: 98887
Ion Ratio Lower Upper
77 100
123 49.1 34.5 51.7
65 15.3 11.1 16.7



#23
Isophorone
Concen: 10.366 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion: 82 Resp: 192164
Ion Ratio Lower Upper
82 100
95 7.0 5.8 8.8
138 18.4 13.8 20.8









#25
2-Nitrophenol
Concen: 9.902 ng/ul
RT: 9.728 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:139	Resp:	48578
Ion Ratio	Lower	Upper
139	100	
65	53.4	43.7 65.5
109	31.6	25.8 38.6



#26
2,4-Dimethylphenol
Concen: 9.480 ng/ul
RT: 9.793 min Scan# 1140
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:107	Resp:	103601
Ion Ratio	Lower	Upper
107	100	
121	51.7	38.3 57.5
122	87.0	66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 9.987 ng/ul
 RT: 10.028 min Scan# 1180
 Delta R.T. 0.000 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010218


Tgt Ion: 93 Resp: 121734
 Ion Ratio Lower Upper
 93 100
 95 32.2 25.4 38.0
 123 10.7 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 9.324 ng/ul
 RT: 10.258 min Scan# 1219
 Delta R.T. -0.006 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09

Tgt Ion: 162 Resp: 81298
 Ion Ratio Lower Upper
 162 100
 164 63.9 52.9 79.3
 98 37.2 29.8 44.8









#30
Naphthalene
Concen: 9.347 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:128 Resp: 302658
Ion Ratio Lower Upper
128 100
129 11.4 8.7 13.1
127 12.8 10.6 16.0



#32
4-Chloroaniline
Concen: 10.007 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:127 Resp: 132781
Ion Ratio Lower Upper
127 100
129 31.9 26.1 39.1









#33
Hexachlorobutadiene
Concen: 9.147 ng/ul
RT: 10.958 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:225 Resp: 47845
Ion Ratio Lower Upper
225 100
223 68.8 51.8 77.6
227 71.1 50.6 76.0



#34
Caprolactam
Concen: 11.409 ng/ul
RT: 11.505 min Scan# 1431
Delta R.T. -0.011 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:113 Resp: 28133
Ion Ratio Lower Upper
113 100
55 173.8 145.5 218.3
56 130.0 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 10.231 ng/ul
RT: 11.899 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:107 Resp: 94420
Ion Ratio Lower Upper
107 100
144 25.4 20.6 31.0
142 73.9 62.2 93.4



#36
2-Methylnaphthalene
Concen: 9.741 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:142 Resp: 203010
Ion Ratio Lower Upper
142 100
141 91.3 70.6 105.8









#37
1-Methylnaphthalene
Concen: 10.039 ng/ul
RT: 12.493 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:142 Resp: 209451
Ion Ratio Lower Upper
142 100
141 90.3 70.9 106.3
116 4.0 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 8.657 ng/ul
RT: 12.646 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:216 Resp: 96214
Ion Ratio Lower Upper
216 100
214 79.0 63.4 95.0
179 22.5 17.8 26.8
108 19.8 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 7.532 ng/ul
RT: 12.634 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:237 Resp: 51508
Ion Ratio Lower Upper
237 100
235 57.1 48.8 73.2
272 11.7 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 8.866 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:196 Resp: 61467
Ion Ratio Lower Upper
196 100
198 95.3 79.8 119.8
200 30.2 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 9.378 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:196 Resp: 70594
Ion Ratio Lower Upper
196 100
198 91.3 77.6 116.4
200 29.9 24.4 36.6



#43
1,1'-Biphenyl
Concen: 8.863 ng/ul
RT: 13.287 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:154 Resp: 280670
Ion Ratio Lower Upper
154 100
153 39.2 31.5 47.3
76 13.6 10.5 15.7





#44

2-Chloronaphthalene

Concen: 8.685 ng/ul

RT: 13.322 min Scan# 1741

Delta R.T. -0.006 min

Lab File: BN020018.D

Acq: 06 Jun 2022 12:09

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:162 Resp: 210152

Ion Ratio Lower Upper

162 100

164 31.7 25.8 38.6

127 41.3 31.4 47.2



#45

2-Nitroaniline

Concen: 10.237 ng/ul

RT: 13.522 min Scan# 1774

Delta R.T. -0.006 min

Lab File: BN020018.D

Acq: 06 Jun 2022 12:09




Tgt Ion: 65 Resp: 59442


Ion Ratio Lower Upper

65 100

92 68.1 54.1 81.1

138 98.0 77.0 115.4









#47
Dimethylphthalate
Concen: 10.141 ng/ul
RT: 13.904 min Scan# 1839
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:163 Resp: 277058
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.0
164 10.3 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 10.439 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:165 Resp: 49202
Ion Ratio Lower Upper
165 100
89 60.4 50.6 75.8
121 23.9 17.9 26.9









#50
Acenaphthylene
Concen: 9.421 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:152 Resp: 360057
Ion Ratio Lower Upper
152 100
151 19.0 16.0 24.0
153 13.4 10.7 16.1



#51
3-Nitroaniline
Concen: 11.103 ng/ul
RT: 14.346 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:138 Resp: 53639
Ion Ratio Lower Upper
138 100
108 11.3 8.6 12.8
92 118.9 92.9 139.3









#52
Acenaphthene
Concen: 9.334 ng/ul
RT: 14.516 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:153 Resp: 233449
Ion Ratio Lower Upper
153 100
152 46.7 38.4 57.6
154 89.2 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 9.198 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:184 Resp: 24048
Ion Ratio Lower Upper
184 100
63 76.1 55.9 83.9
154 72.8 53.1 79.7









#55
4-Nitrophenol
Concen: 10.134 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:109 Resp: 43087
Ion Ratio Lower Upper
109 100
139 128.5 100.1 150.1
65 132.1 104.6 157.0



#56
Dibenzofuran
Concen: 9.516 ng/ul
RT: 14.846 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:168 Resp: 332491
Ion Ratio Lower Upper
168 100
139 40.4 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 11.354 ng/ul
RT: 14.810 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:165 Resp: 77818
Ion Ratio Lower Upper
165 100
63 46.3 36.3 54.5
182 2.2 2.3 3.5#



#58
2,3,4,6-Tetrachlorophenol
Concen: 10.217 ng/ul
RT: 15.075 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:232 Resp: 56103
Ion Ratio Lower Upper
232 100
131 54.0 38.5 57.7
130 2.8 2.2 3.2
166 34.0 25.8 38.8









#59
Diethylphthalate
Concen: 10.668 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:149 Resp: 286663
Ion Ratio Lower Upper
149 100
177 20.7 17.8 26.8
150 12.1 10.5 15.7



#61
Fluorene
Concen: 9.857 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:166 Resp: 267350
Ion Ratio Lower Upper
166 100
165 99.3 77.7 116.5
167 13.5 11.1 16.7





#62

4-Chlorophenyl-phenylether

Concen: 10.109 ng/ul

RT: 15.493 min Scan# 2109

Delta R.T. 0.000 min

Lab File: BN020018.D

Acq: 06 Jun 2022 12:09

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:204 Resp: 123973

Ion Ratio Lower Upper

204 100

206 32.5 26.7 40.1

141 68.2 50.2 75.2



#63

4-Nitroaniline

Concen: 12.859 ng/ul

RT: 15.504 min Scan# 2111

Delta R.T. -0.006 min

Lab File: BN020018.D

Acq: 06 Jun 2022 12:09




Tgt Ion:138 Resp: 53240


Ion Ratio Lower Upper

138 100

92 52.9 41.0 61.6

108 81.5 65.3 97.9









#66
4,6-Dinitro-2-methylphenol
Concen: 9.927 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:198 Resp: 42882
Ion Ratio Lower Upper
198 100
182 4.5 3.0 4.4#
77 26.3 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 9.019 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:169 Resp: 230223
Ion Ratio Lower Upper
169 100
168 69.8 52.4 78.6
167 38.9 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 8.900 ng/ul
RT: 16.387 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:248 Resp: 69558
Ion Ratio Lower Upper
248 100
250 101.9 77.8 116.8
141 90.6 62.6 93.8



#69
Hexachlorobenzene
Concen: 9.315 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:284 Resp: 81885
Ion Ratio Lower Upper
284 100
142 42.3 27.8 41.8#
249 32.9 25.7 38.5









#70
Atrazine
Concen: 9.402 ng/ul
RT: 16.651 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:200 Resp: 76559
Ion Ratio Lower Upper
200 100
173 27.9 20.5 30.7
215 47.9 38.9 58.3



#71
Pentachlorophenol
Concen: 8.014 ng/ul
RT: 16.845 min Scan# 2339
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:266 Resp: 42756
Ion Ratio Lower Upper
266 100
264 64.2 50.2 75.2
268 67.5 49.3 73.9









#72
Phenanthrene
Concen: 9.272 ng/u1
RT: 17.234 min Scan# 2405
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:178 Resp: 431092
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.4
176 19.3 15.2 22.8



#74
Anthracene
Concen: 9.556 ng/u1
RT: 17.322 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:178 Resp: 443801
Ion Ratio Lower Upper
178 100
179 14.9 13.0 19.4
176 18.1 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 8.176 ng/uL
RT: 13.252 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:216 Resp: 104848
Ion Ratio Lower Upper
216 100
214 78.7 63.2 94.8
218 47.2 37.9 56.9



#76
Pentachlorobenzene
Concen: 8.548 ng/uL
RT: 14.769 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:250 Resp: 95504
Ion Ratio Lower Upper
250 100
248 61.5 53.0 79.4
252 59.8 53.4 80.2









#77
 Carbazole
 Concen: 9.619 ng/ul
 RT: 17.592 min Scan# 2466
 Delta R.T. 0.000 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010218


Tgt Ion:167 Resp: 406126
 Ion Ratio Lower Upper
 167 100
 166 21.8 16.1 24.1
 139 14.3 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 10.188 ng/ul
 RT: 18.163 min Scan# 2563
 Delta R.T. 0.000 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09

Tgt Ion:149 Resp: 460120
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.1 10.7
 104 5.4 4.1 6.1









#80
Fluoranthene
Concen: 8.987 ng/u1
RT: 19.251 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:202 Resp: 497687
Ion Ratio Lower Upper
202 100
101 11.7 8.9 13.3
100 9.2 6.7 10.1



#82
Pyrene
Concen: 9.143 ng/u1
RT: 19.610 min Scan# 2809
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:202 Resp: 527406
Ion Ratio Lower Upper
202 100
101 15.2 9.7 14.5#
100 12.4 8.0 12.0#









#83
Butylbenzylphthalate
Concen: 9.883 ng/ul
RT: 20.516 min Scan# 2963
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:149 Resp: 210877
Ion Ratio Lower Upper
149 100
91 72.4 58.8 88.2
206 18.6 18.1 27.1



#84
3,3'-Dichlorobenzidine
Concen: 9.641 ng/ul
RT: 21.298 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:252 Resp: 158227
Ion Ratio Lower Upper
252 100
254 63.5 50.8 76.2
126 13.3 8.2 12.4#









#85
 Benzo(a)anthracene
 Concen: 9.705 ng/ul
 RT: 21.363 min Scan# 3107
 Delta R.T. 0.000 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010218


Tgt Ion:228 Resp: 507957
 Ion Ratio Lower Upper
 228 100
 229 19.8 16.3 24.5
 226 26.6 23.2 34.8



#86
 Bis(2-ethylhexyl)phthalate
 Concen: 9.707 ng/ul
 RT: 21.304 min Scan# 3097
 Delta R.T. 0.000 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09

Tgt Ion:149 Resp: 317812
 Ion Ratio Lower Upper
 149 100
 167 26.7 22.5 33.7
 279 3.2 3.5 5.3





#87

Chrysene

Concen: 9.709 ng/ul

RT: 21.416 min Scan# 3117

Delta R.T. -0.006 min

Lab File: BN020018.D

Acq: 06 Jun 2022 12:09

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:228 Resp: 499163

Ion Ratio Lower Upper

228 100

226 31.9 24.2 36.2

229 19.9 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 9.114 ng/ul




RT: 22.204 min Scan# 3250


Delta R.T. -0.006 min

Lab File: BN020018.D

Acq: 06 Jun 2022 12:09

Tgt Ion:149 Resp: 555710









#90
Benzo(b)fluoranthene
Concen: 9.208 ng/ul
RT: 23.004 min Scan# 3386
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:252 Resp: 530547
Ion Ratio Lower Upper
252 100
253 21.3 16.5 24.7
125 11.8 7.8 11.8#



#91
Benzo(k)fluoranthene
Concen: 9.651 ng/ul
RT: 23.051 min Scan# 3394
Delta R.T. 0.000 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:252 Resp: 504377
Ion Ratio Lower Upper
252 100
253 21.0 18.2 27.4
125 12.0 8.2 12.4









#93
Benzo(a)pyrene
Concen: 9.479 ng/ul
RT: 23.604 min Scan# 3489
Delta R.T. -0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:252 Resp: 518667
Ion Ratio Lower Upper
252 100
253 22.9 18.7 28.1
125 13.7 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 8.785 ng/ul
RT: 26.104 min Scan# 3913
Delta R.T. 0.006 min
Lab File: BN020018.D
Acq: 06 Jun 2022 12:09

Tgt Ion:276 Resp: 550856
Ion Ratio Lower Upper
276 100
138 32.3 19.5 29.3#
277 25.6 21.4 32.0









#95
 Dibenzo(a,h)anthracene
 Concen: 8.920 ng/ul
 RT: 26.116 min Scan# 3916
 Delta R.T. -0.006 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010218

Tgt Ion:278 Resp: 478221
 Ion Ratio Lower Upper
 278 100
 139 18.5 13.8 20.8
 279 21.7 18.3 27.5



#96
 Benzo(g,h,i)perylene
 Concen: 8.283 ng/ul
 RT: 26.833 min Scan# 4037
 Delta R.T. -0.011 min
 Lab File: BN020018.D
 Acq: 06 Jun 2022 12:09

Tgt Ion:276 Resp: 456838
 Ion Ratio Lower Upper
 276 100
 138 27.3 17.0 25.6#
 277 23.7 17.8 26.6

