

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060822\
 Data File : BN020066.D
 Acq On : 08 Jun 2022 17:07
 Operator : CG/JU
 Sample : N3196-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-100-20220602

Quant Time: Jun 09 01:54:14 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 09 01:49:56 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 06/09/2022
 Supervised By :Jagrut Upadhyay 06/09/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.833	152	2939	0.400	ng	0.00
7) Naphthalene-d8	10.616	136	9197	0.400	ng	# 0.00
13) Acenaphthene-d10	14.453	164	5791	0.400	ng	0.01
19) Phenanthrene-d10	17.195	188	13573	0.400	ng	0.01
29) Chrysene-d12	21.378	240	12383	0.400	ng	0.00
35) Perylene-d12	23.715	264	9779	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.413	112	720	0.098	ng	0.00
5) Phenol-d6	7.010	99	440	0.048	ng	0.01
8) Nitrobenzene-d5	8.982	82	1300	0.187	ng	0.01
11) 2-Methylnaphthalene-d10	12.208	152	3440	0.215	ng	0.00
14) 2,4,6-Tribromophenol	15.944	330	278	0.121	ng	0.01
15) 2-Fluorobiphenyl	13.084	172	4907	0.194	ng	0.01
27) Fluoranthene-d10	19.223	212	10238	0.265	ng	0.00
31) Terphenyl-d14	19.822	244	8296	0.284	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.319	88	12072m	3.461	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.297	149	1020	0.054	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060822\
Data File : BN020066.D
Acq On : 08 Jun 2022 17:07
Operator : CG/JU
Sample : N3196-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-100-20220602

Quant Time: Jun 09 01:54:14 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 09 01:49:56 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 06/09/2022
Supervised By :Jagrut Upadhyay 06/09/2022

