

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN060822\  
 Data File : BN020067.D  
 Acq On : 08 Jun 2022 17:44  
 Operator : CG/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Jun 09 01:56:32 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN051322.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jun 09 01:49:56 2022  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|---------|----------------------------|-------|-------|-------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 1.000 | 1.000 | 0.0   | 58    | 0.00     |
| 2       | 1,4-Dioxane                | 0.475 | 0.520 | -9.5  | 64    | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.520 | 0.498 | 4.2   | 56    | 0.00     |
| 4 S     | 2-Fluorophenol             | 1.001 | 1.072 | -7.1  | 62    | 0.00     |
| 5 S     | Phenol-d6                  | 1.254 | 1.330 | -6.1  | 63    | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 1.221 | 1.250 | -2.4  | 59    | 0.00     |
| 7 I     | Naphthalene-d8             | 1.000 | 1.000 | 0.0   | 60    | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.302 | 0.306 | -1.3  | 66    | 0.00     |
| 9       | Naphthalene                | 1.215 | 1.271 | -4.6  | 65    | 0.00     |
| 10      | Hexachlorobutadiene        | 0.221 | 0.226 | -2.3  | 63    | 0.00     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.696 | 0.726 | -4.3  | 65    | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.827 | 0.867 | -4.8  | 66    | 0.00     |
| 13 I    | Acenaphthene-d10           | 1.000 | 1.000 | 0.0   | 62    | 0.01     |
| 14 S    | 2,4,6-Tribromophenol       | 0.158 | 0.142 | 10.1  | 63    | 0.01     |
| 15 S    | 2-Fluorobiphenyl           | 1.746 | 1.768 | -1.3  | 64    | 0.00     |
| 16      | Acenaphthylene             | 1.876 | 1.934 | -3.1  | 71    | 0.00     |
| 17      | Acenaphthene               | 1.388 | 1.404 | -1.2  | 65    | 0.00     |
| 18      | Fluorene                   | 1.754 | 1.823 | -3.9  | 67    | 0.00     |
| 19 I    | Phenanthrene-d10           | 1.000 | 1.000 | 0.0   | 61    | 0.00     |
| 20      | 4,6-Dinitro-2-methylphenol | 0.057 | 0.042 | 26.3# | 48#   | 0.01     |
| 21      | 4-Bromophenyl-phenylether  | 0.244 | 0.243 | 0.4   | 64    | 0.00     |
| 22      | Hexachlorobenzene          | 0.271 | 0.277 | -2.2  | 65    | 0.00     |
| 23      | Atrazine                   | 0.159 | 0.159 | 0.0   | 69    | 0.00     |
| 24      | Pentachlorophenol          | 0.105 | 0.073 | 30.5# | 52    | 0.00     |
| 25      | Phenanthrene               | 1.385 | 1.429 | -3.2  | 66    | 0.00     |
| 26      | Anthracene                 | 1.171 | 1.190 | -1.6  | 67    | 0.00     |
| 27 SURR | Fluoranthene-d10           | 1.137 | 1.209 | -6.3  | 69    | 0.00     |
| 28      | Fluoranthene               | 1.484 | 1.578 | -6.3  | 69    | 0.00     |
| 29 I    | Chrysene-d12               | 1.000 | 1.000 | 0.0   | 66    | 0.00     |
| 30      | Pyrene                     | 1.697 | 1.760 | -3.7  | 70    | 0.00     |
| 31 S    | Terphenyl-d14              | 0.943 | 0.959 | -1.7  | 68    | 0.00     |
| 32      | Benzo(a)anthracene         | 1.458 | 1.477 | -1.3  | 70    | 0.00     |
| 33      | Chrysene                   | 1.635 | 1.724 | -5.4  | 71    | 0.00     |
| 34      | Bis(2-ethylhexyl)phthalate | 0.607 | 0.718 | -18.3 | 88    | 0.00     |
| 35 I    | Perylene-d12               | 1.000 | 1.000 | 0.0   | 63    | 0.00     |
| 36      | Indeno(1,2,3-cd)pyrene     | 1.873 | 1.736 | 7.3   | 62    | 0.00     |
| 37      | Benzo(b)fluoranthene       | 1.746 | 1.832 | -4.9  | 69    | 0.00     |
| 38      | Benzo(k)fluoranthene       | 1.853 | 1.915 | -3.3  | 68    | 0.00     |
| 39 C    | Benzo(a)pyrene             | 1.374 | 1.440 | -4.8  | 70    | 0.00     |
| 40      | Dibenzo(a,h)anthracene     | 1.510 | 1.393 | 7.7   | 62    | 0.00     |
| 41      | Benzo(g,h,i)perylene       | 1.553 | 1.374 | 11.5  | 60    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

**Instrument :**  
BNA\_N  
**LabSampleId :**  
SSTDCCC0.4