

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
 Data File : BN014848.D
 Acq On : 09 Jun 2021 11:34
 Operator : CG/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Jun 09 12:56:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 12:45:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	2305	0.400	ng	0.00
7) Naphthalene-d8	10.496	136	9339	0.400	ng	# 0.01
13) Acenaphthene-d10	14.345	164	6026	0.400	ng	0.00
19) Phenanthrene-d10	17.103	188	12788	0.400	ng	# 0.01
29) Chrysene-d12	21.302	240	14409	0.400	ng	# 0.00
35) Perylene-d12	23.564	264	14845	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.338	112	1311	0.175	ng	0.00
5) Phenol-d6	6.887	99	1448	0.148	ng	0.00
8) Nitrobenzene-d5	8.861	82	1335	0.154	ng	0.00
11) 2-Methylnaphthalene-d10	12.088	152	3685	0.209	ng	0.00
14) 2,4,6-Tribromophenol	15.855	330	461	0.128	ng	0.01
15) 2-Fluorobiphenyl	12.975	172	4613	0.202	ng	0.01
27) Fluoranthene-d10	19.143	212	8956	0.211	ng	0.00
31) Terphenyl-d14	19.748	244	7142	0.203	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	114	0.266	ng	# 83
3) n-Nitrosodimethylamine	3.724	42	66	0.096	ng	# 94
6) bis(2-Chloroethyl)ether	7.154	93	1018	0.136	ng	97
9) Naphthalene	10.547	128	5235	0.203	ng	99
10) Hexachlorobutadiene	10.825	225	1544	0.225	ng	# 100
12) 2-Methylnaphthalene	12.163	142	3567	0.200	ng	97
16) Acenaphthylene	14.066	152	4759	0.190	ng	99
17) Acenaphthene	14.409	154	3448	0.201	ng	98
18) Fluorene	15.398	166	4398	0.205	ng	100
20) 4,6-Dinitro-2-methylph...	15.500	198	69	0.044	ng	# 63
21) 4-Bromophenyl-phenylether	16.300	248	1713	0.190	ng	# 86
22) Hexachlorobenzene	16.410	284	2107	0.221	ng	99
23) Atrazine	16.568	200	854	0.146	ng	# 96
24) Pentachlorophenol	16.762	266	218	0.075	ng	92
25) Phenanthrene	17.140	178	7145	0.199	ng	99
26) Anthracene	17.237	178	5834	0.185	ng	100
28) Fluoranthene	19.173	202	8747	0.201	ng	100
30) Pyrene	19.535	202	8658	0.203	ng	100
32) Benzo(a)anthracene	21.292	228	7269	0.174	ng	96
33) Chrysene	21.345	228	10060	0.217	ng	97
34) Bis(2-ethylhexyl)phtha...	21.228	149	2395	0.171	ng	98
36) Indeno(1,2,3-cd)pyrene	25.843	276	6982	0.125	ng	# 97
37) Benzo(b)fluoranthene	22.927	252	10785	0.224	ng	95
38) Benzo(k)fluoranthene	22.927	252	11065	0.197	ng	# 96
39) Benzo(a)pyrene	23.464	252	8319	0.176	ng	# 72
40) Dibenzo(a,h)anthracene	25.857	278	4960	0.112	ng	# 91
41) Benzo(g,h,i)perylene	26.528	276	7636	0.153	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
Data File : BN014848.D
Acq On : 09 Jun 2021 11:34
Operator : CG/JU
Sample : SSTDIC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Jun 09 12:56:21 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 12:45:46 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.717 min Scan# 579
Delta R.T. 0.009 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:152 Resp: 2305
Ion Ratio Lower Upper
152 100
150 147.3 119.0 178.6
115 76.6 62.2 93.4



1,4-Dioxane
Concen: 0.266 ng
RT: 3.281 min Scan# 98
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion: 88 Resp: 114
Ion Ratio Lower Upper
88 100
43 27.2 12.9 19.3#
58 57.0 38.1 57.1









n-Nitrosodimethylamine
 Concen: 0.096 ng
 RT: 3.724 min Scan# 146
 Delta R.T. 0.018 min
 Lab File: BN014848.D
 Acq: 09 Jun 2021 11:34




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC0.2


Tgt Ion: 42 Resp: 66
 Ion Ratio Lower Upper
 42 100
 74 195.5 149.5 224.3
 44 0.0 4.2 6.2#



2-Fluorophenol
 Concen: 0.175 ng
 RT: 5.338 min Scan# 321
 Delta R.T. 0.009 min
 Lab File: BN014848.D
 Acq: 09 Jun 2021 11:34

Tgt Ion: 112 Resp: 1311
 Ion Ratio Lower Upper
 112 100
 64 47.4 39.0 58.6
 63 25.5 20.6 30.8









Phenol-d6
 Concen: 0.148 ng
 RT: 6.887 min Scan# 489
 Delta R.T. -0.000 min
 Lab File: BN014848.D
 Acq: 09 Jun 2021 11:34




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	99	Resp:	1448
Ion	Ratio	Lower	Upper
99	100		
42	10.8	8.9	13.3
71	44.1	35.6	53.4



bis(2-Chloroethyl)ether
 Concen: 0.136 ng
 RT: 7.154 min Scan# 518
 Delta R.T. 0.009 min
 Lab File: BN014848.D
 Acq: 09 Jun 2021 11:34

Tgt Ion:	93	Resp:	1018
Ion	Ratio	Lower	Upper
93	100		
63	60.5	46.6	70.0
95	37.2	28.6	42.8



Abundance Scan 815 (10.484 min): BN014849.D\data.ms (-807) (-)

Naphthalene-d8

Concen: 0.400 ng

RT: 10.496 min Scan# 816

Delta R.T. 0.012 min

Lab File: BN014848.D


Acq: 09 Jun 2021 11:34

Instrument :

BNA_N

ClientSampled :

SSTDIC0.2



Tgt Ion: 136 Resp: 9339




Ion Ratio Lower Upper

136 100

137 11.7 9.8 14.8

54 4.6 5.4 8.2#

68 4.2 4.9 7.3#



Abundance Scan 687 (8.861 min): BN014849.D\data.ms (-683) (-)

Nitrobenzene-d5

Concen: 0.154 ng

RT: 8.861 min Scan# 687

Delta R.T. -0.000 min

Lab File: BN014848.D

Acq: 09 Jun 2021 11:34




Tgt Ion: 82 Resp: 1335


Ion Ratio Lower Upper

82 100

128 34.7 31.6 47.4

54 42.4 29.0 43.4









Naphthalene
 Concen: 0.203 ng
 RT: 10.547 min Scan# 820
 Delta R.T. 0.012 min
 Lab File: BN014848.D
 Acq: 09 Jun 2021 11:34




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC0.2

Tgt Ion:	128	Resp:	5235
Ion	Ratio	Lower	Upper
128	100		
129	12.0	9.1	13.7
127	14.9	11.8	17.6




Hexachlorobutadiene
 Concen: 0.225 ng
 RT: 10.825 min Scan# 842
 Delta R.T. -0.000 min
 Lab File: BN014848.D
 Acq: 09 Jun 2021 11:34

Tgt Ion:	225	Resp:	1544
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.8	50.2	75.2





Abundance Scan 1024 (12.079 min): BN014849.D\data.ms (-1014) (-)




2-Methylnaphthalene-d10
Concen: 0.209 ng
RT: 12.088 min Scan# 1026
Delta R.T. 0.009 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:152 Resp: 3685
Ion Ratio Lower Upper
152 100
151 20.9 16.7 25.1



Abundance Scan 1026 (12.088 min): BN014848.D\data.ms (-979) (-)





Abundance Scan 1041 (12.155 min): BN014849.D\data.ms (-1031) (-)




2-Methylnaphthalene
Concen: 0.200 ng
RT: 12.163 min Scan# 1043
Delta R.T. 0.009 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:142 Resp: 3567
Ion Ratio Lower Upper
142 100
141 88.1 72.3 108.5
115 48.7 40.7 61.1




Abundance Scan 1043 (12.163 min): BN014848.D\data.ms (-1019) (-)





Abundance Scan 1265 (14.346 min): BN014849.D\data.ms (-1213) (-)

Acenaphthene-d10
Concen: 0.400 ng
RT: 14.345 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion:	164	Resp:	6026
Ion Ratio	Lower	Upper	
164	100		
162	107.7	83.8	125.6
160	49.4	37.4	56.0






Abundance Scan 1265 (14.345 min): BN014848.D\data.ms (-1249) (-)




Abundance Scan 1383 (15.843 min): BN014849.D\data.ms (-1314) (-)

2,4,6-Tribromophenol
Concen: 0.128 ng
RT: 15.855 min Scan# 1384
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Tgt Ion:	330	Resp:	461
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	27.8	22.4	33.6



Abundance Scan 1384 (15.855 min): BN014848.D\data.ms (-1375) (-)






Abundance Scan 1156 (12.963 min): BN014849.D\data.ms (-1156) (-)




2-Fluorobiphenyl
Concen: 0.202 ng
RT: 12.975 min Scan# 1157
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampled :
SSTDIC0.2

Tgt Ion:	172	Resp:	4613
Ion	Ratio	Lower	Upper
172	100		
171	36.7	30.6	45.8
170	24.5	20.2	30.4






Abundance Scan 1243 (14.067 min): BN014849.D\data.ms (-1243) (-)



Acenaphthylene
Concen: 0.190 ng
RT: 14.066 min Scan# 1243
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Tgt Ion:	152	Resp:	4759
Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.7	25.1
153	12.9	10.5	15.7





Abundance Scan 1270 (14.409 min): BN014849.D\data.ms (-1267) (-)

Acenaphthene
Concen: 0.201 ng
RT: 14.409 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion:	154	Resp:	3448
Ion Ratio	Lower	Upper	
154	100		
153	116.4	90.9	136.3
152	56.8	45.8	68.8




Abundance Scan 1270 (14.409 min): BN014848.D\data.ms (-1262) (-)





Abundance Scan 1348 (15.398 min): BN014849.D\data.ms (-1341) (-)


Fluorene
Concen: 0.205 ng
RT: 15.398 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion:	166	Resp:	4398
Ion Ratio	Lower	Upper	
166	100		
165	99.2	79.2	118.8
167	14.3	10.7	16.1



Abundance Scan 1348 (15.398 min): BN014848.D\data.ms (-1332) (-)





Abundance Scan 1485 (17.091 min): BN014849.D\data.ms (-1419) (-)




Phenanthrene-d10
Concen: 0.400 ng
RT: 17.103 min Scan# 1486
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	188	Resp:	12788
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	7.4	9.0	13.4#



Abundance Scan 1486 (17.103 min): BN014848.D\data.ms (-1469) (-)





Abundance Scan 1355 (15.488 min): BN014849.D\data.ms (-1329) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.044 ng
RT: 15.500 min Scan# 1356
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:	198	Resp:	69
Ion	Ratio	Lower	Upper
198	100		
182	66.7	34.5	51.7#
77	0.0	0.0	0.0




Abundance Scan 1356 (15.500 min): BN014848.D\data.ms (-1347) (-)





Abundance Scan 1420 (16.300 min): BN014849.D\data.ms (-1423) (-)

4-Bromophenyl-phenylether
Concen: 0.190 ng
RT: 16.300 min Scan# 1420
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion	Ratio	Lower	Upper
248	100		
250	100.9	86.2	129.4
141	64.8	36.2	54.2#




Abundance Scan 1420 (16.300 min): BN014848.D\data.ms (-1412) (-)





Abundance Scan 1429 (16.410 min): BN014849.D\data.ms (-1423) (-)

Hexachlorobenzene
Concen: 0.221 ng
RT: 16.410 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion	Ratio	Lower	Upper
284	100		
142	27.3	22.2	33.2
249	33.5	26.6	39.8




Abundance Scan 1429 (16.410 min): BN014848.D\data.ms (-1421) (-)





Abundance Scan 1442 (16.568 min): BN014849.D\data.ms (-1428) (-)

Atrazine
Concen: 0.146 ng
RT: 16.568 min Scan# 1442
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion:	200	Resp:	854
Ion Ratio	Lower	Upper	
200	100		
173	32.6	21.7	32.5#
215	53.8	43.1	64.7






Abundance Scan 1442 (16.568 min): BN014848.D\data.ms (-1401) (-)




Abundance Scan 1458 (16.762 min): BN014849.D\data.ms (-1424) (-)

Pentachlorophenol
Concen: 0.075 ng
RT: 16.762 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:	266	Resp:	218
Ion Ratio	Lower	Upper	
266	100		
264	66.5	51.0	76.4
268	72.5	50.6	76.0




Abundance Scan 1458 (16.762 min): BN014848.D\data.ms (-1450) (-)





Abundance Scan 1489 (17.140 min): BN014849.D\data.ms (-1425) (-)

Phenanthrene
Concen: 0.199 ng
RT: 17.140 min Scan# 1489
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.6	23.4
179	15.5	12.1	18.1






Abundance Scan 1489 (17.140 min): BN014848.D\data.ms (-1464) (-)




Abundance Scan 1496 (17.225 min): BN014849.D\data.ms (-1426) (-)

Anthracene
Concen: 0.185 ng
RT: 17.237 min Scan# 1497
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.1	22.7
179	15.3	12.2	18.2



Abundance Scan 1497 (17.237 min): BN014848.D\data.ms (-1488) (-)






Abundance Scan 1787 (19.138 min): BN014849.D\data.ms (-1727) (-)




Fluoranthene-d10
Concen: 0.211 ng
RT: 19.143 min Scan# 1788
Delta R.T. 0.005 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Tgt Ion:	212	Resp:	8956
Ion Ratio	Lower	Upper	
212	100		
106	8.8	7.0	10.6
104	5.1	4.1	6.1






Abundance Scan 1793 (19.168 min): BN014849.D\data.ms (-1728) (-)



Fluoranthene
Concen: 0.201 ng
RT: 19.173 min Scan# 1794
Delta R.T. 0.005 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion:	202	Resp:	8747
Ion Ratio	Lower	Upper	
202	100		
101	8.1	6.6	10.0
203	17.3	13.9	20.9




Abundance Scan 2084 (21.303 min): BN014849.D\data.ms (-2072) (-)

Chrysene-d12
Concen: 0.400 ng
RT: 21.302 min Scan# 2084
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2






Tgt Ion:	240	Resp:	14409
Ion	Ratio	Lower	Upper
240	100		
120	6.6	4.2	6.4#
236	27.1	21.4	32.2




Abundance Scan 1867 (19.535 min): BN014849.D\data.ms (-1836) (-)


Pyrene
Concen: 0.203 ng
RT: 19.535 min Scan# 1867
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34



Tgt Ion:	202	Resp:	8658
Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.9	25.3
203	17.6	13.9	20.9





Abundance Scan 1909 (19.744 min): BN014849.D\data.ms (-1879) (-)




Terphenyl-d14
Concen: 0.203 ng
RT: 19.748 min Scan# 1910
Delta R.T. 0.005 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.8	6.3	9.5
122	8.0	6.2	9.2



Abundance Scan 1910 (19.748 min): BN014848.D\data.ms (-1879) (-)





Abundance Scan 2082 (21.281 min): BN014849.D\data.ms (-2073) (-)




Benzo(a)anthracene
Concen: 0.174 ng
RT: 21.292 min Scan# 2083
Delta R.T. 0.010 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.8	23.0	34.6
229	21.3	15.7	23.5



Abundance Scan 2083 (21.292 min): BN014848.D\data.ms (-2063) (-)






Abundance Scan 2087 (21.334 min): BN014849.D\data.ms (-2087) (-)




Chrysene
Concen: 0.217 ng
RT: 21.345 min Scan# 2088
Delta R.T. 0.010 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	228	Resp:	10060
Ion Ratio	Lower	Upper	
228	100		
226	29.5	25.0	37.6
229	20.8	15.9	23.9






Abundance Scan 2076 (21.218 min): BN014849.D\data.ms (-2076) (-)



Bis(2-ethylhexyl)phthalate
Concen: 0.171 ng
RT: 21.228 min Scan# 2077
Delta R.T. 0.010 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion:	149	Resp:	2395
Ion Ratio	Lower	Upper	
149	100		
167	37.1	28.9	43.3
279	6.2	4.9	7.3




Abundance Scan 2525 (23.561 min): BN014849.D\data.ms (-2535) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.564 min Scan# 2526
Delta R.T. 0.003 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2






Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.3	20.6	31.0
265	26.0	21.4	32.2




Abundance Scan 3190 (25.837 min): BN014849.D\data.ms (-3184) (-)


Indeno(1,2,3-cd)pyrene
Concen: 0.125 ng
RT: 25.843 min Scan# 3192
Delta R.T. 0.006 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34



Tgt Ion	Ratio	Lower	Upper
276	100		
138	9.7	10.0	15.0#
277	26.2	20.4	30.6





Abundance Scan 2327 (22.883 min): BN014849.D\data.ms (-2327) (-)




Benzo(b)fluoranthene
Concen: 0.224 ng
RT: 22.927 min Scan# 2340
Delta R.T. 0.044 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:	252	Resp:	10785
Ion Ratio	Lower	Upper	
252	100		
253	25.4	18.5	27.7
125	8.5	5.8	8.6


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2



Abundance Scan 2340 (22.927 min): BN014848.D\data.ms (-2269) (-)





Abundance Scan 2340 (22.927 min): BN014849.D\data.ms (-2338) (-)




Benzo(k)fluoranthene
Concen: 0.197 ng
RT: 22.927 min Scan# 2340
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:	252	Resp:	11065
Ion Ratio	Lower	Upper	
252	100		
253	25.4	19.1	28.7
125	8.5	5.4	8.0#




Abundance Scan 2340 (22.927 min): BN014848.D\data.ms (-2252) (-)






Abundance Scan 2496 (23.461 min): BN014849.D\data.ms (-2438) (-)

Benzo(a)pyrene
Concen: 0.176 ng
RT: 23.464 min Scan# 2497
Delta R.T. 0.003 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2







Tgt Ion	Ratio	Lower	Upper
252	100		
253	31.7	19.4	29.2#
125	32.4	7.4	11.2#




Abundance Scan 3194 (25.851 min): BN014849.D\data.ms (-3170) (-)

Dibenzo(a,h)anthracene
Concen: 0.112 ng
RT: 25.857 min Scan# 3196
Delta R.T. 0.006 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion	Ratio	Lower	Upper
278	100		
139	15.6	9.0	13.4#
279	30.3	21.0	31.4




Abundance Scan 3391 (26.525 min): BN014849.D\data.ms (-3333) (-)



Benzo(g,h,i)perylene
Concen: 0.153 ng
RT: 26.528 min Scan# 3392
Delta R.T. 0.003 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Abundance Scan 3392 (26.528 min): BN014848.D\data.ms



Tgt Ion:	276	Resp:	7636
Ion Ratio	Lower	Upper	
276	100		
277	25.1	19.7	29.5
138	14.9	9.8	14.6#

Abundance Scan 3392 (26.528 min): BN014848.D\data.ms (-3333) (-)

