

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
 Data File : BN014854.D
 Acq On : 09 Jun 2021 15:49
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060921

Quant Time: Jun 09 16:37:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 16:35:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.327	18.3	94	0.00
3	n-Nitrosodimethylamine	0.400	0.417	-4.2	106	0.00
4 S	2-Fluorophenol	0.400	0.385	3.8	108	0.00
5 S	Phenol-d6	0.400	0.387	3.3	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.372	7.0	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.01
8 S	Nitrobenzene-d5	0.400	0.343	14.2	106	0.00
9	Naphthalene	0.400	0.378	5.5	107	0.00
10	Hexachlorobutadiene	0.400	0.369	7.8	103	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.363	9.3	104	0.00
12	2-Methylnaphthalene	0.400	0.375	6.3	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.400	0.359	10.3	107	0.01
15 S	2-Fluorobiphenyl	0.400	0.382	4.5	107	0.01
16	Acenaphthylene	0.400	0.371	7.3	108	0.00
17	Acenaphthene	0.400	0.373	6.8	105	0.00
18	Fluorene	0.400	0.377	5.8	106	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	105	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.396	1.0	106	0.00
21	4-Bromophenyl-phenylether	0.400	0.366	8.5	107	0.00
22	Hexachlorobenzene	0.400	0.376	6.0	105	0.00
23	Atrazine	0.400	0.359	10.3	110	0.00
24	Pentachlorophenol	0.400	0.458	-14.5	104	0.00
25	Phenanthrene	0.400	0.376	6.0	105	0.00
26	Anthracene	0.400	0.360	10.0	105	0.01
27 SURR	Fluoranthene-d10	0.400	0.350	12.5	105	0.00
28	Fluoranthene	0.400	0.371	7.3	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	106	0.01
30	Pyrene	0.400	0.375	6.3	106	0.00
31 S	Terphenyl-d14	0.400	0.372	7.0	106	0.00
32	Benzo(a)anthracene	0.400	0.368	8.0	111	0.01
33	Chrysene	0.400	0.363	9.3	103	0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.335	16.3	105	0.01
35 I	Perylene-d12	0.400	0.400	0.0	105	0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.352	12.0	108	0.01
37	Benzo(b)fluoranthene	0.400	0.351	12.3	105	0.00
38	Benzo(k)fluoranthene	0.400	0.381	4.8	102	0.00
39 C	Benzo(a)pyrene	0.400	0.370	7.5	107	0.01
40	Dibenzo(a,h)anthracene	0.400	0.359	10.3	109	0.00
41	Benzo(g,h,i)perylene	0.400	0.345	13.8	103	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN060921