

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
 Data File : BN014854.D
 Acq On : 09 Jun 2021 15:49
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060921

Quant Time: Jun 09 16:37:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 16:35:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.116	0.095	18.1	94	0.00
3	n-Nitrosodimethylamine	0.123	0.091	26.0#	106	0.00
4 S	2-Fluorophenol	1.080	1.039	3.8	108	0.00
5 S	Phenol-d6	1.257	1.216	3.3	108	0.00
6	bis(2-Chloroethyl)ether	0.953	0.887	6.9	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.01
8 S	Nitrobenzene-d5	0.330	0.283	14.2	106	0.00
9	Naphthalene	1.095	1.036	5.4	107	0.00
10	Hexachlorobutadiene	0.319	0.294	7.8	103	0.01
11 SURR	2-Methylnaphthalene-d10	0.753	0.684	9.2	104	0.00
12	2-Methylnaphthalene	0.771	0.722	6.4	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.168	0.151	10.1	107	0.01
15 S	2-Fluorobiphenyl	1.512	1.443	4.6	107	0.01
16	Acenaphthylene	1.668	1.547	7.3	108	0.00
17	Acenaphthene	1.128	1.053	6.6	105	0.00
18	Fluorene	1.444	1.362	5.7	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.01
20	4,6-Dinitro-2-methylphenol	0.026	0.013	50.0#	106	0.00
21	4-Bromophenyl-phenylether	0.285	0.260	8.8	107	0.00
22	Hexachlorobenzene	0.324	0.304	6.2	105	0.00
23	Atrazine	0.155	0.139	10.3	110	0.00
24	Pentachlorophenol	0.073	0.042	42.5#	104	0.00
25	Phenanthrene	1.136	1.070	5.8	105	0.00
26	Anthracene	1.006	0.905	10.0	105	0.01
27 SURR	Fluoranthene-d10	1.380	1.207	12.5	105	0.00
28	Fluoranthene	1.438	1.333	7.3	106	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	106	0.01
30	Pyrene	1.191	1.118	6.1	106	0.00
31 S	Terphenyl-d14	0.990	0.921	7.0	106	0.00
32	Benzo(a)anthracene	1.138	1.047	8.0	111	0.01
33	Chrysene	1.335	1.212	9.2	103	0.01
34	Bis(2-ethylhexyl)phthalate	0.351	0.294	16.2	105	0.01
35 I	Perylene-d12	1.000	1.000	0.0	105	0.01
36	Indeno(1,2,3-cd)pyrene	1.202	0.985	18.1	108	0.01
37	Benzo(b)fluoranthene	1.276	1.121	12.1	105	0.00
38	Benzo(k)fluoranthene	1.520	1.447	4.8	102	0.00
39 C	Benzo(a)pyrene	1.210	1.119	7.5	107	0.01
40	Dibenzo(a,h)anthracene	0.920	0.736	20.0	109	0.00
41	Benzo(g,h,i)perylene	1.162	1.002	13.8	103	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
Data File : BN014854.D
Acq On : 09 Jun 2021 15:49
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN060921

Quant Time: Jun 09 16:37:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 16:35:36 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				