

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
 Data File : BN014858.D
 Acq On : 09 Jun 2021 18:37
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 09 19:06:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 16:35:36 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	122	0.00
2	1,4-Dioxane	0.400	0.410	-2.5	135	0.00
3	n-Nitrosodimethylamine	0.400	0.465	-16.3	149	-0.02
4 S	2-Fluorophenol	0.400	0.475	-18.7	152	0.00
5 S	Phenol-d6	0.400	0.487	-21.7	156	0.00
6	bis(2-Chloroethyl)ether	0.400	0.412	-3.0	133	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	123	0.01
8 S	Nitrobenzene-d5	0.400	0.398	0.5	142	0.00
9	Naphthalene	0.400	0.395	1.3	129	0.00
10	Hexachlorobutadiene	0.400	0.385	3.8	125	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.381	4.8	126	0.00
12	2-Methylnaphthalene	0.400	0.394	1.5	128	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.400	0.497	-24.2	172	0.00
15 S	2-Fluorobiphenyl	0.400	0.391	2.3	127	0.01
16	Acenaphthylene	0.400	0.395	1.3	134	0.00
17	Acenaphthene	0.400	0.392	2.0	128	0.00
18	Fluorene	0.400	0.395	1.3	129	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	121	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.416	-4.0	147	0.00
21	4-Bromophenyl-phenylether	0.400	0.388	3.0	132	0.00
22	Hexachlorobenzene	0.400	0.377	5.8	122	0.00
23	Atrazine	0.400	0.456	-14.0	162	0.00
24	Pentachlorophenol	0.400	0.536	-34.0#	187	0.00
25	Phenanthrene	0.400	0.391	2.3	127	0.00
26	Anthracene	0.400	0.397	0.8	135	0.01
27 SURR	Fluoranthene-d10	0.400	0.390	2.5	136	0.00
28	Fluoranthene	0.400	0.405	-1.3	134	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	129	0.00
30	Pyrene	0.400	0.397	0.8	136	0.00
31 S	Terphenyl-d14	0.400	0.387	3.3	134	0.00
32	Benzo(a)anthracene	0.400	0.420	-5.0	154	0.01
33	Chrysene	0.400	0.389	2.8	135	0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.664	-66.0#	255	0.01
35 I	Perylene-d12	0.400	0.400	0.0	139	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.416	-4.0	178	0.00
37	Benzo(b)fluoranthene	0.400	0.405	-1.3	161	0.00
38	Benzo(k)fluoranthene	0.400	0.382	4.5	135	0.00
39 C	Benzo(a)pyrene	0.400	0.404	-1.0	155	0.00
40	Dibenzo(a,h)anthracene	0.400	0.429	-7.2	187	0.00
41	Benzo(g,h,i)perylene	0.400	0.379	5.3	151	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4