

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
 Data File : BN014848.D
 Acq On : 09 Jun 2021 11:34
 Operator : CG/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Manual Integrations
 APPROVED

mohammad
 6/11/2021 1:27:09 PM

Quant Time: Jun 09 12:56:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 12:45:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	2305	0.40	ng	0.00
7) Naphthalene-d8	10.496	136	9339	0.40	ng	# 0.01
13) Acenaphthene-d10	14.345	164	6026	0.40	ng	0.00
19) Phenanthrene-d10	17.103	188	12788	0.40	ng	# 0.01
29) Chrysene-d12	21.302	240	14409	0.40	ng	# 0.00
35) Perylene-d12	23.564	264	14845	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.338	112	1311	0.17	ng	0.00
5) Phenol-d6	6.887	99	1448	0.15	ng	0.00
8) Nitrobenzene-d5	8.861	82	1335	0.15	ng	0.00
11) 2-Methylnaphthalene-d10	12.088	152	3685	0.21	ng	0.00
14) 2,4,6-Tribromophenol	15.855	330	461	0.13	ng	0.01
15) 2-Fluorobiphenyl	12.975	172	4613	0.20	ng	0.01
27) Fluoranthene-d10	19.143	212	8956	0.21	ng	0.00
31) Terphenyl-d14	19.748	244	7142	0.20	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.281	88	114	0.27	ng	# 83
3) n-Nitrosodimethylamine	3.724	42	66	0.10	ng	# 94
6) bis(2-Chloroethyl)ether	7.154	93	1018	0.14	ng	97
9) Naphthalene	10.547	128	5235	0.20	ng	99
10) Hexachlorobutadiene	10.825	225	1544	0.22	ng	# 100
12) 2-Methylnaphthalene	12.163	142	3567	0.20	ng	97
16) Acenaphthylene	14.066	152	4759	0.19	ng	99
17) Acenaphthene	14.409	154	3448	0.20	ng	98
18) Fluorene	15.398	166	4398	0.20	ng	100
20) 4,6-Dinitro-2-methylph...	15.500	198	69	0.04	ng	# 63
21) 4-Bromophenyl-phenylether	16.300	248	1713	0.19	ng	# 86
22) Hexachlorobenzene	16.410	284	2107	0.22	ng	99
23) Atrazine	16.568	200	854	0.15	ng	# 96
24) Pentachlorophenol	16.762	266	218	0.07	ng	92
25) Phenanthrene	17.140	178	7145	0.20	ng	99
26) Anthracene	17.237	178	5834	0.19	ng	100
28) Fluoranthene	19.173	202	8747	0.20	ng	100
30) Pyrene	19.535	202	8658	0.20	ng	100
32) Benzo(a)anthracene	21.292	228	7269	0.17	ng	96
33) Chrysene	21.345	228	10060	0.22	ng	97
34) Bis(2-ethylhexyl)phtha...	21.228	149	2395	0.17	ng	98
36) Indeno(1,2,3-cd)pyrene	25.843	276	6982	0.12	ng	# 97
37) Benzo(b)fluoranthene	22.886	252	8876m	0.18	ng	
38) Benzo(k)fluoranthene	22.927	252	11065	0.20	ng	# 96
39) Benzo(a)pyrene	23.464	252	8319	0.18	ng	# 72
40) Dibenzo(a,h)anthracene	25.857	278	4960	0.11	ng	# 91
41) Benzo(g,h,i)perylene	26.528	276	7636	0.15	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
 Data File : BN014848.D
 Acq On : 09 Jun 2021 11:34
 Operator : CG/JU
 Sample : SSTDIC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.2

Manual Integrations
 APPROVED


mohammad
 6/11/2021 1:27:09 PM

Quant Time: Jun 09 12:56:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 12:45:46 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.717 min Scan# 579
 Delta R.T. 0.009 min
 Lab File: BN014848.D
 Acq: 09 Jun 2021 11:34





Tgt Ion: 152 Resp: 2305
 Ion Ratio Lower Upper
 152 100
 150 147.3 119.0 178.6
 115 76.6 62.2 93.4



#2
 1,4-Dioxane
 Concen: 0.27 ng
 RT: 3.281 min Scan# 98
 Delta R.T. -0.000 min
 Lab File: BN014848.D
 Acq: 09 Jun 2021 11:34




Tgt Ion: 88 Resp: 114
 Ion Ratio Lower Upper
 88 100
 43 27.2 12.9 19.3#
 58 57.0 38.1 57.1



Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Manual Integrations
 APPROVED

mohammad
 6/11/2021 1:27:09 PM







#3
n-Nitrosodimethylamine
Concen: 0.10 ng
RT: 3.724 min Scan# 146
Delta R.T. 0.018 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion: 42 Resp: 66
Ion Ratio Lower Upper
42 100
74 195.5 149.5 224.3
44 0.0 4.2 6.2#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM




#4
2-Fluorophenol
Concen: 0.17 ng
RT: 5.338 min Scan# 321
Delta R.T. 0.009 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion: 112 Resp: 1311
Ion Ratio Lower Upper
112 100
64 47.4 39.0 58.6
63 25.5 20.6 30.8









#5
Phenol-d6
Concen: 0.15 ng
RT: 6.887 min Scan# 489
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34





Tgt Ion	Ratio	Lower	Upper
99	100		
42	10.8	8.9	13.3
71	44.1	35.6	53.4



#6
bis(2-Chloroethyl)ether
Concen: 0.14 ng
RT: 7.154 min Scan# 518
Delta R.T. 0.009 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion	Ratio	Lower	Upper
93	100		
63	60.5	46.6	70.0
95	37.2	28.6	42.8



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Manual Integrations
APPROVED

mohammad
6/11/2021 1:27:09 PM







#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.496 min Scan# 816
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:	136	Resp:	9339
Ion Ratio	Lower	Upper	
136	100		
137	11.7	9.8	14.8
54	4.6	5.4	8.2#
68	4.2	4.9	7.3#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM



#8
Nitrobenzene-d5
Concen: 0.15 ng
RT: 8.861 min Scan# 687
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:	82	Resp:	1335
Ion Ratio	Lower	Upper	
82	100		
128	34.7	31.6	47.4
54	42.4	29.0	43.4









#9
Naphthalene
Concen: 0.20 ng
RT: 10.547 min Scan# 820
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:	128	Resp:	5235
Ion Ratio	Lower	Upper	
128	100		
129	12.0	9.1	13.7
127	14.9	11.8	17.6

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM




#10
Hexachlorobutadiene
Concen: 0.22 ng
RT: 10.825 min Scan# 842
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:	225	Resp:	1544
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	62.8	50.2	75.2









#11
2-Methylnaphthalene-d10
Concen: 0.21 ng
RT: 12.088 min Scan# 1026
Delta R.T. 0.009 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34





Tgt Ion:152 Resp: 3685
Ion Ratio Lower Upper
152 100
151 20.9 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.20 ng
RT: 12.163 min Scan# 1043
Delta R.T. 0.009 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion:142 Resp: 3567
Ion Ratio Lower Upper
142 100
141 88.1 72.3 108.5
115 48.7 40.7 61.1



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Manual Integrations
APPROVED

mohammad
6/11/2021 1:27:09 PM



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.345 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion:164 Resp: 6026
Ion Ratio Lower Upper
164 100
162 107.7 83.8 125.6
160 49.4 37.4 56.0

Manual Integrations
APPROVED



mohammad
6/11/2021 1:27:09 PM




#14
2,4,6-Tribromophenol
Concen: 0.13 ng
RT: 15.855 min Scan# 1384
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34



Tgt Ion:330 Resp: 461
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 27.8 22.4 33.6









#15
2-Fluorobiphenyl
Concen: 0.20 ng
RT: 12.975 min Scan# 1157
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:	172	Resp:	4613
Ion Ratio	Lower	Upper	
172	100		
171	36.7	30.6	45.8
170	24.5	20.2	30.4

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM



#16
Acenaphthylene
Concen: 0.19 ng
RT: 14.066 min Scan# 1243
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34


Tgt Ion:	152	Resp:	4759
Ion Ratio	Lower	Upper	
152	100		
151	20.3	16.7	25.1
153	12.9	10.5	15.7





#17
Acenaphthene
Concen: 0.20 ng
RT: 14.409 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion:154 Resp: 3448
Ion Ratio Lower Upper
154 100
153 116.4 90.9 136.3
152 56.8 45.8 68.8

Manual Integrations
APPROVED



mohammad
6/11/2021 1:27:09 PM




#18
Fluorene
Concen: 0.20 ng
RT: 15.398 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34



Tgt Ion:166 Resp: 4398
Ion Ratio Lower Upper
166 100
165 99.2 79.2 118.8
167 14.3 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.103 min Scan# 1486
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:188 Resp: 12788
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.4 9.0 13.4#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.04 ng
RT: 15.500 min Scan# 1356
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:198 Resp: 69
Ion Ratio Lower Upper
198 100
182 66.7 34.5 51.7#
77 0.0 0.0 0.0








#21
4-Bromophenyl-phenylether
Concen: 0.19 ng
RT: 16.300 min Scan# 1420
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion: 248 Resp: 1713

Ion	Ratio	Lower	Upper
248	100		
250	100.9	86.2	129.4
141	64.8	36.2	54.2#





#22
Hexachlorobenzene
Concen: 0.22 ng
RT: 16.410 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34



Tgt Ion: 284 Resp: 2107


Ion	Ratio	Lower	Upper
284	100		
142	27.3	22.2	33.2
249	33.5	26.6	39.8



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Manual Integrations
APPROVED

mohammad
6/11/2021 1:27:09 PM







#23
Atrazine
Concen: 0.15 ng
RT: 16.568 min Scan# 1442
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion	Ratio	Lower	Upper
200	100		
173	32.6	21.7	32.5#
215	53.8	43.1	64.7

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM



#24
Pentachlorophenol
Concen: 0.07 ng
RT: 16.762 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion	Ratio	Lower	Upper
266	100		
264	66.5	51.0	76.4
268	72.5	50.6	76.0









#25
Phenanthrene
Concen: 0.20 ng
RT: 17.140 min Scan# 1489
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.6	23.4
179	15.5	12.1	18.1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM



#26
Anthracene
Concen: 0.19 ng
RT: 17.237 min Scan# 1497
Delta R.T. 0.012 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.1	22.7
179	15.3	12.2	18.2









#27
Fluoranthene-d10
Concen: 0.21 ng
RT: 19.143 min Scan# 1788
Delta R.T. 0.005 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:212 Resp: 8956
Ion Ratio Lower Upper
212 100
106 8.8 7.0 10.6
104 5.1 4.1 6.1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM




#28
Fluoranthene
Concen: 0.20 ng
RT: 19.173 min Scan# 1794
Delta R.T. 0.005 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:202 Resp: 8747
Ion Ratio Lower Upper
202 100
101 8.1 6.6 10.0
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.302 min Scan# 2084
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34





Tgt Ion:240 Resp: 14409
Ion Ratio Lower Upper
240 100
120 6.6 4.2 6.4#
236 27.1 21.4 32.2



#30
Pyrene
Concen: 0.20 ng
RT: 19.535 min Scan# 1867
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion:202 Resp: 8658
Ion Ratio Lower Upper
202 100
200 21.1 16.9 25.3
203 17.6 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Manual Integrations
APPROVED

mohammad
6/11/2021 1:27:09 PM



#31
Terphenyl-d14
Concen: 0.20 ng
RT: 19.748 min Scan# 1910
Delta R.T. 0.005 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion:244 Resp: 7142
Ion Ratio Lower Upper
244 100
212 8.8 6.3 9.5
122 8.0 6.2 9.2

Manual Integrations
APPROVED



mohammad
6/11/2021 1:27:09 PM




#32
Benzo(a)anthracene
Concen: 0.17 ng
RT: 21.292 min Scan# 2083
Delta R.T. 0.010 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion:228 Resp: 7269
Ion Ratio Lower Upper
228 100
226 26.8 23.0 34.6
229 21.3 15.7 23.5





#33
Chrysene
Concen: 0.22 ng
RT: 21.345 min Scan# 2088
Delta R.T. 0.010 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion:228 Resp: 10060
Ion Ratio Lower Upper
228 100
226 29.5 25.0 37.6
229 20.8 15.9 23.9

Manual Integrations
APPROVED



mohammad
6/11/2021 1:27:09 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.17 ng
RT: 21.228 min Scan# 2077
Delta R.T. 0.010 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion:149 Resp: 2395
Ion Ratio Lower Upper
149 100
167 37.1 28.9 43.3
279 6.2 4.9 7.3





#35
Perylene-d12
Concen: 0.40 ng
RT: 23.564 min Scan# 2526
Delta R.T. 0.003 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.3	20.6	31.0
265	26.0	21.4	32.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.12 ng
RT: 25.843 min Scan# 3192
Delta R.T. 0.006 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion	Ratio	Lower	Upper
276	100		
138	9.7	10.0	15.0#
277	26.2	20.4	30.6









#37
Benzo(b)fluoranthene
Concen: 0.18 ng m
RT: 22.886 min Scan# 2328
Delta R.T. 0.003 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34





Tgt Ion:252 Resp: 8876
Ion Ratio Lower Upper
252 100
253 27.5 18.5 27.7
125 15.6 5.8 8.6#



#38
Benzo(k)fluoranthene
Concen: 0.20 ng
RT: 22.927 min Scan# 2340
Delta R.T. -0.000 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34




Tgt Ion:252 Resp: 11065
Ion Ratio Lower Upper
252 100
253 25.4 19.1 28.7
125 8.5 5.4 8.0#



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Manual Integrations
APPROVED

mohammad
6/11/2021 1:27:09 PM







#39
Benzo(a)pyrene
Concen: 0.18 ng
RT: 23.464 min Scan# 2497
Delta R.T. 0.003 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:252 Resp: 8319
Ion Ratio Lower Upper
252 100
253 31.7 19.4 29.2#
125 32.4 7.4 11.2#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Manual Integrations
APPROVED


mohammad
6/11/2021 1:27:09 PM




#40
Dibenzo(a,h)anthracene
Concen: 0.11 ng
RT: 25.857 min Scan# 3196
Delta R.T. 0.006 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34

Tgt Ion:278 Resp: 4960
Ion Ratio Lower Upper
278 100
139 15.6 9.0 13.4#
279 30.3 21.0 31.4







#41
Benzo(g,h,i)perylene
Concen: 0.15 ng
RT: 26.528 min Scan# 3392
Delta R.T. 0.003 min
Lab File: BN014848.D
Acq: 09 Jun 2021 11:34



Tgt Ion	Ratio	Lower	Upper
276	100		
277	25.1	19.7	29.5
138	14.9	9.8	14.6



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Manual Integrations
APPROVED

mohammad
6/11/2021 1:27:09 PM