

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
 Data File : BN014851.D
 Acq On : 09 Jun 2021 13:22
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Jun 09 13:55:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 13:00:31 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	2268	0.40	ng	0.00
7) Naphthalene-d8	10.483	136	8786	0.40	ng	0.00
13) Acenaphthene-d10	14.345	164	5827	0.40	ng	0.00
19) Phenanthrene-d10	17.103	188	12121	0.40	ng	# 0.01
29) Chrysene-d12	21.302	240	15552	0.40	ng	0.00
35) Perylene-d12	23.560	264	14632	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.329	112	10186	1.38	ng	0.00
5) Phenol-d6	6.887	99	12129	1.26	ng	0.00
8) Nitrobenzene-d5	8.848	82	12723	1.56	ng	-0.01
11) 2-Methylnaphthalene-d10	12.079	152	27473	1.66	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	4875	1.40	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	36919	1.67	ng	0.00
27) Fluoranthene-d10	19.138	212	70854	1.76	ng	0.00
31) Terphenyl-d14	19.744	244	63927	1.68	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.282	88	1077	2.55	ng	Qvalue 98
3) n-Nitrosodimethylamine	3.669	42	1257	1.86	ng	# 87
6) bis(2-Chloroethyl)ether	7.145	93	9659	1.31	ng	97
9) Naphthalene	10.534	128	40178	1.66	ng	99
10) Hexachlorobutadiene	10.825	225	11585	1.79	ng	# 99
12) 2-Methylnaphthalene	12.155	142	28626	1.71	ng	100
16) Acenaphthylene	14.066	152	41845	1.72	ng	100
17) Acenaphthene	14.409	154	27481	1.66	ng	99
18) Fluorene	15.399	166	35531	1.71	ng	100
20) 4,6-Dinitro-2-methylph...	15.475	198	1639	1.11	ng	# 48
21) 4-Bromophenyl-phenylether	16.288	248	15080	1.77	ng	# 58
22) Hexachlorobenzene	16.410	284	15996	1.77	ng	99
23) Atrazine	16.568	200	9705	1.76	ng	95
24) Pentachlorophenol	16.750	266	4007	1.45	ng	99
25) Phenanthrene	17.140	178	57834	1.70	ng	100
26) Anthracene	17.225	178	54343	1.82	ng	99
28) Fluoranthene	19.168	202	76330	1.85	ng	100
30) Pyrene	19.530	202	76547	1.66	ng	99
32) Benzo(a)anthracene	21.292	228	78504	1.74	ng	97
33) Chrysene	21.334	228	85149	1.70	ng	99
34) Bis(2-ethylhexyl)phtha...	21.228	149	26212	1.74	ng	98
36) Indeno(1,2,3-cd)pyrene	25.837	276	84043	1.53	ng	99
37) Benzo(b)fluoranthene	22.883	252	83712	1.76	ng	98
38) Benzo(k)fluoranthene	22.927	252	95349	1.72	ng	96
39) Benzo(a)pyrene	23.465	252	77773	1.67	ng	# 96
40) Dibenzo(a,h)anthracene	25.850	278	66290	1.52	ng	95
41) Benzo(g,h,i)perylene	26.528	276	76775	1.56	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
Data File : BN014851.D
Acq On : 09 Jun 2021 13:22
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Jun 09 13:55:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 13:00:31 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.708 min Scan# 578
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 2268
Ion Ratio Lower Upper
152 100
150 156.6 119.0 178.6
115 80.2 62.2 93.4



#2
1,4-Dioxane
Concen: 2.55 ng
RT: 3.282 min Scan# 98
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion: 88 Resp: 1077
Ion Ratio Lower Upper
88 100
43 15.3 12.9 19.3
58 45.8 38.1 57.1









#3
n-Nitrosodimethylamine
Concen: 1.86 ng
RT: 3.669 min Scan# 140
Delta R.T. -0.037 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 1257
Ion Ratio Lower Upper
42 100
74 206.1 149.5 224.3
44 8.4 4.2 6.2#



#4
2-Fluorophenol
Concen: 1.38 ng
RT: 5.329 min Scan# 320
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion: 112 Resp: 10186
Ion Ratio Lower Upper
112 100
64 48.2 39.0 58.6
63 25.9 20.6 30.8









#5
Phenol-d6
Concen: 1.26 ng
RT: 6.887 min Scan# 489
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	99	Resp:	12129
Ion	Ratio	Lower	Upper
99	100		
42	11.0	8.9	13.3
71	44.0	35.6	53.4



#6
bis(2-Chloroethyl)ether
Concen: 1.31 ng
RT: 7.145 min Scan# 517
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:	93	Resp:	9659
Ion	Ratio	Lower	Upper
93	100		
63	56.4	46.6	70.0
95	33.9	28.6	42.8









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.483 min Scan# 815
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	8786
Ion Ratio	Lower	Upper	
136	100		
137	11.8	9.8	14.8
54	6.8	5.4	8.2
68	5.5	4.9	7.3



#8
Nitrobenzene-d5
Concen: 1.56 ng
RT: 8.848 min Scan# 686
Delta R.T. -0.013 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:	82	Resp:	12723
Ion Ratio	Lower	Upper	
82	100		
128	29.9	31.6	47.4#
54	41.5	29.0	43.4









#9
Naphthalene
Concen: 1.66 ng
RT: 10.534 min Scan# 819
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	128	Resp:	40178
Ion Ratio	Lower	Upper	
128	100		
129	10.9	9.1	13.7
127	14.1	11.8	17.6



#10
Hexachlorobutadiene
Concen: 1.79 ng
RT: 10.825 min Scan# 842
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:	225	Resp:	11585
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.4	50.2	75.2









#11
2-Methylnaphthalene-d10
Concen: 1.66 ng
RT: 12.079 min Scan# 1024
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 27473
Ion Ratio Lower Upper
152 100
151 20.8 16.7 25.1



#12
2-Methylnaphthalene
Concen: 1.71 ng
RT: 12.155 min Scan# 1041
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:142 Resp: 28626
Ion Ratio Lower Upper
142 100
141 90.5 72.3 108.5
115 50.2 40.7 61.1









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.345 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 5827
Ion Ratio Lower Upper
164 100
162 104.8 83.8 125.6
160 47.2 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 1.40 ng
RT: 15.843 min Scan# 1383
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:330 Resp: 4875
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 27.3 22.4 33.6









#15
2-Fluorobiphenyl
Concen: 1.67 ng
RT: 12.963 min Scan# 1156
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 36919
Ion Ratio Lower Upper
172 100
171 37.3 30.6 45.8
170 24.1 20.2 30.4



#16
Acenaphthylene
Concen: 1.72 ng
RT: 14.066 min Scan# 1243
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:152 Resp: 41845
Ion Ratio Lower Upper
152 100
151 20.9 16.7 25.1
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 1.66 ng
RT: 14.409 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 27481
Ion Ratio Lower Upper
154 100
153 114.0 90.9 136.3
152 55.4 45.8 68.8



#18
Fluorene
Concen: 1.71 ng
RT: 15.399 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:166 Resp: 35531
Ion Ratio Lower Upper
166 100
165 98.9 79.2 118.8
167 13.8 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.103 min Scan# 1486
Delta R.T. 0.012 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 12121
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 5.4 9.0 13.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 1.11 ng
RT: 15.475 min Scan# 1354
Delta R.T. -0.013 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:198 Resp: 1639
Ion Ratio Lower Upper
198 100
182 9.7 34.5 51.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.77 ng
RT: 16.288 min Scan# 1419
Delta R.T. -0.012 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 15080
Ion Ratio Lower Upper
248 100
250 88.8 86.2 129.4
141 109.8 36.2 54.2#



#22
Hexachlorobenzene
Concen: 1.77 ng
RT: 16.410 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:284 Resp: 15996
Ion Ratio Lower Upper
284 100
142 28.3 22.2 33.2
249 33.3 26.6 39.8









#23
Atrazine
Concen: 1.76 ng
RT: 16.568 min Scan# 1442
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 9705
Ion Ratio Lower Upper
200 100
173 22.1 21.7 32.5
215 51.5 43.1 64.7



#24
Pentachlorophenol
Concen: 1.45 ng
RT: 16.750 min Scan# 1457
Delta R.T. -0.012 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:266 Resp: 4007
Ion Ratio Lower Upper
266 100
264 62.0 51.0 76.4
268 62.9 50.6 76.0









#25
Phenanthrene
Concen: 1.70 ng
RT: 17.140 min Scan# 1489
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 57834
Ion Ratio Lower Upper
178 100
176 19.5 15.6 23.4
179 15.4 12.1 18.1



#26
Anthracene
Concen: 1.82 ng
RT: 17.225 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:178 Resp: 54343
Ion Ratio Lower Upper
178 100
176 19.1 15.1 22.7
179 15.6 12.2 18.2









#27
Fluoranthene-d10
Concen: 1.76 ng
RT: 19.138 min Scan# 1787
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 70854
Ion Ratio Lower Upper
212 100
106 8.8 7.0 10.6
104 5.1 4.1 6.1



#28
Fluoranthene
Concen: 1.85 ng
RT: 19.168 min Scan# 1793
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:202 Resp: 76330
Ion Ratio Lower Upper
202 100
101 8.2 6.6 10.0
203 17.5 13.9 20.9





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.302 min Scan# 2084

Delta R.T. -0.000 min

Lab File: BN014851.D

Acq: 09 Jun 2021 13:22

Tgt Ion:240 Resp: 15552

Ion Ratio Lower Upper

240 100

120 5.6 4.2 6.4





236 27.3 21.4 32.2

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6



#30

Pyrene

Concen: 1.66 ng

RT: 19.530 min Scan# 1866

Delta R.T. -0.005 min

Lab File: BN014851.D

Acq: 09 Jun 2021 13:22




Tgt Ion:202 Resp: 76547


Ion Ratio Lower Upper

202 100

200 21.4 16.9 25.3

203 17.6 13.9 20.9









#31
 Terphenyl-d14
 Concen: 1.68 ng
 RT: 19.744 min Scan# 1909
 Delta R.T. -0.000 min
 Lab File: BN014851.D
 Acq: 09 Jun 2021 13:22




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244	Resp:	63927
Ion Ratio	Lower	Upper
244	100	
212	7.5	6.3 9.5
122	7.0	6.2 9.2



#32
 Benzo(a)anthracene
 Concen: 1.74 ng
 RT: 21.292 min Scan# 2083
 Delta R.T. 0.010 min
 Lab File: BN014851.D
 Acq: 09 Jun 2021 13:22

Tgt Ion:228	Resp:	78504
Ion Ratio	Lower	Upper
228	100	
226	26.4	23.0 34.6
229	20.1	15.7 23.5









#33
Chrysene
Concen: 1.70 ng
RT: 21.334 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	228	Resp:	85149
Ion Ratio	Lower	Upper	
228	100		
226	31.0	25.0	37.6
229	19.1	15.9	23.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.74 ng
RT: 21.228 min Scan# 2077
Delta R.T. 0.010 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:	149	Resp:	26212
Ion Ratio	Lower	Upper	
149	100		
167	37.0	28.9	43.3
279	6.6	4.9	7.3





#35

Perylene-d12

Concen: 0.40 ng

RT: 23.560 min Scan# 2525

Delta R.T. -0.000 min

Lab File: BN014851.D

Acq: 09 Jun 2021 13:22





Tgt Ion:264 Resp: 14632

Ion Ratio Lower Upper

264 100

260 26.1 20.6 31.0

265 26.6 21.4 32.2



#36

Indeno(1,2,3-cd)pyrene

Concen: 1.53 ng

RT: 25.837 min Scan# 3190

Delta R.T. -0.000 min

Lab File: BN014851.D

Acq: 09 Jun 2021 13:22




Tgt Ion:276 Resp: 84043


Ion Ratio Lower Upper

276 100

138 13.0 10.0 15.0

277 25.6 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 1.76 ng
RT: 22.883 min Scan# 2327
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 83712
Ion Ratio Lower Upper
252 100
253 22.2 18.5 27.7
125 6.0 5.8 8.6



#38
Benzo(k)fluoranthene
Concen: 1.72 ng
RT: 22.927 min Scan# 2340
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:252 Resp: 95349
Ion Ratio Lower Upper
252 100
253 22.1 19.1 28.7
125 5.7 5.4 8.0









#39
Benzo(a)pyrene
Concen: 1.67 ng
RT: 23.465 min Scan# 2497
Delta R.T. 0.003 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 77773
Ion Ratio Lower Upper
252 100
253 22.3 19.4 29.2
125 7.3 7.4 11.2#



#40
Dibenzo(a,h)anthracene
Concen: 1.52 ng
RT: 25.850 min Scan# 3194
Delta R.T. -0.000 min
Lab File: BN014851.D
Acq: 09 Jun 2021 13:22

Tgt Ion:278 Resp: 66290
Ion Ratio Lower Upper
278 100
139 9.3 9.0 13.4
279 23.5 21.0 31.4





#41
 Benzo(g,h,i)perylene
 Concen: 1.56 ng
 RT: 26.528 min Scan# 3392
 Delta R.T. 0.003 min
 Lab File: BN014851.D
 Acq: 09 Jun 2021 13:22

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	276	Resp:	76775
Ion Ratio	Lower	Upper	
276	100		
277	23.7	19.7	29.5
138	11.2	9.8	14.6

