

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
 Data File : BN014852.D
 Acq On : 09 Jun 2021 13:59
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Manual Integrations
 APPROVED

mohammad

6/11/2021 1:27:12 PM

Quant Time: Jun 09 14:28:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 13:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	2371	0.40	ng	0.00
7) Naphthalene-d8	10.483	136	8862	0.40	ng	0.00
13) Acenaphthene-d10	14.346	164	5901	0.40	ng	0.00
19) Phenanthrene-d10	17.103	188	12327	0.40	ng	# 0.01
29) Chrysene-d12	21.302	240	15913	0.40	ng	0.00
35) Perylene-d12	23.564	264	15141	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.329	112	20475	2.65	ng	0.00
5) Phenol-d6	6.887	99	24844	2.46	ng	0.00
8) Nitrobenzene-d5	8.848	82	26825	3.26	ng	-0.01
11) 2-Methylnaphthalene-d10	12.079	152	53559	3.21	ng	0.00
14) 2,4,6-Tribromophenol	0.000	330	0d	0.00	ng	
15) 2-Fluorobiphenyl	12.963	172	71671	3.21	ng	0.00
27) Fluoranthene-d10	19.138	212	142283	3.48	ng	0.00
31) Terphenyl-d14	19.744	244	125262	3.22	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.282	88	2396	5.44	ng	95
3) n-Nitrosodimethylamine	3.651	42	3268m	4.64	ng	
6) bis(2-Chloroethyl)ether	7.145	93	19392	2.52	ng	97
9) Naphthalene	10.534	128	78107	3.19	ng	98
10) Hexachlorobutadiene	10.826	225	22444	3.44	ng	# 98
12) 2-Methylnaphthalene	12.155	142	55550	3.28	ng	100
16) Acenaphthylene	14.066	152	83956	3.42	ng	100
17) Acenaphthene	14.409	154	54155	3.23	ng	99
18) Fluorene	15.399	166	68674	3.26	ng	100
20) 4,6-Dinitro-2-methylph...	15.475	198	4240	2.82	ng	# 43
21) 4-Bromophenyl-phenylether	16.288	248	30141	3.47	ng	# 62
22) Hexachlorobenzene	16.410	284	31565	3.43	ng	100
24) Pentachlorophenol	16.750	266	10068	3.59	ng	99
25) Phenanthrene	17.140	178	114160	3.30	ng	100
26) Anthracene	17.225	178	108196	3.56	ng	100
28) Fluoranthene	19.168	202	149245	3.55	ng	100
30) Pyrene	19.530	202	151430	3.21	ng	99
32) Benzo(a)anthracene	21.281	228	161888	3.51	ng	99
33) Chrysene	21.334	228	165699	3.24	ng	99
36) Indeno(1,2,3-cd)pyrene	25.837	276	182292	3.20	ng	99
37) Benzo(b)fluoranthene	22.883	252	168308	3.43	ng	# 97
38) Benzo(k)fluoranthene	22.927	252	185000	3.22	ng	96
39) Benzo(a)pyrene	23.461	252	159093	3.31	ng	# 95
40) Dibenzo(a,h)anthracene	25.850	278	146320	3.24	ng	# 95
41) Benzo(g,h,i)perylene	26.525	276	160504	3.16	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
Data File : BN014852.D
Acq On : 09 Jun 2021 13:59
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Manual Integrations
APPROVED

mohammad

6/11/2021 1:27:12 PM


Quant Time: Jun 09 14:28:11 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 13:00:31 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.708 min Scan# 578
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2






Tgt Ion:152 Resp: 2371
Ion Ratio Lower Upper
152 100
150 155.8 119.0 178.6
115 80.1 62.2 93.4


Manual Integrations
APPROVED

mohammad



6/11/2021 1:27:12 PM




#2
1,4-Dioxane
Concen: 5.44 ng
RT: 3.282 min Scan# 98
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59



Tgt Ion: 88 Resp: 2396
Ion Ratio Lower Upper
88 100
43 14.1 12.9 19.3
58 44.2 38.1 57.1





#3
n-Nitrosodimethylamine
Concen: 4.64 ng m
RT: 3.651 min Scan# 138
Delta R.T. -0.055 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion: 42 Resp: 3268
Ion Ratio Lower Upper
42 100
74 131.1 149.5 224.3#
44 4.3 4.2 6.2

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM




#4
2-Fluorophenol
Concen: 2.65 ng
RT: 5.329 min Scan# 320
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion: 112 Resp: 20475
Ion Ratio Lower Upper
112 100
64 48.3 39.0 58.6
63 26.2 20.6 30.8









#5
Phenol-d6
Concen: 2.46 ng
RT: 6.887 min Scan# 489
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion:	99	Resp:	24844
Ion	Ratio	Lower	Upper
99	100		
42	11.0	8.9	13.3
71	44.2	35.6	53.4



#6
bis(2-Chloroethyl)ether
Concen: 2.52 ng
RT: 7.145 min Scan# 517
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59



Tgt Ion:	93	Resp:	19392
Ion	Ratio	Lower	Upper
93	100		
63	56.4	46.6	70.0
95	34.1	28.6	42.8




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED

mohammad

6/11/2021 1:27:12 PM



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.483 min Scan# 815
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	136	Resp:	8862
Ion Ratio	Lower	Upper	
136	100		
137	11.6	9.8	14.8
54	6.8	5.4	8.2
68	5.5	4.9	7.3

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#8
Nitrobenzene-d5
Concen: 3.26 ng
RT: 8.848 min Scan# 686
Delta R.T. -0.013 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion:	82	Resp:	26825
Ion Ratio	Lower	Upper	
82	100		
128	30.0	31.6	47.4#
54	40.2	29.0	43.4





#9
Naphthalene
Concen: 3.19 ng
RT: 10.534 min Scan# 819
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion:	128	Resp:	78107
Ion Ratio	Lower	Upper	
128	100		
129	10.8	9.1	13.7
127	14.1	11.8	17.6

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#10
Hexachlorobutadiene
Concen: 3.44 ng
RT: 10.826 min Scan# 842
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion:	225	Resp:	22444
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.1	50.2	75.2





#11
2-Methylnaphthalene-d10
Concen: 3.21 ng
RT: 12.079 min Scan# 1024
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion:152 Resp: 53559
Ion Ratio Lower Upper
152 100
151 21.0 16.7 25.1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM




#12
2-Methylnaphthalene
Concen: 3.28 ng
RT: 12.155 min Scan# 1041
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion:142 Resp: 55550
Ion Ratio Lower Upper
142 100
141 90.2 72.3 108.5
115 50.7 40.7 61.1








#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59




Tgt Ion:164 Resp: 5901
Ion Ratio Lower Upper
164 100
162 102.6 83.8 125.6
160 46.9 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.00 ng
Expected RT: 15.84 min

Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion: 330
Sig Exp Ratio
330 100
332 0.0
141 28.0




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED

mohammad

6/11/2021 1:27:12 PM



#15
2-Fluorobiphenyl
Concen: 3.21 ng
RT: 12.963 min Scan# 1156
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.8	30.6	45.8
170	23.8	20.2	30.4

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#16
Acenaphthylene
Concen: 3.42 ng
RT: 14.066 min Scan# 1243
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59


Tgt Ion	Ratio	Lower	Upper
152	100		
151	21.0	16.7	25.1
153	13.1	10.5	15.7





#17
Acenaphthene
Concen: 3.23 ng
RT: 14.409 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2






Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.3	90.9	136.3
152	54.9	45.8	68.8


Manual Integrations
APPROVED

mohammad



6/11/2021 1:27:12 PM




#18
Fluorene
Concen: 3.26 ng
RT: 15.399 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59



Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.5	79.2	118.8
167	13.7	10.7	16.1





#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.103 min Scan# 1486
Delta R.T. 0.012 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	5.8	9.0	13.4#

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#20
4,6-Dinitro-2-methylphenol
Concen: 2.82 ng
RT: 15.475 min Scan# 1354
Delta R.T. -0.013 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion	Ratio	Lower	Upper
198	100		
182	6.3	34.5	51.7#
77	0.0	0.0	0.0





#21
4-Bromophenyl-phenylether
Concen: 3.47 ng
RT: 16.288 min Scan# 1419
Delta R.T. -0.012 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion	Ratio	Lower	Upper
248	100		
250	90.4	86.2	129.4
141	102.8	36.2	54.2#

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#22
Hexachlorobenzene
Concen: 3.43 ng
RT: 16.410 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion	Ratio	Lower	Upper
284	100		
142	27.4	22.2	33.2
249	33.1	26.6	39.8





#24
Pentachlorophenol
Concen: 3.59 ng
RT: 16.750 min Scan# 1457
Delta R.T. -0.012 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion: 266 Resp: 10068
Ion Ratio Lower Upper
266 100
264 63.5 51.0 76.4
268 64.6 50.6 76.0

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM




#25
Phenanthrene
Concen: 3.30 ng
RT: 17.140 min Scan# 1489
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion: 178 Resp: 114160
Ion Ratio Lower Upper
178 100
176 19.6 15.6 23.4
179 15.3 12.1 18.1









#26
 Anthracene
 Concen: 3.56 ng
 RT: 17.225 min Scan# 1496
 Delta R.T. -0.000 min
 Lab File: BN014852.D
 Acq: 09 Jun 2021 13:59





Tgt Ion:178 Resp: 108196
 Ion Ratio Lower Upper
 178 100
 176 19.0 15.1 22.7
 179 15.5 12.2 18.2



#27
 Fluoranthene-d10
 Concen: 3.48 ng
 RT: 19.138 min Scan# 1787
 Delta R.T. -0.000 min
 Lab File: BN014852.D
 Acq: 09 Jun 2021 13:59



Tgt Ion:212 Resp: 142283
 Ion Ratio Lower Upper
 212 100
 106 8.8 7.0 10.6
 104 5.1 4.1 6.1




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Manual Integrations
 APPROVED

mohammad

6/11/2021 1:27:12 PM



#28
Fluoranthene
Concen: 3.55 ng
RT: 19.168 min Scan# 1793
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion:202 Resp: 149245
Ion Ratio Lower Upper
202 100
101 8.4 6.6 10.0
203 17.5 13.9 20.9

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.302 min Scan# 2084
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion:240 Resp: 15913
Ion Ratio Lower Upper
240 100
120 5.5 4.2 6.4
236 27.1 21.4 32.2





#30
Pyrene
Concen: 3.21 ng
RT: 19.530 min Scan# 1866
Delta R.T. -0.005 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion:	202	Resp:	151430
Ion Ratio	Lower	Upper	
202	100		
200	21.4	16.9	25.3
203	17.6	13.9	20.9

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#31
Terphenyl-d14
Concen: 3.22 ng
RT: 19.744 min Scan# 1909
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion:	244	Resp:	125262
Ion Ratio	Lower	Upper	
244	100		
212	7.6	6.3	9.5
122	6.9	6.2	9.2





#32
Benzo(a)anthracene
Concen: 3.51 ng
RT: 21.281 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion:228 Resp: 161888
Ion Ratio Lower Upper
228 100
226 28.4 23.0 34.6
229 19.1 15.7 23.5

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#33
Chrysene
Concen: 3.24 ng
RT: 21.334 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59


Tgt Ion:228 Resp: 165699
Ion Ratio Lower Upper
228 100
226 31.0 25.0 37.6
229 19.3 15.9 23.9





#35
Perylene-d12
Concen: 0.40 ng
RT: 23.564 min Scan# 2526
Delta R.T. 0.003 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2






Tgt Ion:264 Resp: 15141
Ion Ratio Lower Upper
264 100
260 25.7 20.6 31.0
265 25.6 21.4 32.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#36
Indeno(1,2,3-cd)pyrene
Concen: 3.20 ng
RT: 25.837 min Scan# 3190
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion:276 Resp: 182292
Ion Ratio Lower Upper
276 100
138 12.9 10.0 15.0
277 25.6 20.4 30.6





#37
Benzo(b)fluoranthene
Concen: 3.43 ng
RT: 22.883 min Scan# 2327
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion:252 Resp: 168308
Ion Ratio Lower Upper
252 100
253 21.8 18.5 27.7
125 5.7 5.8 8.6#

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#38
Benzo(k)fluoranthene
Concen: 3.22 ng
RT: 22.927 min Scan# 2340
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion:252 Resp: 185000
Ion Ratio Lower Upper
252 100
253 21.8 19.1 28.7
125 5.6 5.4 8.0





#39
Benzo(a)pyrene
Concen: 3.31 ng
RT: 23.461 min Scan# 2496
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59





Tgt Ion:252 Resp: 159093
Ion Ratio Lower Upper
252 100
253 22.0 19.4 29.2
125 6.6 7.4 11.2#

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED




mohammad


6/11/2021 1:27:12 PM



#40
Dibenzo(a,h)anthracene
Concen: 3.24 ng
RT: 25.850 min Scan# 3194
Delta R.T. -0.000 min
Lab File: BN014852.D
Acq: 09 Jun 2021 13:59

Tgt Ion:278 Resp: 146320
Ion Ratio Lower Upper
278 100
139 8.7 9.0 13.4#
279 23.7 21.0 31.4





#41

Benzo(g,h,i)perylene


Concen: 3.16 ng

RT: 26.525 min Scan# 3391

Delta R.T. -0.000 min

Lab File: BN014852.D

Acq: 09 Jun 2021 13:59





Tgt Ion: 276 Resp: 160504

Ion Ratio Lower Upper

276 100

277 23.8 19.7 29.5

138 11.2 9.8 14.6



Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2

Manual Integrations

APPROVED

mohammad

6/11/2021 1:27:12 PM