

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
 Data File : BN014856.D
 Acq On : 09 Jun 2021 17:01
 Operator : CG/JU
 Sample : PB136126BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB136126BS

Quant Time: Jun 09 17:31:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 16:35:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	2354	0.40	ng	0.00
7) Naphthalene-d8	10.483	136	9435	0.40	ng	0.00
13) Acenaphthene-d10	14.346	164	6294	0.40	ng	0.00
19) Phenanthrene-d10	17.104	188	13265	0.40	ng	# 0.01
29) Chrysene-d12	21.303	240	15562	0.40	ng	0.00
35) Perylene-d12	23.564	264	15006	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.329	112	2450	0.39	ng	0.00
5) Phenol-d6	6.887	99	2818	0.38	ng	0.00
8) Nitrobenzene-d5	8.861	82	2782	0.36	ng	0.00
11) 2-Methylnaphthalene-d10	12.084	152	6559	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	964	0.37	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	8861	0.37	ng	0.00
27) Fluoranthene-d10	19.138	212	15799	0.35	ng	0.00
31) Terphenyl-d14	19.744	244	14290	0.37	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.282	88	224	0.33	ng	# 87
3) n-Nitrosodimethylamine	3.697	42	222	0.42	ng	# 97
6) bis(2-Chloroethyl)ether	7.145	93	2143	0.38	ng	98
9) Naphthalene	10.534	128	9815	0.38	ng	99
10) Hexachlorobutadiene	10.826	225	2875	0.38	ng	# 99
12) 2-Methylnaphthalene	12.159	142	6879	0.38	ng	98
16) Acenaphthylene	14.067	152	9393	0.36	ng	99
17) Acenaphthene	14.409	154	6463	0.36	ng	99
18) Fluorene	15.399	166	8467	0.37	ng	100
20) 4,6-Dinitro-2-methylph...	15.488	198	151	0.39	ng	96
21) 4-Bromophenyl-phenylether	16.300	248	3417	0.36	ng	97
22) Hexachlorobenzene	16.410	284	3997	0.37	ng	99
23) Atrazine	16.568	200	1791	0.35	ng	98
24) Pentachlorophenol	16.763	266	546	0.46	ng	97
25) Phenanthrene	17.140	178	13854	0.37	ng	99
26) Anthracene	17.225	178	11917	0.36	ng	100
28) Fluoranthene	19.168	202	17356	0.36	ng	100
30) Pyrene	19.535	202	17235	0.37	ng	99
32) Benzo(a)anthracene	21.292	228	15675	0.35	ng	96
33) Chrysene	21.345	228	19629	0.38	ng	97
34) Bis(2-ethylhexyl)phtha...	21.228	149	4592	0.34	ng	99
36) Indeno(1,2,3-cd)pyrene	25.840	276	14960	0.36	ng	99
37) Benzo(b)fluoranthene	22.883	252	17307	0.36	ng	99
38) Benzo(k)fluoranthene	22.927	252	20933	0.37	ng	99
39) Benzo(a)pyrene	23.461	252	16783	0.37	ng	99
40) Dibenzo(a,h)anthracene	25.854	278	11205	0.36	ng	99
41) Benzo(g,h,i)perylene	26.528	276	15034	0.34	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060921\
Data File : BN014856.D
Acq On : 09 Jun 2021 17:01
Operator : CG/JU
Sample : PB136126BS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB136126BS

Quant Time: Jun 09 17:31:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 16:35:36 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.708 min Scan# 578
 Delta R.T. -0.000 min
 Lab File: BN014856.D
 Acq: 09 Jun 2021 17:01




Instrument :
 BNA_N
 ClientSampleId :
 PB136126BS


Tgt Ion: 152 Resp: 2354
 Ion Ratio Lower Upper
 152 100
 150 150.9 119.0 178.6
 115 81.5 62.2 93.4



#2
 1,4-Dioxane
 Concen: 0.33 ng
 RT: 3.282 min Scan# 98
 Delta R.T. -0.000 min
 Lab File: BN014856.D
 Acq: 09 Jun 2021 17:01

Tgt Ion: 88 Resp: 224
 Ion Ratio Lower Upper
 88 100
 43 18.3 12.9 19.3
 58 58.5 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.42 ng
RT: 3.697 min Scan# 143
Delta R.T. -0.009 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion: 42 Resp: 222
Ion Ratio Lower Upper
42 100
74 191.0 149.5 224.3
44 0.0 4.2 6.2#



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.329 min Scan# 320
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion: 112 Resp: 2450
Ion Ratio Lower Upper
112 100
64 48.1 39.0 58.6
63 26.0 20.6 30.8









#5
Phenol-d6
Concen: 0.38 ng
RT: 6.887 min Scan# 489
Delta R.T. 0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:	99	Resp:	2818
Ion	Ratio	Lower	Upper
99	100		
42	11.1	8.9	13.3
71	42.9	35.6	53.4



#6
bis(2-Chloroethyl)ether
Concen: 0.38 ng
RT: 7.145 min Scan# 517
Delta R.T. 0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:	93	Resp:	2143
Ion	Ratio	Lower	Upper
93	100		
63	55.9	46.6	70.0
95	35.0	28.6	42.8









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.483 min Scan# 815
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:	136	Resp:	9435
Ion	Ratio	Lower	Upper
136	100		
137	11.8	9.8	14.8
54	7.1	5.4	8.2
68	5.9	4.9	7.3



#8
Nitrobenzene-d5
Concen: 0.36 ng
RT: 8.861 min Scan# 687
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:	82	Resp:	2782
Ion	Ratio	Lower	Upper
82	100		
128	38.3	31.6	47.4
54	37.7	29.0	43.4









#9
Naphthalene
Concen: 0.38 ng
RT: 10.534 min Scan# 819
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:	128	Resp:	9815
Ion	Ratio	Lower	Upper
128	100		
129	11.8	9.1	13.7
127	14.8	11.8	17.6



#10
Hexachlorobutadiene
Concen: 0.38 ng
RT: 10.826 min Scan# 842
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:	225	Resp:	2875
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.5	50.2	75.2









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.084 min Scan# 1025
Delta R.T. 0.004 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:152 Resp: 6559
Ion Ratio Lower Upper
152 100
151 21.0 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.159 min Scan# 1042
Delta R.T. 0.004 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:142 Resp: 6879
Ion Ratio Lower Upper
142 100
141 88.7 72.3 108.5
115 49.3 40.7 61.1









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:164 Resp: 6294
Ion Ratio Lower Upper
164 100
162 102.8 83.8 125.6
160 46.0 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.37 ng
RT: 15.843 min Scan# 1383
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:330 Resp: 964
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 29.0 22.4 33.6









#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 12.963 min Scan# 1156
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:	172	Resp:	8861
Ion Ratio	Lower	Upper	
172	100		
171	37.8	30.6	45.8
170	24.6	20.2	30.4



#16
Acenaphthylene
Concen: 0.36 ng
RT: 14.067 min Scan# 1243
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:	152	Resp:	9393
Ion Ratio	Lower	Upper	
152	100		
151	21.5	16.7	25.1
153	13.3	10.5	15.7









#17
Acenaphthene
Concen: 0.36 ng
RT: 14.409 min Scan# 1270
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:154 Resp: 6463
Ion Ratio Lower Upper
154 100
153 115.8 90.9 136.3
152 57.3 45.8 68.8



#18
Fluorene
Concen: 0.37 ng
RT: 15.399 min Scan# 1348
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:166 Resp: 8467
Ion Ratio Lower Upper
166 100
165 98.6 79.2 118.8
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.104 min Scan# 1486
Delta R.T. 0.012 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:188 Resp: 13265
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.0 9.0 13.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.39 ng
RT: 15.488 min Scan# 1355
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:198 Resp: 151
Ion Ratio Lower Upper
198 100
182 45.7 34.5 51.7
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.300 min Scan# 1420
Delta R.T. 0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:	248	Resp:	3417
Ion	Ratio	Lower	Upper
248	100		
250	106.9	86.2	129.4
141	49.6	36.2	54.2



#22
Hexachlorobenzene
Concen: 0.37 ng
RT: 16.410 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:	284	Resp:	3997
Ion	Ratio	Lower	Upper
284	100		
142	27.2	22.2	33.2
249	32.4	26.6	39.8









#23
Atrazine
Concen: 0.35 ng
RT: 16.568 min Scan# 1442
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:200 Resp: 1791
Ion Ratio Lower Upper
200 100
173 27.6 21.7 32.5
215 51.8 43.1 64.7



#24
Pentachlorophenol
Concen: 0.46 ng
RT: 16.763 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:266 Resp: 546
Ion Ratio Lower Upper
266 100
264 59.5 51.0 76.4
268 64.5 50.6 76.0









#25
Phenanthrene
Concen: 0.37 ng
RT: 17.140 min Scan# 1489
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:178 Resp: 13854
Ion Ratio Lower Upper
178 100
176 19.3 15.6 23.4
179 15.5 12.1 18.1



#26
Anthracene
Concen: 0.36 ng
RT: 17.225 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:178 Resp: 11917
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.35 ng
RT: 19.138 min Scan# 1787
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:212 Resp: 15799
Ion Ratio Lower Upper
212 100
106 8.8 7.0 10.6
104 5.3 4.1 6.1



#28
Fluoranthene
Concen: 0.36 ng
RT: 19.168 min Scan# 1793
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:202 Resp: 17356
Ion Ratio Lower Upper
202 100
101 8.2 6.6 10.0
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.303 min Scan# 2084
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:240 Resp: 15562
Ion Ratio Lower Upper
240 100
120 5.5 4.2 6.4
236 26.8 21.4 32.2



#30
Pyrene
Concen: 0.37 ng
RT: 19.535 min Scan# 1867
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:202 Resp: 17235
Ion Ratio Lower Upper
202 100
200 21.2 16.9 25.3
203 17.7 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.37 ng
 RT: 19.744 min Scan# 1909
 Delta R.T. -0.000 min
 Lab File: BN014856.D
 Acq: 09 Jun 2021 17:01




Instrument :
 BNA_N
 ClientSampleId :
 PB136126BS


Tgt Ion:244	Resp:	14290
Ion Ratio	Lower	Upper
244	100	
212	8.3	6.3 9.5
122	8.1	6.2 9.2



#32
 Benzo(a)anthracene
 Concen: 0.35 ng
 RT: 21.292 min Scan# 2083
 Delta R.T. 0.011 min
 Lab File: BN014856.D
 Acq: 09 Jun 2021 17:01

Tgt Ion:228	Resp:	15675
Ion Ratio	Lower	Upper
228	100	
226	26.4	23.0 34.6
229	20.7	15.7 23.5









#33
Chrysene
Concen: 0.38 ng
RT: 21.345 min Scan# 2088
Delta R.T. 0.011 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampleId :
PB136126BS


Tgt Ion:	228	Resp:	19629
Ion Ratio	Lower	Upper	
228	100		
226	29.4	25.0	37.6
229	20.4	15.9	23.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.34 ng
RT: 21.228 min Scan# 2077
Delta R.T. 0.011 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:	149	Resp:	4592
Ion Ratio	Lower	Upper	
149	100		
167	36.5	28.9	43.3
279	6.5	4.9	7.3





#35

Perylene-d12

Concen: 0.40 ng

RT: 23.564 min Scan# 2526

Delta R.T. 0.003 min

Lab File: BN014856.D

Acq: 09 Jun 2021 17:01





Tgt Ion:264	Resp:	15006
Ion Ratio	Lower	Upper
264	100	
260	25.9	20.6 31.0
265	26.5	21.4 32.2

Instrument :

BNA_N

ClientSampleId :

PB136126BS



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.36 ng




RT: 25.840 min Scan# 3191


Delta R.T. 0.003 min

Lab File: BN014856.D

Acq: 09 Jun 2021 17:01

Tgt Ion:276	Resp:	14960
Ion Ratio	Lower	Upper
276	100	
138	12.9	10.0 15.0
277	26.1	20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.36 ng
RT: 22.883 min Scan# 2327
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01




Instrument :
BNA_N
ClientSampled :
PB136126BS


Tgt Ion:252 Resp: 17307
Ion Ratio Lower Upper
252 100
253 23.7 18.5 27.7
125 7.7 5.8 8.6



#38
Benzo(k)fluoranthene
Concen: 0.37 ng
RT: 22.927 min Scan# 2340
Delta R.T. -0.000 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Tgt Ion:252 Resp: 20933
Ion Ratio Lower Upper
252 100
253 23.1 19.1 28.7
125 6.7 5.4 8.0





#39

Benzo(a)pyrene

Concen: 0.37 ng

RT: 23.461 min Scan# 2496

Delta R.T. -0.000 min

Lab File: BN014856.D

Acq: 09 Jun 2021 17:01





Tgt Ion:	252	Resp:	16783
Ion Ratio	Lower	Upper	
252	100		
253	23.8	19.4	29.2
125	9.7	7.4	11.2

Instrument :

BNA_N

ClientSampleId :

PB136126BS



#40

Dibenzo(a,h)anthracene

Concen: 0.36 ng




RT: 25.854 min Scan# 3195


Delta R.T. 0.003 min

Lab File: BN014856.D

Acq: 09 Jun 2021 17:01

Tgt Ion:	278	Resp:	11205
Ion Ratio	Lower	Upper	
278	100		
139	11.7	9.0	13.4
279	26.4	21.0	31.4





#41
Benzo(g,h,i)perylene
Concen: 0.34 ng
RT: 26.528 min Scan# 3392
Delta R.T. 0.003 min
Lab File: BN014856.D
Acq: 09 Jun 2021 17:01

Instrument :
BNA_N
ClientSampleId :
PB136126BS

Tgt Ion:	276	Resp:	15034
Ion Ratio	Lower	Upper	
276	100		
277	24.7	19.7	29.5
138	12.1	9.8	14.6

