

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060922\
 Data File : BN020070.D
 Acq On : 09 Jun 2022 08:43
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020313

Quant Time: Jun 10 01:04:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	175925	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	825503	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	533819	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	1096303	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	1037452	20.000	ng/u1	0.00
88) Perylene-d12	23.716	264	1048850	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	31274	7.700	ng/uL	0.00
4) Pyridine-d5	3.687	84	219701	19.790	ng/u1	0.00
7) Phenol-d5	6.987	99	283686	19.488	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	186309	20.484	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	240023	20.293	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	249031	19.773	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	124560	19.834	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	132569	19.890	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	235577	20.247	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	369652	19.618	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	764754	19.689	ng/u1	0.00
49) Acenaphthylene-d8	14.140	160	901498	20.014	ng/u1	0.00
54) 4-Nitrophenol-d4	14.640	143	147584	18.962	ng/u1	0.00
60) Fluorene-d10	15.439	176	645093	20.070	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	120523	18.979	ng/u1	0.00
73) Anthracene-d10	17.286	188	965155	20.172	ng/u1	0.00
81) Pyrene-d10	19.580	212	1092125	19.418	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.563	264	1039581	20.258	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	31782	7.412	ng/uL	91
5) Pyridine	3.705	79	231008	19.849	ng/u1	90
6) Benzaldehyde	6.958	77	172469	25.144	ng/u1	93
8) Phenol	7.011	94	310347	20.021	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.240	93	244672	20.167	ng/u1	97
12) 2-Chlorophenol	7.381	128	251863	20.371	ng/u1	99
13) 2-Methylphenol	8.258	108	238135	19.983	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.352	45	383394	20.152	ng/u1	98
16) Acetophenone	8.634	105	393723	20.556	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.628	70	199666	20.407	ng/u1	95
18) 4-Methylphenol	8.587	108	267040	20.303	ng/u1	94
19) Hexachloroethane	8.899	117	100457	20.193	ng/u1	87
22) Nitrobenzene	9.010	77	289648	19.891	ng/u1	97
23) Isophorone	9.540	82	558445	19.423	ng/u1	98
25) 2-Nitrophenol	9.722	139	149053	20.238	ng/u1	98
26) 2,4-Dimethylphenol	9.787	107	301860	19.784	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.022	93	348846	19.827	ng/u1	99
29) 2,4-Dichlorophenol	10.257	162	244942	20.351	ng/u1	99
30) Naphthalene	10.657	128	844444	19.738	ng/u1	99
32) 4-Chloroaniline	10.763	127	370101	19.638	ng/u1	100
33) Hexachlorobutadiene	10.951	225	135019	19.539	ng/u1	97
34) Caprolactam	11.516	113	51564	10.971	ng/u1	93
35) 4-Chloro-3-methylphenol	11.893	107	287861	20.255	ng/u1	99
36) 2-Methylnaphthalene	12.269	142	580190	20.059	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060922\
 Data File : BN020070.D
 Acq On : 09 Jun 2022 08:43
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020313

Quant Time: Jun 10 01:04:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.493	142	589692	19.979	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.646	216	271932	19.973	ng/ul	98
40) Hexachlorocyclopentadiene	12.628	237	145528	17.572	ng/ul	99
41) 2,4,6-Trichlorophenol	12.881	196	190748	20.319	ng/ul	91
42) 2,4,5-Trichlorophenol	12.951	196	209679	20.167	ng/ul	99
43) 1,1'-Biphenyl	13.281	154	786272	19.891	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	590834	19.703	ng/ul	97
45) 2-Nitroaniline	13.522	65	188152	20.168	ng/ul	96
47) Dimethylphthalate	13.904	163	753235	19.264	ng/ul	98
48) 2,6-Dinitrotoluene	14.022	165	156125	20.422	ng/ul	98
50) Acenaphthylene	14.169	152	993368	19.919	ng/ul	100
51) 3-Nitroaniline	14.345	138	168374	20.833	ng/ul	99
52) Acenaphthene	14.510	153	626399	19.229	ng/ul	99
53) 2,4-Dinitrophenol	14.557	184	77446	15.822	ng/ul	98
55) 4-Nitrophenol	14.657	109	126489	19.039	ng/ul	98
56) Dibenzofuran	14.845	168	902413	19.663	ng/ul	97
57) 2,4-Dinitrotoluene	14.810	165	227278	20.078	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.075	232	157800	19.334	ng/ul#	89
59) Diethylphthalate	15.269	149	783824	19.471	ng/ul	99
61) Fluorene	15.492	166	695575	19.478	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.487	204	321780	19.566	ng/ul	93
63) 4-Nitroaniline	15.504	138	165185	21.872	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.575	198	123137	19.125	ng/ul	98
67) N-Nitrosodiphenylamine	15.704	169	635179	20.210	ng/ul	98
68) 4-Bromophenyl-phenylether	16.381	248	192620	20.176	ng/ul#	87
69) Hexachlorobenzene	16.504	284	215692	19.654	ng/ul#	84
70) Atrazine	16.651	200	224600	20.481	ng/ul	96
71) Pentachlorophenol	16.845	266	131815	19.284	ng/ul	97
72) Phenanthrene	17.228	178	1147934	19.841	ng/ul	100
74) Anthracene	17.322	178	1177910	20.248	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.251	216	284430	19.375	ng/uL	98
76) Pentachlorobenzene	14.769	250	262059	20.197	ng/uL	94
77) Carbazole	17.586	167	1102904	20.839	ng/ul	97
78) Di-n-butylphthalate	18.163	149	1330837	20.279	ng/ul	99
80) Fluoranthene	19.251	202	1317842	19.401	ng/ul	97
82) Pyrene	19.610	202	1355290	19.448	ng/ul#	92
83) Butylbenzylphthalate	20.516	149	628823	20.050	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.298	252	416456	21.017	ng/ul#	98
85) Benzo(a)anthracene	21.363	228	1263036	19.280	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.304	149	914244	20.660	ng/ul	95
87) Chrysene	21.416	228	1292897	20.173	ng/ul	99
89) Di-n-octyl phthalate	22.204	149	1618040	18.954	ng/ul	100
90) Benzo(b)fluoranthene	23.004	252	1314733	19.972	ng/ul#	97
91) Benzo(k)fluoranthene	23.051	252	1290766	20.598	ng/ul	96
93) Benzo(a)pyrene	23.610	252	1267087	19.885	ng/ul	95
94) Indeno(1,2,3-cd)pyrene	26.104	276	1299663	19.824	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.121	278	1108360	19.793	ng/ul#	94
96) Benzo(g,h,i)perylene	26.839	276	1091996	19.724	ng/ul#	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060922\
 Data File : BN020070.D
 Acq On : 09 Jun 2022 08:43
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020313

Quant Time: Jun 10 01:04:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.412 ng/uL
RT: 3.317 min Scan# 39
Delta R.T. 0.001 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion: 88 Resp: 31782
Ion Ratio Lower Upper
88 100
43 31.7 22.4 33.6
58 84.2 61.0 91.4



#5
Pyridine
Concen: 19.849 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion: 79 Resp: 231008
Ion Ratio Lower Upper
79 100
52 92.8 65.0 97.4
51 42.4 31.6 47.4









#6
Benzaldehyde
Concen: 25.144 ng/ul
RT: 6.958 min Scan# 61
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion: 77 Resp: 172469
Ion Ratio Lower Upper
77 100
105 98.8 74.7 112.1
106 96.6 70.2 105.4



#8
Phenol
Concen: 20.021 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion: 94 Resp: 310347
Ion Ratio Lower Upper
94 100
65 29.8 23.6 35.4
66 40.6 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 20.167 ng/ul
RT: 7.240 min Scan# 707
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion: 93 Resp: 244672
Ion Ratio Lower Upper
93 100
63 82.7 63.4 95.0
95 33.7 26.2 39.2



#12
2-Chlorophenol
Concen: 20.371 ng/ul
RT: 7.381 min Scan# 730
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:128 Resp: 251863
Ion Ratio Lower Upper
128 100
64 47.9 38.6 57.8
130 32.6 25.0 37.4









#13
2-Methylphenol
Concen: 19.983 ng/ul
RT: 8.258 min Scan# 81
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:108 Resp: 238135
Ion Ratio Lower Upper
108 100
107 89.3 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.152 ng/ul
RT: 8.352 min Scan# 895
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion: 45 Resp: 383394
Ion Ratio Lower Upper
45 100
77 12.7 10.9 16.3
79 9.7 8.6 13.0









#16
Acetophenone
Concen: 20.556 ng/ul
RT: 8.634 min Scan# 944
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:105 Resp: 393723
Ion Ratio Lower Upper
105 100
77 77.6 63.5 95.3
51 30.9 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 20.407 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion: 70 Resp: 199666
Ion Ratio Lower Upper
70 100
42 63.9 47.8 71.6
101 10.0 7.8 11.6
130 22.4 15.8 23.8









#18
4-Methylphenol
Concen: 20.303 ng/ul
RT: 8.587 min Scan# 911
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:108 Resp: 267040
Ion Ratio Lower Upper
108 100
107 116.6 88.0 132.0



#19
Hexachloroethane
Concen: 20.193 ng/ul
RT: 8.899 min Scan# 988
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:117 Resp: 100457
Ion Ratio Lower Upper
117 100
201 87.6 83.2 124.8
199 56.1 50.2 75.4









#22
Nitrobenzene
Concen: 19.891 ng/ul
RT: 9.010 min Scan# 1008
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion: 77 Resp: 289648
Ion Ratio Lower Upper
77 100
123 45.4 34.5 51.7
65 14.5 11.1 16.7



#23
Isophorone
Concen: 19.423 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion: 82 Resp: 558445
Ion Ratio Lower Upper
82 100
95 7.5 5.8 8.8
138 18.2 13.8 20.8





#25

2-Nitrophenol

Concen: 20.238 ng/ul

RT: 9.722 min Scan# 1129

Delta R.T. -0.006 min

Lab File: BN020070.D

Acq: 09 Jun 2022 08:43

Instrument :

BNA_N

ClientSampleId :

SSTD020313





Tgt Ion:139 Resp: 149053

Ion Ratio Lower Upper

139 100

65 55.3 43.7 65.5

109 33.7 25.8 38.6



#26

2,4-Dimethylphenol

Concen: 19.784 ng/ul

RT: 9.787 min Scan# 1139

Delta R.T. -0.006 min

Lab File: BN020070.D

Acq: 09 Jun 2022 08:43




Tgt Ion:107 Resp: 301860


Ion Ratio Lower Upper

107 100

121 49.0 38.3 57.5

122 84.6 66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.827 ng/ul
 RT: 10.022 min Scan# 1179
 Delta R.T. -0.006 min
 Lab File: BN020070.D
 Acq: 09 Jun 2022 08:43




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020313


Tgt Ion: 93 Resp: 348846
 Ion Ratio Lower Upper
 93 100
 95 31.5 25.4 38.0
 123 11.3 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 20.351 ng/ul
 RT: 10.257 min Scan# 1219
 Delta R.T. -0.006 min
 Lab File: BN020070.D
 Acq: 09 Jun 2022 08:43

Tgt Ion: 162 Resp: 244942
 Ion Ratio Lower Upper
 162 100
 164 65.3 52.9 79.3
 98 37.3 29.8 44.8









#30
Naphthalene
Concen: 19.738 ng/ul
RT: 10.657 min Scan# 1287
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:128 Resp: 844444
Ion Ratio Lower Upper
128 100
129 10.6 8.7 13.1
127 12.8 10.6 16.0



#32
4-Chloroaniline
Concen: 19.638 ng/ul
RT: 10.763 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:127 Resp: 370101
Ion Ratio Lower Upper
127 100
129 32.8 26.1 39.1









#33
Hexachlorobutadiene
Concen: 19.539 ng/ul
RT: 10.951 min Scan# 1337
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:225 Resp: 135019
Ion Ratio Lower Upper
225 100
223 63.0 51.8 77.6
227 61.1 50.6 76.0



#34
Caprolactam
Concen: 10.971 ng/ul
RT: 11.516 min Scan# 1433
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:113 Resp: 51564
Ion Ratio Lower Upper
113 100
55 167.8 145.5 218.3
56 129.4 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 20.255 ng/ul
RT: 11.893 min Scan# 1497
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:107 Resp: 287861
Ion Ratio Lower Upper
107 100
144 25.2 20.6 31.0
142 77.0 62.2 93.4



#36
2-Methylnaphthalene
Concen: 20.059 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:142 Resp: 580190
Ion Ratio Lower Upper
142 100
141 88.5 70.6 105.8









#37
1-Methylnaphthalene
Concen: 19.979 ng/ul
RT: 12.493 min Scan# 1599
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:142 Resp: 589692
Ion Ratio Lower Upper
142 100
141 92.2 70.9 106.3
116 3.8 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.973 ng/ul
RT: 12.646 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:216 Resp: 271932
Ion Ratio Lower Upper
216 100
214 77.3 63.4 95.0
179 22.8 17.8 26.8
108 20.3 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 17.572 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:237 Resp: 145528
Ion Ratio Lower Upper
237 100
235 60.1 48.8 73.2
272 11.8 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 20.319 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:196 Resp: 190748
Ion Ratio Lower Upper
196 100
198 89.3 79.8 119.8
200 29.3 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 20.167 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:196 Resp: 209679
Ion Ratio Lower Upper
196 100
198 96.4 77.6 116.4
200 29.6 24.4 36.6



#43
1,1'-Biphenyl
Concen: 19.891 ng/ul
RT: 13.281 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:154 Resp: 786272
Ion Ratio Lower Upper
154 100
153 40.3 31.5 47.3
76 13.1 10.5 15.7









#44
2-Chloronaphthalene
Concen: 19.703 ng/ul
RT: 13.322 min Scan# 1740
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:162 Resp: 590834
Ion Ratio Lower Upper
162 100
164 32.8 25.8 38.6
127 42.6 31.4 47.2



#45
2-Nitroaniline
Concen: 20.168 ng/ul
RT: 13.522 min Scan# 1774
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion: 65 Resp: 188152
Ion Ratio Lower Upper
65 100
92 67.0 54.1 81.1
138 102.3 77.0 115.4









#47
Dimethylphthalate
Concen: 19.264 ng/ul
RT: 13.904 min Scan# 1839
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:163 Resp: 753235
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.0
164 10.4 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 20.422 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:165 Resp: 156125
Ion Ratio Lower Upper
165 100
89 61.6 50.6 75.8
121 22.5 17.9 26.9









#50
Acenaphthylene
Concen: 19.919 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:152 Resp: 993368
Ion Ratio Lower Upper
152 100
151 20.0 16.0 24.0
153 13.1 10.7 16.1



#51
3-Nitroaniline
Concen: 20.833 ng/ul
RT: 14.345 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:138 Resp: 168374
Ion Ratio Lower Upper
138 100
108 11.2 8.6 12.8
92 114.8 92.9 139.3





#52

Acenaphthene

Concen: 19.229 ng/ul

RT: 14.510 min Scan# 1943

Delta R.T. -0.006 min

Lab File: BN020070.D

Acq: 09 Jun 2022 08:43

Instrument :

BNA_N

ClientSampleId :

SSTD020313





Tgt Ion:153 Resp: 626399

Ion Ratio Lower Upper

153 100

152 47.5 38.4 57.6

154 88.1 69.7 104.5



#53

2,4-Dinitrophenol

Concen: 15.822 ng/ul

RT: 14.557 min Scan# 1950

Delta R.T. 0.000 min

Lab File: BN020070.D

Acq: 09 Jun 2022 08:43




Tgt Ion:184 Resp: 77446


Ion Ratio Lower Upper

184 100

63 72.2 55.9 83.9

154 67.7 53.1 79.7









#55
4-Nitrophenol
Concen: 19.039 ng/ul
RT: 14.657 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:109 Resp: 126489
Ion Ratio Lower Upper
109 100
139 128.9 100.1 150.1
65 129.8 104.6 157.0



#56
Dibenzofuran
Concen: 19.663 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:168 Resp: 902413
Ion Ratio Lower Upper
168 100
139 40.6 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 20.078 ng/ul
RT: 14.810 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:165 Resp: 227278
Ion Ratio Lower Upper
165 100
63 47.7 36.3 54.5
182 2.9 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.334 ng/ul
RT: 15.075 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:232 Resp: 157800
Ion Ratio Lower Upper
232 100
131 57.1 38.5 57.7
130 3.3 2.2 3.2#
166 37.1 25.8 38.8









#59
Diethylphthalate
Concen: 19.471 ng/ul
RT: 15.269 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:149 Resp: 783824
Ion Ratio Lower Upper
149 100
177 21.2 17.8 26.8
150 13.1 10.5 15.7



#61
Fluorene
Concen: 19.478 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:166 Resp: 695575
Ion Ratio Lower Upper
166 100
165 98.0 77.7 116.5
167 13.4 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 19.566 ng/ul
RT: 15.487 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:204 Resp: 321780
Ion Ratio Lower Upper
204 100
206 30.7 26.7 40.1
141 68.6 50.2 75.2



#63
4-Nitroaniline
Concen: 21.872 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:138 Resp: 165185
Ion Ratio Lower Upper
138 100
92 51.6 41.0 61.6
108 82.1 65.3 97.9









#66
4,6-Dinitro-2-methylphenol
Concen: 19.125 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:198 Resp: 123137
Ion Ratio Lower Upper
198 100
182 4.1 3.0 4.4
77 26.3 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 20.210 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:169 Resp: 635179
Ion Ratio Lower Upper
169 100
168 63.9 52.4 78.6
167 35.2 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 20.176 ng/ul
RT: 16.381 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:248 Resp: 192620
Ion Ratio Lower Upper
248 100
250 97.6 77.8 116.8
141 102.6 62.6 93.8#



#69
Hexachlorobenzene
Concen: 19.654 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:284 Resp: 215692
Ion Ratio Lower Upper
284 100
142 48.0 27.8 41.8#
249 36.4 25.7 38.5









#70
Atrazine
Concen: 20.481 ng/ul
RT: 16.651 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:200 Resp: 224600
Ion Ratio Lower Upper
200 100
173 29.3 20.5 30.7
215 50.0 38.9 58.3



#71
Pentachlorophenol
Concen: 19.284 ng/ul
RT: 16.845 min Scan# 2339
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:266 Resp: 131815
Ion Ratio Lower Upper
266 100
264 66.3 50.2 75.2
268 60.3 49.3 73.9









#72
Phenanthrene
Concen: 19.841 ng/ul
RT: 17.228 min Scan# 2404
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:178 Resp: 1147934
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 18.9 15.2 22.8



#74
Anthracene
Concen: 20.248 ng/ul
RT: 17.322 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:178 Resp: 1177910
Ion Ratio Lower Upper
178 100
179 15.4 13.0 19.4
176 18.4 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.375 ng/uL
RT: 13.251 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:216 Resp: 284430
Ion Ratio Lower Upper
216 100
214 80.3 63.2 94.8
218 48.8 37.9 56.9



#76
Pentachlorobenzene
Concen: 20.197 ng/uL
RT: 14.769 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:250 Resp: 262059
Ion Ratio Lower Upper
250 100
248 61.5 53.0 79.4
252 61.6 53.4 80.2









#77
 Carbazole
 Concen: 20.839 ng/ul
 RT: 17.586 min Scan# 2465
 Delta R.T. -0.006 min
 Lab File: BN020070.D
 Acq: 09 Jun 2022 08:43




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020313


Tgt Ion:167 Resp: 1102904
 Ion Ratio Lower Upper
 167 100
 166 21.5 16.1 24.1
 139 13.8 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 20.279 ng/ul
 RT: 18.163 min Scan# 2563
 Delta R.T. 0.000 min
 Lab File: BN020070.D
 Acq: 09 Jun 2022 08:43

Tgt Ion:149 Resp: 1330837
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.1 10.7
 104 5.3 4.1 6.1









#80
Fluoranthene
Concen: 19.401 ng/u1
RT: 19.251 min Scan# 2748
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:202 Resp: 1317842
Ion Ratio Lower Upper
202 100
101 12.3 8.9 13.3
100 9.6 6.7 10.1



#82
Pyrene
Concen: 19.448 ng/u1
RT: 19.610 min Scan# 2809
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:202 Resp: 1355290
Ion Ratio Lower Upper
202 100
101 15.6 9.7 14.5#
100 12.4 8.0 12.0#





#83

Butylbenzylphthalate

Concen: 20.050 ng/ul

RT: 20.516 min Scan# 2963

Delta R.T. -0.006 min

Lab File: BN020070.D

Acq: 09 Jun 2022 08:43

Instrument :

BNA_N

ClientSampleId :

SSTD020313





Tgt Ion:149 Resp: 628823

Ion Ratio Lower Upper

149 100

91 71.2 58.8 88.2

206 19.8 18.1 27.1



#84

3,3'-Dichlorobenzidine

Concen: 21.017 ng/ul

RT: 21.298 min Scan# 3096

Delta R.T. 0.000 min

Lab File: BN020070.D

Acq: 09 Jun 2022 08:43




Tgt Ion:252 Resp: 416456


Ion Ratio Lower Upper

252 100

254 62.3 50.8 76.2

126 13.0 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 19.280 ng/ul
RT: 21.363 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:228 Resp: 1263036
Ion Ratio Lower Upper
228 100
229 20.1 16.3 24.5
226 26.6 23.2 34.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.660 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:149 Resp: 914244
Ion Ratio Lower Upper
149 100
167 25.5 22.5 33.7
279 3.5 3.5 5.3









#87
Chrysene
Concen: 20.173 ng/ul
RT: 21.416 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:228 Resp: 1292897
Ion Ratio Lower Upper
228 100
226 29.6 24.2 36.2
229 20.4 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 18.954 ng/ul
RT: 22.204 min Scan# 3250
Delta R.T. -0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:149 Resp: 1618040





#90

Benzo(b)fluoranthene

Concen: 19.972 ng/u1

RT: 23.004 min Scan# 3386

Delta R.T. 0.000 min

Lab File: BN020070.D

Acq: 09 Jun 2022 08:43

Instrument :

BNA_N

ClientSampleId :

SSTD020313





Tgt Ion:252 Resp: 1314733

Ion Ratio Lower Upper

252 100

253 21.5 16.5 24.7

125 12.0 7.8 11.8#



#91

Benzo(k)fluoranthene

Concen: 20.598 ng/u1

RT: 23.051 min Scan# 3394

Delta R.T. 0.000 min

Lab File: BN020070.D

Acq: 09 Jun 2022 08:43




Tgt Ion:252 Resp: 1290766


Ion Ratio Lower Upper

252 100

253 20.7 18.2 27.4

125 11.9 8.2 12.4









#93
Benzo(a)pyrene
Concen: 19.885 ng/ul
RT: 23.610 min Scan# 3489
Delta R.T. 0.000 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43




Instrument :
BNA_N
ClientSampleId :
SSTD020313


Tgt Ion:252 Resp: 1267087
Ion Ratio Lower Upper
252 100
253 21.0 18.7 28.1
125 13.7 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.824 ng/ul
RT: 26.104 min Scan# 3913
Delta R.T. 0.006 min
Lab File: BN020070.D
Acq: 09 Jun 2022 08:43

Tgt Ion:276 Resp: 1299663
Ion Ratio Lower Upper
276 100
138 30.3 19.5 29.3#
277 24.0 21.4 32.0









#95
 Dibenzo(a,h)anthracene
 Concen: 19.793 ng/ul
 RT: 26.121 min Scan# 3915
 Delta R.T. 0.000 min
 Lab File: BN020070.D
 Acq: 09 Jun 2022 08:43

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020313

Tgt Ion:278 Resp: 1108360
 Ion Ratio Lower Upper
 278 100
 139 20.9 13.8 20.8#
 279 25.1 18.3 27.5



#96
 Benzo(g,h,i)perylene
 Concen: 19.724 ng/ul
 RT: 26.839 min Scan# 4038
 Delta R.T. -0.006 min
 Lab File: BN020070.D
 Acq: 09 Jun 2022 08:43

Tgt Ion:276 Resp: 1091996
 Ion Ratio Lower Upper
 276 100
 138 27.8 17.0 25.6#
 277 25.2 17.8 26.6

