

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060922\
 Data File : BN020071.D
 Acq On : 09 Jun 2022 09:19
 Operator : CG/JU
 Sample : PB145381BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK381

Quant Time: Jun 10 01:04:17 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	136013	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	640540	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	429144	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.187	188	894271	20.000	ng/u1	0.00
79) Chrysene-d12	21.375	240	861502	20.000	ng/u1	# 0.00
88) Perylene-d12	23.710	264	812416	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	17458	5.559	ng/uL	0.00
4) Pyridine-d5	3.687	84	248290	28.928	ng/u1	0.00
7) Phenol-d5	6.987	99	302269	26.857	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	199597	28.385	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	259072	28.330	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	265609	27.278	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	130540	26.788	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	144176	27.878	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	241894	26.793	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	413844	28.306	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	879613	28.170	ng/u1	0.00
49) Acenaphthylene-d8	14.140	160	1005359	27.763	ng/u1	0.00
54) 4-Nitrophenol-d4	14.640	143	151683	24.242	ng/u1	0.00
60) Fluorene-d10	15.440	176	745366	28.847	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	121552	23.465	ng/u1	0.00
73) Anthracene-d10	17.287	188	1172458	30.041	ng/u1	0.00
81) Pyrene-d10	19.581	212	1345253	28.804	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.557	264	1153050	29.008	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060922\
Data File : BN020071.D
Acq On : 09 Jun 2022 09:19
Operator : CG/JU
Sample : PB145381BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK381

Quant Time: Jun 10 01:04:17 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 06 15:14:59 2022
Response via : Initial Calibration

