

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060922\
 Data File : BN020073.D
 Acq On : 09 Jun 2022 10:32
 Operator : CG/JU
 Sample : N3212-06
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F5A55

Quant Time: Jun 10 01:04:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	131588	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	624948	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	414243	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.187	188	880781	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	837315	20.000	ng/u1	0.00
88) Perylene-d12	23.704	264	830173	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	15594	5.133	ng/uL	0.00
4) Pyridine-d5	3.687	84	89719	10.804	ng/u1	0.00
7) Phenol-d5	6.987	99	77159	7.086	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	202231	29.727	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	210040	23.741	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	147436	15.651	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	136905	28.795	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	144201	28.578	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	229398	26.042	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	356750	25.009	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	923886	30.652	ng/u1	0.00
49) Acenaphthylene-d8	14.140	160	1032206	29.530	ng/u1	0.00
54) 4-Nitrophenol-d4	14.640	143	36378	6.023	ng/u1	0.00
60) Fluorene-d10	15.440	176	796605	31.939	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	124593	24.420	ng/u1	0.00
73) Anthracene-d10	17.287	188	1359723	35.373	ng/u1	0.00
81) Pyrene-d10	19.581	212	1510001	33.265	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.557	264	1428739	35.175	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060922\
Data File : BN020073.D
Acq On : 09 Jun 2022 10:32
Operator : CG/JU
Sample : N3212-06
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F5A55

Quant Time: Jun 10 01:04:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 06 15:14:59 2022
Response via : Initial Calibration

