




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037203.D
Acq On : 09 Jun 2025 22:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 10 06:02:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 10 06:01:37 2025
Response via : Initial Calibration



TIC: BN037203.D\data.ms

(70) Pentachlorophenol (C)




16.639min (0.000) 275.85 ng

response 102519

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.86
264.00	61.60	57.89
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037203.D
Acq On : 09 Jun 2025 22:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 10 06:02:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 10 06:01:37 2025
Response via : Initial Calibration



TIC: BN037203.D\data.ms

(77) Benzidine

19.245min (0.000) 0.00 ng

response 432254

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.09
183.00	13.20	12.58
0.00	0.00	0.00